




KMCT
MEDICAL COLLEGE

Approved by NMC

Affiliated to Kerala University of Health Sciences (KUHS)

TO WHOMSOEVER IT MAY CONCERN

This is to certify that the information in the attached documents is verified by me and is true to the best of my knowledge.




Principal
KMCT Medical College
Manassery P.O.
Kozhikode

COMPLAINT DATE	COMPLAINT LOG TIME	DEPARTMENT	NAME OF EQUIPMENT	MAKE
6/1/2023	9:30AM	W&C OBG WARD	CTG MACHINE	BPL
6/1/2023	9:00AM	W&C NICU	VENTILATOR	DRAGGER
6/1/2023	4:20PM	ICCU	VENTILATOR	SKANRAY
6/1/2023	3:15PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
6/1/2023	3:00PM	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
6/2/2023	2:30PM	RADIO DIAGNOSIS	MAMMOGRAPHY MACHINE	GE
6/2/2023	2:00PM	RO PLANT	RO WATER	NA
6/2/2023	12:30PM	DERMATOLOGY OT	CENTRIFUGE	KEMI
6/2/2023	12:00PM	NMICU	BIPAP	PHILIPS
6/3/2023	11:00AM	MICU	VENTILATOR	NEUMOVENT
6/3/2023	9:45AM	RICU	MULTIPARA MONITOR	SKANRAY
6/3/2023	9:40PM	W&C COVID WARD	VENTILATOR	SKANRAY
6/3/2023	9:30AM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
6/3/2023	9:00AM	ICCU	VENTILATOR	NEUMOVENT
6/4/2023	8:50PM	CSSD	AUTOCLAVE MACHINE	SURGITECH
6/4/2023	8:15PM	ICCU	MULTIPARA MONITOR	SCHILLER
6/4/2023	7:30PM	ED	BIPAP	BME
6/5/2023	6:00PM	ICCU	VENTILATOR	SIEMENS
6/5/2023	8:00AM	ICCU	VENTILATOR	SIEMENS
6/5/2023	6:00AM	RO PLANT	RO WATER	NA
6/5/2023	4:00PM	RO PLANT	RO WATER	NA
6/5/2023	3:00PM	ED	ABG MACHINE	GEM
6/5/2023	10:30AM	RADIOLOGY	CR	AGFA
6/5/2023	10:0AM	NSICU	VENTILATOR	NEUMOVENT
6/5/2023	1:00PM	RADIOLOGY	CR	AGFA
6/5/2023	9:30PM	CVTS ICU	ABG MACHINE	GEM
6/5/2023	9:00AM	RADIOLOGY	CR	AGFA
6/5/2023	02:30PM	CVTS ICU	ABG MACHINE	GEM
6/6/2023	2:00PM	USG ROOM	USG MACHINE	GE
6/6/2023	9:00am	RADIO DIAGNOSIS	X RAY MACHINE (MOBILE-C	SKANRAY
6/6/2023	11:15AM	RADIOLOGY	CR	AGFA
6/7/2023	1:50PM	EMERGENCY DEPARTM	ABG MACHINE	GEM
6/7/2023	6:20PM	EMERGENCY DEPARTM	LARYNGOSCOPE (ADL)	NA
6/7/2023	9:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
6/7/2023	10:00AM	NMICU	VENTILATOR	SIEMENS
6/8/2023	10:30AM	RADIOLOGY	CT MACHINE	WIPRO GE
6/8/2023	11:00AM	RADIOLOGY	MRI MACHINE	WIPRO GE
6/8/2023	12:00PM	MAJOR OT	CAUTRY PATIENT PLATE CAB	NA
6/8/2023	1:00PM	SICU	PATIENT MONITOR	MINDRAY
6/8/2023	1:30PM	MAJOR OT	CAUTRY FOOT SWITCH	NA
6/9/2023	3:00PM	RO PLANT	RO WATER	NA
6/9/2023	3:30PM	W&C OT	CAUTRY PATIENT PLATE CAB	NA

6/9/2023	09:00AM	RO PLANT	NA	NA
6/9/2023	09:00AM	MICU	MULTIPARA MONITOR	EDAN
6/9/2023	10:20AM	LABOUR ROOM	USG MACHINE	SAMSUNG
6/9/2023	11:00AM	NSICU	MULTIPARA MONITOR	L&T
6/10/2023	12:00PM	LABOUR ROOM	CTG MACHINE	MS TECHNOLOGIE
6/10/2023	03:00PM	SICU	SYRINGE PUMP	L&T
6/10/2023	4:15PM	FMW	ECG MACHINE	EDAN
6/10/2023	09:00AM	EMERGENCY DEPARTM	ABG MACHINE	GEM
6/10/2023	10:30AM	MICU	ECG MACHINE	EDAN
6/11/2023	01:30PM	SICU	MULTIPARA MONITOR	MINDRAY
6/11/2023	03:45PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
6/11/2023	09:00AM	MOT	C ARM	SKANRAY
6/11/2023	10:30AM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
6/11/2023	11:15AM	URO ICU	MULTIPARAMONITOR	AMIGO
6/11/2023	01:30PM	ECHO	CARDIOLOGY	ECHO MACHINE
6/12/2023	03:45PM	NSICU	MULTIPARA MONITOR	AMIGO
6/12/2023	04:15PM	LABOUR ROOM	USG MACHINE	SAMSUNG
6/12/2023	04:45PM	RO PLANT	RO WATER	NA
6/12/2023	09:00AM	NMICU	MULTIPARA MONITOR	MINDRAY
6/12/2023	12:00PM	MICU	VENTILATOR	
6/12/2023	02:45PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
6/12/2023	06:00PM	RICU	BIPAP	PHILIPS
6/12/2023	09:00AM	MOT	UPS	CONSUL NEOWAT
6/12/2023	10:05AM	MOT	BP APPARATUS	ELKOMETER
6/13/2023	11:12AM	ED	ABG MACHINE	GEM
6/13/2023	12:50PM	CVTS ICU	ABG MACHINE	GEM
6/13/2023	09:15AM	RICU	GLUCOMETER	GLUCOCARD
6/13/2023	10:10AM	LABOUR ROOM	O2 FLOWMETER	NA
6/14/2023	11:20AM	MOT	CAUTRY MACHINE	ALAN
6/14/2023	01:50PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
6/14/2023	02:45PM	NSICU	MULTIPARA MONITOR	AMIGO
6/14/2023	03:45PM	W&C OT	WARMER	PRAGATHI
6/14/2023	09:20AM	ED	BIPAP	PHILIPS
6/14/2023	10:20PM	MEDICINE OP	BIO THESIOMETER	DIABATIC FOOT CA
6/15/2023	12:35PM	ED	ABG MACHINE	GEM
6/15/2023	02:50PM	ED	ABG MACHINE	GEM
6/15/2023	04:00PM	CSSD	ETO STERILIZER	3M
6/15/2023	09:30AM	W&C COVID WARD	FLOW METER	NA
6/15/2023	10:20PM	ICCU	VENTILATOR	SIEMENS
6/15/2023	11:35AM	RO PLANT	RO WATER	NA
6/15/2023	01:40PM	CSSD	AUTOCLAVE MACHINE	SURGITECH
6/15/2023	03:00PM	NSICU	ECG MACHINE	SKANRAY
6/16/2023	04:20PM	MOT	PORTABLE SUCTION	NA
6/16/2023	09:15PM	NSICU	VENTILATOR	NEUMOVENT
6/16/2023	10:20PM	RADIOLOGY	CR	AGFA

6/16/2023	11:30PM	CVTS ICU	ABG MACHINE	GEM
6/16/2023	02:10PM	RADIOLOGY	CR	AGFA
6/16/2023	03:20PM	SICU	MONITOR	SKANRAY
6/16/2023	06:00PM	RICU	VENTILATOR	SKANRAY
6/17/2023	09:30AM	RO PLANT	NA	NA
6/17/2023	10:20AM	CENTRAL LAB	UPS	NA
6/17/2023	11:30AM	EMERGENCY DEPARTM	DEFIBRILLATOR	PHILIPS
6/17/2023	02:20PM	EMERGENCY DEPARTM	ECG MACHINE	EDAN
6/17/2023	03:50PM	NMICU	SIEMENS	SERVO VENTILATO
6/17/2023	04:55PM	RO PLANT	NA	NA
6/17/2023	7:10PM	EMERGENCY DEPARTM	ECG MACHINE	EDAN
6/18/2023	09:10AM	SURGERY WARD	SUCTION	NA
6/18/2023	10:15AM	CSSD	SEALING MACHINE	SEPACK
6/18/2023	11:30AM	RICU	DEFIBRILLATOR	HP
6/18/2023	12:30PM	EMERGENCY DEPARTM	X RAY MACHINE (MOBILE	SKANRAY
6/18/2023	01:45PM	RO PLANT	NA	NA
6/18/2023	02:30PM	MOT	OT LIGHT	HOSPEDIA
6/18/2023	03:40PM	NICU	WARMER	PRAGATHI
6/19/2023	04:50PM	CVTS	VENTILATOR	NEUMOVENT
6/19/2023	09:50AM	USG ROOM	USG MACHINE	GE
6/19/2023	11:30AM	W AND C OT	OT TABLE	HOSPITECH
6/19/2023	02:20PM	W&C SICU	ECG MACHINE	AMIGO
6/19/2023	10:10AM	RO PLANT	RO WATER	NA
6/19/2023	11:20AM	NMICU	MONITOR	MINDRAY
6/19/2023	01:50PM	ICCU	SYRINGE PUMP	B BRAUN
6/19/2023	02:45PM	ED	ECG MACHINE	EDAN
6/20/2023	03:45PM	CSSD	SEALINHG MACHINE	SEPACK
6/20/2023	09:20AM	SICU	VENTILATOR	SKANRAY
6/20/2023	10:20PM	ICCU	SYRINGE PUMP	SCHILLER
6/20/2023	12:35PM	SURGERY WARD	VACCUM	NA
6/20/2023	02:50PM	W&C SICU	VENTILATOR	SIEMENS
6/20/2023	04:00PM	RO UNIT	RO WATER	NA
6/20/2023	09:30AM	RICU	BIPAP	PHILIPS
6/20/2023	10:20PM	SICU	VENTILATOR	SKANRAY
6/21/2023	11:35AM	CENTRAL LAB	ELECTROLYTE ANALYSER	B&E
6/21/2023	01:40PM	ED	MONITOR	BLUESTAR
6/21/2023	03:00PM	ED	FLOW METER	NA
6/21/2023	04:20PM	ED	DEFIBRILLATOR	PHILIPS
6/21/2023	09:15PM	PULMO WARD	GLUCOMETER	ALEREGI
6/21/2023	10:20PM	PULMO WARD	BIPAP	PHILIPS
6/21/2023	11:30PM	DIALYSIS	MONITOR	AMIGO
6/21/2023	02:10PM	MOT	LIGHT SOURCE	STORZ
6/22/2023	03:20PM	RO PLANT	RO WATER	NA
6/22/2023	06:00PM	CVTS OT	SEALINHG MACHINE	NA
6/22/2023	09:30AM	W&C OT	CAUOTARY MACHINE	CAL-SURG

6/22/2023	10:20AM	PULMO WARD	BIPAP	RESMED
6/22/2023	11:30AM	RICU	VENTILATOR	NEUMOVENT
6/22/2023	01:00PM	AYURVEDA COVID HOS	CENTRAL FLOW METER	NA
6/23/2023	02:20PM	ED	PORTABLE VENTILLATOR	PHILIPS
6/23/2023	3:00pm	ED	ECG MACHINE	EDAN
6/23/2023	55:0pm	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
6/23/2023	5:30pm	RO PLANT	RO WATER	NA
6/23/2023	9:00am	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
6/23/2023	10:00am	nmiCU	coat	janak
6/23/2023	11:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
6/23/2023	12:00PM	W&C PICU	MULTIPARA MONITOR	SCURE
6/24/2023	2:00PM	EMERGENCY DEPART	ABG MACHINE	GEM
6/24/2023	3:00PM	CVTS	COAT	NA
6/24/2023	5:00PM	RO PLANT	RO WATER	NA
6/24/2023	9:00PM	MAJOR OT	CAUTERY MACHINE	ALAN
6/24/2023	10:00AM	RICU	MULTIPARA MONITOR	EDAN
6/24/2023	11:00AM	DIALYSIS	RO WATER	NA
6/24/2023	12:00PM	MICU	BIPAP	PHILIPS
6/24/2023	10:00AM	SICU	MULTIPARA MONITOR	MINDRAY
6/25/2023	11:00AM	MICROBIOLOGY	BACTERIA CULTURE	BIOMERIEUX
6/25/2023	12:00PM	ICCU	VENTILATOR	SIEMENS
6/25/2023	3:00PM	ED	ECG MACHINE	EDAN
6/25/2023	5:00PM	RICU	MONITOR	L&T
6/25/2023	10:00AM	CVTS ICU	ABG MACHINE	GEM
6/25/2023	11:00AM	CENTRAL LAB	AUTO HEMATOLOGY ANAL	HORIBA
6/25/2023	12:00PM	MRI	MRI WORK STATION	GE
6/25/2023	1:00PM	OBG WARD	SLIT LAMP	NA
6/26/2023	2:00PM	PAY WARD	PORTABEL SUCTION	NA
6/26/2023	2:30PM	PAY WARD	O2 FLOW METER	NA
6/26/2023	4:00PM	RADIOLOGY	PRINTER	AGFA
6/26/2023	5:00PM	MICROBIOLOGY LAB	CENTRIFUGE	REMI
6/26/2023	2:00AM	USG ROOM	USG MACHINE	GE
6/26/2023	5:00AM	RO PLANT	RO WATER	NA
6/26/2023	7:00AM	RADIOLOGY	CR 35	AGFA
6/27/2023	10:30AM	RADIOLOGY	CR 35	AGFA
6/27/2023	1:00PM	RO PLANT	RO WATER	NA
6/27/2023	11:30AM	MICROBIOLOGY	CENTRIFUGE	REMI
6/27/2023	10:15AM	RADIOLOGY	CR 35	AGFA
6/27/2023	11:15PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
6/27/2023	10:00AM	ED	ABG MACHINE	GEM
6/27/2023	9:30AM	FMW	ECG MACHINE	EDAN
6/28/2023	12:10PM	OPHTHALMOLOGY	SLIT LAMP	APPASAMY
6/28/2023	4:15PM	OPHTHALMOLOGY	OCT MACHINE	APPASAMY ASSOC

6/28/2023	9:45AM	ED	DEFIBRILLATOR	PHILIPS
6/28/2023	9:30AM	CVTS ICU	VENTILATOR	NEUMOVENT
6/28/2023	9:00AM	CVTS ICU	VENTILATOR	IIT KANPUR
6/28/2023	4:20PM	RADIOLOGY	CR30	AGFA
6/28/2023	4:00PM	ED	SYRINGE PUMP	B BRAUN
6/28/2023	3:20PM	ED	ABG MACHINE	GEM
3/29/2024	2:30PM	ECHO ROOM	ECG MACHINE	GE
3/29/2024	1:42PM	RO PLANT	RO WATER	NA
3/29/2024	1:15PM	RADIOLOGY	CR30	AGFA
3/29/2024	1:10PM	W&C COVID	DIGITAL BP	SCURE
3/29/2024	9:30AM	ICCU	SYRINGE PUMP	B BRAUN
3/29/2024	9:00PM	ENT OT	ENT VIDEO SCOPE	NASIO
3/29/2024	9:00AM	EMERGENCY DEPART	ECG MACHINE	EDAN
3/29/2024	8:00PM	ENT OT	MICROSCOPE	CARL ZEISS
3/30/2024	7:00PM	ICCU	PORTABLE XRAY	SKANRAY
3/30/2024	6:00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
3/30/2024	5:15PM	EMERGENCY DEPART	INFUSION PUMP	B BRAUN
3/30/2024	5:00PM	CSSD	ETO STERILIZER	3M
3/30/2024	4:00PM	MICU	PORTABLE XRAY	SKANRAY
3/30/2024	3:10PM	DIALYSIS	BP APPARATUS	ELKOMETER
3/30/2024	12:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
3/30/2024	11:30PM	CVTS ICU	MONITOR	MINDRAY
4/1/2024	11:00PM	CVTS ICU	SYRINGE PUMP	B BRAUN
4/1/2024	10:0AM	COH2	PULSE OXYMETER	CHOICE MED
4/1/2024	10:0AM	COH2	DIGITAL BP	SCURE
4/1/2024	10:00PM	ENT OT	MICROSCOPE	CARL ZEISS
4/1/2024	8:40PM	NMICU	MONITOR	MINDRAY
4/1/2024	6:55PM	RO PLANT	RO WATER	
4/1/2024	5:00AM	ICCU	SYRINGE PUMP	SCHILLER
4/1/2024	2:00AM	NMICU	SYRINGE PUMP	B BRAUN
4/3/2024	12:30PM	DIALYSIS	MULTIPARA MONITOR	AMIGO
4/3/2024	12:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
4/3/2024	11:00AM	MICU	BIPAP	PHILIPS
4/3/2024	10:00AM	CVTS ICU	VENTILATOR	NEUMOVENT
4/3/2024	1:00PM	NEWROWARD	BP APPARATUS	NA
4/3/2024	9:00AM	CVTS OT	OT TABLE	ACCURA
4/3/2024	7:35PM	SICU	MONITOR	MINDRAY
4/3/2024	7:00AM	RO PLANT	RO WATER	NA
4/6/2024	3:00PM	ED	DEFIBRILLATOR	PHILIPS
4/6/2024	3:00PM	RADIOLOGY	X- RAY UNITE	NA
4/6/2024	9:30AM	MOT	LIGHT SOURCE	KARL STORES
4/6/2024	9:00AM	MOT	LAPROSCOPE	STORZ
4/6/2024	8:10PM	RICU	BIPAP	PHILIPS
4/6/2024	7:00PM	ICCU	BIPAP	PHILIPS
4/6/2024	2:00PM	PULMO WARD	BIPAP	NA

4/8/2024	10:30AM	EMERGENCY DEPART	MONITOR	MINDRAY
4/8/2024	1:00PM	MICU	MONITOR	MINDRAY
4/8/2024	12:00PM	RO PLANT	RO WATER	NA
4/8/2024	11:00AM	W&C	VAPOURISER	PENLON
4/8/2024	10:30AM	RADIOLOGY	CR 35	AGFA
4/8/2024	10:00AM	USG DEPARTMENT	USG MACHINE	GE
4/8/2024	9:30AM	USG DEPARTMENT	USG MACHINE	GE
4/11/2024	9:30AM	USG ROOM	USG MACHINE	GE
4/11/2024	10:30AM	CSSD	AUTOCLAVE	JOSHY ENGINEERIN
4/11/2024	1:30PM	W&C CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
4/11/2024	9:30PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
4/11/2024	3:30PM	RO PLANT	RO WATER	NA
4/11/2024	11:30AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
4/11/2024	10:30PM	CSSD	ETO STERILIZER	3M
4/15/2024	3:00PM	RO PLANT	RO WATER	NA
4/15/2024	12:30PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
4/15/2024	11:00AM	RICU	BIPAP	PHILIPS
4/15/2024	10:00AM	ED	ABG MACHINE	GEM
4/15/2024	4:00PM	USG DEPARTMENT	USG MACHINE	GE
4/15/2024	3:50PM	MICU	BIPAP	BME
4/15/2024	12:30PM	BLOOD BANK	REFRIGERATED CENTRIFUG	ELTEK
4/15/2024	11:30AM	ED	ABG MACHINE	GEM
4/18/2024	10:00AM	MICU	X RAY MACHINE (MOBILE-C	SKANRAY
4/18/2024	2:00PM	RO PLANT	RO WATER	NA
4/18/2024	11:00AM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
4/18/2024	10:00PM	ICCU	VENTILATOR	SKANRAY
4/18/2024	6:30AM	RO PLANT	RO WATER	NA
4/18/2024	11:00AM	EMERGENCY DEPART	ECG MACHINE	EDAN
4/18/2024	9:00am	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
4/18/2024	10:00am	nmiCU	coat	janak
4/18/2024	11:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
4/18/2024	12:00PM	W&C PICU	MULTIPARA MONITOR	SCURE
4/21/2024	2:00PM	EMERGENCY DEPART	ABG MACHINE	GEM
4/21/2024	3:00PM	CVTS	COAT	NA
4/21/2024	5:00PM	RO PLANT	RO WATER	NA
4/21/2024	9:00PM	MAJOR OT	CAUTERY MACHINE	ALAN
4/21/2024	10:00AM	RICU	MULTIPARA MONITOR	EDAN
4/21/2024	11:00AM	DIALYSIS	RO WATER	NA
4/21/2024	12:00PM	MICU	BIPAP	PHILIPS
4/21/2024	10:00AM	SICU	MULTIPARA MONITOR	MINDRAY
4/24/2024	11:00AM	MICROBIOLOGY	BACTERIA CULTURE	BIOMERIEUX
4/24/2024	12:00PM	ICCU	VENTILATOR	SIEMENS
4/24/2024	3:00PM	ED	ECG MACHINE	EDAN
4/24/2024	5:00PM	RICU	MONITOR	L&T

4/24/2024	10:00AM	CVTS ICU	ABG MACHINE	GEM
4/24/2024	11:00AM	CENTRAL LAB	AUTO HEMATOLOGY ANAL	HORIBA
4/24/2024	12:00PM	MRI	MRI WORK STATION	GE
4/24/2024	1:00PM	OBG WARD	SLIT LAMP	NA
4/29/2024	4:00PM	RADIOLOGY	PRINTER	AGFA
4/29/2024	5:00PM	MICROBIOLOGY LAB	CENTRIFUGE	REMI
4/29/2024	2:00AM	USG ROOM	USG MACHINE	GE
4/29/2024	5:00AM	RO PLANT	RO WATER	NA
4/29/2024	7:00AM	RADIOLOGY	CR 35	AGFA
4/29/2024	10:30AM	RADIOLOGY	CR 35	AGFA
4/29/2024	1:00PM	RO PLANT	RO WATER	NA
4/29/2024	11:30AM	MICROBIOLOGY	CENTRIFUGE	REMI
4/29/2024	10:15AM	RADIOLOGY	CR 35	AGFA
4/29/2024	11:15PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
4/29/2024	10:00AM	ED	ABG MACHINE	GEM
4/29/2024	9:30AM	FMW	ECG MACHINE	EDAN
4/29/2024	12:10PM	OPHTHALMOLOGY	SLIT LAMP	APPASAMY
4/30/2024	4:15PM	OPHTHALMOLOGY	OCT MACHINE	APPASAMY ASSOC
4/30/2024	9:45AM	ED	DEFIBRILLATOR	PHILIPS
4/30/2024	9:30AM	CVTS ICU	VENTILATOR	NEUMOVENT
4/30/2024	9:00AM	CVTS ICU	VENTILATOR	IIT KANPUR
4/30/2024	4:20PM	RADIOLAGY	CR30	AGFA
4/30/2024	4:00PM	ED	SYRINGE PUMP	B BRAUN
4/30/2024	3:20PM	ED	ABG MACHINE	GEM
4/30/2024	2:30PM	ECHO ROOM	ECG MACHINE	GE
4/30/2024	1:42PM	RO PLANT	RO WATER	NA
4/30/2024	1:15PM	RADIOLOGY	CR30	AGFA
4/30/2024	1:10PM	W&C COVID	DIGITAL BP	SCURE
4/30/2024	9:30AM	ICCU	SYRINGE PUMP	B BRAUN
4/30/2024	9:00PM	ENT OT	ENT VIDEO SCOPE	NASIO
4/30/2024	9:00AM	EMERGENCY DEPART	ECG MACHINE	EDAN
4/30/2024	8:00PM	ENT OT	MICROSCOPE	CARL ZEISS
5/1/2024	7:00PM	ICCU	PORTABLE XRAY	SKANRAY
5/1/2024	6:00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
5/1/2024	5:15PM	EMERGENCY DEPART	INFUSION PUMP	B BRAUN
5/1/2024	5:00PM	CSSD	ETO STERILIZER	3M
5/1/2024	4:00PM	MICU	PORTABLE XRAY	SKANRAY
5/1/2024	3:10PM	DIALYSIS	BP APPARATUS	ELKOMETER
5/1/2024	12:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
5/1/2024	11:30PM	CVTS ICU	MONITOR	MINDRAY
5/3/2024	11:00PM	CVTS ICU	SYRINGE PUMP	B BRAUN
5/3/2024	10:0AM	COH2	PULSE OXYMETER	CHOICE MED
5/3/2024	10:0AM	COH2	DIGITAL BP	SCURE
5/3/2024	10:00PM	ENT OT	MICROSCOPE	CARL ZEISS

5/3/2024	8:40PM	NMICU	MONITOR	MINDRAY
5/3/2024	6:55PM	RO PLANT	RO WATER	
5/3/2024	5:00AM	ICCU	SYRINGE PUMP	SCHILLER
5/3/2024	2:00AM	NMICU	SYRINGE PUMP	B BRAUN
5/5/2024	12:30PM	DIALYSIS	MULTIPARA MONITOR	AMIGO
5/5/2024	12:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
5/5/2024	11:00AM	MICU	BIPAP	PHILIPS
5/5/2024	10:00AM	CVTS ICU	VENTILATOR	NEUMOVENT
5/5/2024	1:00PM	NEWROWARD	BP APPARATUS	NA
5/5/2024	9:00AM	CVTS OT	OT TABLE	ACCURA
5/8/2024	7:35PM	SICU	MONITOR	MINDRAY
5/8/2024	7:00AM	RO PLANT	RO WATER	NA
5/8/2024	3:00PM	ED	DEFIBRILLATOR	PHILIPS
5/8/2024	3:00PM	RADIOLOGY	X- RAY UNITE	NA
5/8/2024	9:30AM	MOT	LIGHT SOURCE	KARL STORES
5/8/2024	9:00AM	MOT	LAPROSCOPE	STORZ
5/10/2024	8:10PM	RICU	BIPAP	PHILIPS
5/10/2024	7:00PM	ICCU	BIPAP	PHILIPS
5/10/2024	2:00PM	PULMO WARD	BIPAP	NA
5/10/2024	10:30AM	EMERGENCY DEPART	MONITOR	MINDRAY
5/10/2024	1:00PM	MICU	MONITOR	MINDRAY
5/13/2024	12:00PM	RO PLANT	RO WATER	NA
5/13/2024	11:00AM	W&C	VAPOURISER	PENLON
5/13/2024	10:30AM	RADIOLOGY	CR 35	AGFA
5/13/2024	10:00AM	USG DEPARTMENT	USG MACHINE	GE
5/13/2024	9:30AM	USG DEPARTMENT	USG MACHINE	GE
5/13/2024	9:30AM	USG ROOM	USG MACHINE	GE
5/13/2024	10:30AM	CSSD	AUTOCLAVE	JOSHY ENGINEERIN
5/15/2024	1:30PM	W&C CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
5/15/2024	9:30PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
5/15/2024	3:30PM	RO PLANT	RO WATER	NA
5/15/2024	11:30AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
5/15/2024	10:30PM	CSSD	ETO STERILIZER	3M
5/15/2024	3:00PM	RO PLANT	RO WATER	NA
5/15/2024	12:30PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
5/17/2024	11:00AM	RICU	BIPAP	PHILIPS
5/17/2024	10:00AM	ED	ABG MACHINE	GEM
5/17/2024	4:00PM	USG DEPARTMENT	USG MACHINE	GE
5/17/2024	3:50PM	MICU	BIPAP	BME
5/17/2024	12:30PM	BLOOD BANK	REFRIGERATED CENTRIFUG	ELTEK
5/17/2024	11:30AM	ED	ABG MACHINE	GEM
5/19/2024	10:00AM	MICU	X RAY MACHINE (MOBILE-C	SKANRAY
5/19/2024	3:00PM	ED	FLOWMETER	NA
5/19/2024	10:00am	USG DEPARTMENT	USG MACHINE	GE
5/19/2024	11:00AM	ECHO ROOM	ECHO MACHINE	VIVID E9

5/19/2024	1:00PM	RO PLANT	RO WATER	NA
5/19/2024	3:00PM	CVTS ICU	SYRINGE PUMP	B BRAUN
5/21/2024	4:00PM	CARDIOLOGY	TMT	GE
5/21/2024	5:00PM	MEDICAL ICU	MULTIPARA MONITOR	EDAN
5/21/2024	9:00AM	ED	DEFIBRILLATOR	PHILIPS
5/21/2024	3:30PM	CENTRAL LAB	CENTRIFUGE	ROTEK
5/21/2024	12:30PM	ED	MONITOR	MINDRAY
5/25/2024	10:30AM	RADIOLOGY	CR 30	C
5/25/2024	10:15AM	CVTS	AUTOCLAVE	NA
5/25/2024	1:00PM	BLOOD BANK	CENTRIFUGE	ROTEK
5/25/2024	9:30AM	ED	X RAY MACHINE (MOBILE-C	SKANRAY
5/25/2024	9:00AM	NSICU	VENTILATOR	NEUMOVENT
5/25/2024	7:30AM	MICU	VENTILATOR	NEUMOVENT
5/25/2024	7:00AM	MOT	LAPROSCOPE	STORZ
5/25/2024	3:30PM	NEURO WARD	SYRINGE PUMP	B BRAUN
5/25/2024	2:30PM	ED	X RAY MACHINE (MOBILE-C	SKANRAY
5/25/2024	2:00PM	CENTRAL LAB	CENTRIFUGE	ROTEK
5/25/2024	2:00PM	ED	MULTIPARA MONITOR	MINDRAY
5/25/2024	12:00PM	PHYSIOTHERAPY	ULTRA SOUND PROBE	TECHNOMED ELEC
5/25/2024	9:30AM	W&C CENTRAL LAB	AUTO HEAMATOLOGY ANA	HEMAX
5/25/2024	4:30pm	ICCU	BIPAP	PHILIPS
5/25/2024	4:00PM	MICU	MULTIPARA MONITOR	MINDRAY
5/28/2024	2:45PM	SICU	VENTILATOR	NEUMOVENT
5/28/2024	12:00PM	RO PLANT	RO WATER	NA
5/28/2024	11:30AM	ICCU	BIPAP	PHILIPS
5/28/2024	11:00AM	ICCU	SYRINGE PUMP	SCHILLER
5/28/2024	10:5AM	W&C OT	ANAESTHESIA	NA
5/28/2024	10:15AM	BLOOD BANK	URINE ANALYSER	SYSMEX
5/28/2024	3:30PM	NMICU	VENTILATOR	SIEMENS
5/29/2024	12:30PM	ICCU	SYRINGE PUMP	B BRAUN
5/29/2024	12:15pm	RO PLANT	RO WATER	NA
5/29/2024	11:15AM	RADIOLOGY	CR 30	C
5/29/2024	11:00AM	CENTRAL LAB	ELECTROLYTE ANALYSER	B&E
5/29/2024	10:30AM	CENTRAL LAB	CENTRIFUGE	ROTEK
5/29/2024	10:00PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
5/29/2024	11:30AM	ICCU	BIPAP	PHILIPS
5/29/2024	11:00AM	ICCU	SYRINGE PUMP	SCHILLER
5/29/2024	10:5AM	W&C OT	ANAESTHESIA	NA
5/29/2024	11:30AM	SP PULMO WARD	NEBULISER	ELKO
5/29/2024	10:15AM	BLOOD BANK	URINE ANALYSER	SYSMEX
5/30/2024	11:00AM	MEDICAL ICU	SYRINGE PUMP	B BRAUN
5/30/2024	1:00PM	CENTRAL LAB	MICROSCOPE	LABOMED
5/30/2024	4:00PM	W&C PICU	PORTABLE SUCTION	SURGI CRAFT
5/30/2024	10:5AM	DIALYSIS	MULTIPARA MONITOR	AMIGO
5/30/2024	10:15AM	CVTS ICU	VENTILATOR	NEUMOVENT

5/30/2024	10:00PM	MICU	BIPAP	PHILIPS
5/30/2024	11:00AM	CVTS ICU	VENTILATOR	NEUMOVENT
5/31/2024	1:00PM	NEWROWARD	BP APPARATUS	NA
5/31/2024	4:00PM	CVTS OT	OT TABLE	ACCURA
5/31/2024	10:00AM	RO PLANT	RO WATER	NA
5/31/2024	11:00AM	CVTS OT	SEALINHG MACHINE	NA
5/31/2024	12:00PM	W&C OT	QUOTARY MACHINE	CAL-SURG
5/31/2024	1:00PM	PULMO WARD	BIPAP	RESMED
5/31/2024	2:00PM	RICU	VENTILATOR	NEUMOVENT
5/31/2024	4:00PM	AYURVEDA COVID HOS	CENTRAL FLOW METER	NA
2/1/2024	10:00AM	ED	PORTABLE VENTILLATOR	PHILIPS
2/1/2024	11:00AM	RADIOLOGY	CR	AGFA
2/1/2024	12:00PM	NSICU	VENTILATOR	NEUMOVENT
2/1/2024	10:00AM	BME	DOCUMENTATION	
2/1/2024	2:00PM	CVTS ICU	ABG MACHINE	GEM
2/1/2024	4:00PM	RADIOLOGY	CR	AGFA
2/1/2024	5:00PM	SICU	MONITOR	SKANRAY
2/1/2024	10:00AM	RO UNIT	RO WATER	NA
8/2/2023	9:30AM	RICU	BIPAP	PHILIPS
8/2/2023	10:30AM	SICU	VENTILATOR	SKANRAY
8/2/2023	1:30PM	CENTRAL LAB	ELECTROLYTE ANALYSER	B&E
8/2/2023	9:30PM	ED	MONITOR	BLUESTAR
8/2/2023	11:30AM	ED	DEFIBRILLATOR	PHILIPS
8/2/2023	10:30PM	PULMO WARD	GLUCOMETER	ALEREGI
8/2/2023	3:00PM	PULMO WARD	BIPAP	PHILIPS
8/3/2023	12:30PM	DIALYSIS	MONITOR	AMIGO
8/3/2023	11:00AM	MICU	BIPAP	RESMED
8/3/2023	10:00AM	CVTS ICU	VENTILATOR	NEUMOVENT
8/3/2023	3:50PM	CENTRAL LAB	BIOCHEMISTRY ANALYSER	RANDOX
8/3/2023	12:30PM	NSICU	ECG MACHINE	SKANRAY
8/3/2023	11:30AM	MOT	ETCO2 MODULE	
8/3/2023	10:00AM	RADIOLOGY	CR35X	AGFA
8/3/2023	3:00PM	EMERGENCY DEPART	MULTIPARAMONITOR	AMIGO
8/3/2023	2:00PM	CENTRAL LAB	ELECTROLYTE ANALYSER	B&E
8/3/2023	11:00AM	SICU	VENTILATOR	SKANRAY
8/4/2023	11:00AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/4/2023	12:00PM	MICU	VENTILATOR	NA
8/4/2023	9:00AM	OT	OT DRILL	NA
8/4/2023	11:00AM	CSSD	SEALING MACHINE	SEPACK
8/4/2023	1:00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/4/2023	3:00PM	NEW BLOCK MICU	ECG MACHINE	NA
8/4/2023	11:00AM	OPHTHALMOLOGY	SLIT LAMP	APPASAMY

8/5/2023	12:00PM	NMICU	PATIENT BED	NA
8/5/2023	1:00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/5/2023	2:00PM	RO PLANT	RO WATER	NA
8/5/2023	9:156PM	EMERGENCY DEPART	MULTIPARA MONITOR	MINDRAY
8/5/2023	8:00AM	RO PLANT	RO WATER	NA
8/6/2023	9:00AM	ENT WARD	PORTABLE SUCTION	G I
8/6/2023	9:20AM	CENTRAL LAB	CENTRIFUGE	KEMI
8/6/2023	9:45AM	ICCU	GLUCOMETER	ON CALL PLUS
8/6/2023	10:30AM	MALE MEDICINE WA	BP APPARATUS	LIFE LINE
8/6/2023	11:00AM	NEURO SURGERY ICU	MULTIPARA MONITOR	L&T
8/6/2023	11:30AM	CSSD	SEALER	SEPACK
8/6/2023	10:00AM	RICU	VENTILATOR	NEUMOVENT
8/6/2023	10:45AM	BLOOD BANK	URINE ANALYSER	SYSMEX
8/6/2023	12:00PM	EMERGENCY DEPART	ECG MACHINE	EDAN
8/7/2023	1:00PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
8/7/2023	4:00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/7/2023	8/7/2023	RO PLANT	RO WATER	NA
8/7/2023	10:00AM	MEDICAL ICU	MULTIPARA MONITOR	MINDRAY
8/7/2023	12:30PM	RO PLANT	RO WATER	NA
8/7/2023	10:00AM	NSICU	MULTIPARA MONITOR	L&T
8/7/2023	12:00PM	EMERGENCY DEPARTM	FLOWMETER	NA
8/7/2023	5:00PM	MEDICAL ICU	VENTILATOR	SKANRAY
8/8/2023	7:00AM	RO PLANT	RO WATER	NA
8/8/2023	1:00PM	MICU	PATIENT MONITOR	EDAN
8/8/2023	10:30AM	OPHTHALMOLOGY	SLIT LAMP	WHELCH ALLYN
8/8/2023	10:00AM	CVTS ICU	ABG MACHINE	ROCHE
8/8/2023	1:00PM	ECG ROOM	ECG MACHINE	EDAN
8/8/2023	9:30AM	RICU	BIPAP	BMC
8/8/2023	7:30PM	CENTRAL LAB	HORMON ANALYSER	BECKMAN COULTER
8/8/2023	6:00PM	RO PLANT	RO WATER	NA
8/8/2023	3:50PM	NICU	PHOTO THERAPY	I BISS
8/8/2023	3:30PM	W&C COVID	WARMER	I BISS
8/9/2023	3:00PM	NSICU	INFUSION PUMP	B BRAUN
8/9/2023	2:20PM	CENTRAL LAB	ELECTROLYTE ANALYSER	CORNLEY
8/9/2023	2:00PM	CENTRAL LAB	BIOCHEMISTRY ANALYSER	RANDOX
8/9/2023	2:00PM	CSSD	ETO STERILIZER	3M
8/9/2023	12:30PM	GASTOLOGY	APC	SHALYA
8/9/2023	12:00PM	BRONCHO SCOPY	SUCTION	NA
8/9/2023	11:30PM	NSOT	NEURO MICROSCOPE	CARL ZEISS
8/9/2023	8:00AM	CATH LAB	UPS	EATON
8/10/2023	7:00AM	RO PLANT	RO WATER	NA
8/10/2023	3:00PM	NEURO SURGERY ICU	INFUSION PUMP	B BRAUN
8/10/2023	2:00PM	ICCU	SYRINGE PUMP	SCHILLER

8/10/2023	2:00AM	ED	ECG MACHINE	EDAN
8/10/2023	12:30AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANAL	RANDOX
8/10/2023	11:20PM	RO PLANT	RO WATER	NA
8/10/2023	10:00AM	EMERGENCY DEPART	ECG MACHINE	EDAN
8/11/2023	1:30AM	CVTS ICU	ABG MACHINE	GEM
8/11/2023	1:20PM	ICCU	SYRINGE PUMP	SCHILLER
8/11/2023	8:00PM	ED	ECG MACHINE	EDAN
8/11/2023	7:00PM	NSICU	ECG MACHINE	SKANRAY
8/11/2023	12:00PM	RO PLANT	RO WATER	NA
8/11/2023	1:00PM	NEURO SURGERY ICU	VENTILATOR	NEUMOVENT
8/11/2023	9:00AM	RO PLANT	RO WATER	NA
8/11/2023	8:00PM	CENTRAL LAB	AUTO HEMATOLOGY ANAL	HORIBA
8/12/2023	6:00PM	RO PLANT	RO WATER	NA
8/12/2023	3:00PM	NMICU	ECG MACHINE	NIDEX
8/12/2023	12.30PM	W&C OT	LAPROSCOPE	STORZ
8/12/2023	1.00PM	EMERGENCY DEPARTM	ECG MACHINE	EDAN
8/12/2023	2.30PM	CENTRAL LAB	HORMON ANALYSER	BECKMAN COULTER
8/12/2023	4.30PM	CENTRAL LAB	CENTRIFUGE	ROTEK
8/12/2023	2.00pm	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/12/2023	3.00PM	RICU	MULTIPARA MONITOR	AMIGO
8/13/2023	11.30AM	MICROBIOLOGY	MICROSCOPE	LABOMED
8/13/2023	9.30AM	RO PLANT	RO WATER	NA
8/13/2023	9.00AM	SICU	VENTILATOR	SKANRAY
8/13/2023	7.00PM	W&C LAB	HEMATOLOGY ANALYZER	HEMAX
8/13/2023	9.25PM	EMERGENCY DEPARTM	ABG MACHINE	GEM
8/13/2023	10.20pm	EMERGENCY DEPARTM	LARYNGOSCOPE	NA
8/13/2023	8.40PM	RICU	VENTILATOR	SKANRAY
8/14/2023	9.40pm	RADIO DIAGNOSIS	CR	AGFA
8/14/2023	11.40pm	PULMO WARD	BIPAP	PHILIPS
8/14/2023	4.40PM	RO PLANT	RO WATER	NA
8/14/2023	8.00PM	NSICU	MULTIPARA MONITOR	L&T
8/14/2023	10.00PM	CSSD	ETO STERILIZER	3M
8/14/2023	8.00PM	PAYWARD	SYRINGE PUMP	B BRAUN
8/14/2023	10.00PM	EMERGENCY DEPARTM	ECG MACHINE	EDAN
8/15/2023	10.15PM	RO PLANT	RO WATER	NA
8/15/2023	10.55am	DIALYSIS	DIALYSIS MACHINE	NIPRO
8/15/2023	11.30AM	RO PLANT	RO WATER	NA
8/15/2023	10:00AM	CVTS ICU	VENTILATOR	SCHILLER
8/15/2023	11:00AM	MAJOR OT	CAUTERY MACHINE	ALAN
8/15/2023	9:00AM	RO PLANT	RO WATER	NA
8/15/2023	10:40AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/15/2023	1:10PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
8/16/2023	11:00AM	PAYWARD	SYRINGE PUMP	B BRAUN
8/16/2023	2:00PM	RO PLANT	RO WATER	NA
8/16/2023	11:00AM	RICU	SYRINGE PUMP	OMYA

8/16/2023	12:00PM	MICROBIOLOGY	BACTERIA CULTURE	BIOMERIEUX
8/16/2023	10.55am	RADIO DIAGNOSIS	CR	AGFA
8/16/2023	11.00am	RO PLANT	RO WATER	NA
8/16/2023	9.25am	MICU	MULTIPARA MONITOR	MINDRAY
8/17/2023	1.10pm	RO PLANT	RO WATER	NA
8/17/2023	8.50AM	ICCU	INFUSSION PUMP	AKAS
8/17/2023	1.00PM	W&C OT	OT WALL	Na
8/17/2023	10.0AM	PICU	HFNC	fisher & paykel
8/17/2023	2.00PM	RICU	VENTILATOR	SKANRAY
8/17/2023	9.50PM	RO PLANT	RO WATER	NA
8/17/2023	11.00AM	RADIO DIAGNOSIS	CR	AGFA
8/18/2023	8.00AM	CENTRAL LAB	CENTRIFUGE	ROTEK
8/18/2023	9.00AM	RICU	VENTILATOR	SKANRAY
8/18/2023	10.00am	EMERGENCY DEPARTM	ABG MACHINE	GEM
8/18/2023	11.00am	w&c lab	HEMATOLOGY ANALYZER	HEMAX
8/18/2023	8.00am	USG	USG MACHINE	GE
8/18/2023	11.30am	RICU	VENTILATOR	SKANRAY
8/18/2023	10.00am	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/18/2023	11.30am	CENTRAL LAB	HEMATOLOGY ANALYZER	HORIBA
8/19/2023	12.00pm	EMERGENCY DEPARTM	ABG MACHINE	GEM
8/19/2023	4.30pm	RICU	VENTILATOR	SKANRAY
8/19/2023	12.55pm	CENTRAL LAB	HORMON ANALYSER	BECKMAN COULTE
8/19/2023	9.00PM	RICU	VENTILATOR	SKANRAY
8/19/2023	10.00AM	MICU	VENTILATOR	SKANRAY
8/19/2023	11.30AM	W&C OT	CAUTERY MACHINE	L&T
8/19/2023	12.00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/19/2023	12.30PM	ED	FLOWMETER	NA
8/19/2023	2.30PM	NSOT	OT LIGHT	ACCURA
8/20/2023	10.00AM	MICU	COT	NA
8/20/2023	10.30AM	RO PLANT	RO WATER	NA
8/20/2023	11.00AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/20/2023	12.10PM	RICU	VENTILATOR	SKANRAY
8/20/2023	1.15PM	RADIODIAGNOSIS	CR PRINTER	AGFA
8/20/2023	5.30PM	MICU	VENTILATOR	SKANRAY
8/20/2023	6.0PM	NICU	MONITOR	MINDRAY
8/21/2023	7.00PM	RICU	BIPAP	philips
8/21/2023	8.15PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/21/2023	10.00PM	CENTRAL LAB	URINE ANALYSER	SYSMEX
8/21/2023	9.45AM	NMICU	MONITOR	MINDRAY
8/21/2023	11.05AM	W&C LAB	HEMATOLOGY ANALYZER	HEMAX
8/21/2023	12.30PM	MICROBIOLOGY	AUTOCLAVE	OSWORLD
8/21/2023	1.45PM	ICCU	DEFIBRILLATOR	PHILIPS
8/21/2023	2.45PM	BLOOD BANK	CENTRIFUGE	KEMI
8/22/2023	4.00PM	RADIODIAGNOSIS	MAMMOGRAPHY MACHINE	GE
8/22/2023	9.30AM	OPHTHALMOLOGY	SLIT LAMP	APPASAMY

8/22/2023	6.30AM	CENTRAL LAB	HORMON ANALYSER	BECKMAN COULTER
8/22/2023	7.45AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/22/2023	8.30AM	W&C OT	LARYNGOSCOPIC UNIT	KARL STORZ
8/22/2023	11.00AM	RO PLANT	RO WATER	NA
8/23/2023	11.15AM	EMERGENCY DEPARTMENT	ABG MACHINE	GEM
8/23/2023	7:40AM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
8/23/2023	6:50AM	NSICU	VENTILATOR	NUEMOVENT
8/23/2023	6:00PM	MICU	VENTILATOR	SKANRAY
8/23/2023	11:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANALY	RANDOX
8/23/2023	10:35PM	PAYWARD	NEBULISER	NA
8/23/2023	10:10PM	CENTRAL LAB	FULLY BIOCHEMISTRY ANALY	RANDOX
8/23/2023	8:00AM	CVTS ICU	LARYNGOSCOPE (ADL)	NA
8/23/2023	6:30PM	W&C OT	FUMIGATION MACHINE	NA
8/24/2023	5:20PM	CSSD	ETO STERILIZER	3M
8/24/2023	4:00PM	W&C OT	FUMIGATION MACHINE	NA
8/24/2023	2:15AM	DIALYSIS	BP APPARATUS	ELKOMETER
8/24/2023	8:55AM	CVTS ICU	ABG MACHINE	ROCHE
8/24/2023	8:45AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/24/2023	8:20AM	USG	USG MACHINE	GE
8/25/2023	6:30AM	RO PLANT	RO WATER	NA
8/25/2023	5:50AM	DIALYSIS	RO PLANT	NA
8/25/2023	1:40AM	CVTS ICU	VENTILATOR	NUEMOVENT
8/25/2023	9:00PM	CENTRAL LAB	ELECTROLYTE ANALYSER	CORNLEY
8/25/2023	9:00AM	PICU	MONITOR,FLOW METER,SUCT	SCURE
8/25/2023	8:30PM	MICU	MULTIPARA MONITOR	AMIGO
8/25/2023	8:00PM	COVID HOSPITAL	ECG MACHINE	EDAN
8/26/2023	10:00AM	ICCU	SYRINGE PUMP	SCHILLER
8/26/2023	12:00PM	CENTRAL LAB	CENTRIFUGE	ROTEK
8/26/2023	10:00AM	MAJOR OT	LAPROSCOPE	STORZ
8/26/2023	11:00PM	EMERGENCY DEPARTMENT	DEFIBRILLATOR	PHILIPS
8/26/2023	12:00PM	OPHTHALMOLOGY	OCT MACHINE	ZEISS
8/26/2023	1:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
8/26/2023	2:00PM	CENTRAL LAB	FULLY BIOCHEMISTRY ANALY	RANDOX
8/27/2023	3:00PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
8/27/2023	5:00PM	CATH LAB	UPS	EATON
8/27/2023	11:00AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/27/2023	12:00PM	RO PLANT	RO WATER	NA
8/27/2023	12:00PM	W&C OT	PATIENT PLATE	NA
8/27/2023	7:00PM	RICU	MONITOR	AMIGO
8/27/2023	10:00PM	CHEST&TB WARD	BIPAP	RESMED
8/27/2023	11:00AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
8/28/2023	12:00PM	RO PLANT	RO WATER	NA
8/28/2023	12:00PM	W&C OT	PATIENT PLATE	NA
8/28/2023	7:00PM	RICU	MONITOR	AMIGO
8/28/2023	10:00PM	CHEST&TB WARD	BIPAP	RESMED

8/28/2023	11:00AM	CVTS OT	AUTOCLAVE	NA
8/28/2023	7:00AM	RO PLANT	RO WATER	NA
8/28/2023	9:30AM	DIALYSIS	BP APPARATUS	ELKOMETER
8/28/2023	9:00AM	RADIOLOGY	X RAY MACHINE (MOBILE-C	SKANRAY
8/29/2023	8:00PM	ED	ABG MACHINE	GEM
8/29/2023	7:20PM	CSSD	ETO STERILIZER	3M
8/29/2023	7:00PM	ED	VENTILATOR	SKANRAY
8/29/2023	6:30PM	RICU	MULTIPARAMONITOR	AMIGO
8/29/2023	10:00 AM	CVTS OT	OT LIGHT	ACCURA
8/29/2023	9:30 AM	CENTRAL LAB	URINE ANALYSER	SYSMEX
8/29/2023	9:10PM	ED	VENTILATOR	SKANRAY
8/30/2023	8:30PM	ED	VENTILATOR	PHILIPS
8/30/2023	8:30AM	ED	GLUCOMETER	NO CODING ONE P
8/30/2023	7:45AM	CENTRAL LAB	BIOCHEMISTRY ANALYSER	RANDOX
8/30/2023	7:00PM	MOT	FOGGING MACHINE	FUGOSTAR
8/30/2023	9:00am	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
8/30/2023	10:00am	nmiCU	coat	janak
8/30/2023	11:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
8/30/2023	12:00PM	W&C PICU	MULTIPARA MONITOR	SCURE
9/1/2023	2:00PM	EMERGENCY DEPART	ABG MACHINE	GEM
9/1/2023	3:00PM	CVTS	COAT	NA
9/1/2023	9:30AM	CENTRAL LAB	ELECTROLYTE ANALYSER	CORNLEY
9/1/2023	9:00AM	W&C OT	CAUTERY MACHINE	L&T
9/1/2023	11:30AM	RO PLANT	RO WATER	NA
9/1/2023	12:45PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
9/1/2023	02:00PM	MEDICAL ICU	SYRINGE PUMP	B BRAUN
9/1/2023	11:45AM	CENTRAL LAB	ELECTROLYTE ANALYSER	B&E
9/1/2023	10:30AM	SICU	MONITOR	MINDRAY
9/2/2023	1:45PM	W&C LAB	AUTO HEMATOLOGY ANAL	HEMAX
9/2/2023	1:15PM	RICU	VENTILATOR	NEUMOVENT
9/2/2023	9:30AM	NMICU	INFUSION PUMP	B BRAUN
9/2/2023	9:00AM	RADIOLOGY	PRINTER	AGFA
9/2/2023	3:00PM	CENTRAL LAB	HORMON ANALYSER	BECKMAN COULTER
9/2/2023	2:15PM	SICU	VENTILATOR	NEUMOVENT
9/2/2023	2:00PM	NMICU	VENTILATOR	SIEMENS
9/2/2023	11:00AM	USG ROOM	USG MACHINE	GE
9/2/2023	1:00PM	CENTRAL LAB	ESR MACHINE	YHLO
9/3/2023	9:30AM	SICU	MONITOR	MINDRAY
9/3/2023	9:15AM	MICU	DEFIBRILLATOR	HP
9/3/2023	3:35PM	W&C OT	COLPOSCOPE	BPL
9/3/2023	10:45AM	ED	ECG MACHINE	EDAN
9/3/2023	7:30AM	RO PLANT	RO WATER	NA
9/3/2023	6:00PM	ED	ABG MACHINE	GEM
9/3/2023	3:50PM	W&C OT	MULTIPARA MONITOR	SCURE

9/4/2023	2:45PM	CVTS OT	HEART LUNG MACHINE	SORIN STOKERT
9/4/2023	3:00PM	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
9/4/2023	11:30AM	SUCTION	NA	NA
9/4/2023	10:00AM	MOT	LAPROSCOPE	KARLSTORZ
9/4/2023	9:30PM	CVTS ICU	VENTILATOR	NEUMOVENT
9/4/2023	8:00PM	DIALYSIS	UPS	CONSOL NEOWATT
9/4/2023	6:30PM	RO PLANT	RO PLANT	NA
9/4/2023	12:30PM	LABOUR ROOM	NST MACHINE	MS TECHNOLOGIES
9/5/2023	11:00AM	SICU	MULTIPARA MONITOR	MINDRY
9/5/2023	6:00 PM	RADIOLOGY	CR	AGFA
9/5/2023	10:30 AM	ED	THERMOMETER	NA
9/5/2023	9:30 AM	RICU	MULTIPARA MONITOR	EDAN
9/5/2023	9:20	MICU	VENTILATOR	NEUMOVENT
9/5/2023	7:30AM	LABOUR ROOM	CTG MACHIBE	EDAN
9/5/2023	7:00AM	MICU	BIPAP	RESMED
9/5/2023	7:00AM	RICU	MULTIPARA MONITOR	EDAN
9/5/2023	6:15AM	LABOUR ROOM	CTG MACHIBE	MS TECHNOLOGIES
9/6/2023	2:15PM	SURGERY WARD	AIR VALVE	NA
9/6/2023	12:24PM	CENTRAL LAB	HEAMATOLOGY ANALYSER	HORIBA
9/6/2023	12:15PM	CENTRAL LAB	HEAMATOLOGY ANALYSER	HORIBA
9/6/2023	11:10AM	CENTRAL LAB	HORMONE ANALYSER	BECKMAN COULTER
9/6/2023	1:35PM	ED	ABG MACHINE	GEM
9/6/2023	1:10PM	MICU	MULTIPARA MONITOR	MINDRAY
9/6/2023	9:50PM	W&C OT	UPS	SOCOMEK INNOVAT
9/6/2023	9:25AM	MICU	ECG MACHINE	EDAN
9/6/2023	8:20PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
9/7/2023	8:15PM	RO PLANT	RO PLANT	NA
9/7/2023	8:15PM	SICU	ECG MACHINE	EDAN
9/7/2023	7:25PM	RICU	BIPAP	NA
9/7/2023	7:15PM	ED	ABG MACHINE	GEM
9/7/2023	7:00PM	ED	ABG MACHINE	GEM
9/7/2023	6:00PM	LABOUR ROOM	CTG MACHINE	EDAN
9/7/2023	6:55PM	RICU	VENTILATOR	SKANRAY
9/8/2023	5:05PM	RO PLANT	RO PLANT	NA
9/8/2023	3:10PM	MICU	MULTIPARA MONITOR	MINDRAY
9/8/2023	2:00PM	ED	VENTILATOR	SKANRAY
9/8/2023	12:00PM	RO PLANT	RO PLANT	NA
9/8/2023	11:45PM	OFW	BP APPARATUS	ELKOMETER
9/8/2023	10:50AM	MICU	MULTIPARA MONITOR	MINDRAY
9/8/2023	10:30PM	CSSD	AUTOCLAVE	SURGITECH
9/8/2023	10:10AM	ED	INFUSION PUMP	B BRAUN
9/8/2023	1:30PM	ED	MULTIPARA MONITOR	MINDRAY
9/9/2023	5:10 PM	RO PLANT	RO PLANT	NA
9/9/2023	12:05 PM	CSSD	AUTOCLAVE	SURGITECH

9/9/2023	10:50AM	CENTRAL LAB	CENTRIFUGE	ROTEK
9/9/2023	10:40AM	MICU	BP APPARATUS	ELKOMETER
9/9/2023	10:30AM	RADIOLOGY	READER	AGFA
9/9/2023	10:00AM	PEDIATRIC WARD	BP APPARATUS	ELKOMETER
9/10/2023	10:00AM	RICU	VENTILATOR	SKANRAY
9/10/2023	1:00PM	EMERGENCY OT	CAUTERY MACHINE	CALL SURGE
9/10/2023	9:30AM	RO PLANT	NA	NA
9/10/2023	8:50AM	RADIOLOGY	X RAY	ALLENGERS
9/10/2023	5:00PM	RO PLANT	RO WATER	NA
9/10/2023	3:50PM	DIALYSIS	UPS	CONSUL NEOWAT
9/10/2023	12:00PM	RICU	MONITOR	AMIGO
9/10/2023	11:00AM	CVTS ICU	VENTILATOR	SKANRAY
9/11/2023	10:30PM	ED	VENTILATOR	SKANRAY
9/11/2023	10:30AM	MICU	BIPAP	RESMED
9/11/2023	10:15AM	CENTRAL LAB	ESR MACHINE	YHLO
9/11/2023	10:00AM	RO PLANT	RO WATER	NA
9/11/2023	1:00PM	RO PLANT	RO WATER	NA
9/11/2023	1:00AM	ICCU	BIPAP	PHILIPS
9/11/2023	9:00AM	BLOOD BANK	BLOOD COLLECTION MONITOR	GENESYS
9/11/2023	6:30AM	CENTRAL LAB	ESR MACHINE	YHLO
9/11/2023	6:00PM	RO PLANT	RO WATER	NA
9/12/2023	4:00PM	ED	ABG MACHINE	GEM
9/12/2023	12:30PM	RICU	VENTILATOR	SKANRAY
9/12/2023	11:00AM	RO PLANT	RO WATER	NA
9/12/2023	11:00AM	W&C OT	CAUTERY MACHINE	SKANRAY
9/12/2023	10:00AM	W&C OT	CAUTERY MACHINE	SKANRAY
9/12/2023	10:00AM	W&C OT	OT LIGHT	SKANRAY
9/12/2023	1:00PM	ED	ABG MACHINE	GEM
9/12/2023	1:00PM	MOT	UPS	CONSUL NEOWAT
9/12/2023	1:00AM	FMW	BIPAP	BMC
9/13/2023	8:30PM	ICCU	VENTILATOR	SIEMENS
9/13/2023	8:15PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
9/13/2023	8:15PM	RO PLANT	RO WATER	NA
9/13/2023	7:15PM	ED	DEFIBRILLATOR	PHILIPS
9/13/2023	9:00AM	RADIOLOGY	DR	AGFA
9/13/2023	6:00PM	CSSD	ETO STERILIZER	3M
9/13/2023	3:35PM	ED	MONITOR	MINDRAY
9/13/2023	10:00PM	CVTS	VENTILATOR	NEUMOVENT
9/13/2023	1:15PM	ED	INFUSION PUMP	B BRAUN
9/13/2023	8:00PM	MICU	ECG MACHINE	EDAN
9/14/2023	3:00PM	CVTS	COAT	NA
9/14/2023	9:30AM	CENTRAL LAB	ELECTROLYTE ANALYSER	CORNLEY
9/14/2023	9:00AM	W&C OT	CAUTERY MACHINE	L&T
9/14/2023	11:30AM	RO PLANT	RO WATER	NA
9/14/2023	12:45PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS

9/14/2023	02:00PM	MEDICAL ICU	SYRINGE PUMP	B BRAUN
9/14/2023	11:45AM	CENTRAL LAB	ELECTROLYTE ANALYSER	B&E
9/14/2023	10:30AM	SICU	MONITOR	MINDRAY
9/14/2023	1:45PM	W&C LAB	AUTO HEMATOLOGY ANAL	HEMAX
9/14/2023	1:15PM	RICU	VENTILATOR	NEUMOVENT
15--09-2023	9:30AM	NMICU	INFUSION PUMP	B BRAUN
15--09-2023	9:00AM	RADIOLOGY	PRINTER	AGFA
15--09-2023	3:00PM	CENTRAL LAB	HORMON ANALYSER	BECKMAN COULTER
15--09-2023	2:15PM	SICU	VENTILATOR	NEUMOVENT
15--09-2023	2:00PM	NMICU	VENTILATOR	SIEMENS
15--09-2023	11:00AM	USG ROOM	USG MACHINE	GE
15--09-2023	1:00PM	CENTRAL LAB	ESR MACHINE	YHLO
15--09-2023	3:50PM	SPOT	MULTIPARA MONITOR	BIOLIGHT
15--09-2023	9:30AM	SICU	MONITOR	MINDRAY
15--09-2023	9:15AM	MICU	DEFIBRILLATOR	HP
16--09-2023	3:35PM	W&C OT	COLPOSCOPE	BPL
16--09-2023	10:45AM	ED	ECG MACHINE	EDAN
16--09-2023	7:30AM	RO PLANT	RO WATER	NA
16--09-2023	6:00PM	ED	ABG MACHINE	GEM
16--09-2023	3:50PM	W&C OT	MULTIPARA MONITOR	SCURE
16--09-2023	2:45PM	CVTS OT	HEART LUNG MACHINE	SORIN STOKERT
9/17/2023	3:00PM	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
9/17/2023	11:30AM	SUCTION	NA	NA
9/17/2023	10:00AM	MOT	LAPROSCOPE	KARLSTORZ
9/17/2023	1:00PM	RICU	MULTIPARAMONITOR	AMIGO
9/17/2023	10:00AM	7th Floor PAY WARD	ECG MACHINE	BPL
9/17/2023	5:30PM	ED	ABG MACHINE	GEM
9/17/2023	6:45PM	RICU	BIPAP	PHILIPS
9/17/2023	8:00PM	RADIOLOGY	XRAY MACHINE	ALLENGERS
9/17/2023	8:00AM	USG	USG MACHINE	GE
9/18/2023	6:10PM	ED	PORTABLE XRAY	SKANRAY
9/18/2023	8:20PM	NSICU	MULTIPARA MONITOR	L&T
9/18/2023	7:00PM	RICU	VENTILATOR	SKANRAY
9/18/2023	10:00AM	ICCU	ECG MACHINE	GE
9/18/2023	6:00PM	NSICU	ECG MACHINE	SKANRAY
9/18/2023	7:45PM	ED	ABG MACHINE	GEM
9/18/2023	9:00PM	NSOT	DRILL	NA
9/18/2023	10:00PM	ED	DEFIBRILLATOR	PHILIPS
9/19/2023	07:00PM	MOT	OT LIGHT	NA
9/19/2023	9:30PM	CVTS	SEALING MACHINE	NA
9/18/2023	10:00PM	SPRICU	DEFIBRILLATOR	NIHON KOHDEN
9/19/2023	11:0PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
9/19/2023	12:00PM	ED	BIPAP	BME
9/19/2023	6:00PM	ICCU	VENTILATOR	SIEMENS

9/19/2023	8:00AM	ICCU	VENTILATOR	SIEMENS
9/19/2023	6:00AM	RO PLANT	RO WATER	NA
9/19/2023	4:00PM	RO PLANT	RO WATER	NA
9/20/2023	3:00PM	ED	ABG MACHINE	GEM
9/20/2023	10:30AM	RADIOLOGY	CR	AGFA
9/20/2023	10:0AM	NSICU	VENTILATOR	NEUMOVENT
9/20/2023	1:00PM	RADIOLOGY	CR	AGFA
9/20/2023	9:30PM	CVTS ICU	ABG MACHINE	GEM
9/20/2023	11:30AM	SP PULMO WARD	NEBULISER	ELKO
9/20/2023	9:00AM	RADIOLOGY	CR	AGFA
9/20/2023	02:30PM	CVTS ICU	ABG MACHINE	GEM
9/20/2023	2:00PM	USG ROOM	USG MACHINE	GE
9/21/2023	9:00am	RADIO DIAGNOSIS	X RAY MACHINE (MOBILE-6	SKANRAY
9/21/2023	11:15AM	RADIOLOGY	CR	AGFA
9/21/2023	1:50PM	EMERGENCY DEPARTM	ABG MACHINE	GEM
9/21/2023	6:20PM	EMERGENCY DEPARTM	LARYNGOSCOPE (ADL)	NA
9/21/2023	9:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
9/21/2023	10:00AM	NMICU	VENTILATOR	SIEMENS
9/21/2023	10:30AM	RADIOLOGY	CT MACHINE	WIPRO GE
9/21/2023	10:00AM	SPMICU	VENTILATOR	HAMILTON
9/21/2023	10:30AM	VITALS PEDIATRICS	WEINGING MACHINE	EAGLE
9/21/2023	11:00AM	RADIOLOGY	MRI MACHINE	WIPRO GE
9/22/2023	12:00PM	MAJOR OT	CAUTRY PATIENT PLATE CAB	NA
9/22/2023	1:00PM	SICU	PATIENT MONITOR	MINDRAY
9/22/2023	1:30PM	MAJOR OT	CAUTRY FOOT SWITCH	NA
9/22/2023	3:00PM	RO PLANT	RO WATER	NA
9/22/2023	3:30PM	W&C OT	CAUTRY PATIENT PLATE CAB	NA
9/22/2023	09:00AM	RO PLANT	NA	NA
9/22/2023	1:30PM	SPOT	LAPROSCOPE	NA
9/22/2023	09:00AM	MICU	MULTIPARA MONITOR	EDAN
9/22/2023	10:20AM	LABOUR ROOM	USG MACHINE	SAMSUNG
9/22/2023	09:00AM	SPMICU	MULTIPARA MONITOR	BIOLIGHT
9/23/2023	11:00AM	NSICU	MULTIPARA MONITOR	L&T
9/23/2023	12:00PM	LABOUR ROOM	CTG MACHINE	MS TECHNOLOGIE
9/23/2023	03:00PM	SICU	SYRINGE PUMP	L&T
9/23/2023	9:00AM	RO PLANT	RO WATER	NA
9/23/2023	10:40AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
9/23/2023	1:10PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
9/23/2023	11:00AM	PAYWARD	SYRINGE PUMP	B BRAUN
9/23/2023	2:00PM	RO PLANT	RO WATER	NA
9/24/2023	11:00AM	RICU	SYRINGE PUMP	OMYA
9/24/2023	12:00PM	MICROBIOLOGY	BACTERIA CULTURE	BIOMERIEUX
9/24/2023	10.55am	RADIO DIAGNOSIS	CR	AGFA
9/24/2023	10.55am	RADIO DIAGNOSIS	DR	AGFA
9/24/2023	11.00am	RO PLANT	RO WATER	NA

9/24/2023	9.25am	MICU	MULTIPARA MONITOR	MINDRAY
9/24/2023	1.10pm	RO PLANT	RO WATER	NA
9/24/2023	8.50AM	ICCU	INFUSSION PUMP	AKAS
9/25/2023	1.00PM	W&C OT	OT WALL	Na
9/25/2023	10.0AM	PICU	HFNC	fisher & paykel
9/25/2023	2.00PM	RICU	VENTILATOR	SKANRAY
9/25/2023	9.50PM	RO PLANT	RO WATER	NA
9/25/2023	11.00AM	RADIO DIAGNOSIS	CR	AGFA
9/25/2023	8.00AM	CENTRAL LAB	CENTRIFUGE	ROTEK
9/25/2023	9.00AM	RICU	VENTILATOR	SKANRAY
9/25/2023	10.00am	EMERGENCY DEPARTM	ABG MACHINE	GEM
9/26/2023	9:45AM	SPOT	MULTIPARA MONITOR	BIOLIGHT
9/26/2023	1:30AM	SPOT	OT TABLE	SURGITECH
9/26/2023	1:45 PM	NEW BLOCK PEDI VITALS	WEINGING MACHINE	EAGLE
9/26/2023	10:20AM	OPHTHAL OT	BIPOLAR COAGULATOR	MICRO SURGICAL
9/26/2023	11:00AM	3RD MOT	MULTIPARA MONITOR	BIOLIGHT
9/26/2023	12:03PM	SICU	DEFIBRILLATOR	NIHON KOHDEN
9/26/2023	2:00PM	CSSD	ETO STERILIZER	3M
9/26/2023	6:00pm	ED	BP APPARATUS	ELKOMETER
9/26/2023	9:00PM	XRAY CR	CASSETTE READER	AGFA
9/26/2023	8:00AM	SPOT	OT TABLE	STAAN
9/27/2023	11:30 AM	SPOT	SHAVER S/M	STRYKER
9/27/2023	9:45 AM	RICU	VENTILATOR	SKANRAY
9/27/2023	12:00 PM	NEW BLOCK	DR	AGFA
9/27/2023	7:00PM	SICU	MONITOR	SKANRAY
9/27/2023	10:00PM	ICCU	VENTILATOR	SKANRAY
9/27/2023	2:00AM	SURGERY WARD	ECG MACHINE	EDAN
9/28/2023	7:00AM	CENTRAL LAB	SEMI BIOCHEMISTRY	ERBA
9/28/2023	8:00AM	CVTS ICU	VENTILATOR	SCHILLER
9/28/2023	10:00AM	NMICU	VENTILATOR	SKANRAY
10/1/2023	10:00AM	ED	PORTABLE VENTILLATOR	PHILIPS
10/1/2023	11:00AM	RADIOLOGY	CR	AGFA
10/1/2023	12:00PM	NSICU	VENTILATOR	NEUMOVENT
10/1/2023	10:00AM	BME	DOCUMENTATION	
10/1/2023	2:00PM	CVTS ICU	ABG MACHINE	GEM
10/1/2023	4:00PM	RADIOLOGY	CR	AGFA
10/1/2023	5:00PM	SICU	MONITOR	SKANRAY
10/1/2023	10:00AM	RO UNIT	RO WATER	NA
10/2/2023	9:30AM	RICU	BIPAP	PHILIPS
10/2/2023	10:30AM	SICU	VENTILATOR	SKANRAY
10/2/2023	1:30PM	CENTRAL LAB	ELECTROLYTE ANALYSER	B&E
10/2/2023	9:30PM	ED	MONITOR	BLUESTAR

10/2/2023	11:30AM	ED	DEFIBRILLATOR	PHILIPS
10/2/2023	10:30PM	PULMO WARD	GLUCOMETER	ALEREGL
10/2/2023	3:00PM	PULMO WARD	BIPAP	PHILIPS
10/3/2023	12:30PM	DIALYSIS	MONITOR	AMIGO
10/3/2023	11:00AM	MICU	BIPAP	RESMED
10/3/2023	10:00AM	CVTS ICU	VENTILATOR	NEUMOVENT
10/3/2023	3:50PM	CENTRAL LAB	BIOCHEMISTRY ANALYSER	RANDOX
10/3/2023	12:30PM	NSICU	ECG MACHINE	SKANRAY
10/3/2023	11:30AM	MOT	ETCO2 MODULE	
10/3/2023	10:00AM	RADIOLOGY	CR35X	AGFA
10/3/2023	3:00PM	EMERGENCY DEPART	MULTIPARAMONITOR	AMIGO
10/3/2023	2:00PM	CENTRAL LAB	ELECTROLYTE ANALYSER	B&E
10/3/2023	11:00AM	SICU	VENTILATOR	SKANRAY
10/4/2023	11:00AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/4/2023	12:00PM	MICU	VENTILATOR	NA
10/4/2023	9:00AM	OT	OT DRILL	NA
10/4/2023	11:00AM	CSSD	SEALING MACHINE	SEPACK
10/4/2023	1:00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/4/2023	3:00PM	NEW BLOCK MICU	ECG MACHINE	NA
10/4/2023	11:00AM	OPHTHALMOLOGY	SLIT LAMP	APPASAMY
10/5/2023	12:00PM	NMICU	PATIENT BED	NA
10/5/2023	1:00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/5/2023	2:00PM	RO PLANT	RO WATER	NA
10/5/2023	9:156PM	EMERGENCY DEPART	MULTIPARA MONITOR	MINDRAY
10/5/2023	8:00AM	RO PLANT	RO WATER	NA
10/6/2023	9:00AM	ENT WARD	PORTABLE SUCTION	G I
10/6/2023	9:20AM	CENTRAL LAB	CENTRIFUGE	KEMI
10/6/2023	9:45AM	ICCU	GLUCOMETER	ON CALL PLUS
10/6/2023	10:30AM	MALE MEDICINE WA	BP APPARATUS	LIFE LINE
10/6/2023	11:00AM	NEURO SURGERY ICU	MULTIPARA MONITOR	L&T
10/6/2023	11:30AM	CSSD	SEALER	SEPACK
10/6/2023	10:00AM	RICU	VENTILATOR	NEUMOVENT
10/6/2023	10:45AM	BLOOD BANK	URINE ANALYSER	SYSMEX
10/6/2023	12:00PM	EMERGENCY DEPART	ECG MACHINE	EDAN
10/7/2023	1:00PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
10/7/2023	4:00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/7/2023	5:00PM	RO PLANT	RO WATER	NA
10/7/2023	10:00AM	MEDICAL ICU	MULTIPARA MONITOR	MINDRAY
10/7/2023	12:30PM	RO PLANT	RO WATER	NA
10/7/2023	10:00AM	NSICU	MULTIPARA MONITOR	L&T
10/7/2023	12:00PM	EMERGENCY DEPARTM	FLOWMETER	NA

10/7/2023	5:00PM	MEDICAL ICU	VENTILATOR	SKANRAY
10/8/2023	7:00AM	RO PLANT	RO WATER	NA
10/8/2023	1:00PM	MICU	PATIENT MONITOR	EDAN
10/8/2023	10:30AM	OPHTHALMOLOGY	SLIT LAMP	WHELCH ALLYN
10/8/2023	10:00AM	CVTS ICU	ABG MACHINE	ROCHE
10/8/2023	1:00PM	ECG ROOM	ECG MACHINE	EDAN
10/8/2023	9:30AM	RICU	BIPAP	BMC
10/8/2023	7:30PM	CENTRAL LAB	HORMON ANALYSER	BECKMAN COULTER
10/8/2023	6:00PM	RO PLANT	RO WATER	NA
10/8/2023	3:50PM	NICU	PHOTO THERAPY	I BISS
10/8/2023	3:30PM	W&C COVID	WARMER	I BISS
10/9/2023	3:00PM	NSICU	INFUSION PUMP	B BRAUN
10/9/2023	2:20PM	CENTRAL LAB	ELECTROLYTE ANALYSER	CORNLEY
10/9/2023	2:00PM	CENTRAL LAB	BIOCHEMISTRY ANALYSER	RANDOX
10/9/2023	2:00PM	CSSD	ETO STERILIZER	3M
10/9/2023	12:30PM	GASTOLOGY	APC	SHALYA
10/9/2023	12:00PM	BRONCHO SCOPY	SUCTION	NA
10/9/2023	11:30PM	NSOT	NEURO MICROSCOPE	CARL ZEISS
10/9/2023	8:00AM	CATH LAB	UPS	EATON
10/10/2023	7:00AM	RO PLANT	RO WATER	NA
10/10/2023	3:00PM	NEURO SURGERY ICU	INFUSION PUMP	B BRAUN
10/10/2023	2:00PM	ICCU	SYRINGE PUMP	SCHILLER
10/10/2023	2:00AM	ED	ECG MACHINE	EDAN
10/10/2023	12:30AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANAL	RANDOX
10/10/2023	11:20PM	RO PLANT	RO WATER	NA
10/10/2023	10:00AM	EMERGENCY DEPART	ECG MACHINE	EDAN
10/11/2023	1:30AM	CVTS ICU	ABG MACHINE	GEM
10/11/2023	1:20PM	ICCU	SYRINGE PUMP	SCHILLER
10/11/2023	8:00PM	ED	ECG MACHINE	EDAN
10/11/2023	7:00PM	NSICU	ECG MACHINE	SKANRAY
10/11/2023	12:00PM	RO PLANT	RO WATER	NA
10/11/2023	1:00PM	NEURO SURGERY ICU	VENTILATOR	NEUMOVENT
10/11/2023	9:00AM	RO PLANT	RO WATER	NA
10/11/2023	8:00PM	CENTRAL LAB	AUTO HEMATOLOGY ANAL	HORIBA
10/12/2023	6:00PM	RO PLANT	RO WATER	NA
10/12/2023	3:00PM	NMICU	ECG MACHINE	NIDEK
10/12/2023	12.30PM	W&C OT	LAPROSCOPE	STORZ
10/12/2023	1.00PM	EMERGENCY DEPARTM	ECG MACHINE	EDAN
10/12/2023	2.30PM	CENTRAL LAB	HORMON ANALYSER	BECKMAN COULTER
10/12/2023	4.30PM	CENTRAL LAB	CENTRIFUGE	ROTEK
10/12/2023	2.00pm	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/12/2023	3.00PM	RICU	MULTIPARA MONITOR	AMIGO
10/13/2023	11.30AM	MICROBIOLOGY	MICROSCOPE	LABOMED
10/13/2023	9.30AM	RO PLANT	RO WATER	NA
10/13/2023	9.00AM	SICU	VENTILATOR	SKANRAY

10/13/2023	7.00PM	W&C LAB	HEMATOLOGY ANALYZER	HEMAX
10/13/2023	9.25PM	EMERGENCY DEPARTM	ABG MACHINE	GEM
10/13/2023	10.20pm	EMERGENCY DEPARTM	LARYNGOSCOPE	NA
10/13/2023	8.40PM	RICU	VENTILATOR	SKANRAY
10/14/2023	9.40pm	RADIO DIAGNOSIS	CR	AGFA
10/14/2023	11.40pm	PULMO WARD	BIPAP	PHILIPS
10/14/2023	4.40PM	RO PLANT	RO WATER	NA
10/14/2023	8.00PM	NSICU	MULTIPARA MONITOR	L&T
10/14/2023	10.00PM	CSSD	ETO STERILIZER	3M
10/14/2023	8.00PM	PAYWARD	SYRINGE PUMP	B BRAUN
10/14/2023	10.00PM	EMERGENCY DEPARTM	ECG MACHINE	EDAN
10/15/2023	10.15PM	RO PLANT	RO WATER	NA
10/15/2023	10.55am	DIALYSIS	DIALYSIS MACHINE	NIPRO
10/15/2023	11.30AM	RO PLANT	RO WATER	NA
10/15/2023	10:00AM	CVTS ICU	VENTILATOR	SCHILLER
10/15/2023	11:00AM	MAJOR OT	CAUTERY MACHINE	ALAN
10/15/2023	9:00AM	RO PLANT	RO WATER	NA
10/15/2023	10:40AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/15/2023	1:10PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
10/16/2023	11:00AM	PAYWARD	SYRINGE PUMP	B BRAUN
10/16/2023	2:00PM	RO PLANT	RO WATER	NA
10/16/2023	11:00AM	RICU	SYRINGE PUMP	OMYA
10/16/2023	12:00PM	MICROBIOLOGY	BACTERIA CULTURE	BIOMERIEUX
10/16/2023	10.55am	RADIO DIAGNOSIS	CR	AGFA
10/16/2023	11.00am	RO PLANT	RO WATER	NA
10/16/2023	9.25am	MICU	MULTIPARA MONITOR	MINDRAY
10/17/2023	1.10pm	RO PLANT	RO WATER	NA
10/17/2023	8.50AM	ICCU	INFUSSION PUMP	AKAS
10/17/2023	1.00PM	W&C OT	OT WALL	Na
10/17/2023	10.0AM	PICU	HFNC	fisher & paykel
10/17/2023	2.00PM	RICU	VENTILATOR	SKANRAY
10/17/2023	9.50PM	RO PLANT	RO WATER	NA
10/17/2023	11.00AM	RADIO DIAGNOSIS	CR	AGFA
10/18/2023	8.00AM	CENTRAL LAB	CENTRIFUGE	ROTEK
10/18/2023	9.00AM	RICU	VENTILATOR	SKANRAY
10/18/2023	10.00am	EMERGENCY DEPARTM	ABG MACHINE	GEM
10/18/2023	11.00am	w&c lab	HEMATOLOGY ANALYZER	HEMAX
10/18/2023	8.00am	USG	USG MACHINE	GE
10/18/2023	11.30am	RICU	VENTILATOR	SKANRAY
10/18/2023	10.00am	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/18/2023	11.30am	CENTRAL LAB	HEMATOLOGY ANALYZER	HORIBA
10/19/2023	12.00pm	EMERGENCY DEPARTM	ABG MACHINE	GEM
10/19/2023	4.30pm	RICU	VENTILATOR	SKANRAY
10/19/2023	12.55pm	CENTRAL LAB	HORMON ANALYSER	BECKMAN COULTER
10/19/2023	9.00PM	RICU	VENTILATOR	SKANRAY

10/19/2023	10.00AM	MICU	VENTILATOR	SKANRAY
10/19/2023	11.30AM	W&C OT	CAUTERY MACHINE	L&T
10/19/2023	12.00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/19/2023	12.30PM	ED	FLOWMETER	NA
10/19/2023	2.30PM	NSOT	OT LIGHT	ACCURA
10/20/2023	10.00AM	MICU	COT	NA
10/20/2023	10.30AM	RO PLANT	RO WATER	NA
10/20/2023	11.00AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/20/2023	12.10PM	RICU	VENTILATOR	SKANRAY
10/20/2023	1.15PM	RADIODIAGNOSIS	CR PRINTER	AGFA
10/20/2023	5.30PM	MICU	VENTILATOR	SKANRAY
10/20/2023	6.0PM	NICU	MONITOR	MINDRAY
10/21/2023	7.00PM	RICU	BIPAP	philips
10/21/2023	8.15PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/21/2023	10.00PM	CENTRAL LAB	URINE ANALYSER	SYSMEX
10/21/2023	9.45AM	NMICU	MONITOR	MINDRAY
10/21/2023	11.05AM	W&C LAB	HEMATOLOGY ANALYZER	HEMAX
10/21/2023	12.30PM	MICROBIOLOGY	AUTOClave	OSWORLD
10/21/2023	1.45PM	ICCU	DEFIBRILLATOR	PHILIPS
10/21/2023	2.45PM	BLOOD BANK	CENTRIFUGE	KEMI
10/22/2023	4.00PM	RADIODIAGNOSIS	MAMMOGRAPHY MACHINE	GE
10/22/2023	9.30AM	OPHTHALMOLOGY	SLIT LAMP	APPASAMY
10/22/2023	6.30AM	CENTRAL LAB	HORMON ANALYSER	BECKMAN COULTER
10/22/2023	7.45AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/22/2023	8.30AM	W&C OT	LARYNGOSCOPIC UNIT	KARL STORZ
10/22/2023	11.00AM	RO PLANT	RO WATER	NA
10/23/2023	11.15AM	EMERGENCY DEPARTMENT	ABG MACHINE	GEM
10/23/2023	7:40AM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
10/23/2023	6:50AM	NSICU	VENTILATOR	NUEMOVENT
10/23/2023	6:00PM	MICU	VENTILATOR	SKANRAY
10/23/2023	11:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANALY	RANDOX
10/23/2023	10:35PM	PAYWARD	NEBULISER	NA
10/23/2023	10:10PM	CENTRAL LAB	FULLY BIOCHEMISTRY ANALY	RANDOX
10/23/2023	8:00AM	CVTS ICU	LARYNGOSCOPE (ADL)	NA
10/23/2023	6:30PM	W&C OT	FUMIGATION MACHINE	NA
10/24/2023	5:20PM	CSSD	ETO STERILIZER	3M
10/24/2023	4:00PM	W&C OT	FUMIGATION MACHINE	NA
10/24/2023	2:15AM	DIALYSIS	BP APPARATUS	ELKOMETER
10/24/2023	8:55AM	CVTS ICU	ABG MACHINE	ROCHE
10/24/2023	8:45AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/24/2023	8:20AM	USG	USG MACHINE	GE
10/25/2023	6:30AM	RO PLANT	RO WATER	NA
10/25/2023	5:50AM	DIALYSIS	RO PLANT	NA
10/25/2023	1:40AM	CVTS ICU	VENTILATOR	NUEMOVENT
10/25/2023	9:00PM	CENTRAL LAB	ELECTROLYTE ANALYSER	CORNLEY

10/25/2023	9:00AM	PICU	MONITOR, FLOW METER, SUC	SCURE
10/25/2023	8:30PM	MICU	MULTIPARA MONITOR	AMIGO
10/25/2023	8:00PM	COVID HOSPITAL	ECG MACHINE	EDAN
10/26/2023	10:00AM	ICCU	SYRINGE PUMP	SCHILLER
10/26/2023	12:00PM	CENTRAL LAB	CENTRIFUGE	ROTEK
10/26/2023	10:00AM	MAJOR OT	LAPROSCOPE	STORZ
10/26/2023	11:00PM	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
10/26/2023	12:00PM	OPHTHALMOLOGY	OCT MACHINE	ZEISS
10/26/2023	1:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
10/26/2023	2:00PM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
10/27/2023	3:00PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
10/27/2023	5:00PM	CATH LAB	UPS	EATON
10/27/2023	11:00AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/27/2023	12:00PM	RO PLANT	RO WATER	NA
10/27/2023	12:00PM	W&C OT	PATIENT PLATE	NA
10/27/2023	7:00PM	RICU	MONITOR	AMIGO
10/27/2023	10:00PM	CHEST&TB WARD	BIPAP	RESMED
10/27/2023	11:00AM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
10/28/2023	12:00PM	RO PLANT	RO WATER	NA
10/28/2023	12:00PM	W&C OT	PATIENT PLATE	NA
10/28/2023	7:00PM	RICU	MONITOR	AMIGO
10/28/2023	10:00PM	CHEST&TB WARD	BIPAP	RESMED
10/28/2023	11:00AM	CVTS OT	AUTOCLAVE	NA
10/28/2023	7:00AM	RO PLANT	RO WATER	NA
10/28/2023	9:30AM	DIALYSIS	BP APPARATUS	ELKOMETER
10/28/2023	9:00AM	RADIOLOGY	X RAY MACHINE (MOBILE-6	SKANRAY
10/29/2023	8:00PM	ED	ABG MACHINE	GEM
10/29/2023	7:20PM	CSSD	ETO STERILIZER	3M
10/29/2023	7:00PM	ED	VENTILATOR	SKANRAY
10/29/2023	6:30PM	RICU	MULTIPARAMONITOR	AMIGO
10/29/2023	10:00 AM	CVTS OT	OT LIGHT	ACCURA
10/29/2023	9:30 AM	CENTRAL LAB	URINE ANALYSER	SYSMEX
10/29/2023	9:10PM	ED	VENTILATOR	SKANRAY
10/30/2023	8:30PM	ED	VENTILATOR	PHILIPS
10/30/2023	8:30AM	ED	GLUCOMETER	NO CODING ONE P
10/30/2023	7:45AM	CENTRAL LAB	BIOCHEMISTRY ANALYSER	RANDOX
10/30/2023	7:00PM	MOT	FOGGING MACHINE	FUGOSTAR
10/30/2023	9:00am	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
10/30/2023	10:00am	nmiCU	coat	janak
10/30/2023	11:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
10/30/2023	12:00PM	W&C PICU	MULTIPARA MONITOR	SCURE
11/1/2023	9:30AM	W&C OBG WARD	CTG MACHINE	BPL
11/1/2023	9:00AM	W&C NICU	VENTILATOR	DRAGGER
11/1/2023	4:20PM	ICCU	VENTILATOR	SKANRAY
11/1/2023	3:15PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA

11/1/2023	3:00PM	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
11/2/2023	2:30PM	RADIO DIAGNOSIS	MAMMOGRAPHY MACHINE	GE
11/2/2023	2:00PM	RO PLANT	RO WATER	NA
11/2/2023	12:30PM	DERMATOLOGY OT	CENTRIFUGE	KEMI
11/2/2023	12:00PM	NMICU	BIPAP	PHILIPS
11/3/2023	11:00AM	MICU	VENTILATOR	NEUMOVENT
11/3/2023	9:45AM	RICU	MULTIPARA MONITOR	SKANRAY
11/3/2023	9:40PM	W&C COVID WARD	VENTILATOR	SKANRAY
11/3/2023	9:30AM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
11/3/2023	9:00AM	ICCU	VENTILATOR	NEUMOVENT
11/4/2023	8:50PM	CSSD	AUTOCLAVE MACHINE	SURGITECH
11/4/2023	8:15PM	ICCU	MULTIPARA MONITOR	SCHILLER
11/4/2023	7:30PM	ED	BIPAP	BME
11/5/2023	6:00PM	ICCU	VENTILATOR	SIEMENS
11/5/2023	8:00AM	ICCU	VENTILATOR	SIEMENS
11/5/2023	6:00AM	RO PLANT	RO WATER	NA
11/5/2023	4:00PM	RO PLANT	RO WATER	NA
11/5/2023	3:00PM	ED	ABG MACHINE	GEM
11/5/2023	10:30AM	RADIOLOGY	CR	AGFA
11/5/2023	10:0AM	NSICU	VENTILATOR	NEUMOVENT
11/5/2023	1:00PM	RADIOLOGY	CR	AGFA
11/5/2023	9:30PM	CVTS ICU	ABG MACHINE	GEM
11/5/2023	9:00AM	RADIOLOGY	CR	AGFA
11/5/2023	02:30PM	CVTS ICU	ABG MACHINE	GEM
11/6/2023	2:00PM	USG ROOM	USG MACHINE	GE
11/6/2023	9:00am	RADIO DIAGNOSIS	X RAY MACHINE (MOBILE-C	SKANRAY
11/6/2023	11:15AM	RADIOLOGY	CR	AGFA
11/7/2023	1:50PM	EMERGENCY DEPARTM	ABG MACHINE	GEM
11/7/2023	6:20PM	EMERGENCY DEPARTM	LARYNGOSCOPE (ADL)	NA
11/7/2023	9:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
11/7/2023	10:00AM	NMICU	VENTILATOR	SIEMENS
11/8/2023	10:30AM	RADIOLOGY	CT MACHINE	WIPRO GE
11/8/2023	11:00AM	RADIOLOGY	MRI MACHINE	WIPRO GE
11/8/2023	12:00PM	MAJOR OT	CAUTRY PATIENT PLATE CAB	NA
11/8/2023	1:00PM	SICU	PATIENT MONITOR	MINDRAY
11/8/2023	1:30PM	MAJOR OT	CAUTRY FOOT SWITCH	NA
11/9/2023	3:00PM	RO PLANT	RO WATER	NA
11/9/2023	3:30PM	W&C OT	CAUTRY PATIENT PLATE CAB	NA
11/9/2023	09:00AM	RO PLANT	NA	NA
11/9/2023	09:00AM	MICU	MULTIPARA MONITOR	EDAN
11/9/2023	10:20AM	LABOUR ROOM	USG MACHINE	SAMSUNG
11/9/2023	11:00AM	NSICU	MULTIPARA MONITOR	L&T
4/24/2024	12:00PM	LABOUR ROOM	CTG MACHINE	MS TECHNOLOGIE
4/24/2024	03:00PM	SICU	SYRINGE PUMP	L&T
4/24/2024	4:15PM	FMW	ECG MACHINE	EDAN

4/24/2024	09:00AM	EMERGENCY DEPARTM	ABG MACHINE	GEM
4/24/2024	10:30AM	MICU	ECG MACHINE	EDAN
4/24/2024	01:30PM	SICU	MULTIPARA MONITOR	MINDRAY
4/24/2024	03:45PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
4/24/2024	09:00AM	MOT	C ARM	SKANRAY
4/24/2024	10:30AM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
4/24/2024	11:15AM	URO ICU	MULTIPARAMONITOR	AMIGO
4/24/2024	01:30PM	ECHO	CARDIOLOGY	ECHO MACHINE
4/29/2024	03:45PM	NSICU	MULTIPARA MONITOR	AMIGO
4/29/2024	04:15PM	LABOUR ROOM	USG MACHINE	SAMSUNG
4/29/2024	04:45PM	RO PLANT	RO WATER	NA
4/29/2024	09:00AM	NMICU	MULTIPARA MONITOR	MINDRAY
4/29/2024	12:00PM	MICU	VENTILATOR	
4/29/2024	02:45PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
4/29/2024	06:00PM	RICU	BIPAP	PHILIPS
4/29/2024	09:00AM	MOT	UPS	CONSUL NEOWAT
4/29/2024	10:05AM	MOT	BP APPARATUS	ELKOMETER
11/13/2023	11:12AM	ED	ABG MACHINE	GEM
11/13/2023	12:50PM	CVTS ICU	ABG MACHINE	GEM
11/13/2023	09:15AM	RICU	GLUCOMETER	GLUCOCARD
11/13/2023	10:10AM	LABOUR ROOM	O2 FLOWMETER	NA
11/14/2023	11:20AM	MOT	CAUTRY MACHINE	ALAN
11/14/2023	01:50PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
11/14/2023	02:45PM	NSICU	MULTIPARA MONITOR	AMIGO
11/14/2023	03:45PM	W&C OT	WARMER	PRAGATHI
11/14/2023	09:20AM	ED	BIPAP	PHILIPS
11/14/2023	10:20PM	MEDICINE OP	BIO THESIOMETER	DIABATIC FOOT CA
11/5/2023	12:35PM	ED	ABG MACHINE	GEM
11/5/2023	02:50PM	ED	ABG MACHINE	GEM
11/5/2023	04:00PM	CSSD	ETO STERILIZER	3M
11/5/2023	09:30AM	W&C COVID WARD	FLOW METER	NA
11/5/2023	10:20PM	ICCU	VENTILATOR	SIEMENS
11/5/2023	11:35AM	RO PLANT	RO WATER	NA
11/5/2023	01:40PM	CSSD	AUTOCLAVE MACHINE	SURGITECH
11/5/2023	03:00PM	NSICU	ECG MACHINE	SKANRAY
11/16/2023	04:20PM	MOT	PORTABLE SUCTION	NA
11/16/2023	09:15PM	NSICU	VENTILATOR	NEUMOVENT
11/16/2023	10:20PM	RADIOLOGY	CR	AGFA
11/16/2023	11:30PM	CVTS ICU	ABG MACHINE	GEM
11/16/2023	02:10PM	RADIOLOGY	CR	AGFA
11/16/2023	03:20PM	SICU	MONITOR	SKANRAY
11/16/2023	06:00PM	RICU	VENTILATOR	SKANRAY
11/17/2023	09:30AM	RO PLANT	NA	NA
11/17/2023	10:20AM	CENTRAL LAB	UPS	NA
11/17/2023	11:30AM	EMERGENCY DEPARTM	DEFIBRILLATOR	PHILIPS

11/17/2023	02:20PM	EMERGENCY DEPARTM	ECG MACHINE	EDAN
11/17/2023	03:50PM	NMICU	SIEMENS	SERVO VENTILATO
11/17/2023	04:55PM	RO PLANT	NA	NA
11/17/2023	7:10PM	EMERGENCY DEPARTM	ECG MACHINE	EDAN
11/30/2023	09:10AM	SURGERY WARD	SUCTION	NA
11/30/2023	10:15AM	CSSD	SEALING MACHINE	SEPACK
11/30/2023	11:30AM	RICU	DEFIBRILLATOR	HP
11/30/2023	12:30PM	EMERGENCY DEPARTM	X RAY MACHINE (MOBILE 6	SKANRAY
11/30/2023	01:45PM	RO PLANT	NA	NA
11/30/2023	02:30PM	MOT	OT LIGHT	HOSPEDIA
11/30/2023	03:40PM	NICU	WARMER	PRAGATHI
12/1/2023	04:50PM	CVTS	VENTILATOR	NEUMOVENT
12/1/2023	09:50AM	USG ROOM	USG MACHINE	GE
12/1/2023	11:30AM	W AND C OT	OT TABLE	HOSPITECH
12/1/2023	02:20PM	W&C SICU	ECG MACHINE	AMIGO
12/1/2023	10:10AM	RO PLANT	RO WATER	NA
12/1/2023	11:20AM	NMICU	MONITOR	MINDRAY
12/1/2023	01:50PM	ICCU	SYRINGE PUMP	B BRAUN
12/1/2023	02:45PM	ED	ECG MACHINE	EDAN
5/1/2024	03:45PM	CSSD	SEALINHG MACHINE	SEPACK
5/1/2024	09:20AM	SICU	VENTILATOR	SKANRAY
5/1/2024	10:20PM	ICCU	SYRINGE PUMP	SCHILLER
5/1/2024	12:35PM	SURGERY WARD	VACCUM	NA
5/1/2024	02:50PM	W&C SICU	VENTILATOR	SIEMENS
5/1/2024	04:00PM	RO UNIT	RO WATER	NA
5/1/2024	09:30AM	RICU	BIPAP	PHILIPS
5/1/2024	10:20PM	SICU	VENTILATOR	SKANRAY
5/3/2024	11:35AM	CENTRAL LAB	ELECTROLYTE ANALYSER	B&E
5/3/2024	01:40PM	ED	MONITOR	BLUESTAR
5/3/2024	03:00PM	ED	FLOW METER	NA
5/3/2024	04:20PM	ED	DEFIBRILLATOR	PHILIPS
5/3/2024	09:15PM	PULMO WARD	GLUCOMETER	ALEREGI
5/3/2024	10:20PM	PULMO WARD	BIPAP	PHILIPS
5/3/2024	11:30PM	DIALYSIS	MONITOR	AMIGO
5/3/2024	02:10PM	MOT	LIGHT SOURCE	STORZ
5/5/2024	03:20PM	RO PLANT	RO WATER	NA
5/5/2024	06:00PM	CVTS OT	SEALINHG MACHINE	NA
5/5/2024	09:30AM	W&C OT	CAUOTARY MACHINE	CAL-SURG
5/5/2024	10:20AM	PULMO WARD	BIPAP	RESMED
5/5/2024	11:30AM	RICU	VENTILATOR	NEUMOVENT
5/5/2024	01:00PM	AYURVEDA COVID HOS	CENTRAL FLOW METER	NA
11/23/2023	02:20PM	ED	PORTABLE VENTILLATOR	PHILIPS
11/23/2023	3:00pm	ED	ECG MACHINE	EDAN
11/23/2023	55:0pm	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
11/23/2023	5:30pm	RO PLANT	RO WATER	NA

11/23/2023	9:00am	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
11/23/2023	10:00am	nmiCU	coat	janak
11/23/2023	11:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
11/23/2023	12:00PM	W&C PICU	MULTIPARA MONITOR	SCURE
5/8/2024	2:00PM	EMERGENCY DEPART	ABG MACHINE	GEM
5/8/2024	3:00PM	CVTS	COAT	NA
5/8/2024	5:00PM	RO PLANT	RO WATER	NA
5/8/2024	9:00PM	MAJOR OT	CAUTERY MACHINE	ALAN
5/8/2024	10:00AM	RICU	MULTIPARA MONITOR	EDAN
5/8/2024	11:00AM	DIALYSIS	RO WATER	NA
5/8/2024	12:00PM	MICU	BIPAP	PHILIPS
5/8/2024	10:00AM	SICU	MULTIPARA MONITOR	MINDRAY
11/25/2023	11:00AM	MICROBIOLOGY	BACTERIA CULTURE	BIOMERIEUX
11/25/2023	12:00PM	ICCU	VENTILATOR	SIEMENS
11/25/2023	3:00PM	ED	ECG MACHINE	EDAN
11/25/2023	5:00PM	RICU	MONITOR	L&T
11/25/2023	10:00AM	CVTS ICU	ABG MACHINE	GEM
11/25/2023	11:00AM	CENTRAL LAB	AUTO HEMATOLOGY ANAL	HORIBA
11/25/2023	12:00PM	MRI	MRI WORK STATION	GE
11/25/2023	1:00PM	OBG WARD	SLIT LAMP	NA
5/10/2024	2:00PM	PAY WARD	PORTABEL SUCTION	NA
5/10/2024	2:30PM	PAY WARD	O2 FLOW METER	NA
5/10/2024	4:00PM	RADIOLOGY	PRINTER	AGFA
5/10/2024	5:00PM	MICROBIOLOGY LAB	CENTRIFUGE	REMI
5/10/2024	2:00AM	USG ROOM	USG MACHINE	GE
5/10/2024	5:00AM	RO PLANT	RO WATER	NA
5/10/2024	7:00AM	RADIOLOGY	CR 35	AGFA
11/27/2023	10:30AM	RADIOLOGY	CR 35	AGFA
11/27/2023	1:00PM	RO PLANT	RO WATER	NA
11/27/2023	11:30AM	MICROBIOLOGY	CENTRIFUGE	REMI
11/27/2023	10:15AM	RADIOLOGY	CR 35	AGFA
11/27/2023	11:15PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
11/27/2023	10:00AM	ED	ABG MACHINE	GEM
11/27/2023	9:30AM	FMW	ECG MACHINE	EDAN
11/28/2023	12:10PM	OPHTHALMOLOGY	SLIT LAMP	APPASAMY
11/28/2023	4:15PM	OPHTHALMOLOGY	OCT MACHINE	APPASAMY ASSOC
11/28/2023	9:45AM	ED	DEFIBRILLATOR	PHILIPS
11/28/2023	9:30AM	CVTS ICU	VENTILATOR	NEUMOVENT
11/28/2023	9:00AM	CVTS ICU	VENTILATOR	IIT KANPUR
11/28/2023	4:20PM	RADIOLAGY	CR30	AGFA
11/28/2023	4:00PM	ED	SYRINGE PUMP	B BRAUN
11/28/2023	3:20PM	ED	ABG MACHINE	GEM
11/29/2023	10:20AM	CENTRAL LAB	UPS	NA
11/29/2023	11:30AM	EMERGENCY DEPARTM	DEFIBRILLATOR	PHILIPS
11/29/2023	02:20PM	EMERGENCY DEPARTM	ECG MACHINE	EDAN

11/29/2023	03:50PM	NMICU	SIEMENS	SERVO VENTILATO
11/29/2023	04:55PM	RO PLANT	NA	NA
11/29/2023	7:10PM	EMERGENCY DEPARTM	ECG MACHINE	EDAN
11/30/2023	09:10AM	SURGERY WARD	SUCTION	NA
11/30/2023	10:15AM	CSSD	SEALING MACHINE	SEPACK
11/30/2023	11:30AM	RICU	DEFIBRILLATOR	HP
11/30/2023	12:30PM	EMERGENCY DEPARTM	X RAY MACHINE (MOBILE	SKANRAY
11/30/2023	01:45PM	RO PLANT	NA	NA
11/30/2023	02:30PM	MOT	OT LIGHT	HOSPEDIA
11/30/2023	03:40PM	NICU	WARMER	PRAGATHI
12/1/2023	04:50PM	CVTS	VENTILATOR	NEUMOVENT
12/1/2023	09:50AM	USG ROOM	USG MACHINE	GE
12/1/2023	11:30AM	W AND C OT	OT TABLE	HOSPITECH
12/1/2023	02:20PM	W&C SICU	ECG MACHINE	AMIGO
12/1/2023	10:10AM	RO PLANT	RO WATER	NA
12/1/2023	11:20AM	NMICU	MONITOR	MINDRAY
12/1/2023	01:50PM	ICCU	SYRINGE PUMP	B BRAUN
12/1/2023	02:45PM	ED	ECG MACHINE	EDAN
3/30/2024	6:00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
3/30/2024	5:15PM	EMERGENCY DEPART	INFUSION PUMP	B BRAUN
3/30/2024	5:00PM	CSSD	ETO STERILIZER	3M
3/30/2024	4:00PM	MICU	PORTABLE XRAY	SKANRAY
3/30/2024	3:10PM	DIALYSIS	BP APPARATUS	ELKOMETER
3/30/2024	12:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
3/30/2024	11:30PM	CVTS ICU	MONITOR	MINDRAY
4/1/2024	11:00PM	CVTS ICU	SYRINGE PUMP	B BRAUN
4/1/2024	10:0AM	COH2	PULSE OXYMETER	CHOICE MED
4/1/2024	10:0AM	COH2	DIGITAL BP	SCURE
4/1/2024	10:00PM	ENT OT	MICROSCOPE	CARL ZEISS
4/1/2024	8:40PM	NMICU	MONITOR	MINDRAY
4/1/2024	6:55PM	RO PLANT	RO WATER	
4/1/2024	5:00AM	ICCU	SYRINGE PUMP	SCHILLER
4/1/2024	2:00AM	NMICU	SYRINGE PUMP	B BRAUN
4/3/2024	12:30PM	DIALYSIS	MULTIPARA MONITOR	AMIGO
4/3/2024	12:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
4/3/2024	11:00AM	MICU	BIPAP	PHILIPS
4/3/2024	10:00AM	CVTS ICU	VENTILATOR	NEUMOVENT
4/3/2024	1:00PM	NEWROWARD	BP APPARATUS	NA
4/3/2024	9:00AM	CVTS OT	OT TABLE	ACCURA
4/3/2024	7:35PM	SICU	MONITOR	MINDRAY
4/3/2024	7:00AM	RO PLANT	RO WATER	NA
4/6/2024	3:00PM	ED	DEFIBRILLATOR	PHILIPS
4/6/2024	3:00PM	RADIOLOGY	X- RAY UNITE	NA
4/6/2024	9:30AM	MOT	LIGHT SOURCE	KARL STORES
4/6/2024	9:00AM	MOT	LAPROSCOPE	STORZ

4/6/2024	8:10PM	RICU	BIPAP	PHILIPS
4/6/2024	7:00PM	ICCU	BIPAP	PHILIPS
4/6/2024	2:00PM	PULMO WARD	BIPAP	NA
4/8/2024	10:30AM	EMERGENCY DEPART	MONITOR	MINDRAY
4/8/2024	1:00PM	MICU	MONITOR	MINDRAY
4/8/2024	12:00PM	RO PLANT	RO WATER	NA
4/8/2024	11:00AM	W&C	VAPOURISER	PENLON
4/8/2024	10:30AM	RADIOLOGY	CR 35	AGFA
4/8/2024	10:00AM	USG DEPARTMENT	USG MACHINE	GE
4/8/2024	9:30AM	USG DEPARTMENT	USG MACHINE	GE
4/11/2024	9:30AM	USG ROOM	USG MACHINE	GE
4/11/2024	10:30AM	CSSD	AUTOCLAVE	JOSHY ENGINEERIN
4/11/2024	1:30PM	W&C CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
4/11/2024	9:30PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
4/11/2024	3:30PM	RO PLANT	RO WATER	NA
4/11/2024	11:30AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
4/11/2024	10:30PM	CSSD	ETO STERILIZER	3M
4/15/2024	3:00PM	RO PLANT	RO WATER	NA
4/15/2024	12:30PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
4/15/2024	11:00AM	RICU	BIPAP	PHILIPS
4/15/2024	10:00AM	ED	ABG MACHINE	GEM
4/15/2024	4:00PM	USG DEPARTMENT	USG MACHINE	GE
4/15/2024	3:50PM	MICU	BIPAP	BME
4/15/2024	12:30PM	BLOOD BANK	REFRIGERATED CENTRIFUG	ELTEK
4/15/2024	11:30AM	ED	ABG MACHINE	GEM
4/18/2024	10:00AM	MICU	X RAY MACHINE (MOBILE-C	SKANRAY
4/18/2024	2:00PM	RO PLANT	RO WATER	NA
4/18/2024	11:00AM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
4/18/2024	10:00PM	ICCU	VENTILATOR	SKANRAY
4/18/2024	6:30AM	RO PLANT	RO WATER	NA
4/18/2024	11:00AM	EMERGENCY DEPART	ECG MACHINE	EDAN
4/18/2024	9:00am	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
4/18/2024	10:00am	nmiCU	coat	janak
4/18/2024	11:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
4/18/2024	12:00PM	W&C PICU	MULTIPARA MONITOR	SCURE
4/21/2024	2:00PM	EMERGENCY DEPART	ABG MACHINE	GEM
4/21/2024	3:00PM	CVTS	COAT	NA
4/21/2024	5:00PM	RO PLANT	RO WATER	NA
4/21/2024	9:00PM	MAJOR OT	CAUTERY MACHINE	ALAN
4/21/2024	10:00AM	RICU	MULTIPARA MONITOR	EDAN
4/21/2024	11:00AM	DIALYSIS	RO WATER	NA
4/21/2024	12:00PM	MICU	BIPAP	PHILIPS
4/21/2024	10:00AM	SICU	MULTIPARA MONITOR	MINDRAY
4/24/2024	11:00AM	MICROBIOLOGY	BACTERIA CULTURE	BIOMERIEUX

4/24/2024	12:00PM	ICCU	VENTILATOR	SIEMENS
4/24/2024	3:00PM	ED	ECG MACHINE	EDAN
4/24/2024	5:00PM	RICU	MONITOR	L&T
4/24/2024	10:00AM	CVTS ICU	ABG MACHINE	GEM
4/24/2024	11:00AM	CENTRAL LAB	AUTO HEMATOLOGY ANAL	HORIBA
4/24/2024	12:00PM	MRI	MRI WORK STATION	GE
4/24/2024	1:00PM	OBG WARD	SLIT LAMP	NA
4/29/2024	4:00PM	RADIOLOGY	PRINTER	AGFA
4/29/2024	5:00PM	MICROBIOLOGY LAB	CENTRIFUGE	REMI
4/29/2024	2:00AM	USG ROOM	USG MACHINE	GE
4/29/2024	5:00AM	RO PLANT	RO WATER	NA
4/29/2024	7:00AM	RADIOLOGY	CR 35	AGFA
4/29/2024	10:30AM	RADIOLOGY	CR 35	AGFA
4/29/2024	1:00PM	RO PLANT	RO WATER	NA
4/29/2024	11:30AM	MICROBIOLOGY	CENTRIFUGE	REMI
4/29/2024	10:15AM	RADIOLOGY	CR 35	AGFA
4/29/2024	11:15PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
4/29/2024	10:00AM	ED	ABG MACHINE	GEM
4/29/2024	9:30AM	FMW	ECG MACHINE	EDAN
4/29/2024	12:10PM	OPHTHALMOLOGY	SLIT LAMP	APPASAMY
4/30/2024	4:15PM	OPHTHALMOLOGY	OCT MACHINE	APPASAMY ASSOC
4/30/2024	9:45AM	ED	DEFIBRILLATOR	PHILIPS
4/30/2024	9:30AM	CVTS ICU	VENTILATOR	NEUMOVENT
4/30/2024	9:00AM	CVTS ICU	VENTILATOR	IIT KANPUR
4/30/2024	4:20PM	RADIOLAGY	CR30	AGFA
4/30/2024	4:00PM	ED	SYRINGE PUMP	B BRAUN
4/30/2024	3:20PM	ED	ABG MACHINE	GEM
4/30/2024	2:30PM	ECHO ROOM	ECG MACHINE	GE
4/30/2024	1:42PM	RO PLANT	RO WATER	NA
4/30/2024	1:15PM	RADIOLOGY	CR30	AGFA
4/30/2024	1:10PM	W&C COVID	DIGITAL BP	SCURE
4/30/2024	9:30AM	ICCU	SYRINGE PUMP	B BRAUN
4/30/2024	9:00PM	ENT OT	ENT VIDEO SCOPE	NASIO
4/30/2024	9:00AM	EMERGENCY DEPART	ECG MACHINE	EDAN
4/30/2024	8:00PM	ENT OT	MICROSCOPE	CARL ZEISS
5/1/2024	7:00PM	ICCU	PORTABLE XRAY	SKANRAY
5/1/2024	6:00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
5/1/2024	5:15PM	EMERGENCY DEPART	INFUSION PUMP	B BRAUN
5/1/2024	5:00PM	CSSD	ETO STERILIZER	3M
5/1/2024	4:00PM	MICU	PORTABLE XRAY	SKANRAY
5/1/2024	3:10PM	DIALYSIS	BP APPARATUS	ELKOMETER
5/1/2024	12:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
5/1/2024	11:30PM	CVTS ICU	MONITOR	MINDRAY
5/3/2024	11:00PM	CVTS ICU	SYRINGE PUMP	B BRAUN

5/3/2024	10:0AM	COH2	PULSE OXYMETER	CHOICE MED
5/3/2024	10:0AM	COH2	DIGITAL BP	SCURE
5/3/2024	10:00PM	ENT OT	MICROSCOPE	CARL ZEISS
5/3/2024	8:40PM	NMICU	MONITOR	MINDRAY
5/3/2024	6:55PM	RO PLANT	RO WATER	
5/3/2024	5:00AM	ICCU	SYRINGE PUMP	SCHILLER
5/3/2024	2:00AM	NMICU	SYRINGE PUMP	B BRAUN
5/5/2024	12:30PM	DIALYSIS	MULTIPARA MONITOR	AMIGO
5/5/2024	12:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
5/5/2024	11:00AM	MICU	BIPAP	PHILIPS
5/5/2024	10:00AM	CVTS ICU	VENTILATOR	NEUMOVENT
5/5/2024	1:00PM	NEWROWARD	BP APPARATUS	NA
5/5/2024	9:00AM	CVTS OT	OT TABLE	ACCURA
5/8/2024	7:35PM	SICU	MONITOR	MINDRAY
5/8/2024	7:00AM	RO PLANT	RO WATER	NA
5/8/2024	3:00PM	ED	DEFIBRILLATOR	PHILIPS
5/8/2024	3:00PM	RADIOLOGY	X- RAY UNITE	NA
5/8/2024	9:30AM	MOT	LIGHT SOURCE	KARL STORES
5/8/2024	9:00AM	MOT	LAPROSCOPE	STORZ
5/10/2024	8:10PM	RICU	BIPAP	PHILIPS
5/10/2024	7:00PM	ICCU	BIPAP	PHILIPS
5/10/2024	2:00PM	PULMO WARD	BIPAP	NA
5/10/2024	10:30AM	EMERGENCY DEPART	MONITOR	MINDRAY
5/10/2024	1:00PM	MICU	MONITOR	MINDRAY
5/13/2024	12:00PM	RO PLANT	RO WATER	NA
5/13/2024	11:00AM	W&C	VAPOURISER	PENLON
5/13/2024	10:30AM	RADIOLOGY	CR 35	AGFA
5/13/2024	10:00AM	USG DEPARTMENT	USG MACHINE	GE
5/13/2024	9:30AM	USG DEPARTMENT	USG MACHINE	GE
5/13/2024	9:30AM	USG ROOM	USG MACHINE	GE
5/13/2024	10:30AM	CSSD	AUTOCLAVE	JOSHY ENGINEERIN
5/15/2024	1:30PM	W&C CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
5/15/2024	9:30PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
5/15/2024	3:30PM	RO PLANT	RO WATER	NA
5/15/2024	11:30AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
5/15/2024	10:30PM	CSSD	ETO STERILIZER	3M
5/15/2024	3:00PM	RO PLANT	RO WATER	NA
5/15/2024	12:30PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
5/17/2024	11:00AM	RICU	BIPAP	PHILIPS
5/17/2024	10:00AM	ED	ABG MACHINE	GEM
5/17/2024	4:00PM	USG DEPARTMENT	USG MACHINE	GE
5/17/2024	3:50PM	MICU	BIPAP	BME
5/17/2024	12:30PM	BLOOD BANK	REFRIGERATED CENTRIFUG	ELTEK
5/17/2024	11:30AM	ED	ABG MACHINE	GEM
5/19/2024	10:00AM	MICU	X RAY MACHINE (MOBILE-C	SKANRAY

5/19/2024	3:00PM	ED	FLOWMETER	NA
5/19/2024	10:00am	USG DEPARTMENT	USG MACHINE	GE
5/19/2024	11:00AM	ECHO ROOM	ECHO MACHINE	VIVID E9
5/19/2024	1:00PM	RO PLANT	RO WATER	NA
5/19/2024	3:00PM	CVTS ICU	SYRINGE PUMP	B BRAUN
5/21/2024	4:00PM	CARDIOLOGY	TMT	GE
5/21/2024	5:00PM	MEDICAL ICU	MULTIPARA MONITOR	EDAN
5/21/2024	9:00AM	ED	DEFIBRILLATOR	PHILIPS
5/21/2024	3:30PM	CENTRAL LAB	CENTRIFUGE	ROTEK
5/21/2024	12:30PM	ED	MONITOR	MINDRAY
5/25/2024	10:30AM	RADIOLOGY	CR 30	C
5/25/2024	10:15AM	CVTS	AUTOCLAVE	NA
5/25/2024	1:00PM	BLOOD BANK	CENTRIFUGE	ROTEK
5/25/2024	9:30AM	ED	X RAY MACHINE (MOBILE-6	SKANRAY
5/25/2024	9:00AM	NSICU	VENTILATOR	NEUMOVENT
5/25/2024	7:30AM	MICU	VENTILATOR	NEUMOVENT
5/25/2024	7:00AM	MOT	LAPROSCOPE	STORZ
5/25/2024	3:30PM	NEURO WARD	SYRINGE PUMP	B BRAUN
5/25/2024	2:30PM	ED	X RAY MACHINE (MOBILE-6	SKANRAY
5/25/2024	2:00PM	CENTRAL LAB	CENTRIFUGE	ROTEK
5/25/2024	2:00PM	ED	MULTIPARA MONITOR	MINDRAY
5/25/2024	12:00PM	PHYSIOTHERAPY	ULTRA SOUND PROBE	TECHNOMED ELEC
5/25/2024	9:30AM	W&C CENTRAL LAB	AUTO HEAMATOLOGY ANAL	HEMAX
5/25/2024	4:30pm	ICCU	BIPAP	PHILIPS
5/25/2024	4:00PM	MICU	MULTIPARA MONITOR	MINDRAY
5/28/2024	2:45PM	SICU	VENTILATOR	NEUMOVENT
5/28/2024	12:00PM	RO PLANT	RO WATER	NA
5/28/2024	11:30AM	ICCU	BIPAP	PHILIPS
5/28/2024	11:00AM	ICCU	SYRINGE PUMP	SCHILLER
5/28/2024	10:5AM	W&C OT	ANAESTHESIA	NA
5/28/2024	10:15AM	BLOOD BANK	URINE ANALYSER	SYSMEX
5/28/2024	3:30PM	NMICU	VENTILATOR	SIEMENS
5/29/2024	12:30PM	ICCU	SYRINGE PUMP	B BRAUN
5/29/2024	12:15pm	RO PLANT	RO WATER	NA
5/29/2024	11:15AM	RADIOLOGY	CR 30	C
5/29/2024	11:00AM	CENTRAL LAB	ELECTROLYTE ANALYSER	B&E
5/29/2024	10:30AM	CENTRAL LAB	CENTRIFUGE	ROTEK
5/29/2024	10:00PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
5/29/2024	11:30AM	ICCU	BIPAP	PHILIPS
5/29/2024	11:00AM	ICCU	SYRINGE PUMP	SCHILLER
5/29/2024	10:5AM	W&C OT	ANAESTHESIA	NA
5/29/2024	11:30AM	SP PULMO WARD	NEBULISER	ELKO
5/29/2024	10:15AM	BLOOD BANK	URINE ANALYSER	SYSMEX
5/30/2024	11:00AM	MEDICAL ICU	SYRINGE PUMP	B BRAUN
5/30/2024	1:00PM	CENTRAL LAB	MICROSCOPE	LABOMED

5/30/2024	4:00PM	W&C PICU	PORTABLE SUCTION	SURGI CRAFT
5/30/2024	10:5AM	DIALYSIS	MULTIPARA MONITOR	AMIGO
5/30/2024	10:15AM	CVTS ICU	VENTILATOR	NEUMOVENT
5/30/2024	10:00PM	MICU	BIPAP	PHILIPS
5/30/2024	11:00AM	CVTS ICU	VENTILATOR	NEUMOVENT
5/31/2024	1:00PM	NEWROWARD	BP APPARATUS	NA
5/31/2024	4:00PM	CVTS OT	OT TABLE	ACCURA
5/31/2024	10:00AM	RO PLANT	RO WATER	NA
5/31/2024	11:00AM	CVTS OT	SEALINHG MACHINE	NA
5/31/2024	12:00PM	W&C OT	QUOTARY MACHINE	CAL-SURG
5/31/2024	1:00PM	PULMO WARD	BIPAP	RESMED
5/31/2024	2:00PM	RICU	VENTILATOR	NEUMOVENT
5/31/2024	4:00PM	AYURVEDA COVID HOS	CENTRAL FLOW METER	NA
2/1/2024	10:00AM	ED	PORTABLE VENTILLATOR	PHILIPS
2/2/2024	12:00PM	LABOUR ROOM	CTG MACHINE	MS TECHNOLOGIE
2/2/2024	03:00PM	SICU	SYRINGE PUMP	L&T
2/2/2024	4:15PM	FMW	ECG MACHINE	EDAN
2/2/2024	09:00AM	EMERGENCY DEPARTM	ABG MACHINE	GEM
2/2/2024	10:30AM	MICU	ECG MACHINE	EDAN
2/3/2024	01:30PM	SICU	MULTIPARA MONITOR	MINDRAY
2/3/2024	03:45PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
2/3/2024	09:00AM	MOT	C ARM	SKANRAY
2/3/2024	10:30AM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
2/3/2024	11:15AM	URO ICU	MULTIPARAMONITOR	AMIGO
2/3/2024	01:30PM	ECHO	CARDIOLOGY	ECHO MACHINE
2/6/2024	03:45PM	NSICU	MULTIPARA MONITOR	AMIGO
2/6/2024	04:15PM	LABOUR ROOM	USG MACHINE	SAMSUNG
2/6/2024	04:45PM	RO PLANT	RO WATER	NA
2/6/2024	09:00AM	NMICU	MULTIPARA MONITOR	MINDRAY
2/6/2024	12:00PM	MICU	VENTILATOR	
2/6/2024	02:45PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
2/6/2024	06:00PM	RICU	BIPAP	PHILIPS
2/6/2024	09:00AM	MOT	UPS	CONSUL NEOWAT
2/6/2024	10:05AM	MOT	BP APPARATUS	ELKOMETER
2/10/2024	11:12AM	ED	ABG MACHINE	GEM
2/10/2024	12:50PM	CVTS ICU	ABG MACHINE	GEM
2/10/2024	09:15AM	RICU	GLUCOMETER	GLUCOCARD
2/10/2024	10:10AM	LABOUR ROOM	O2 FLOWMETER	NA
2/14/2024	11:20AM	MOT	CAUTRY MACHINE	ALAN
2/14/2024	01:50PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
2/14/2024	02:45PM	NSICU	MULTIPARA MONITOR	AMIGO
2/14/2024	03:45PM	W&C OT	WARMER	PRAGATHI
2/14/2024	09:20AM	ED	BIPAP	PHILIPS
2/14/2024	10:20PM	MEDICINE OP	BIO THESIOMETER	DIABATIC FOOT CA
2/18/2024	12:35PM	ED	ABG MACHINE	GEM

2/18/2024	02:50PM	ED	ABG MACHINE	GEM
2/18/2024	04:00PM	CSSD	ETO STERILIZER	3M
2/18/2024	09:30AM	W&C COVID WARD	FLOW METER	NA
2/18/2024	10:20PM	ICCU	VENTILATOR	SIEMENS
2/18/2024	11:35AM	RO PLANT	RO WATER	NA
2/18/2024	01:40PM	CSSD	AUTOCLAVE MACHINE	SURGITECH
2/18/2024	03:00PM	NSICU	ECG MACHINE	SKANRAY
2/20/2024	04:20PM	MOT	PORTABLE SUCTION	NA
2/20/2024	09:15PM	NSICU	VENTILATOR	NEUMOVENT
2/20/2024	10:20PM	RADIOLOGY	CR	AGFA
2/20/2024	11:30PM	CVTS ICU	ABG MACHINE	GEM
2/20/2024	02:10PM	RADIOLOGY	CR	AGFA
2/20/2024	03:20PM	SICU	MONITOR	SKANRAY
2/20/2024	06:00PM	RICU	VENTILATOR	SKANRAY
2/23/2024	09:30AM	RO PLANT	NA	NA
2/23/2024	10:20AM	CENTRAL LAB	UPS	NA
2/23/2024	11:30AM	EMERGENCY DEPARTM	DEFIBRILLATOR	PHILIPS
2/23/2024	02:20PM	EMERGENCY DEPARTM	ECG MACHINE	EDAN
2/23/2024	03:50PM	NMICU	SIEMENS	SERVO VENTILATO
2/23/2024	04:55PM	RO PLANT	NA	NA
2/23/2024	7:10PM	EMERGENCY DEPARTM	ECG MACHINE	EDAN
2/26/2024	09:10AM	SURGERY WARD	SUCTION	NA
2/26/2024	10:15AM	CSSD	SEALING MACHINE	SEPACK
2/26/2024	11:30AM	RICU	DEFIBRILLATOR	HP
2/26/2024	12:30PM	EMERGENCY DEPARTM	X RAY MACHINE (MOBILE C	SKANRAY
2/26/2024	01:45PM	RO PLANT	NA	NA
2/26/2024	02:30PM	MOT	OT LIGHT	HOSPEDIA
2/26/2024	03:40PM	NICU	WARMER	PRAGATHI
2/27/2024	04:50PM	CVTS	VENTILATOR	NEUMOVENT
2/27/2024	09:50AM	USG ROOM	USG MACHINE	GE
2/27/2024	11:30AM	W AND C OT	OT TABLE	HOSPITECH
2/27/2024	02:20PM	W&C SICU	ECG MACHINE	AMIGO
2/27/2024	10:10AM	RO PLANT	RO WATER	NA
2/27/2024	11:20AM	NMICU	MONITOR	MINDRAY
2/27/2024	01:50PM	ICCU	SYRINGE PUMP	B BRAUN
2/27/2024	02:45PM	ED	ECG MACHINE	EDAN
3/1/2024	03:45PM	CSSD	SEALINHG MACHINE	SEPACK
3/1/2024	09:20AM	SICU	VENTILATOR	SKANRAY
3/1/2024	10:20PM	ICCU	SYRINGE PUMP	SCHILLER
3/1/2024	12:35PM	SURGERY WARD	VACCUM	NA
3/1/2024	02:50PM	W&C SICU	VENTILATOR	SIEMENS
3/1/2024	04:00PM	RO UNIT	RO WATER	NA
3/1/2024	09:30AM	RICU	BIPAP	PHILIPS
3/1/2024	10:20PM	SICU	VENTILATOR	SKANRAY
3/6/2024	11:35AM	CENTRAL LAB	ELECTROLYTE ANALYSER	B&E

3/6/2024	01:40PM	ED	MONITOR	BLUESTAR
3/6/2024	03:00PM	ED	FLOW METER	NA
3/6/2024	04:20PM	ED	DEFIBRILLATOR	PHILIPS
3/6/2024	09:15PM	PULMO WARD	GLUCOMETER	ALEREGI
3/6/2024	10:20PM	PULMO WARD	BIPAP	PHILIPS
3/6/2024	11:30PM	DIALYSIS	MONITOR	AMIGO
3/6/2024	02:10PM	MOT	LIGHT SOURCE	STORZ
3/8/2024	03:20PM	RO PLANT	RO WATER	NA
3/8/2024	06:00PM	CVTS OT	SEALINHG MACHINE	NA
3/8/2024	09:30AM	W&C OT	CAUOTARY MACHINE	CAL-SURG
3/8/2024	10:20AM	PULMO WARD	BIPAP	RESMED
3/8/2024	11:30AM	RICU	VENTILATOR	NEUMOVENT
3/8/2024	01:00PM	AYURVEDA COVID HOS	CENTRAL FLOW METER	NA
3/10/2024	02:20PM	ED	PORTABLE VENTILLATOR	PHILIPS
3/10/2024	3:00pm	ED	ECG MACHINE	EDAN
3/10/2024	55:0pm	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
3/10/2024	5:30pm	RO PLANT	RO WATER	NA
3/10/2024	9:00am	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
3/10/2024	10:00am	nmiCU	coat	janak
3/10/2024	11:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
3/10/2024	12:00PM	W&C PICU	MULTIPARA MONITOR	SCURE
3/14/2024	2:00PM	EMERGENCY DEPART	ABG MACHINE	GEM
3/14/2024	3:00PM	CVTS	COAT	NA
3/14/2024	5:00PM	RO PLANT	RO WATER	NA
3/14/2024	9:00PM	MAJOR OT	CAUTERY MACHINE	ALAN
3/14/2024	10:00AM	RICU	MULTIPARA MONITOR	EDAN
3/14/2024	11:00AM	DIALYSIS	RO WATER	NA
3/14/2024	12:00PM	MICU	BIPAP	PHILIPS
3/14/2024	10:00AM	SICU	MULTIPARA MONITOR	MINDRAY
3/17/2024	11:00AM	MICROBIOLOGY	BACTERIA CULTURE	BIOMERIEUX
3/17/2024	12:00PM	ICCU	VENTILATOR	SIEMENS
3/17/2024	3:00PM	ED	ECG MACHINE	EDAN
3/17/2024	5:00PM	RICU	MONITOR	L&T
3/17/2024	10:00AM	CVTS ICU	ABG MACHINE	GEM
3/17/2024	11:00AM	CENTRAL LAB	AUTO HEMATOLOGY ANAL	HORIBA
3/17/2024	12:00PM	MRI	MRI WORK STATION	GE
3/17/2024	1:00PM	OBG WARD	SLIT LAMP	NA
3/19/2024	2:00PM	PAY WARD	PORTABEL SUCTION	NA
3/19/2024	2:30PM	PAY WARD	O2 FLOW METER	NA
3/19/2024	4:00PM	RADIOLOGY	PRINTER	AGFA
3/19/2024	5:00PM	MICROBIOLOGY LAB	CENTRIFUGE	REMI
3/19/2024	2:00AM	USG ROOM	USG MACHINE	GE
3/19/2024	5:00AM	RO PLANT	RO WATER	NA

3/19/2024	7:00AM	RADIOLOGY	CR 35	AGFA
3/21/2024	10:30AM	RADIOLOGY	CR 35	AGFA
3/21/2024	1:00PM	RO PLANT	RO WATER	NA
3/21/2024	11:30AM	MICROBIOLOGY	CENTRIFUGE	REMI
3/21/2024	10:15AM	RADIOLOGY	CR 35	AGFA
3/21/2024	11:15PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
3/21/2024	10:00AM	ED	ABG MACHINE	GEM
3/21/2024	9:30AM	FMW	ECG MACHINE	EDAN
3/25/2024	12:10PM	OPHTHALMOLOGY	SLIT LAMP	APPASAMY
3/25/2024	4:15PM	OPHTHALMOLOGY	OCT MACHINE	APPASAMY ASSOC
3/25/2024	9:45AM	ED	DEFIBRILLATOR	PHILIPS
3/25/2024	9:30AM	CVTS ICU	VENTILATOR	NEUMOVENT
3/25/2024	9:00AM	CVTS ICU	VENTILATOR	IIT KANPUR
3/25/2024	4:20PM	RADIOLAGY	CR30	AGFA
3/25/2024	4:00PM	ED	SYRINGE PUMP	B BRAUN
3/25/2024	3:20PM	ED	ABG MACHINE	GEM
3/29/2024	2:30PM	ECHO ROOM	ECG MACHINE	GE
3/29/2024	1:42PM	RO PLANT	RO WATER	NA
3/29/2024	1:15PM	RADIOLOGY	CR30	AGFA
3/29/2024	1:10PM	W&C COVID	DIGITAL BP	SCURE
3/29/2024	9:30AM	ICCU	SYRINGE PUMP	B BRAUN
3/29/2024	9:00PM	ENT OT	ENT VIDEO SCOPE	NASIO
3/29/2024	9:00AM	EMERGENCY DEPART	ECG MACHINE	EDAN
3/29/2024	8:00PM	ENT OT	MICROSCOPE	CARL ZEISS
3/30/2024	7:00PM	ICCU	PORTABLE XRAY	SKANRAY
3/30/2024	6:00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
3/30/2024	5:15PM	EMERGENCY DEPART	INFUSION PUMP	B BRAUN
3/30/2024	5:00PM	CSSD	ETO STERILIZER	3M
3/30/2024	4:00PM	MICU	PORTABLE XRAY	SKANRAY
3/30/2024	3:10PM	DIALYSIS	BP APPARATUS	ELKOMETER
3/30/2024	12:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
3/30/2024	11:30PM	CVTS ICU	MONITOR	MINDRAY
4/1/2024	11:00PM	CVTS ICU	SYRINGE PUMP	B BRAUN
4/1/2024	10:0AM	COH2	PULSE OXYMETER	CHOICE MED
4/1/2024	10:0AM	COH2	DIGITAL BP	SCURE
4/1/2024	10:00PM	ENT OT	MICROSCOPE	CARL ZEISS
4/1/2024	8:40PM	NMICU	MONITOR	MINDRAY
4/1/2024	6:55PM	RO PLANT	RO WATER	
4/1/2024	5:00AM	ICCU	SYRINGE PUMP	SCHILLER
4/1/2024	2:00AM	NMICU	SYRINGE PUMP	B BRAUN
4/3/2024	12:30PM	DIALYSIS	MULTIPARA MONITOR	AMIGO
4/3/2024	12:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
4/3/2024	11:00AM	MICU	BIPAP	PHILIPS
4/3/2024	10:00AM	CVTS ICU	VENTILATOR	NEUMOVENT
4/3/2024	1:00PM	NEWROWARD	BP APPARATUS	NA

4/3/2024	9:00AM	CVTS OT	OT TABLE	ACCURA
4/3/2024	7:35PM	SICU	MONITOR	MINDRAY
4/3/2024	7:00AM	RO PLANT	RO WATER	NA
4/6/2024	3:00PM	ED	DEFIBRILLATOR	PHILIPS
4/6/2024	3:00PM	RADIOLOGY	X- RAY UNITE	NA
4/6/2024	9:30AM	MOT	LIGHT SOURCE	KARL STORES
4/6/2024	9:00AM	MOT	LAPROSCOPE	STORZ
4/6/2024	8:10PM	RICU	BIPAP	PHILIPS
4/6/2024	7:00PM	ICCU	BIPAP	PHILIPS
4/6/2024	2:00PM	PULMO WARD	BIPAP	NA
4/8/2024	10:30AM	EMERGENCY DEPART	MONITOR	MINDRAY
4/8/2024	1:00PM	MICU	MONITOR	MINDRAY
4/8/2024	12:00PM	RO PLANT	RO WATER	NA
4/8/2024	11:00AM	W&C	VAPOURISER	PENLON
4/8/2024	10:30AM	RADIOLOGY	CR 35	AGFA
4/8/2024	10:00AM	USG DEPARTMENT	USG MACHINE	GE
4/8/2024	9:30AM	USG DEPARTMENT	USG MACHINE	GE
4/11/2024	9:30AM	USG ROOM	USG MACHINE	GE
4/11/2024	10:30AM	CSSD	AUTOCLAVE	JOSHY ENGINEERIN
4/11/2024	1:30PM	W&C CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
4/11/2024	9:30PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
4/11/2024	3:30PM	RO PLANT	RO WATER	NA
4/11/2024	11:30AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
4/11/2024	10:30PM	CSSD	ETO STERILIZER	3M
4/15/2024	3:00PM	RO PLANT	RO WATER	NA
4/15/2024	12:30PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
4/15/2024	11:00AM	RICU	BIPAP	PHILIPS
4/15/2024	10:00AM	ED	ABG MACHINE	GEM
4/15/2024	4:00PM	USG DEPARTMENT	USG MACHINE	GE
4/15/2024	3:50PM	MICU	BIPAP	BME
4/15/2024	12:30PM	BLOOD BANK	REFRIGERATED CENTRIFUG	ELTEK
4/15/2024	11:30AM	ED	ABG MACHINE	GEM
4/18/2024	10:00AM	MICU	X RAY MACHINE (MOBILE-C	SKANRAY
4/18/2024	2:00PM	RO PLANT	RO WATER	NA
4/18/2024	11:00AM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA
4/18/2024	10:00PM	ICCU	VENTILATOR	SKANRAY
4/18/2024	6:30AM	RO PLANT	RO WATER	NA
4/18/2024	11:00AM	EMERGENCY DEPART	ECG MACHINE	EDAN
4/18/2024	9:00am	EMERGENCY DEPART	DEFIBRILLATOR	PHILIPS
4/18/2024	10:00am	nmiCU	coat	janak
4/18/2024	11:00AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
4/18/2024	12:00PM	W&C PICU	MULTIPARA MONITOR	SCURE
4/21/2024	2:00PM	EMERGENCY DEPART	ABG MACHINE	GEM
4/21/2024	3:00PM	CVTS	COAT	NA

4/21/2024	5:00PM	RO PLANT	RO WATER	NA
4/21/2024	9:00PM	MAJOR OT	CAUTERY MACHINE	ALAN
4/21/2024	10:00AM	RICU	MULTIPARA MONITOR	EDAN
4/21/2024	11:00AM	DIALYSIS	RO WATER	NA
4/21/2024	12:00PM	MICU	BIPAP	PHILIPS
4/21/2024	10:00AM	SICU	MULTIPARA MONITOR	MINDRAY
4/24/2024	11:00AM	MICROBIOLOGY	BACTERIA CULTURE	BIOMERIEUX
4/24/2024	12:00PM	ICCU	VENTILATOR	SIEMENS
4/24/2024	3:00PM	ED	ECG MACHINE	EDAN
4/24/2024	5:00PM	RICU	MONITOR	L&T
4/24/2024	10:00AM	CVTS ICU	ABG MACHINE	GEM
4/24/2024	11:00AM	CENTRAL LAB	AUTO HEMATOLOGY ANAL	HORIBA
4/24/2024	12:00PM	MRI	MRI WORK STATION	GE
4/24/2024	1:00PM	OBG WARD	SLIT LAMP	NA
4/29/2024	4:00PM	RADIOLOGY	PRINTER	AGFA
4/29/2024	5:00PM	MICROBIOLOGY LAB	CENTRIFUGE	REMI
4/29/2024	2:00AM	USG ROOM	USG MACHINE	GE
4/29/2024	5:00AM	RO PLANT	RO WATER	NA
4/29/2024	7:00AM	RADIOLOGY	CR 35	AGFA
4/29/2024	10:30AM	RADIOLOGY	CR 35	AGFA
4/29/2024	1:00PM	RO PLANT	RO WATER	NA
4/29/2024	11:30AM	MICROBIOLOGY	CENTRIFUGE	REMI
4/29/2024	10:15AM	RADIOLOGY	CR 35	AGFA
4/29/2024	11:15PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS
4/29/2024	10:00AM	ED	ABG MACHINE	GEM
4/29/2024	9:30AM	FMW	ECG MACHINE	EDAN
4/29/2024	12:10PM	OPHTHALMOLOGY	SLIT LAMP	APPASAMY
4/30/2024	4:15PM	OPHTHALMOLOGY	OCT MACHINE	APPASAMY ASSOC
4/30/2024	9:45AM	ED	DEFIBRILLATOR	PHILIPS
4/30/2024	9:30AM	CVTS ICU	VENTILATOR	NEUMOVENT
4/30/2024	9:00AM	CVTS ICU	VENTILATOR	IIT KANPUR
4/30/2024	4:20PM	RADIOLAGY	CR30	AGFA
4/30/2024	4:00PM	ED	SYRINGE PUMP	B BRAUN
4/30/2024	3:20PM	ED	ABG MACHINE	GEM
4/30/2024	2:30PM	ECHO ROOM	ECG MACHINE	GE
4/30/2024	1:42PM	RO PLANT	RO WATER	NA
4/30/2024	1:15PM	RADIOLOGY	CR30	AGFA
4/30/2024	1:10PM	W&C COVID	DIGITAL BP	SCURE
4/30/2024	9:30AM	ICCU	SYRINGE PUMP	B BRAUN
4/30/2024	9:00PM	ENT OT	ENT VIDEO SCOPE	NASIO
4/30/2024	9:00AM	EMERGENCY DEPART	ECG MACHINE	EDAN
4/30/2024	8:00PM	ENT OT	MICROSCOPE	CARL ZEISS
5/1/2024	7:00PM	ICCU	PORTABLE XRAY	SKANRAY
5/1/2024	6:00PM	DIALYSIS	DIALYSIS MACHINE	FRESENIUS

5/1/2024	5:15PM	EMERGENCY DEPART	INFUSION PUMP	B BRAUN
5/1/2024	5:00PM	CSSD	ETO STERILIZER	3M
5/1/2024	4:00PM	MICU	PORTABLE XRAY	SKANRAY
5/1/2024	3:10PM	DIALYSIS	BP APPARATUS	ELKOMETER
5/1/2024	12:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
5/1/2024	11:30PM	CVTS ICU	MONITOR	MINDRAY
5/3/2024	11:00PM	CVTS ICU	SYRINGE PUMP	B BRAUN
5/3/2024	10:0AM	COH2	PULSE OXYMETER	CHOICE MED
5/3/2024	10:0AM	COH2	DIGITAL BP	SCURE
5/3/2024	10:00PM	ENT OT	MICROSCOPE	CARL ZEISS
5/3/2024	8:40PM	NMICU	MONITOR	MINDRAY
5/3/2024	6:55PM	RO PLANT	RO WATER	
5/3/2024	5:00AM	ICCU	SYRINGE PUMP	SCHILLER
5/3/2024	2:00AM	NMICU	SYRINGE PUMP	B BRAUN
5/5/2024	12:30PM	DIALYSIS	MULTIPARA MONITOR	AMIGO
5/5/2024	12:00PM	CVTS ICU	VENTILATOR	NEUMOVENT
5/5/2024	11:00AM	MICU	BIPAP	PHILIPS
5/5/2024	10:00AM	CVTS ICU	VENTILATOR	NEUMOVENT
5/5/2024	1:00PM	NEWROWARD	BP APPARATUS	NA
5/5/2024	9:00AM	CVTS OT	OT TABLE	ACCURA
5/8/2024	7:35PM	SICU	MONITOR	MINDRAY
5/8/2024	7:00AM	RO PLANT	RO WATER	NA
5/8/2024	3:00PM	ED	DEFIBRILLATOR	PHILIPS
5/8/2024	3:00PM	RADIOLOGY	X- RAY UNITE	NA
5/8/2024	9:30AM	MOT	LIGHT SOURCE	KARL STORES
5/8/2024	9:00AM	MOT	LAPROSCOPE	STORZ
5/10/2024	8:10PM	RICU	BIPAP	PHILIPS
5/10/2024	7:00PM	ICCU	BIPAP	PHILIPS
5/10/2024	2:00PM	PULMO WARD	BIPAP	NA
5/10/2024	10:30AM	EMERGENCY DEPART	MONITOR	MINDRAY
5/10/2024	1:00PM	MICU	MONITOR	MINDRAY
5/13/2024	12:00PM	RO PLANT	RO WATER	NA
5/13/2024	11:00AM	W&C	VAPOURISER	PENLON
5/13/2024	10:30AM	RADIOLOGY	CR 35	AGFA
5/13/2024	10:00AM	USG DEPARTMENT	USG MACHINE	GE
5/13/2024	9:30AM	USG DEPARTMENT	USG MACHINE	GE
5/13/2024	9:30AM	USG ROOM	USG MACHINE	GE
5/13/2024	10:30AM	CSSD	AUTOCLAVE	JOSHY ENGINEERIN
5/15/2024	1:30PM	W&C CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
5/15/2024	9:30PM	RADIOLOGY	COMPUTED RADIOGRAPHY	AGFA
5/15/2024	3:30PM	RO PLANT	RO WATER	NA
5/15/2024	11:30AM	CENTRAL LAB	FULLY BIOCHEMISTRY ANA	RANDOX
5/15/2024	10:30PM	CSSD	ETO STERILIZER	3M
5/15/2024	3:00PM	RO PLANT	RO WATER	NA
5/15/2024	12:30PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA

5/17/2024	11:00AM	RICU	BIPAP	PHILIPS
5/17/2024	10:00AM	ED	ABG MACHINE	GEM
5/17/2024	4:00PM	USG DEPARTMENT	USG MACHINE	GE
5/17/2024	3:50PM	MICU	BIPAP	BME
5/17/2024	12:30PM	BLOOD BANK	REFRIGERATED CENTRIFUGE	ELTEK
5/17/2024	11:30AM	ED	ABG MACHINE	GEM
5/19/2024	10:00AM	MICU	X RAY MACHINE (MOBILE-6	SKANRAY
5/19/2024	3:00PM	ED	FLOWMETER	NA
5/19/2024	10:00am	USG DEPARTMENT	USG MACHINE	GE
5/19/2024	11:00AM	ECHO ROOM	ECHO MACHINE	VIVID E9
5/19/2024	1:00PM	RO PLANT	RO WATER	NA
5/19/2024	3:00PM	CVTS ICU	SYRINGE PUMP	B BRAUN
5/21/2024	4:00PM	CARDIOLOGY	TMT	GE
5/21/2024	5:00PM	MEDICAL ICU	MULTIPARA MONITOR	EDAN
5/21/2024	9:00AM	ED	DEFIBRILLATOR	PHILIPS
5/21/2024	3:30PM	CENTRAL LAB	CENTRIFUGE	ROTEK
5/21/2024	12:30PM	ED	MONITOR	MINDRAY
5/25/2024	10:30AM	RADIOLOGY	CR 30	C
5/25/2024	10:15AM	CVTS	AUTOCLAVE	NA
5/25/2024	1:00PM	BLOOD BANK	CENTRIFUGE	ROTEK
5/25/2024	9:30AM	ED	X RAY MACHINE (MOBILE-6	SKANRAY
5/25/2024	9:00AM	NSICU	VENTILATOR	NEUMOVENT
5/25/2024	7:30AM	MICU	VENTILATOR	NEUMOVENT
5/25/2024	7:00AM	MOT	LAPROSCOPE	STORZ
5/25/2024	3:30PM	NEURO WARD	SYRINGE PUMP	B BRAUN
5/25/2024	2:30PM	ED	X RAY MACHINE (MOBILE-6	SKANRAY
5/25/2024	2:00PM	CENTRAL LAB	CENTRIFUGE	ROTEK
5/25/2024	2:00PM	ED	MULTIPARA MONITOR	MINDRAY
5/25/2024	12:00PM	PHYSIOTHERAPY	ULTRA SOUND PROBE	TECHNOMED ELEC
5/25/2024	9:30AM	W&C CENTRAL LAB	AUTO HEAMATOLOGY ANAL	HEMAX
5/25/2024	4:30pm	ICCU	BIPAP	PHILIPS
5/25/2024	4:00PM	MICU	MULTIPARA MONITOR	MINDRAY
5/28/2024	2:45PM	SICU	VENTILATOR	NEUMOVENT
5/28/2024	12:00PM	RO PLANT	RO WATER	NA
5/28/2024	11:30AM	ICCU	BIPAP	PHILIPS
5/28/2024	11:00AM	ICCU	SYRINGE PUMP	SCHILLER
5/28/2024	10:5AM	W&C OT	ANAESTHESIA	NA
5/28/2024	10:15AM	BLOOD BANK	URINE ANALYSER	SYSMEX
5/28/2024	3:30PM	NMICU	VENTILATOR	SIEMENS
5/29/2024	12:30PM	ICCU	SYRINGE PUMP	B BRAUN
5/29/2024	12:15pm	RO PLANT	RO WATER	NA
5/29/2024	11:15AM	RADIOLOGY	CR 30	C
5/29/2024	11:00AM	CENTRAL LAB	ELECTROLYTE ANALYSER	B&E
5/29/2024	10:30AM	CENTRAL LAB	CENTRIFUGE	ROTEK
5/29/2024	10:00PM	RADIO DIAGNOSIS	COMPUTED RADIOGRAPHY	AGFA

[illegible]

[illegible]

[illegible]

[illegible]

MODEL	SERIAL NUMBER	BME ID
DIVA	EATBTB1498	KMCT/BME/W OBGW/0783
BABYLOG8000	ARNB80008	KMCT/BME/W NICU/0768
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
CR 30-X	33130	KMCT/BME/RD/1226
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
ALPHA ST	8.30039E+13	KMCT/BME/RD/1047
NA	NA	NA
C 48	13253	KMCT/BME/DOT/0084
RESPICARE	NA	NA
GRAPHNET-TS	NA	KMCT/BME/NSICU/0045
PLANET 45	H19AR2002	KMCT/BME/RICU/1127
SKANRESPIRO	MR20CB2979	KMCT/BME/SICU/1231
CR 30-X	33130	KMCT/BME/RD/1226
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
MEDIPHARM	438	KMCT/BME/CSSD/1245
TRUESCOPE 2	M003E000129	KMCT/BME/ICCU/0133
NA	NA	NA
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
NA	NA	NA
NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
CR 35-X	5165	KMCT/BME/RD/0714
GRAPHNET-TS	NA	KMCT/BME/NSICU/0045
CR 35-X	5165	KMCT/BME/RD/0714
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
CR 35-X	5165	KMCT/BME/RD/0714
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
LOGIC P5	5372508	KMCT/BME/USG/0026
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
CR 35-X	5165	KMCT/BME/RD/0714
PREMIER 3000	NA	KMCT/BME/ED/0037
NA		KMCT/BME/ED/0907
RX IMOLA	72010997	KMCT/BME/CL/0354
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
REVELUTION ACTS	32003BG6	KMCT/BME/RD/0713
1.5T HDxt	NA	KMCT/BME/RD/0712
NA	NA	NA
MEC1201	CC11116596	KMCT/BME/SICU/0111
NA	NA	NA
NA	NA	NA
NA	NA	NA

NA	NA	NA
IM8	M16606320007	KMCT/BME/MICU/0179
SONOACE R7	30QQM3HG80034T	KMCT/BME/USG/1051
STAR 50N	F11S1163	KMCT/BME/NSICU/0041
FM301	510118765 816	KMCT/BME/W LR/1170
SP102	407122	KMCT/BME/SICU/0121
SE-3	M17901470001	KMCT/BME/FMW/0630
PREMIER 3000	24429	KMCT/BME/ED/0037
SE3	7602880005	KMCT/BME/MICU/0192
MEC1200	CC11116587	KMCT/BME/SICU/0110
CR 35-X	5165	KMCT/BME/RD/0714
SKAN C	MY2718RO158	KMCT/BME/MOT/0509
CR 35-X	5165	KMCT/BME/RD/0714
A9000	SPR201102111705	KMCT/BME/URO/1250
GE	VIVID E9	KMCT/BME/CRD/0255
A9000	SPR201102111745	KMCT/BME/NSICU/1255
SONOACE R7	30QQM3HG80034T	KMCT/BME/USG/1051
NA	NA	NA
MEC1000	AQ1C163873	KMCT/BME/NMICU/0087
SKANRAY	SKANRESPIRO	MR20CA3558
CR 35-X	5165	KMCT/BME/RD/0714
NA	NA	NA
NA	20163400012 122	NA
NA		NA
PREMIER 3000	24429	KMCT/BME/ED/0037
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
GT-1070	1070L7600270	KMCT/BME/RICU/0889
NA	NA	NA
ELSY 360+	1709326	KMCT/BME/MOT/0460
4008S	NA	NA
NA	NA	NA
NA	NA	NA
NA	NA	NA
VIBRO DROP	VD1909181	KMCT/BME/MOP/1179
PREMIER 3000	24429	KMCT/BME/ED/0037
PREMIER 3000	24429	KMCT/BME/ED/0037
STERI VAC	359642	KMCT/BME/CSSD/0521
NA	NA	NA
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
NA	NA	NA
MEDIPHARM	438	KMCT/BME/CSSD/1245
ORION PLUS	R14WB1182	KMCT/BME/NSICU/1220
NA	NA	KMCT/BME/MOT/0503
GRAPHNET-TS	NA	KMCT/BME/NSICU/0045
CR 35-X	5165	KMCT/BME/RD/0714

PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
CR 35-X	5165	KMCT/BME/RD/0714
PLANET 45	NA	NA
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
NA	NA	NA
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
SE1200	M17708420001	KMCT/BME/ED/0009
10231		KMCT/BME/NMICU/0274
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
	CS3984	KMCT/BME/CSSD/0526
CODEMASTER	3304A13416	KMCT/BME/RICU/0275
	MY2619R0207	KMCT/BME/RD/1156
NA	NA	NA
SURGICO 60H	10K857	KMCT/BME/MOT/0471
PRESERVIA XXL		KMCT/BME/W LR/0787
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
LOGIC F6	601660WX0	KMCT/BME/USG/1049
		KMCT/BME/W OT/1011
EC13	1907143	KMCT/BME/W SICU /1188
NA	NA	NA
MEC1000	NA	KMCT/BME/NMICU/0092
PERFUSER COMPACT	204288	KMCT/BME/W PICU/0857
SE1200	M17708420001	KMCT/BME/ED/0009
CS31+V	CS10571	KMCT/BME/CSSD/1086
SKANRESPIRO PLUS	MR214805	KMCT/BME/SICU/1231
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
NA	NA	NA
SERVO VENTILATOR	10420S11	KMCT/BME/W SICU/0803
NA	NA	NA
NA	NA	NA
SKANRESPIRO	MR20CB2979	KMCT/BME/SICU/1231
CBS-300	3-Mar-16	KMCT/BME/CL/0356
MPM 100	AQ16156662	KMCT/BME/ED/0005
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
	EB16JA02538	KMCT/BME/CTW/0952
NA	NA	NA
NA	NA	NA
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
NA	NA	NA
NA	NA	NA
250B		KMCT/BME/W OT/1002

S9	22131740542	KMCT/BME/MICU/0191
GRAPHNET-TS	120910021083A1V	KMCT/BME/NSICU/0044
NA	NA	NA
TRILLOGY 202	TV01601130D	KMCT/BME/MBICU/0038
SE1200	M17708420001	KMCT/BME/ED/0009
OCM 4008S	2SXA8128	KMCT/BME/DYLS/0327
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
na	na	na
RX IMOLA	72010997	KMCT/BME/CL/0354
MP 1210	AX1405200156	KMCT/BME/W PICU/0852
PREMIER 3000	24429	KMCT/BME/ED/0037
NA	NA	NA
NA	NA	NA
ELSY 360+	1709326	KMCT/BME/MOT/0460
IM8	M15B03300002	KMCT/BME/RICU/0270
NA	NA	NA
NA	NA	NA
MEC1204	CC09114837	KMCT/BME/SICU/0114
VITECT2 COMPACT	VK2C13334	KMCT/BME/MBL/0209
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
SE1200	M17708420001	KMCT/BME/ED/0009
STAR 50N	NA	NA
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
PENTRA XL 80	907PXL8808	KMCT/BME/CL/1169
NA	NA	NA
NA	NA	NA
NA	NA	NA
NA	NA	NA
AGFA35	NA	NA
		KMCT/BME/MBL/0226
LOGIQ F6	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
NA	NA	NA
NA	NA	KMCT/BME/MBL/0226
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
OCM 4008S	2SXA7824	KMCT/BME/DYLS/0326
PREMIER 3000	24429	KMCT/BME/ED/0037
SE-3	M17901470001	KMCT/BME/FMW/0630
	1521207003	KMCT/BME/OPH/0642
APPA OCT 2018	IROCT2202140	KMCT/BME/OPH/1324

EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	DEMO
CR30X	33130	KMCT/BME/RD/1226
PERFUSER COMPACT	234754	KMCT/BME/ED/0011
PREMIER 3000	24429	KMCT/BME/ED/0037
VIVID E9	NA	KMCT/BME/CRD/0255
NA	NA	NA
CR30X	33130	KMCT/BME/RD/1226
NA	NA	NA
PERFUSER COMPACT	215272	KMCT/BME/ICCU/0152
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
MICROSCAN	MY121600142	KMCT/BME/RD/1039
OCM 4008S	2SXA8128	KMCT/BME/DYLS/0327
INFUSOMAT P	120638	KMCT/BME/ED/0013
STERI VAC	359642	KMCT/BME/CSSD/0521
MICROSCAN	MY121600142	KMCT/BME/RD/1039
NA	354679	KMCT/BME/DYLS/0349
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
BENEVIEW T8	CF26111807	KMCT/BME/CVTSICU/0372
PERFUSER COMPACT	NA	NA
NA	NA	NA
NA	NA	
MEDITEC AG	6627500010	KMCT/BME/MOT/0485
MEC1000	AQ1C163866	KMCT/BME/NMICU/0091
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
PERFUSER COMPACT	167255	KMCT/BME/NMICU/0100
A9000	SPR20110211778	KMCT/BME/DYLS/1225
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
NA	NA	NA
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	NA
AT 202		KMCT/BME/CVTSOT/1064
MEC1200	CC11116587	KMCT/BME/SICU/0110
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
NA	NA	NA
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
NA	NA	NA

UMEC10	KN-75009246	KMCT/BME/ED/0004
MEC1200	CC11116595	KMCT/BME/MICU/0173
NA	NA	NA
SIGMADELTA	D0307-0501	KMCT/BME/W OT/1010
CR 35-X	5165	KMCT/BME/RD/0714
LOGIC P5	5372508	KMCT/BME/USG/0026
LOGIC F6.R2	601133WX0	KMCT/BME/USG/1050
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
JG WORK		KMCT/BME/CSSD/0522
RX DAYTONA PLUS	72410266	KMCT/BME/W CL/1163
CR 35-X	5165	KMCT/BME/RD/0714
NA	NA	NA
RX IMOLA	72010997	KMCT/BME/CL/0354
STERI VAC	359642	KMCT/BME/CSSD/0521
NA	NA	NA
CR 30-X	33130	KMCT/BME/RD/1226
NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
RC75800	DAC18	KMCT/BME/BB/0313
PREMIER 3000	24429	KMCT/BME/ED/0037
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
NA	NA	NA
CR 35-X	5165	KMCT/BME/RD/0714
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
na	na	na
RX IMOLA	72010997	KMCT/BME/CL/0354
MP 1210	AX1405200156	KMCT/BME/W PICU/0852
PREMIER 3000	24429	KMCT/BME/ED/0037
NA	NA	NA
NA	NA	NA
ELSY 360+	1709326	KMCT/BME/MOT/0460
IM8	M15B03300002	KMCT/BME/RICU/0270
NA	NA	NA
NA	NA	NA
MEC1204	CC09114837	KMCT/BME/SICU/0114
VITECT2 COMPACT	VK2C13334	KMCT/BME/MBL/0209
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
SE1200	M17708420001	KMCT/BME/ED/0009
STAR 50N	NA	NA

PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
PENTRA XL 80	907PXL8808	KMCT/BME/CL/1169
NA	NA	NA
NA	NA	NA
AGFA35	NA	NA
		KMCT/BME/MBL/0226
LOGIQ F6	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
NA	NA	NA
NA	NA	KMCT/BME/MBL/0226
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
OCM 4008S	2SXA7824	KMCT/BME/DYLS/0326
PREMIER 3000	24429	KMCT/BME/ED/0037
SE-3	M17901470001	KMCT/BME/FMW/0630
	1521207003	KMCT/BME/OPH/0642
APPA OCT 2018	IROCT2202140	KMCT/BME/OPH/1324
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	DEMO
CR30X	33130	KMCT/BME/RD/1226
PERFUSER COMPACT	234754	KMCT/BME/ED/0011
PREMIER 3000	24429	KMCT/BME/ED/0037
VIVID E9	NA	KMCT/BME/CRD/0255
NA	NA	NA
CR30X	33130	KMCT/BME/RD/1226
NA	NA	NA
PERFUSER COMPACT	215272	KMCT/BME/ICCU/0152
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
MICROSCAN	MY121600142	KMCT/BME/RD/1039
OCM 4008S	2SXA8128	KMCT/BME/DYLS/0327
INFUSOMAT P	120638	KMCT/BME/ED/0013
STERI VAC	359642	KMCT/BME/CSSD/0521
MICROSCAN	MY121600142	KMCT/BME/RD/1039
NA	354679	KMCT/BME/DYLS/0349
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
BENEVIEW T8	CF26111807	KMCT/BME/CVTSICU/0372
PERFUSER COMPACT	NA	NA
NA	NA	NA
NA	NA	
MEDITEC AG	6627500010	KMCT/BME/MOT/0485

MEC1000	AQ1C163866	KMCT/BME/NMICU/0091
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
PERFUSER COMPACT	167255	KMCT/BME/NMICU/0100
A9000	SPR20110211778	KMCT/BME/DYLS/1225
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
NA	NA	NA
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	NA
AT 202		KMCT/BME/CVTSOT/1064
MEC1200	CC11116587	KMCT/BME/SICU/0110
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
NA	NA	NA
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
NA	NA	NA
UMEC10	KN-75009246	KMCT/BME/ED/0004
MEC1200	CC11116595	KMCT/BME/MICU/0173
NA	NA	NA
SIGMADELTA	D0307-0501	KMCT/BME/W OT/1010
CR 35-X	5165	KMCT/BME/RD/0714
LOGIC P5	5372508	KMCT/BME/USG/0026
LOGIC F6.R2	601133WX0	KMCT/BME/USG/1050
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
URG WORK		KMCT/BME/CSSD/0522
RX DAYTONA PLUS	72410266	KMCT/BME/W CL/1163
CR 35-X	5165	KMCT/BME/RD/0714
NA	NA	NA
RX IMOLA	72010997	KMCT/BME/CL/0354
STERI VAC	359642	KMCT/BME/CSSD/0521
NA	NA	NA
CR 30-X	33130	KMCT/BME/RD/1226
NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
RC75800	DAC18	KMCT/BME/BB/0313
PREMIER 3000	24429	KMCT/BME/ED/0037
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
NA	NA	NA
LOGIC F6	601660WX0	KMCT/BME/USG/1049
NA	NA	NA

NA	NA	NA
PERFUSER COMPACT	174046	KMCT/BME/CVTSICU/0391
T2100	SBC10326654SAQ	KMCT/BME/CRD/0256
IM8	M15306420004	KMCT/BME/MICU/0178
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
M8ME	9382	KMCT/BME/CL/1173
UMEC10	KN-75009241	KMCT/BME/ED/0001
CR 30-X	33130	KMCT/BME/RD/1226
NA	NA	NA
M8ME	9382	KMCT/BME/CL/1173
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
GRAPHNET-TS		KMCT/BME/NSICU/0045
GRAPHNET-TS		KMCT/BME/NSICU/0044
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
PERFUSER COMPACT	167149	KMCT/BME/NMICU/0096
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
M8ME	9382	KMCT/BME/CL/1173
UMEC10	KN-75009239	KMCT/BME/ED/0002
ELECTROSON-608	TU1-6688	KMCT/BME/PHY/0527
HEMAX 530AL	H53AD1E12	KMCT/BME/W CL/0358
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
MEC1200	CC11116595	KMCT/BME/MICU/0173
GRAPHNET-TS		KMCT/BME/NSICU/0044
NA	NA	NA
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
NA	NA	NA
UF-4000	12068	KMCT/BME/CL/1211
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
PERFUSER COMPACT	167334	KMCT/BME/NMICU/0099
NA	NA	NA
CR 30-X	33130	KMCT/BME/RD/1226
CBS-300	3-Mar-16	KMCT/BME/CL/0356
M8ME	9382	KMCT/BME/CL/1173
CR 35-X	5165	KMCT/BME/RD/0714
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
NA	NA	NA
NA	NA	PFWNB 17
UF-4000	12068	KMCT/BME/CL/1211
PERFUSER COMPACT	234810	KMCT/BME/MICU/0183
VISION2000	151084228	KMCT/BME/CL/0364
		KMCT/BME/W PICU/0864
A9000	SPR20110211778	KMCT/BME/DYLS/1225
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378

NA	NA	NA
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	NA
AT 202		KMCT/BME/CVTSOT/1064
NA	NA	NA
NA	NA	NA
250B		KMCT/BME/W OT/1002
S9	22131740542	KMCT/BME/MICU/0191
GRAPHNET-TS	120910021083A1V	KMCT/BME/NSICU/0044
NA	NA	NA
TRILLOGY 202	TV01601130D	KMCT/BME/MBICU/0038
CR 35-X	5165	KMCT/BME/RD/0714
GRAPHNET-TS	NA	KMCT/BME/NSICU/0045
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
CR 35-X	5165	KMCT/BME/RD/0714
PLANET 45	NA	NA
NA	NA	NA
NA	NA	NA
SKANRESPIRO	MR20CB2979	KMCT/BME/SICU/1231
CBS-300	3-Mar-16	KMCT/BME/CL/0356
MPM 100	AQ16156662	KMCT/BME/ED/0005
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
	EB16JA02538	KMCT/BME/CTW/0952
NA	NA	NA
NA	NA	NA
NA	NA	NA
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
RX IMOLA	72010997	KMCT/BME/CL/0354
ORION PLUS	R14WB1182	KMCT/BME/NSICU/1220
PLANET 55	T07PR1026	KMCT/BME/MOT/0432
CR 35-X	5165	KMCT/BME/RD/0714
A9000	SPR201102111616	KMCT/BME/RICU/1246
CBS-300	3-Mar-16	KMCT/BME/CL/0356
SKANRESPIRO	MR20CA3558	KMCT/BME/MICU/1232
OCM 4008S	5SXAFG51	KMCT/BME/DYLS/0331
SKANRAY	SKANRESPIRO	MR20CA3558
NA	NA	NA
	CS3984	KMCT/BME/CSSD/0526
OCM 4008S	4SXAU833	KMCT/BME/DYLS/0330
NA	NA	NA
	1521207003	KMCT/BME/OPH/0642

NA	NA	NA
OCM 4008S	5SXAFG51	KMCT/BME/DYLS/0331
NA	NA	NA
UMEC10	KN-75009241	KMCT/BME/ED/0001
NA	NA	NA
		KMCT/BME/ENT W/0205
C8C AC7	23462	KMCT/BME/CL/0365
	103E200E538	KMCT/BME/ICCU/0901
EXEL	9421877	KMCT/BME/MMW/0625
STAR 50N	F11S1163	KMCT/BME/NSICU/0041
	93HW10459	KMCT/BME/CSSD/0525
GRAPHNET-TS	120910021083A1V	KMCT/BME/NSICU/0044
UF-4000	12068	KMCT/BME/CL/1211
SE1200	M17708420001	KMCT/BME/ED/0009
CR 30-X	33130	KMCT/BME/RD/1226
OCM 4008S	2SXA7824	KMCT/BME/DYLS/0326
NA	NA	NA
MEC1200	CC11116595	KMCT/BME/MICU/0173
NA	NA	NA
STAR 50N	F11S1186	KMCT/BME/NSICU/0039
NA	NA	NA
SKANRESPIRO	MR20CA3558	NA
NA	NA	NA
NA	NA	NA
AA107	11720	KMCT/BME/OPH/0643
COBAS 121	9715	KMCT/BME/CVTSICU/0382
SE-3	SE3B22211213818	KMCT/BME/ECG/1052
NA	NA	NA
ACCESS 2	504192	KMCT/BME/CL/1244
NA	NA	NA
IREX 10	81071103	KMCT/BME/W NICU/0753
I CARE 10	70071109	KMCT/BME/W NICU/0760
INFUSOMAT P	72557	KMCT/BME/NSICU/0048
ACCULYTE 3P	DJ05L17010008	KMCT/BME/CL/0357
RX IMOLA	72010997	KMCT/BME/CL/0354
STERI VAC	359642	KMCT/BME/CSSD/0521
ARGON SHALYA	NA	NA
NA	NA	NA
OPMI VARIO S8	6629303063	KMCT/BME/NSOT/0421
160 KVA	9390	NA
NA	NA	NA
INFUSOMAT P	72561	KMCT/BME/NSICU/0049
EVADROP SP300	11041280	KMCT/BME/ICCU/0145

SE1200	M17708420001	KMCT/BME/ED/0009
RX IMOLA	72010997	KMCT/BME/CL/0354
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
SE1200	M17708420001	KMCT/BME/ED/0009
ORION PLUS	R14WB1182	KMCT/BME/NSICU/1220
NA	NA	NA
GRAPHNET-TS	120610111083A1V	KMCT/BME/NSICU/0045
NA	NA	NA
PENTRA XL 80	907PXL8808	KMCT/BME/CL/1169
NA	NA	NA
ECG703	ECG7031409016	KMCT/BME/RICU/0289
		KMCT/BME/W OT/1280
SE1200	M17708420001	KMCT/BME/ED/0009
ACCESS 2	504192	KMCT/BME/CL/1244
M8ME	9382	KMCT/BME/CL/1173
OCM 4008S	5SXAFG51	KMCT/BME/DYLS/0331
A9000	SPR201102111614	KMCT/BME/RICU/1247
VISION2000	90825739	KMCT/BME/MBL/0214
NA	NA	NA
SKAN RESPIRO PLUS	MR20CB2979	KMCT/BME/SICU/1231
HEMAX 530AL	H53AD1E12	KMCT/BME/W CL/0358
PRIMER 3000	24429	KMCT/BME/ED/0037
NA	NA	NA
SKANRESPIRO	MR20CB2971	KMCT/BME/RICU/1233
CR 35-X	5165	KMCT/BME/RD/0714
NA	NA	NA
NA	NA	NA
STAR 50 N	F11S1186	KMCT/BME/NSICU/0039
STERI VAC	359642	KMCT/BME/CSSD/0521
PERFUSER COMPACT	234753	KMCT/BME/ED/0010
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
SURDIAL SS PLUS	20CX20317	KMCT/BME/DYLS/1300
NA	NA	NA
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
ELSY 360+	1709326	KMCT/BME/MOT/0460
NA	NA	NA
OCM 4008S	5SXAFL29	KMCT/BME/DYLS/0332
CR 35-X	5165	KMCT/BME/RD/0714
PERFUSER COMPACT	167151	KMCT/BME/NSICU/0051
NA	NA	NA
S430I	S430I0914106	KMCT/BME/RICU/0279

VITECT2 COMPACT	VK2C13334	KMCT/BME/MBL/0209
CR 30-X	33130	KMCT/BME/RD/1226
NA	NA	NA
MEC1200	CC11116599	KMCT/BME/MICU/0174
NA	NA	NA
INFU 505	C02091925	KMCT/BME/MICU/0187
NA	NA	NA
RESPIRONICS	MR20CB2971	KMCT/BME/RICU/1233
NA	NA	NA
CR 30-X	33130	KMCT/BME/RD/1226
M8ME	9382	KMCT/BME/CL/1173
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
PRIMER 3000	24429	KMCT/BME/ED/0037
HEMAX 530AL	H53AD1E12	KMCT/BME/W CL/0358
Logic f6 R2	601660WX0	KMCT/BME/USG/1049
RESPIRONICS	MR20CB2971	KMCT/BME/RICU/1233
OCM 4008S	2SXA8127	KMCT/BME/DYLS/0325
PENTRA XL 80	907PXL8808	KMCT/BME/CL/1169
PRIMER 3000	24429	KMCT/BME/ED/0037
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
ACCESS 2	504192	KMCT/BME/CL/1244
RESPIRONICS	MR21CA3748	KMCT/BME/RICU/1242
SKAN RESPIRO	MR21CA4943	KMCT/BME/MICU/1272
DIGITAL 400	E0953695	KMCT/BME/W OT/1003
4008SNG	1SXA2LQL	KMCT/BME/DYLS/1277
NA	NA	NA
AL700+AL700	NA	KMCT/BME/NSOT/0416
NA	NA	NA
NA	NA	NA
OCM 4008S	4SXAU833	KMCT/BME/DYLS/0330
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
SKAN RESPIRO	MR21CA4943	KMCT/BME/MICU/1272
MEC1200	CC11116592	KMCT/BME/W NICU/0743
RESPIRONICS	C19381813323B	KMCT/BME/RICU/0276
OCM 4008S	2SXA8129	KMCT/BME/DYLS/0328
UF-4000	12068	KMCT/BME/CL/1211
MEC1000	AQ1C163860	KMCT/BME/NMICU/0086
HEMAX 530AL	H53AD1E12	KMCT/BME/W CL/0358
39	3907	KMCT/BME/MBL/0229
EFFICIA DFM 100	CN32604085	KMCT/BME/ICCU/0162
C8C AC7	23463	KMCT/BME/BB/0316
ALPHA ST	83003866 234314	KMCT/BME/RD/1047
	1521207003	KMCT/BME/OPH/0642

ACCESS 2	504192	KMCT/BME/CL/1244
OCM 4008S	2SXA7821	KMCT/BME/DYLS/0329
NA	NA	KMCT/BME/W OT/1280
NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
CR35X	5165	KMCT/BME/RD/0714
GRAPHNET		NA
SKANRESPIRO	NA	NA
RX IMOLA	72010997	KMCT/BME/CL/0354
NA	NA	NA
RX IMOLA	72010997	KMCT/BME/CL/0354
NA	NA	NA
NA	NA	NA
STERI VAC	359642	KMCT/BME/CSSD/0521
NA	NA	NA
NA	NA	NA
COBAS 121	9715	KMCT/BME/CVTSICU/0382
4008S	NA	NA
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
NA	NA	NA
GRAPHNET	NA	NA
ACCULYTE 3P	DJ05L17010008	KMCT/BME/CL/0357
MP 1210	AX1405200156	KMCT/BME/W PICU/0852
A9000	SPR201102111699	KMCT/BME/MICU/1254
SE3	7602880005	KMCT/BME/MICU/0192
EVADROP SP300	11041313	KMCT/BME/ICCU/0150
M8ME	9382	KMCT/BME/CL/1173
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
STRATUS OCT 3000	3001-2197	KMCT/BME/OPH/1176
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
RX IMOLA	72010997	KMCT/BME/CL/0354
CR 35-X	5165	KMCT/BME/RD/0714
160 KVA	9390	NA
OCM 4008S	2SXA8127	KMCT/BME/DYLS/0325
NA	NA	NA
NA	NA	NA
A9000	SPR201102111616	KMCT/BME/RICU/1246
S9	22141107877	KMCT/BME/CTW/0692
OCM 4008S	2SXA8127	KMCT/BME/DYLS/0325
NA	NA	NA
NA	NA	NA
A9000	SPR201102111616	KMCT/BME/RICU/1246
S9	22141107877	KMCT/BME/CTW/0692

NA	NA	NA
NA	NA	NA
NA	354679	KMCT/BME/DYLS/0349
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
PREMIER 3000	24429	KMCT/BME/ED/0037
STERI VAC	359642	KMCT/BME/CSSD/0521
SKANRESPIRO	MR20CB2971	KMCT/BME/COVID/1233
A9000	SPR201102111718	KMCT/BME/RICU/1249
AL 700+AL 700	NA	KMCT/BME/CVTSOT/1065
UC 3500	NA	NA
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
A40	V1473253980E0	KMCT/BME/ED/0023
GM 01BAA	NA	KMCT/BME/ED/0905
RX IMOLA	72010997	KMCT/BME/CL/0354
ULV	NA	KMCT/BME/MOT/0502
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
na	na	na
RX IMOLA	72010997	KMCT/BME/CL/0354
MP 1210	AX1405200156	KMCT/BME/W PICU/0852
PREMIER 3000	24429	KMCT/BME/ED/0037
NA	NA	NA
ACCULYTE 3P	DJ05L17010008	KMCT/BME/CL/0357
DIGITAL 400	E0953695	KMCT/BME/W OT/1003
NA	NA	NA
OCM 4008S	2SXA8128	KMCT/BME/DYLS/0327
PERFUSER COMPACT	234723	KMCT/BME/MICU/0182
CBS-300	3-Mar-16	KMCT/BME/CL/0356
MEC1200	CC11116587	KMCT/BME/SICU/0110
HEMAX 530AL	H53AD1E12	KMCT/BME/W CL/0358
GRAPHNET-TS		KMCT/BME/NSICU/0044
INFUSOMAT P	72557	KMCT/BME/NSICU/0048
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
ACCESS 2	501611	KMCT/BME/CL/0353
GRAPHNET-TS		KMCT/BME/NSICU/0045
SERVO 300	10231	KMCT/BME/NMICU/0274
LOGIC F6R2	601133WX0	KMCT/BME/USG/1050
VISION PRO B		KMCT/BME/CL/1187
MEC1200	CC11116587	KMCT/BME/SICU/0110
CODEMASTER	3714A11364	KMCT/BME/MICU/0190
CS9805C	DVC0060616	KMCT/BME/W OT/1001
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
PRIMER 3000	24429	KMCT/BME/ED/0037
MP1210	AX1405200014	KMCT/BME/W OT/0996

S5	48E02095	KMCT/BME/CVTSOT/1061
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
NA	NA	NA
NA	NA	KMCT/BME/MOT/1282
GRAPHNET -TS	120510151083AIV	KMCT/BME/CVTSICU/0379
HERCULES	C1201702000524	NA
NA	NA	NA
FM301	510118765 816	KMCT/BME/W LR/1170
MEC1200	CC11116587	KMCT/BME/SICU/0110
CR 35X	5165	
NA	NA	NA
IM8	M15B03300002	KMCT/BME/RICU/0270
GRAPHNET -TS	120910021083A1V	KMCT/BME/NSICU/0044
F3	FB2021110368DG	KMCT/BME/W LR/0782
NA	NA	NA
IM8	M15B03300002	KMCT/BME/RICU/0270
FM301	510118765 816	KMCT/BME/W LR/1170
NA	NA	NA
PENTRA XL 80	907PXL8808	KMCT/BME/CL/1169
PENTRA XL 80	907PXL8808	KMCT/BME/CL/1169
ACCESS 2	504192	KMCT/BME/CL/1244
PREMIER 3000	KN-75009241	KMCT/BME/ED/0037
MEC1200	CC11116599	KMCT/BME/MICU/0174
ITY2-TW110XCB	NA	NA
SE-3	7602880005	KMCT/BME/MICU/0192
CR 35-X	5165	KMCT/BME/RD/0714
NA	NA	NA
SE3	7602880005	KMCT/BME/MICU/0192
NA	NA	NA
PREMIER 3000	KN-75009241	KMCT/BME/ED/0037
PREMIER 3000	KN-75009241	KMCT/BME/ED/0037
F3	FB2021110368DG	KMCT/BME/W LR/0782
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
NA	NA	NA
MEC1200	CC11116588	KMCT/BME/MICU/0177
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
NA	NA	NA
NA	448791	KMCT/BME/FMW/1171
MEC1200	CC11116588	KMCT/BME/MICU/0177
MEDIPHARM	438	KMCT/BME/CSSD/0524
INFUSOMAT P	72486	KMCT/BME/ED/1138
UMEC10	KN-75009241	KMCT/BME/ED/0001
NA	NA	NA
MEDIPHARM	438	KMCT/BME/CSSD/1245

M8ME	9382	KMCT/BME/CL/1173
NA	197	NA
CR 35X	5165	KMCT/BME/RD/0714
NA	1218	NA
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
250B	NA	KMCT/BME/MOT/0463
NA	NA	NA
MARS 50	2K16010013-dx/HF	KMCT/BME/RD/0719
NA	NA	NA
HEXALES	C1201702000524	NA
A9000	SPR201102111616	KMCT/BME/RICU/1246
SKANRESPIRO	MR20CB2979	KMCT/BME/SICU/1231
SKANRESPIRO	MR20CB2979	KMCT/BME/SICU/1231
S9	NA	NA
VISION PRO B	NA	KMCT/BME/CL/1187
NA	NA	NA
NA	NA	NA
RESMED	NA	NA
GBCM	2931322878	KMCT/BME/BB/1235
VISION PRO B	NA	KMCT/BME/CL/1187
NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
NA	NA	NA
SURGISKAN 200	E18FA0066	KMCT/BME/W OT/1033
SURGISKAN 200	E18FA0066	KMCT/BME/W OT/1033
NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
NA	20163400012 122	NA
NA	NA	NA
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
OCM 4008S	2SXA8128	KMCT/BME/DYLS/0327
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
VALORY	NA	RDDR 01
STERI VAC	359642	KMCT/BME/CSSD/0521
MEC-1000	AQ1C163870	KMCT/BME/NMICU/0090
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
INFUSOMAT P	120638	KMCT/BME/ED/0013
SE3	7602880005	KMCT/BME/MICU/0192
NA	NA	NA
ACCULYTE 3P	DJ05L17010008	KMCT/BME/CL/0357
DIGITAL 400	E0953695	KMCT/BME/W OT/1003
NA	NA	NA
OCM 4008S	2SXA8128	KMCT/BME/DYLS/0327

PERFUSER COMPACT	234723	KMCT/BME/MICU/0182
CBS-300	3-Mar-16	KMCT/BME/CL/0356
MEC1200	CC11116587	KMCT/BME/SICU/0110
HEMAX 530AL	H53AD1E12	KMCT/BME/W CL/0358
GRAPHNET-TS		KMCT/BME/NSICU/0044
INFUSOMAT P	72557	KMCT/BME/NSICU/0048
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
ACCESS 2	501611	KMCT/BME/CL/0353
GRAPHNET-TS		KMCT/BME/NSICU/0045
SERVO 300	10231	KMCT/BME/NMICU/0274
LOGIC F6R2	601133WX0	KMCT/BME/USG/1050
VISION PRO B		KMCT/BME/CL/1187
S12	S71B0000200	MICUMP01
MEC1200	CC11116587	KMCT/BME/SICU/0110
CODEMASTER	3714A11364	KMCT/BME/MICU/0190
CS9805C	DVC0060616	KMCT/BME/W OT/1001
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
PRIMER 3000	24429	KMCT/BME/ED/0037
MP1210	AX1405200014	KMCT/BME/W OT/0996
S5	48E02095	KMCT/BME/CVTSOT/1061
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
NA	NA	NA
NA	NA	KMCT/BME/MOT/1282
A9000	SPR201102111616	KMCT/BME/RICU/1246
6208 VIEW PLUS	DETE2F25380	PMWECG 06
PREMIER 3000	24429	KMCT/BME/ED/0037
NA	NA	NA
100DX	NA	KMCT/BME/RD/1042
LOGIC F6	601660WX0	KMCT/BME/USG/1049
NA	MY2619R0207	KMCT/BME/RD/1156
STAR 50N	F11S1186	KMCT/BME/NSICU/0039
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
MAC1200ST	550057088	KMCT/BME/ICCU/0161
ORION PLUS	R14WB1182	KMCT/BME/NSICU/1220
PREMIER 3000	24429	KMCT/BME/ED/0037
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
NA	NA	NA
NA	NA	NA
CARDIOLIFE	9263	RICUDF06
OCM 4008S	5SXAFL29	KMCT/BME/DYLS/0332
NA	NA	NA
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160

SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
NA	NA	NA
NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
CR 35-X	5165	KMCT/BME/RD/0714
GRAPHNET-TS	NA	KMCT/BME/NSICU/0045
CR 35-X	5165	KMCT/BME/RD/0714
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
NA	NA	PFWNB 17
CR 35-X	5165	KMCT/BME/RD/0714
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
LOGIC P5	5372508	KMCT/BME/USG/0026
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
CR 35-X	5165	KMCT/BME/RD/0714
PREMIER 3000	NA	KMCT/BME/ED/0037
NA		KMCT/BME/ED/0907
RX IMOLA	72010997	KMCT/BME/CL/0354
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
REVELUTION ACTS	32003BG6	KMCT/BME/RD/0713
C1	NA	KMCT/BME/ICCU/0160
NA		MICUVL02
1.5T HDxt	NA	KMCT/BME/RD/0712
NA	NA	NA
MEC1201	CC11116596	KMCT/BME/SICU/0111
NA	NA	NA
NA	NA	NA
NA	NA	NA
NA	NA	NA
NA	NA	NA
IM8	M16606320007	KMCT/BME/MICU/0179
SONOACE R7	30QQM3HG80034T	KMCT/BME/USG/1051
S12	NA	RICUMP 28
STAR 50N	F11S1163	KMCT/BME/NSICU/0041
FM301	510118765 816	KMCT/BME/W LR/1170
SP102	407122	KMCT/BME/SICU/0121
NA	NA	NA
OCM 4008S	5SXAF129	KMCT/BME/DYLS/0332
CR 35-X	5165	KMCT/BME/RD/0714
PERFUSER COMPACT	167151	KMCT/BME/NSICU/0051
NA	NA	NA
S430I	S430I0914106	KMCT/BME/RICU/0279
VITECT2 COMPACT	VK2C13334	KMCT/BME/MBL/0209
CR 30-X	33130	KMCT/BME/RD/1226
VALORY	NA	KMCT/BME/DR01
NA	NA	NA

MEC1200	CC11116599	KMCT/BME/MICU/0174
NA	NA	NA
INFU 505	C02091925	KMCT/BME/MICU/0187
NA	NA	NA
RESPIRONICS	MR20CB2971	KMCT/BME/RICU/1233
NA	NA	NA
CR 30-X	33130	KMCT/BME/RD/1226
M8ME	9382	KMCT/BME/CL/1173
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
PRIMER 3000	24429	KMCT/BME/ED/0037
S12	S72B0000153	SPOTMP23
STAAN	202212176	SPOTOT 01
EMC 8100A	E06824	PDOPWM 15
ACC001	271080011	KMCT/BME/OPHOT/0514
S12		
CARDIOLIFE		
STERI VAC	359642	KMCT/BME/CSSD/0521
CR 30-X		
WARRIOR W1		
1188	11F014264	KMCT/BME/MOT/0494
SKANRESPIRO	MR20CB2971	KMCT/BME/URO/1233
VALORY	10/13/1902	RDDR 01
PLANET45		
SKANRESPIRO		
SE-3		
CHEM 5 PLUS		
NEUMOVENT		
SKANRESPIRO		
TRILLOGY 202	TV01601130D	KMCT/BME/MBICU/0038
CR 35-X	5165	KMCT/BME/RD/0714
GRAPHNET-TS	NA	KMCT/BME/NSICU/0045
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
CR 35-X	5165	KMCT/BME/RD/0714
PLANET 45	NA	NA
NA	NA	NA
NA	NA	NA
SKANRESPIRO	MR20CB2979	KMCT/BME/SICU/1231
CBS-300	3-Mar-16	KMCT/BME/CL/0356
MPM 100	AQ16156662	KMCT/BME/ED/0005

EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
	EB16JA02538	KMCT/BME/CTW/0952
NA	NA	NA
NA	NA	NA
NA	NA	NA
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
RX IMOLA	72010997	KMCT/BME/CL/0354
ORION PLUS	R14WB1182	KMCT/BME/NSICU/1220
PLANET 55	T07PR1026	KMCT/BME/MOT/0432
CR 35-X	5165	KMCT/BME/RD/0714
A9000	SPR201102111616	KMCT/BME/RICU/1246
CBS-300	3-Mar-16	KMCT/BME/CL/0356
SKANRESPIRO	MR20CA3558	KMCT/BME/MICU/1232
OCM 4008S	5SXAFG51	KMCT/BME/DYLS/0331
SKANRAY	SKANRESPIRO	MR20CA3558
NA	NA	NA
	CS3984	KMCT/BME/CSSD/0526
OCM 4008S	4SXAU833	KMCT/BME/DYLS/0330
NA	NA	NA
	1521207003	KMCT/BME/OPH/0642
NA	NA	NA
OCM 4008S	5SXAFG51	KMCT/BME/DYLS/0331
NA	NA	NA
UMEC10	KN-75009241	KMCT/BME/ED/0001
NA	NA	NA
		KMCT/BME/ENT W/0205
C8C AC7	23462	KMCT/BME/CL/0365
	103E200E538	KMCT/BME/ICCU/0901
EXEL	9421877	KMCT/BME/MMW/0625
STAR 50N	F11S1163	KMCT/BME/NSICU/0041
	93HW10459	KMCT/BME/CSSD/0525
GRAPHNET-TS	120910021083A1V	KMCT/BME/NSICU/0044
UF-4000	12068	KMCT/BME/CL/1211
SE1200	M17708420001	KMCT/BME/ED/0009
CR 30-X	33130	KMCT/BME/RD/1226
OCM 4008S	2SXA7824	KMCT/BME/DYLS/0326
NA	NA	NA
MEC1200	CC11116595	KMCT/BME/MICU/0173
NA	NA	NA
STAR 50N	F11S1186	KMCT/BME/NSICU/0039
NA	NA	NA

SKANRESPIRO	MR20CA3558	NA
NA	NA	NA
NA	NA	NA
AA107	11720	KMCT/BME/OPH/0643
COBAS 121	9715	KMCT/BME/CVTSICU/0382
SE-3	SE3B22211213818	KMCT/BME/ECG/1052
NA	NA	NA
ACCESS 2	504192	KMCT/BME/CL/1244
NA	NA	NA
IREX 10	81071103	KMCT/BME/W NICU/0753
I CARE 10	70071109	KMCT/BME/W NICU/0760
INFUSOMAT P	72557	KMCT/BME/NSICU/0048
ACCULYTE 3P	DJ05L17010008	KMCT/BME/CL/0357
RX IMOLA	72010997	KMCT/BME/CL/0354
STERI VAC	359642	KMCT/BME/CSSD/0521
ARGON SHALYA	NA	NA
NA	NA	NA
OPMI VARIO S8	6629303063	KMCT/BME/NSOT/0421
160 KVA	9390	NA
NA	NA	NA
INFUSOMAT P	72561	KMCT/BME/NSICU/0049
EVADROP SP300	11041280	KMCT/BME/ICCU/0145
SE1200	M17708420001	KMCT/BME/ED/0009
RX IMOLA	72010997	KMCT/BME/CL/0354
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
SE1200	M17708420001	KMCT/BME/ED/0009
ORION PLUS	R14WB1182	KMCT/BME/NSICU/1220
NA	NA	NA
GRAPHNET-TS	120610111083A1V	KMCT/BME/NSICU/0045
NA	NA	NA
PENTRA XL 80	907PXL8808	KMCT/BME/CL/1169
NA	NA	NA
ECG703	ECG7031409016	KMCT/BME/RICU/0289
		KMCT/BME/W OT/1280
SE1200	M17708420001	KMCT/BME/ED/0009
ACCESS 2	504192	KMCT/BME/CL/1244
M8ME	9382	KMCT/BME/CL/1173
OCM 4008S	5SXAFG51	KMCT/BME/DYLS/0331
A9000	SPR201102111614	KMCT/BME/RICU/1247
VISION2000	90825739	KMCT/BME/MBL/0214
NA	NA	NA
SKAN RESPIRO PLUS	MR20CB2979	KMCT/BME/SICU/1231

HEMAX 530AL	H53AD1E12	KMCT/BME/W CL/0358
PRIMER 3000	24429	KMCT/BME/ED/0037
NA	NA	NA
SKANRESPIRO	MR20CB2971	KMCT/BME/RICU/1233
CR 35-X	5165	KMCT/BME/RD/0714
NA	NA	NA
NA	NA	NA
STAR 50 N	F11S1186	KMCT/BME/NSICU/0039
STERI VAC	359642	KMCT/BME/CSSD/0521
PERFUSER COMPACT	234753	KMCT/BME/ED/0010
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
SURDIAL SS PLUS	20CX20317	KMCT/BME/DYLS/1300
NA	NA	NA
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
ELSY 360+	1709326	KMCT/BME/MOT/0460
NA	NA	NA
OCM 4008S	5SXAFL29	KMCT/BME/DYLS/0332
CR 35-X	5165	KMCT/BME/RD/0714
PERFUSER COMPACT	167151	KMCT/BME/NSICU/0051
NA	NA	NA
S430I	S430I0914106	KMCT/BME/RICU/0279
VITECT2 COMPACT	VK2C13334	KMCT/BME/MBL/0209
CR 30-X	33130	KMCT/BME/RD/1226
NA	NA	NA
MEC1200	CC11116599	KMCT/BME/MICU/0174
NA	NA	NA
INFU 505	C02091925	KMCT/BME/MICU/0187
NA	NA	NA
RESPIRONICS	MR20CB2971	KMCT/BME/RICU/1233
NA	NA	NA
CR 30-X	33130	KMCT/BME/RD/1226
M8ME	9382	KMCT/BME/CL/1173
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
PRIMER 3000	24429	KMCT/BME/ED/0037
HEMAX 530AL	H53AD1E12	KMCT/BME/W CL/0358
Logic f6 R2	601660WX0	KMCT/BME/USG/1049
RESPIRONICS	MR20CB2971	KMCT/BME/RICU/1233
OCM 4008S	2SXA8127	KMCT/BME/DYLS/0325
PENTRA XL 80	907PXL8808	KMCT/BME/CL/1169
PRIMER 3000	24429	KMCT/BME/ED/0037
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
ACCESS 2	504192	KMCT/BME/CL/1244
RESPIRONICS	MR21CA3748	KMCT/BME/RICU/1242

SKAN RESPIRO	MR21CA4943	KMCT/BME/MICU/1272
DIGITAL 400	E0953695	KMCT/BME/W OT/1003
4008SNG	1SXA2LQL	KMCT/BME/DYLS/1277
NA	NA	NA
AL700+AL700	NA	KMCT/BME/NSOT/0416
NA	NA	NA
NA	NA	NA
OCM 4008S	4SXAU833	KMCT/BME/DYLS/0330
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
SKAN RESPIRO	MR21CA4943	KMCT/BME/MICU/1272
MEC1200	CC11116592	KMCT/BME/W NICU/0743
RESPIRONICS	C19381813323B	KMCT/BME/RICU/0276
OCM 4008S	2SXA8129	KMCT/BME/DYLS/0328
UF-4000	12068	KMCT/BME/CL/1211
MEC1000	AQ1C163860	KMCT/BME/NMICU/0086
HEMAX 530AL	H53AD1E12	KMCT/BME/W CL/0358
39	3907	KMCT/BME/MBL/0229
EFFICIA DFM 100	CN32604085	KMCT/BME/ICCU/0162
C8C AC7	23463	KMCT/BME/BB/0316
ALPHA ST	83003866 234314	KMCT/BME/RD/1047
	1521207003	KMCT/BME/OPH/0642
ACCESS 2	504192	KMCT/BME/CL/1244
OCM 4008S	2SXA7821	KMCT/BME/DYLS/0329
NA	NA	KMCT/BME/W OT/1280
NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
CR35X	5165	KMCT/BME/RD/0714
GRAPHNET		NA
SKANRESPIRO	NA	NA
RX IMOLA	72010997	KMCT/BME/CL/0354
NA	NA	NA
RX IMOLA	72010997	KMCT/BME/CL/0354
NA	NA	NA
NA	NA	NA
STERI VAC	359642	KMCT/BME/CSSD/0521
NA	NA	NA
NA	NA	NA
COBAS 121	9715	KMCT/BME/CVTSICU/0382
4008S	NA	NA
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
NA	NA	NA
GRAPHNET	NA	NA
ACCULYTE 3P	DJ05L17010008	KMCT/BME/CL/0357

MP 1210	AX1405200156	KMCT/BME/W PICU/0852
A9000	SPR201102111699	KMCT/BME/MICU/1254
SE3	7602880005	KMCT/BME/MICU/0192
EVADROP SP300	11041313	KMCT/BME/ICCU/0150
M8ME	9382	KMCT/BME/CL/1173
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
STRATUS OCT 3000	3001-2197	KMCT/BME/OPH/1176
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
RX IMOLA	72010997	KMCT/BME/CL/0354
CR 35-X	5165	KMCT/BME/RD/0714
160 KVA	9390	NA
OCM 4008S	2SXA8127	KMCT/BME/DYLS/0325
NA	NA	NA
NA	NA	NA
A9000	SPR201102111616	KMCT/BME/RICU/1246
S9	22141107877	KMCT/BME/CTW/0692
OCM 4008S	2SXA8127	KMCT/BME/DYLS/0325
NA	NA	NA
NA	NA	NA
A9000	SPR201102111616	KMCT/BME/RICU/1246
S9	22141107877	KMCT/BME/CTW/0692
NA	NA	NA
NA	NA	NA
NA	354679	KMCT/BME/DYLS/0349
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
PREMIER 3000	24429	KMCT/BME/ED/0037
STERI VAC	359642	KMCT/BME/CSSD/0521
SKANRESPIRO	MR20CB2971	KMCT/BME/COVID/1233
A9000	SPR201102111718	KMCT/BME/RICU/1249
AL 700+AL 700	NA	KMCT/BME/CVTSOT/1065
UC 3500	NA	NA
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
A40	V1473253980E0	KMCT/BME/ED/0023
GM 01BAA	NA	KMCT/BME/ED/0905
RX IMOLA	72010997	KMCT/BME/CL/0354
ULV	NA	KMCT/BME/MOT/0502
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
na	na	na
RX IMOLA	72010997	KMCT/BME/CL/0354
MP 1210	AX1405200156	KMCT/BME/W PICU/0852
DIVA	EATBTB1498	KMCT/BME/W OBGW/0783
BABYLOG8000	ARNB80008	KMCT/BME/W NICU/0768
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
CR 30-X	33130	KMCT/BME/RD/1226

EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
ALPHA ST	8.30039E+13	KMCT/BME/RD/1047
NA	NA	NA
C 48	13253	KMCT/BME/DOT/0084
RESPICARE	NA	NA
GRAPHNET-TS	NA	KMCT/BME/NSICU/0045
PLANET 45	H19AR2002	KMCT/BME/RICU/1127
SKANRESPIRO	MR20CB2979	KMCT/BME/SICU/1231
CR 30-X	33130	KMCT/BME/RD/1226
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
MEDIPHARM	438	KMCT/BME/CSSD/1245
TRUESCOPE 2	M003E000129	KMCT/BME/ICCU/0133
NA	NA	NA
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
NA	NA	NA
NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
CR 35-X	5165	KMCT/BME/RD/0714
GRAPHNET-TS	NA	KMCT/BME/NSICU/0045
CR 35-X	5165	KMCT/BME/RD/0714
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
CR 35-X	5165	KMCT/BME/RD/0714
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
LOGIC P5	5372508	KMCT/BME/USG/0026
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
CR 35-X	5165	KMCT/BME/RD/0714
PREMIER 3000	NA	KMCT/BME/ED/0037
NA		KMCT/BME/ED/0907
RX IMOLA	72010997	KMCT/BME/CL/0354
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
REVELUTION ACTS	32003BG6	KMCT/BME/RD/0713
1.5T HDxt	NA	KMCT/BME/RD/0712
NA	NA	NA
MEC1201	CC11116596	KMCT/BME/SICU/0111
NA	NA	NA
NA	NA	NA
NA	NA	NA
NA	NA	NA
IM8	M16606320007	KMCT/BME/MICU/0179
SONOACE R7	30QQM3HG80034T	KMCT/BME/USG/1051
STAR 50N	F11S1163	KMCT/BME/NSICU/0041
FM301	510118765 816	KMCT/BME/W LR/1170
SP102	407122	KMCT/BME/SICU/0121
SE-3	M17901470001	KMCT/BME/FMW/0630

PREMIER 3000	24429	KMCT/BME/ED/0037
SE3	7602880005	KMCT/BME/MICU/0192
MEC1200	CC11116587	KMCT/BME/SICU/0110
CR 35-X	5165	KMCT/BME/RD/0714
SKAN C	MY2718RO158	KMCT/BME/MOT/0509
CR 35-X	5165	KMCT/BME/RD/0714
A9000	SPR201102111705	KMCT/BME/URO/1250
GE	VIVID E9	KMCT/BME/CRD/0255
A9000	SPR201102111745	KMCT/BME/NSICU/1255
SONOACE R7	30QQM3HG80034T	KMCT/BME/USG/1051
NA	NA	NA
MEC1000	AQ1C163873	KMCT/BME/NMICU/0087
SKANRAY	SKANRESPIRO	MR20CA3558
CR 35-X	5165	KMCT/BME/RD/0714
NA	NA	NA
NA	20163400012 122	NA
NA		NA
PREMIER 3000	24429	KMCT/BME/ED/0037
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
GT-1070	1070L7600270	KMCT/BME/RICU/0889
NA	NA	NA
ELSY 360+	1709326	KMCT/BME/MOT/0460
4008S	NA	NA
NA	NA	NA
NA	NA	NA
NA	NA	NA
VIBRO DROP	VD1909181	KMCT/BME/MOP/1179
PREMIER 3000	24429	KMCT/BME/ED/0037
PREMIER 3000	24429	KMCT/BME/ED/0037
STERI VAC	359642	KMCT/BME/CSSD/0521
NA	NA	NA
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
NA	NA	NA
MEDIPHARM	438	KMCT/BME/CSSD/1245
ORION PLUS	R14WB1182	KMCT/BME/NSICU/1220
NA	NA	KMCT/BME/MOT/0503
GRAPHNET-TS	NA	KMCT/BME/NSICU/0045
CR 35-X	5165	KMCT/BME/RD/0714
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
CR 35-X	5165	KMCT/BME/RD/0714
PLANET 45	NA	NA
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
NA	NA	NA
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008

SE1200	M17708420001	KMCT/BME/ED/0009
10231		KMCT/BME/NMICU/0274
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
	CS3984	KMCT/BME/CSSD/0526
CODEMASTER	3304A13416	KMCT/BME/RICU/0275
	MY2619R0207	KMCT/BME/RD/1156
NA	NA	NA
SURGICO 60H	10K857	KMCT/BME/MOT/0471
PRESERVIA XXL		KMCT/BME/W LR/0787
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
LOGIC F6	601660WX0	KMCT/BME/USG/1049
		KMCT/BME/W OT/1011
EC13	1907143	KMCT/BME/W SICU /1188
NA	NA	NA
MEC1000	NA	KMCT/BME/NMICU/0092
PERFUSER COMPACT	204288	KMCT/BME/W PICU/0857
SE1200	M17708420001	KMCT/BME/ED/0009
CS31+V	CS10571	KMCT/BME/CSSD/1086
SKANRESPIRO PLUS	MR214805	KMCT/BME/SICU/1231
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
NA	NA	NA
SERVO VENTILATOR	10420S11	KMCT/BME/W SICU/0803
NA	NA	NA
NA	NA	NA
SKANRESPIRO	MR20CB2979	KMCT/BME/SICU/1231
CBS-300	3-Mar-16	KMCT/BME/CL/0356
MPM 100	AQ16156662	KMCT/BME/ED/0005
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
	EB16JA02538	KMCT/BME/CTW/0952
NA	NA	NA
NA	NA	NA
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
NA	NA	NA
NA	NA	NA
250B		KMCT/BME/W OT/1002
S9	22131740542	KMCT/BME/MICU/0191
GRAPHNET-TS	120910021083A1V	KMCT/BME/NSICU/0044
NA	NA	NA
TRILLOGY 202	TV01601130D	KMCT/BME/MBICU/0038
SE1200	M17708420001	KMCT/BME/ED/0009
OCM 4008S	2SXA8128	KMCT/BME/DYLS/0327
NA	NA	NA

EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
na	na	na
RX IMOLA	72010997	KMCT/BME/CL/0354
MP 1210	AX1405200156	KMCT/BME/W PICU/0852
PREMIER 3000	24429	KMCT/BME/ED/0037
NA	NA	NA
NA	NA	NA
ELSY 360+	1709326	KMCT/BME/MOT/0460
IM8	M15B03300002	KMCT/BME/RICU/0270
NA	NA	NA
NA	NA	NA
MEC1204	CC09114837	KMCT/BME/SICU/0114
VITECT2 COMPACT	VK2C13334	KMCT/BME/MBL/0209
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
SE1200	M17708420001	KMCT/BME/ED/0009
STAR 50N	NA	NA
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
PENTRA XL 80	907PXL8808	KMCT/BME/CL/1169
NA	NA	NA
NA	NA	NA
NA	NA	NA
NA	NA	NA
AGFA35	NA	NA
		KMCT/BME/MBL/0226
LOGIQ F6	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
NA	NA	NA
NA	NA	KMCT/BME/MBL/0226
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
OCM 4008S	2SXA7824	KMCT/BME/DYLS/0326
PREMIER 3000	24429	KMCT/BME/ED/0037
SE-3	M17901470001	KMCT/BME/FMW/0630
	1521207003	KMCT/BME/OPH/0642
APPA OCT 2018	IROCT2202140	KMCT/BME/OPH/1324
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	DEMO
CR30X	33130	KMCT/BME/RD/1226
PERFUSER COMPACT	234754	KMCT/BME/ED/0011
PREMIER 3000	24429	KMCT/BME/ED/0037
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
SE1200	M17708420001	KMCT/BME/ED/0009

10231		KMCT/BME/NMICU/0274
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
	CS3984	KMCT/BME/CSSD/0526
CODEMASTER	3304A13416	KMCT/BME/RICU/0275
	MY2619R0207	KMCT/BME/RD/1156
NA	NA	NA
SURGICO 60H	10K857	KMCT/BME/MOT/0471
PRESERVIA XXL		KMCT/BME/W LR/0787
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
LOGIC F6	601660WX0	KMCT/BME/USG/1049
		KMCT/BME/W OT/1011
EC13	1907143	KMCT/BME/W SICU /1188
NA	NA	NA
MEC1000	NA	KMCT/BME/NMICU/0092
PERFUSER COMPAC	204288	KMCT/BME/W PICU/0857
SE1200	M17708420001	KMCT/BME/ED/0009
OCM 4008S	2SXA8128	KMCT/BME/DYLS/0327
INFUSOMAT P	120638	KMCT/BME/ED/0013
STERI VAC	359642	KMCT/BME/CSSD/0521
MICROSCAN	MY121600142	KMCT/BME/RD/1039
NA	354679	KMCT/BME/DYLS/0349
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
BENEVIEW T8	CF26111807	KMCT/BME/CVTSICU/0372
PERFUSER COMPAC	NA	NA
NA	NA	NA
NA	NA	
MEDITEC AG	6627500010	KMCT/BME/MOT/0485
MEC1000	AQ1C163866	KMCT/BME/NMICU/0091
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
PERFUSER COMPAC	167255	KMCT/BME/NMICU/0100
A9000	SPR20110211778	KMCT/BME/DYLS/1225
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
NA	NA	NA
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	NA
AT 202		KMCT/BME/CVTSOT/1064
MEC1200	CC11116587	KMCT/BME/SICU/0110
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
NA	NA	NA
LED NOVA 150	SS5909	KMCT/BME/MOT/0501

RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
NA	NA	NA
UMEC10	KN-75009246	KMCT/BME/ED/0004
MEC1200	CC11116595	KMCT/BME/MICU/0173
NA	NA	NA
SIGMADELTA	D0307-0501	KMCT/BME/W OT/1010
CR 35-X	5165	KMCT/BME/RD/0714
LOGIC P5	5372508	KMCT/BME/USG/0026
LOGIC F6.R2	601133WX0	KMCT/BME/USG/1050
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
JG WORK		KMCT/BME/CSSD/0522
RX DAYTONA PLUS	72410266	KMCT/BME/W CL/1163
CR 35-X	5165	KMCT/BME/RD/0714
NA	NA	NA
RX IMOLA	72010997	KMCT/BME/CL/0354
STERI VAC	359642	KMCT/BME/CSSD/0521
NA	NA	NA
CR 30-X	33130	KMCT/BME/RD/1226
NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
RC75800	DAC18	KMCT/BME/BB/0313
PREMIER 3000	24429	KMCT/BME/ED/0037
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
NA	NA	NA
CR 35-X	5165	KMCT/BME/RD/0714
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
na	na	na
RX IMOLA	72010997	KMCT/BME/CL/0354
MP 1210	AX1405200156	KMCT/BME/W PICU/0852
PREMIER 3000	24429	KMCT/BME/ED/0037
NA	NA	NA
NA	NA	NA
ELSY 360+	1709326	KMCT/BME/MOT/0460
IM8	M15B03300002	KMCT/BME/RICU/0270
NA	NA	NA
NA	NA	NA
MEC1204	CC09114837	KMCT/BME/SICU/0114
VITECT2 COMPACT	VK2C13334	KMCT/BME/MBL/0209

SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
SE1200	M17708420001	KMCT/BME/ED/0009
STAR 50N	NA	NA
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
PENTRA XL 80	907PXL8808	KMCT/BME/CL/1169
NA	NA	NA
NA	NA	NA
AGFA35	NA	NA
		KMCT/BME/MBL/0226
LOGIQ F6	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
NA	NA	NA
NA	NA	KMCT/BME/MBL/0226
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
OCM 4008S	2SXA7824	KMCT/BME/DYLS/0326
PREMIER 3000	24429	KMCT/BME/ED/0037
SE-3	M17901470001	KMCT/BME/FMW/0630
	1521207003	KMCT/BME/OPH/0642
APPA OCT 2018	IROCT2202140	KMCT/BME/OPH/1324
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	DEMO
CR30X	33130	KMCT/BME/RD/1226
PERFUSER COMPACT	234754	KMCT/BME/ED/0011
PREMIER 3000	24429	KMCT/BME/ED/0037
VIVID E9	NA	KMCT/BME/CRD/0255
NA	NA	NA
CR30X	33130	KMCT/BME/RD/1226
NA	NA	NA
PERFUSER COMPACT	215272	KMCT/BME/ICCU/0152
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
MICROSCAN	MY121600142	KMCT/BME/RD/1039
OCM 4008S	2SXA8128	KMCT/BME/DYLS/0327
INFUSOMAT P	120638	KMCT/BME/ED/0013
STERI VAC	359642	KMCT/BME/CSSD/0521
MICROSCAN	MY121600142	KMCT/BME/RD/1039
NA	354679	KMCT/BME/DYLS/0349
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
BENEVIEW T8	CF26111807	KMCT/BME/CVTSICU/0372
PERFUSER COMPACT	NA	NA

NA	NA	NA
NA	NA	
MEDITEC AG	6627500010	KMCT/BME/MOT/0485
MEC1000	AQ1C163866	KMCT/BME/NMICU/0091
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
PERFUSER COMPACT	167255	KMCT/BME/NMICU/0100
A9000	SPR20110211778	KMCT/BME/DYLS/1225
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
NA	NA	NA
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	NA
AT 202		KMCT/BME/CVTSOT/1064
MEC1200	CC11116587	KMCT/BME/SICU/0110
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
NA	NA	NA
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
NA	NA	NA
UMEC10	KN-75009246	KMCT/BME/ED/0004
MEC1200	CC11116595	KMCT/BME/MICU/0173
NA	NA	NA
SIGMADELTA	D0307-0501	KMCT/BME/W OT/1010
CR 35-X	5165	KMCT/BME/RD/0714
LOGIC P5	5372508	KMCT/BME/USG/0026
LOGIC F6.R2	601133WX0	KMCT/BME/USG/1050
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
URG WORK		KMCT/BME/CSSD/0522
RX DAYTONA PLUS	72410266	KMCT/BME/W CL/1163
CR 35-X	5165	KMCT/BME/RD/0714
NA	NA	NA
RX IMOLA	72010997	KMCT/BME/CL/0354
STERI VAC	359642	KMCT/BME/CSSD/0521
NA	NA	NA
CR 30-X	33130	KMCT/BME/RD/1226
NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
RC75800	DAC18	KMCT/BME/BB/0313
PREMIER 3000	24429	KMCT/BME/ED/0037
MICRO SCAN	MY121600142	KMCT/BME/RD/1039

NA	NA	NA
LOGIC F6	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
NA	NA	NA
PERFUSER COMPACT	174046	KMCT/BME/CVTSICU/0391
T2100	SBC10326654SAQ	KMCT/BME/CRD/0256
IM8	M15306420004	KMCT/BME/MICU/0178
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
M8ME	9382	KMCT/BME/CL/1173
UMEC10	KN-75009241	KMCT/BME/ED/0001
CR 30-X	33130	KMCT/BME/RD/1226
NA	NA	NA
M8ME	9382	KMCT/BME/CL/1173
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
GRAPHNET-TS		KMCT/BME/NSICU/0045
GRAPHNET-TS		KMCT/BME/NSICU/0044
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
PERFUSER COMPACT	167149	KMCT/BME/NMICU/0096
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
M8ME	9382	KMCT/BME/CL/1173
UMEC10	KN-75009239	KMCT/BME/ED/0002
ELECTROSON-608	TU1-6688	KMCT/BME/PHY/0527
HEMAX 530AL	H53AD1E12	KMCT/BME/W CL/0358
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
MEC1200	CC11116595	KMCT/BME/MICU/0173
GRAPHNET-TS		KMCT/BME/NSICU/0044
NA	NA	NA
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
NA	NA	NA
UF-4000	12068	KMCT/BME/CL/1211
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
PERFUSER COMPACT	167334	KMCT/BME/NMICU/0099
NA	NA	NA
CR 30-X	33130	KMCT/BME/RD/1226
CBS-300	3-Mar-16	KMCT/BME/CL/0356
M8ME	9382	KMCT/BME/CL/1173
CR 35-X	5165	KMCT/BME/RD/0714
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
NA	NA	NA
NA	NA	PFWNB 17
UF-4000	12068	KMCT/BME/CL/1211
PERFUSER COMPACT	234810	KMCT/BME/MICU/0183
VISION2000	151084228	KMCT/BME/CL/0364

		KMCT/BME/W PICU/0864
A9000	SPR20110211778	KMCT/BME/DYLS/1225
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
NA	NA	NA
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	NA
AT 202		KMCT/BME/CVTSOT/1064
NA	NA	NA
NA	NA	NA
250B		KMCT/BME/W OT/1002
S9	22131740542	KMCT/BME/MICU/0191
GRAPHNET-TS	120910021083A1V	KMCT/BME/NSICU/0044
NA	NA	NA
TRILLOGY 202	TV01601130D	KMCT/BME/MBICU/0038
FM301	510118765 816	KMCT/BME/W LR/1170
SP102	407122	KMCT/BME/SICU/0121
SE-3	M17901470001	KMCT/BME/FMW/0630
PREMIER 3000	24429	KMCT/BME/ED/0037
SE3	7602880005	KMCT/BME/MICU/0192
MEC1200	CC11116587	KMCT/BME/SICU/0110
CR 35-X	5165	KMCT/BME/RD/0714
SKAN C	MY2718RO158	KMCT/BME/MOT/0509
CR 35-X	5165	KMCT/BME/RD/0714
A9000	SPR201102111705	KMCT/BME/URO/1250
GE	VIVID E9	KMCT/BME/CRD/0255
A9000	SPR201102111745	KMCT/BME/NSICU/1255
SONOACE R7	30QQM3HG80034T	KMCT/BME/USG/1051
NA	NA	NA
MEC1000	AQ1C163873	KMCT/BME/NMICU/0087
SKANRAY	SKANRESPIRO	MR20CA3558
CR 35-X	5165	KMCT/BME/RD/0714
NA	NA	NA
NA	20163400012 122	NA
NA		NA
PREMIER 3000	24429	KMCT/BME/ED/0037
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
GT-1070	1070L7600270	KMCT/BME/RICU/0889
NA	NA	NA
ELSY 360+	1709326	KMCT/BME/MOT/0460
4008S	NA	NA
NA	NA	NA
NA	NA	NA
NA	NA	NA
VIBRO DROP	VD1909181	KMCT/BME/MOP/1179
PREMIER 3000	24429	KMCT/BME/ED/0037

PREMIER 3000	24429	KMCT/BME/ED/0037
STERI VAC	359642	KMCT/BME/CSSD/0521
NA	NA	NA
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
NA	NA	NA
MEDIPHARM	438	KMCT/BME/CSSD/1245
ORION PLUS	R14WB1182	KMCT/BME/NSICU/1220
NA	NA	KMCT/BME/MOT/0503
GRAPHNET-TS	NA	KMCT/BME/NSICU/0045
CR 35-X	5165	KMCT/BME/RD/0714
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
CR 35-X	5165	KMCT/BME/RD/0714
PLANET 45	NA	NA
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
NA	NA	NA
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
SE1200	M17708420001	KMCT/BME/ED/0009
10231		KMCT/BME/NMICU/0274
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
	CS3984	KMCT/BME/CSSD/0526
CODEMASTER	3304A13416	KMCT/BME/RICU/0275
	MY2619R0207	KMCT/BME/RD/1156
NA	NA	NA
SURGICO 60H	10K857	KMCT/BME/MOT/0471
PRESERVIA XXL		KMCT/BME/W LR/0787
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
LOGIC F6	601660WX0	KMCT/BME/USG/1049
		KMCT/BME/W OT/1011
EC13	1907143	KMCT/BME/W SICU /1188
NA	NA	NA
MEC1000	NA	KMCT/BME/NMICU/0092
PERFUSER COMPACT	204288	KMCT/BME/W PICU/0857
SE1200	M17708420001	KMCT/BME/ED/0009
CS31+V	CS10571	KMCT/BME/CSSD/1086
SKANRESPIRO PLUS	MR214805	KMCT/BME/SICU/1231
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
NA	NA	NA
SERVO VENTILATOR	10420S11	KMCT/BME/W SICU/0803
NA	NA	NA
NA	NA	NA
SKANRESPIRO	MR20CB2979	KMCT/BME/SICU/1231
CBS-300	3-Mar-16	KMCT/BME/CL/0356

MPM 100	AQ16156662	KMCT/BME/ED/0005
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
	EB16JA02538	KMCT/BME/CTW/0952
NA	NA	NA
NA	NA	NA
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
NA	NA	NA
NA	NA	NA
250B		KMCT/BME/W OT/1002
S9	22131740542	KMCT/BME/MICU/0191
GRAPHNET-TS	120910021083A1V	KMCT/BME/NSICU/0044
NA	NA	NA
TRILLOGY 202	TV01601130D	KMCT/BME/MBICU/0038
SE1200	M17708420001	KMCT/BME/ED/0009
OCM 4008S	2SXA8128	KMCT/BME/DYLS/0327
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
na	na	na
RX IMOLA	72010997	KMCT/BME/CL/0354
MP 1210	AX1405200156	KMCT/BME/W PICU/0852
PREMIER 3000	24429	KMCT/BME/ED/0037
NA	NA	NA
NA	NA	NA
ELSY 360+	1709326	KMCT/BME/MOT/0460
IM8	M15B03300002	KMCT/BME/RICU/0270
NA	NA	NA
NA	NA	NA
MEC1204	CC09114837	KMCT/BME/SICU/0114
VITECT2 COMPACT	VK2C13334	KMCT/BME/MBL/0209
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
SE1200	M17708420001	KMCT/BME/ED/0009
STAR 50N	NA	NA
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
PENTRA XL 80	907PXL8808	KMCT/BME/CL/1169
NA	NA	NA
NA	NA	NA
NA	NA	NA
NA	NA	NA
AGFA35	NA	NA
		KMCT/BME/MBL/0226
LOGIQ F6	601660WX0	KMCT/BME/USG/1049
NA	NA	NA

DRYSTAR AXYS	10177	KMCT/BME/RD/0715
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
NA	NA	NA
NA	NA	KMCT/BME/MBL/0226
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
OCM 4008S	2SXA7824	KMCT/BME/DYLS/0326
PREMIER 3000	24429	KMCT/BME/ED/0037
SE-3	M17901470001	KMCT/BME/FMW/0630
	1521207003	KMCT/BME/OPH/0642
APPA OCT 2018	IROCT2202140	KMCT/BME/OPH/1324
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	DEMO
CR30X	33130	KMCT/BME/RD/1226
PERFUSER COMPACT	234754	KMCT/BME/ED/0011
PREMIER 3000	24429	KMCT/BME/ED/0037
VIVID E9	NA	KMCT/BME/CRD/0255
NA	NA	NA
CR30X	33130	KMCT/BME/RD/1226
NA	NA	NA
PERFUSER COMPACT	215272	KMCT/BME/ICCU/0152
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
MICROSCAN	MY121600142	KMCT/BME/RD/1039
OCM 4008S	2SXA8128	KMCT/BME/DYLS/0327
INFUSOMAT P	120638	KMCT/BME/ED/0013
STERI VAC	359642	KMCT/BME/CSSD/0521
MICROSCAN	MY121600142	KMCT/BME/RD/1039
NA	354679	KMCT/BME/DYLS/0349
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
BENEVIEW T8	CF26111807	KMCT/BME/CVTSICU/0372
PERFUSER COMPACT	NA	NA
NA	NA	NA
NA	NA	
MEDITEC AG	6627500010	KMCT/BME/MOT/0485
MEC1000	AQ1C163866	KMCT/BME/NMICU/0091
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
PERFUSER COMPACT	167255	KMCT/BME/NMICU/0100
A9000	SPR20110211778	KMCT/BME/DYLS/1225
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
NA	NA	NA
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	NA

AT 202		KMCT/BME/CVTSOT/1064
MEC1200	CC11116587	KMCT/BME/SICU/0110
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
NA	NA	NA
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
NA	NA	NA
UMEC10	KN-75009246	KMCT/BME/ED/0004
MEC1200	CC11116595	KMCT/BME/MICU/0173
NA	NA	NA
SIGMADELTA	D0307-0501	KMCT/BME/W OT/1010
CR 35-X	5165	KMCT/BME/RD/0714
LOGIC P5	5372508	KMCT/BME/USG/0026
LOGIC F6.R2	601133WX0	KMCT/BME/USG/1050
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
NG WORK		KMCT/BME/CSSD/0522
RX DAYTONA PLUS	72410266	KMCT/BME/W CL/1163
CR 35-X	5165	KMCT/BME/RD/0714
NA	NA	NA
RX IMOLA	72010997	KMCT/BME/CL/0354
STERI VAC	359642	KMCT/BME/CSSD/0521
NA	NA	NA
CR 30-X	33130	KMCT/BME/RD/1226
NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
RC75800	DAC18	KMCT/BME/BB/0313
PREMIER 3000	24429	KMCT/BME/ED/0037
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
NA	NA	NA
CR 35-X	5165	KMCT/BME/RD/0714
SKANRESPIRO PLUS	MR21CA3748	KMCT/BME/RICU/1242
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
na	na	na
RX IMOLA	72010997	KMCT/BME/CL/0354
MP 1210	AX1405200156	KMCT/BME/W PICU/0852
PREMIER 3000	24429	KMCT/BME/ED/0037
NA	NA	NA

NA	NA	NA
ELSY 360+	1709326	KMCT/BME/MOT/0460
IM8	M15B03300002	KMCT/BME/RICU/0270
NA	NA	NA
NA	NA	NA
MEC1204	CC09114837	KMCT/BME/SICU/0114
VITECT2 COMPACT	VK2C13334	KMCT/BME/MBL/0209
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
SE1200	M17708420001	KMCT/BME/ED/0009
STAR 50N	NA	NA
PRIMER 3000	NA	KMCT/BME/CVTS ICU/1209
PENTRA XL 80	907PXL8808	KMCT/BME/CL/1169
NA	NA	NA
NA	NA	NA
AGFA35	NA	NA
		KMCT/BME/MBL/0226
LOGIQ F6	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
NA	NA	NA
NA	NA	KMCT/BME/MBL/0226
DRYSTAR AXYS	10177	KMCT/BME/RD/0715
OCM 4008S	2SXA7824	KMCT/BME/DYLS/0326
PREMIER 3000	24429	KMCT/BME/ED/0037
SE-3	M17901470001	KMCT/BME/FMW/0630
	1521207003	KMCT/BME/OPH/0642
APPA OCT 2018	IROCT2202140	KMCT/BME/OPH/1324
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	DEMO
CR30X	33130	KMCT/BME/RD/1226
PERFUSER COMPACT	234754	KMCT/BME/ED/0011
PREMIER 3000	24429	KMCT/BME/ED/0037
VIVID E9	NA	KMCT/BME/CRD/0255
NA	NA	NA
CR30X	33130	KMCT/BME/RD/1226
NA	NA	NA
PERFUSER COMPACT	215272	KMCT/BME/ICCU/0152
NA	NA	NA
SE1200	M17708420001	KMCT/BME/ED/0009
NA	NA	NA
MICROSCAN	MY121600142	KMCT/BME/RD/1039
OCM 4008S	2SXA8128	KMCT/BME/DYLS/0327

INFUSOMAT P	120638	KMCT/BME/ED/0013
STERI VAC	359642	KMCT/BME/CSSD/0521
MICROSCAN	MY121600142	KMCT/BME/RD/1039
NA	354679	KMCT/BME/DYLS/0349
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
BENEVIEW T8	CF26111807	KMCT/BME/CVTSICU/0372
PERFUSER COMPACT	NA	NA
NA	NA	NA
NA	NA	
MEDITEC AG	6627500010	KMCT/BME/MOT/0485
MEC1000	AQ1C163866	KMCT/BME/NMICU/0091
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
PERFUSER COMPACT	167255	KMCT/BME/NMICU/0100
A9000	SPR20110211778	KMCT/BME/DYLS/1225
GRAPHNET-TS	120510651083AIV	KMCT/BME/CVTSICU/0378
NA	NA	NA
GRAPHNET-TS	120510151083AIV	KMCT/BME/CVTSICU/0379
NA	NA	NA
AT 202		KMCT/BME/CVTSOT/1064
MEC1200	CC11116587	KMCT/BME/SICU/0110
NA	NA	NA
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
NA	NA	NA
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
NA	NA	NA
UMEC10	KN-75009246	KMCT/BME/ED/0004
MEC1200	CC11116595	KMCT/BME/MICU/0173
NA	NA	NA
SIGMADELTA	D0307-0501	KMCT/BME/W OT/1010
CR 35-X	5165	KMCT/BME/RD/0714
LOGIC P5	5372508	KMCT/BME/USG/0026
LOGIC F6.R2	601133WX0	KMCT/BME/USG/1050
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
URG WORK		KMCT/BME/CSSD/0522
RX DAYTONA PLUS	72410266	KMCT/BME/W CL/1163
CR 35-X	5165	KMCT/BME/RD/0714
NA	NA	NA
RX IMOLA	72010997	KMCT/BME/CL/0354
STERI VAC	359642	KMCT/BME/CSSD/0521
NA	NA	NA
CR 30-X	33130	KMCT/BME/RD/1226

NA	NA	NA
PREMIER 3000	24429	KMCT/BME/ED/0037
LOGIC F6.R2	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
RC75800	DAC18	KMCT/BME/BB/0313
PREMIER 3000	24429	KMCT/BME/ED/0037
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
NA	NA	NA
LOGIC F6	601660WX0	KMCT/BME/USG/1049
NA	NA	NA
NA	NA	NA
PERFUSER COMPACT	174046	KMCT/BME/CVTSICU/0391
T2100	SBC10326654SAQ	KMCT/BME/CRD/0256
IM8	M15306420004	KMCT/BME/MICU/0178
EFFICIA DFM 100	32605475	KMCT/BME/ED/0008
M8ME	9382	KMCT/BME/CL/1173
UMEC10	KN-75009241	KMCT/BME/ED/0001
CR 30-X	33130	KMCT/BME/RD/1226
NA	NA	NA
M8ME	9382	KMCT/BME/CL/1173
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
GRAPHNET-TS		KMCT/BME/NSICU/0045
GRAPHNET-TS		KMCT/BME/NSICU/0044
LED NOVA 150	SS5909	KMCT/BME/MOT/0501
PERFUSER COMPACT	167149	KMCT/BME/NMICU/0096
MICRO SCAN	MY121600142	KMCT/BME/RD/1039
M8ME	9382	KMCT/BME/CL/1173
UMEC10	KN-75009239	KMCT/BME/ED/0002
ELECTROSON-608	TU1-6688	KMCT/BME/PHY/0527
HEMAX 530AL	H53AD1E12	KMCT/BME/W CL/0358
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
MEC1200	CC11116595	KMCT/BME/MICU/0173
GRAPHNET-TS		KMCT/BME/NSICU/0044
NA	NA	NA
RESPIRONICS	C17953779695B	KMCT/BME/ICCU/0164
EVADROP SP300	11041279	KMCT/BME/ICCU/0144
NA	NA	NA
UF-4000	12068	KMCT/BME/CL/1211
SERVO VENTILATOR	18566S11	KMCT/BME/ICCU/0160
PERFUSER COMPACT	167334	KMCT/BME/NMICU/0099
NA	NA	NA
CR 30-X	33130	KMCT/BME/RD/1226
CBS-300	3-Mar-16	KMCT/BME/CL/0356
M8ME	9382	KMCT/BME/CL/1173
CR 35-X	5165	KMCT/BME/RD/0714

[illegible]

[illegible]

[illegible]

[illegible]

NATURE OF COMPLAINT
Printer issue
Need to check
High airway pressure alarm
Cassette jam
CALIBRATION OUT
Need to change bucky
Need to refill
Not working
Need to change EPAP valve
Need to calibrate
SPO2 probe not working
Need to check&shift to muthalam hospital
Cassette jam
Tube disconnect alarm
Dry phase not getting started
Not getting ECG&SPO2
Need to setup the machine
Airway pressure alarm
Need to set ventilator
Low water level
Low water level
Software error
Cassete stuck
Low o2 concentration
cassette jam
Need to change catridge
Cassatte jam
NEED TO READ BARCODE
IMAGE NOT CLEAR
not working
Cassatte jam
OUT OF PAPER
BULB WAS NOT WORKING
Error showing
Expiratory minute volume zero
Software issue
Printer issue
Not working
Display not working
Not working
Low water level
Not working

LOW WATER LEVEL
ECG NOT WORKING
Printer not working
ECG NOT WORKING
PROBE NOT WORKING
NOT WORKING PROPERLY
ECG NOT WORKING
CATRIDGE COMPLAINT
LEAD NOT WORKING
NIBP NOT WORKING
CASSATTE JAM
NOT WORKING
CASSATTE JAM
NEEDS MONITOR
SOME KEYS IN ALPHA NUMERIC NOT WORKING
SPO2 NOT WORKING
Printer not working
Low water level
SPO2 probe not working
LOW TIDAL VOLUME
CASSATTE JAM
Need to arrange &set
Alarm
Bag leakage
Power issuse
Not working
E1 ERROR
o2 reading tube broken
bipolar not working
Low water level
SPO2 probe not working
temperature cannot set in warmer
need a standby for emergency patient shifting
Need to set and check
Power fuilare need to calibrate
Software issue
E25 error
Gauge damage
Airway pressurec alarm
Low water level
Leak from the watertank
R3 lead off
Not getting enough pressure
Low o2 concentration
cassette jam

Need to change catridge
Cassatte jam
SPO2 probe not working
Flow sensor complaint
LOW WATER LEVEL
UPS DOWN
NO ECG WAVE
LEAD NOT READY
NEED TO SET
LOW WATER LEVEL
WAVE NOT GETTING
NEED TO ADJUST PRESSURE
BELT LOOSE
SHOWING ERROR
SHOWING ERROR
LOW WATER LEVEL
NOT WORKING
TEMP ISSUE
O2 LEAK
MEMORY FULL
REMOTE NOT WORKING
WAVE NOT GETTING
Low water level
NIBP ISSUE
NEED CALIBRATION
V1 LEAD NOT GETTING
NOT WORKING
TECHNICAL ISSUE
NOT WORKING
LOW PRESSURE
NEED TO SETUP
NEED TO REFILL
SHOW ERROR
WANT TO SETUP
TEST FAILED
SPO2 NOT WORKING
LEAK
NOT WORKING
NOT WORKING
NEED TO SETUP
NIBP NOT WORKING
NOT WORKING
Low water level
BELT ISSUE
PATIENT PLATE LOOSE CONTACT

LEAK
Need to arrange
Not working properly
Check circuit down
Lead off
E13 Error
Low water level
need pappe
need to chek
PRESSURE ADJUST
ECG NOT WORKING
CHANGING CARTIAGE
CHEKKING
NEED WATER
BIPOLAR NOT WORKING
NOT WORKING
NEED WATER
NEED TO BIPAP
NIBP not working
Not ready mode
Need to arrange for emergency purpose.
Not working
Showing ECG variation
Catridge expired
Thermostat compartment temperature out of range
NOT WORKING
LAMP NOT WORKING
NOT WORKING
NOT WORKING
CASET JAM
NOT WORKING
MEMORY FULL
RO WATER LEVEL LOW
CASSETTE jam
CASSETTE jam
RO WATER LEVEL LOW
NOT WORKING
CASSETTE jam
MACHINE DOWN
NEED TO CHANGE CATRIDGE
ECG NOT GETTING
ALIT LAMP CARRIER ALIGNMENT NOT WORKING PROPERLY
CHIN SET TABLE ALIGNMENT NOT WORKING PROPERLY

NEED TO CALIBRATE
O2 CELL NOT SHOWING VALUES
NEED TO SET
Cassatte jam
Not working
catrige expired
KEYPAD NOT WORKIND
Low water level
Cassette jam
showing error
NOT WORKING
FOR DEMO PURPOSE
Artifact
Need new holder
Not working
ERROR 050
Air alarm
C3 ,E12 ERROR
Continuousn expose
Need to change bulbb
ARRANGE FROM ICCU
ARRANGED FROM NSOT
ARRANGED FROM MOT
NOT WORKING
NOT WORKING
lamp not working
NEED TO ARRAGED
Low water level
NEED TO ARRAGED
NEED TO ARRAGED
ECG&SPO2 not on display
O2 leak
Need to set
Caliberation failure
NEED TO ARRAGED
Not working
not connected to pressure
Low water level
Calibration needed
FUSE NEED TO CHANGE
Lamb fuse
Display problem
Need to set
Need to set
NEED TO ARRANGE AND SET

NOT WORKING
NEED TO ARRANGE AND SET ,NIBP NOT WORKING
Low water level
NOT WORKING
Cassette jam
LINEAR PROBE COMPLAINT
Machine stuck automatically
NEED TO SET PRINTER
Coil damage..need to replace coil
Not working
Cassatte jam
Need to refilling
Water alarm.ro plant issue
E25 Error
Need to refilling
Cassatte jam
Need to set
PRINTER ISSUE
Artifact
Need to set
Diode complaint
Need to insert new printer paper
CN010 ERROR
Need to refilling
Cassatte jam
Need to set
Need to refilling
Lead off
need pappe
need to chek
PRESSSURE ADJEST
ECG NOT WORKING
CHANGING CARTIAGE
CHEKKING
NEED WATER
BIPOLAR NOT WORKING
NOT WORKING
NEED WATER
NEED TO BIPAP
NIBP not working
Not ready mode
Need to arrange for emergency purpose.
Not working
Showing ECG variation

Catridge expired
Thermostat compartment temperature out of range
NOT WORKING
LAMP NOT WORKING
CASET JAM
NOT WORKING
MEMORY FULL
RO WATER LEVEL LOW
CASSETTE jam
CASSETTE jam
RO WATER LEVEL LOW
NOT WORKING
CASSETTE jam
MACHINE DOWN
NEED TO CHANGE CATRIDGE
ECG NOT GETTING
ALIT LAMP CARRIER ALLIGNMENT NOT WORKING PROPERLY
CHIN SET TABLE ALLIGNMENT NOT WORKING PROPERLY
NEED TO CALLIBRATE
O2 CELL NOT SHOWING VALUES
NEED TO SET
Cassatte jam
Not working
catrige expired
KEYPAD NOT WORKIND
Low water level
Cassette jam
showing error
NOT WORKING
FOR DEMO PURPOSE
Artifact
Need new holder
Not working
ERROR 050
Air alarm
C3 ,E12 ERROR
Continuousn expose
Need to change bulbb
ARRANGE FROM ICCU
ARRANGED FROM NSOT
ARRANGED FROM MOT
NOT WORKING
NOT WORKING
lamp not working

NEED TO ARRAGED
Low water level
NEED TO ARRAGED
NEED TO ARRAGED
ECG&SPO2 not on display
O2 leak
Need to set
Caliberation failure
NEED TO ARRAGED
Not working
not connected to pressure
Low water level
Calibration needed
FUSE NEED TO CHANGE
Lamb fuse
Display problem
Need to set
Need to set
NEED TO ARRANGE AND SET
NOT WORKING
NEED TO ARRANGE AND SET ,NIBP NOT WORKIN
Low water level
NOT WORKING
Cassette jam
LINEAR PROBE COMPLAINT
Machine stuck automatically
NEED TO SET PRINTER
Coil damage.need to replace coil
Not working
Cassatte jam
Need to refilling
Water alarm.ro plant issue
E25 Error
Need to refilling
Cassatte jam
Need to set
PRINTER ISSUE
Artifact
Need to set
Diode complaint
Need to insert new printer paper
CN010 ERROR
Leakage
MEMORY FULL
COLAER INFORMED TO SERVICER

NEED TO WATER
MOT WORKING
FILLTER PROBLEM
SPO2 NOT WORKING
Need to calibration
Need to change carbon brush
ECG cable not working
Cassatte jam
Need to change valve
Speed sensor not working properly
CN010 ERROR
Calibration failure
Need to calibration
Monitor issue
Need to checke
KB010 Disconnected
Loud noise due improper alignment
ECG not working
Wire disconnected
Not working need to check
Need to bipap
ECG not working
Need to calibrate
Water level low
Power failure
To change unit
O2 Level need to check
Not working
Low o2 alarm expiratory mint volume
Low battery
Low water level
Casset jam
showing error
Loud noise
Cassete jam
Power failure
To change unit
O2 Level need to check
Wire burning
Not working
NOT WORKING
NOT CLARITY
NOT WORKING
ECG&SPO2 not on display
O2 leak

Need to set
Caliberation failure
NEED TO ARRAGED
Not working
Low water level
BELT ISSUE
PATIENT PLATE LOOSE CONTACT
LEAK
Need to arrange
Not working properly
Check circuit down
Cassete stuck
Low o2 concentration
NEW BLOCK FILING
Need to change catridge
Cassatte jam
SPO2 probe not working
NEED TO REFILL
SHOW ERROR
WANT TO SETUP
TEST FAILED
SPO2 NOT WORKING
NOT WORKING
NOT WORKING
NEED TO SETUP
NIBP NOT WORKING
Need to set and arrange
AIR out let complante
System stuck
Not getting current ecg waveform
ETCO2 Module
Cassette jam
Need to arrange monitor
NEED TO CHANGE PACK
LOW O2 CONCENTRATION
BLOOD IN THE REANSDOTER
LOW AIR WAY PRESSURE
NEED TO CONECT CO2
BELT CAME OUT
BLOOD ON THE TRAMSUCTION
LL OFF
ADJESTINH PROB NOT WRKING PROPPERLLY

SCREW BROKEN
BLOOD IN THE REANSDOTER
NEED TO WATER
LOOSE CONTACT
REFILLED
SUCTION WAS NOT GETTING
CARBON BRUSH GONE
Display problem
MERCURY LEAK
NIB CABLE DAMAGE
HEAT WAS NOT GETTING
Need to arrange
Not working
Lead off
CASET JAM
SERVICER COME FOR INSTALLATION
Low water level
NIBP NOT WORKING
FILTER NNED TO CHANGE
SPO2 probe not working
O2 LEAK
NEED TO SET
LOW WATER LEVEL
SPO2 NOT WORKING
lamp not working properly
Catridge expired
Not getting on
Tube disconnect alarm
RV clean out error
Low water level
Need to check all phototherapy
Showing low pressure,
Need to set
showing error
Error showing
Error showing ,C5, C4 ,C11
Not working
Pressure low
Lamp not working properly
UPS down
Low water level
Continuous alarm showing
Syringe error showing

Lead off
Purified water producing device is lacking in water su
Flow not proper
Waves not showing
Printer issue
Display not showing
Waves not showing
Waves not showing
Low water level
Need to setup humidifier
Low water level
Sample struck.
Low water level
Printer problem
low gas pressure in insufflator
Leads not getting
data base error
Need to change carbon brush
conductivity error
ECG waves forms not getting properly
lense not clear
Low water level
Need to set
Sample stuck
Power failure
LOOSE CONTACT bulb
LOW AIRWAY PRESSURE
Cassette jam
loose connection
Low water level
Power failure
C5,C12 ERROR
DIGITAL PROBLEM
Waves not showing
Low water level
Showing conductivity error
Low water level
NEED TO SET
BIPOLAR NOT WORKING
Low water level
SERVICER COME FOR INSTALLATION
Cassette jam
NEED TO SET
Low water level
SHOWING ALARM

Host failure alarm
Cassatte jam
Low water level
SPO2 probe not working
Low water level
Need to set
OT wall damaged
need to set
low air supply
Low water level
Cassette jam
Not working
LOW AIR WAY PRESSURE
parameter not showing
need cleaning
STORAGE FULL
Airway pressure
LOW Tem
SHOWING ALARM
NEED TO CHANGE CATRIDGE
High airway pressure alarm
Tray not removing from system
Flow sensor complaint
High airway pressure alarm
MONO POLAR NOT WORKING
LOWER TEMPERTURE
LEAK
Lamp not working properly
UP AND DOWN POSITION NOT WORKING
Low water level
LOWER TEMPERTURE
PRESSURE HIGH
FILM JAM
FOR SETTING VENTILATOR
NIBP NOT WORKING
LEAK ISSUE
BLOOD ENTERING TO THE PROTECTOR
PRESSURE VARIATION
LOOSE CUFF
showing error
POWER CORD DAMAGED
SHOCK NOT DELIVERED
NOT WORKING
NOT GETTING POWER SUPPLY
BULB NOT WORKING

SOFTWARE ISSUE
BLOOD ENTERING TO THE PROTECTOR
INSUFFLATOR NOT WORKING
Low water level
NEED TO CHANGE PRINTER PAPER
Cassete stuck
low air pressure alarm
to check the flow sensor
To check
not working
Database error
not working
fumes not coming
C11 ALARM
not working
Hinges broken
PCO2 failed to calibrate
Display not working
Keyboard not working
water level too low
To ON the plant for cleaning
O2 low alarm
Calibration out and need to change pack
Need to arrange and set
ECG lead off and settings full changed.
Printing problem and printer door not closed.
HIGH PRESSURE
NOT WORKING
NOT WORKING
FOR CALIBRATION
Software issue
Low air pressure
RE substance capacity error
Cassette jam
Need to change battery
NOT WORKING
Low water level
NOT GETTING CONTINUITY
SPO2 NOT WORKING
NEED TO SETUP
NOT WORKING
Low water level
NOT GETTING CONTINUITY
SPO2 NOT WORKING
NEED TO SETUP

NOT WORKING
low water level
Bag leake
Error CN010
Printer issue
ERROR 25
Arrange&set
Not getting ECG
Wire desoldered.
Sample stuck
for set it for emergency treatment
Low air pressure alarm
Error showing while switching ON
Full capacity sensor of sub tank doesnot give ON sign
Not working
need pappe
need to chek
PRESSURE ADJEST
ECG NOT WORKING
CHANGING CARTIAGE
CHEKKING
ERROR MOVE A
NEED TO CHECK
Low water level
NEED TO CHECK
NOT WORKING
SHOWING ERROR
SPO2 not working
SHOWING ERROR
CALIBRATION OUT
NEED TO SET
FILM JAM
RV BLOCK
NEED TO CALIBRATE
TECHNICAL ALARM , FUSE DAMAGED
NEED TO INSTALL NEW PROBE
SAMPLE STUCK
ECG cable not working
CALIBRATION OUT
NEED TO FIX CABLE CONNECTION
NOT GETTING PROPER WAVE
Low water level
MAXIMUM SAMPLE REACHED
ECG cable not working

NEED TO FIX LIGHT HANDLE
CALIBRATION FAILURE
NOT GETTING ENOUGH PRESSURE
NITROUS CYLINDER LEAK
Showing low pressure
Need to check all battery
Water refilling
Adapter complaint
Not getting ECG
cassate jam
Termometer battery weak
Need to check ECG cable
Calibration failed
Printer issue
Need to arrange emergency
ECG Waves are not proper
Not working
Air valve leakage
Mode unvisible
Mode unvisible
INSTALLATION
All electrode closed
Need to check SPO2
Showing battery alarm
Need to set rhythm lead
cassate jam
Low water level
Mode changed
Need to arrange emergency
Calibration failed
Result showing low value
Wave speed low
Need to arrange emergency
Low level of water
Monitor display not working
Need to set ventilator
Low water level
BP cuff and bulb damaged
ECG Waves are not proper
temperature not working
Need to set properly
Need to conect new ECG cable
Low water level
installation

Carbon sensor damaged&sensor loose
Mercury not raising properly
Cassatte jam
Mercury leak
Need to check
Error patient plate
Need to change a filter
Chest x ray stanel screw missing
Low water level
Battery drain and machine automatically off
monitor automatically
change connector set
Need to set ventilator
need to set
Machine showing error
Low water level
Low water level
Need to arrange for emergency
Not adjust pressure
Sample stuck
Low water level
Need to change ABG paper
Need to set
Low water level
Bipolar cable not working.
Monopolar not working
handle problem
Power failure
Over load
Need to arrange emergency
Expiratory minute volume error
water alarm
Low water level
ECG error
Detector connection failed
E29 error
Need to set monitor
O2 wave alarm
Need to set
Printer issue
CHEKKING
ERROR MOVE A
NEED TO CHECK
Low water level
NEED TO CHECK

NOT WORKING
SHOWING ERROR
SPO2 not working
SHOWING ERROR
CALIBRATION OUT
NEED TO SET
FILM JAM
RV BLOCK
NEED TO CALIBRATE
TECHNICAL ALARM , FUSE DAMAGED
NEED TO INSTALL NEW PROBE
SAMPLE STUCK
ETCO2 not working
ECG cable not working
CALIBRATION OUT
NEED TO FIX CABLE CONNECTION
NOT GETTING PROPER WAVE
Low water level
MAXIMUM SAMPLE REACHED
ECG cable not working
NEED TO FIX LIGHT HANDLE
CALIBRATION FAILURE
NOT GETTING ENOUGH PRESSURE
NITROUS CYLINDER LEAK
NOT WORKING ECG
LEADS PORT COMPLAINT
SHOWING ERROR
NEED TO SETUP
HANDPIECE NOT WORKING
STORAGE FULL
REMOTE NOT WORKING
SPO2 NOT WORKING
LOW O2 SUPPLY
LEADS not getting
LEADS not getting
CATRIDGE EXPIRED
SHOWING ERRORS
CALIBRATION FAILURE
NOT WORKING
NOT WORKING PROPERLY
Calibration failed
FLOW ALARM
Need to setup the machine
Airway pressure alarm

Need to set ventilator
Low water level
Low water level
Software error
Cassete stuck
Low o2 concentration
cassette jam
Need to change catridge
Wire burning
Cassatte jam
NEED TO READ BARCODE
IMAGE NOT CLEAR
not working
Cassatte jam
OUT OF PAPER
BULB WAS NOT WORKING
Error showing
Expiratory minute volume zero
Software issue
Expiratory minute volume zero
Flow sensor complaint
Printer issue
Not working
Display not working
Not working
Low water level
Not working
LOW WATER LEVEL
WHITE BALANCE
ECG NOT WORKING
Printer not working
ECG NOT WORKING
ECG NOT WORKING
PROBE NOT WORKING
NOT WORKING PROPERLY
Low water level
SERVICER COME FOR INSTALLATION
Cassette jam
NEED TO SET
Low water level
SHOWING ALARM
Host failure alarm
Cassatte jam
Work station hang
Low water level

SPO2 probe not working
Low water level
Need to set
OT wall damaged
need to set
low air supply
Low water level
Cassette jam
Not working
LOW AIR WAY PRESSURE
parameter not showing
ETCO2 WAVE FORM ISSUE
POWER SUPPLY PLUG PIN ISSUE
NEED TO INSTALLATION

FOOT SWICTH COMPLAINT
LEADS ERROR IN E.C.G WAVES
PRINTING ISSUE
SENSOR ISSUE
CUFF CHANGED
PRINT ERASE ISSUE
SOCKET ISSUE
HAND PIECE CABLE LOOSE CONTACT ISSUE
Flow sensor complaint
ALLIGMENT ISSUE
IBP SETTING ISSUE
AIR KEY NEEDS TO CHANGED
PRINT ISSUE
ASPIRATION PROBLEM
OXYGEN CONCENTRATION PROBLEM
AIR KEY NEEDS TO CHANGED
Check circuit down
Cassete stuck
Low o2 concentration
NEW BLOCK FILING
Need to change catridge
Cassatte jam
SPO2 probe not working
NEED TO REFILL
SHOW ERROR
WANT TO SETUP
TEST FAILED
SPO2 NOT WORKING

NOT WORKING
NOT WORKING
NEED TO SETUP
NIBP NOT WORKING
Need to set and arrange
AIR out let complante
System stuck
Not getting current ecg waveform
ETCO2 Module
Cassette jam
Need to arrange monitor
NEED TO CHANGE PACK
LOW O2 CONCENTRATION
BLOOD IN THE REANSDOTER
LOW AIR WAY PRESSURE
NEED TO CONECT CO2
BELT CAME OUT
BLOOD ON THE TRAMSUCTION
LL OFF
ADJESTINH PROB NOT WRKING PROPPERLLY
SCREW BROKEN
BLOOD IN THE REANSDOTER
NEED TO WATER
LOOSE CONTACT
REFILLED
SUCTION WAS NOT GETTING
CARBON BRUSH GONE
Display problem
MERCURY LEAK
NIB CABLE DAMAGE
HEAT WAS NOT GETTING
Need to arrange
Not working
Lead off
CASET JAM
SERVICER COME FOR INSTALLATION
Low water level
NIBP NOT WORKING
FILTER NNED TO CHANGE
SPO2 probe not working
O2 LEAK

NEED TO SET
LOW WATER LEVEL
SPO2 NOT WORKING
lamp not working properly
Catridge expired
Not getting on
Tube disconnect alarm
RV clean out error
Low water level
Need to check all phototherapy
Showing low pressure,
Need to set
showing error
Error showing
Error showing ,C5, C4 ,C11
Not working
Pressure low
Lamp not working properly
UPS down
Low water level
Continuous alarm showing
Syringe error showing
Lead off
Purified water producing device is lacking in water su
Flow not proper
Waves not showing
Printer issue
Display not showing
Waves not showing
Waves not showing
Low water level
Need to setup humidifier
Low water level
Sample struck.
Low water level
Printer problem
low gas pressure in insufflator
Leads not getting
data base error
Need to change carbon brush
conductivity error
ECG waves forms not getting properly
lense not clear
Low water level
Need to set

Sample stuck
Power failure
LOOSE CONTACT bulb
LOW AIRWAY PRESSURE
Cassette jam
loose connection
Low water level
Power failure
C5,C12 ERROR
DIGITAL PROBLEM
Waves not showing
Low water level
Showing conductivity error
Low water level
NEED TO SET
BIPOLAR NOT WORKING
Low water level
SERVICER COME FOR INSTALLATION
Cassette jam
NEED TO SET
Low water level
SHOWING ALARM
Host failure alarm
Cassatte jam
Low water level
SPO2 probe not working
Low water level
Need to set
OT wall damaged
need to set
low air supply
Low water level
Cassette jam
Not working
LOW AIR WAY PRESSURE
parameter not showing
need cleaning
STORAGE FULL
Airway pressure
LOW Tem
SHOWING ALARM
NEED TO CHANGE CATRIDGE
High airway pressure alarm
Tray not removing from system
Flow sensor complaint

High airway pressure alarm
MONO POLAR NOT WORKING
LOWER TEMPATURE
LEAK
Lamp not working properly
UP AND DOWN POSITION NOT WORKING
Low water level
LOWER TEMPATURE
PRESSURE HIGH
FILM JAM
FOR SETTING VENTILATOR
NIBP NOT WORKING
LEAK ISSUE
BLOOD ENTERING TO THE PROTECTOR
PRESSURE VARIATION
LOOSE CUFF
showing error
POWER CORD DAMAGED
SHOCK NOT DELIVERED
NOT WORKING
NOT GETTING POWER SUPPLY
BULB NOT WORKING
SOFTWARE ISSUE
BLOOD ENTERING TO THE PROTECTOR
INSUFFLATOR NOT WORKING
Low water level
NEED TO CHANGE PRINTER PAPER
Cassete stuck
low air pressure alarm
to check the flow sensor
To check
not working
Database error
not working
fumes not coming
C11 ALARM
not working
Hinges broken
PCO2 failed to calibrate
Display not working
Keyboard not working
water level too low
To ON the plant for cleaning
O2 low alarm
Calibration out and need to change pack

Need to arrange and set
ECG lead off and settings full changed.
Printing problem and printer door not closed.
HIGH PRESSURE
NOT WORKING
NOT WORKING
FOR CALIBRATION
Software issue
Low air pressure
RE substance capacity error
Cassette jam
Need to change battery
NOT WORKING
Low water level
NOT GETTING CONTINUITY
SPO2 NOT WORKING
NEED TO SETUP
NOT WORKING
Low water level
NOT GETTING CONTINUITY
SPO2 NOT WORKING
NEED TO SETUP
NOT WORKING
low water level
Bag leake
Error CN010
Printer issue
ERROR 25
Arrange&set
Not getting ECG
Wire desoldered.
Sample stuck
for set it for emergency treatment
Low air pressure alarm
Error showing while switching ON
Full capacity sensor of sub tank doesnot give ON sign
Not working
need pappe
need to chek
PRESSSURE ADJEST
ECG NOT WORKING
Printer issue
Need to check
High airway pressure alarm
Cassette jam

CALIBERATION OUT
Need to change bucky
Need to refill
Not working
Need to change EPAP valve
Need to calibrate
SPO2 probe not working
Need to check&shift to muthalam hospital
Cassette jam
Tube disconnect alarm
Dry phase not getting started
Not getting ECG&SPO2
Need to setup the machine
Airway pressure alarm
Need to set ventilator
Low water level
Low water level
Software error
Cassete stuck
Low o2 concentration
cassette jam
Need to change catridge
Cassatte jam
NEED TO READ BARCODE
IMAGE NOT CLEAR
not working
Cassatte jam
OUT OF PAPER
BULB WAS NOT WORKING
Error showing
Expiratory minute volume zero
Software issue
Printer issue
Not working
Display not working
Not working
Low water level
Not working
LOW WATER LEVEL
ECG NOT WORKING
Printer not working
ECG NOT WORKING
PROBE NOT WORKING
NOT WORKING PROPERLY
ECG NOT WORKING

CATRIDGE COMPLAINT
LEAD NOT WORKING
NIBP NOT WORKING
CASSATTE JAM
NOT WORKING
CASSATTE JAM
NEEDS MONITOR
SOME KEYS IN ALPHA NUMERIC NOT WORKING
SPO2 NOT WORKING
Printer not working
Low water level
SPO2 probe not working
LOW TIDAL VOLUME
CASSATTE JAM
Need to arrange &set
Alarm
Bag leakage
Power isssue
Not working
E1 ERROR
o2 reading tube broken
bipolar not working
Low water level
SPO2 probe not working
temperature cannot set in warmer
need a standby for emergency patient shifting
Need to set and check
Power fuilare need to calibrate
Software issue
E25 error
Gauge damage
Airway pressurec alarm
Low water level
Leak from the watertank
R3 lead off
Not getting enough pressure
Low o2 concentration
cassette jam
Need to change catridge
Cassatte jam
SPO2 probe not working
Flow sensor complaint
LOW WATER LEVEL
UPS DOWN
NO ECG WAVE

LEAD NOT READY
NEED TO SET
LOW WATER LEVEL
WAVE NOT GETTING
NEED TO ADJUST PRESSURE
BELT LOOSE
SHOWING ERROR
SHOWING ERROR
LOW WATER LEVEL
NOT WORKING
TEMP ISSUE
O2 LEAK
MEMORY FULL
REMOTE NOT WORKING
WAVE NOT GETTING
Low water level
NIBP ISSUE
NEED CALIBRATION
V1 LEAD NOT GETTING
NOT WORKING
TECHNICAL ISSUE
NOT WORKING
LOW PRESSURE
NEED TO SETUP
NEED TO REFILL
SHOW ERROR
WANT TO SETUP
TEST FAILED
SPO2 NOT WORKING
LEAK
NOT WORKING
NOT WORKING
NEED TO SETUP
NIBP NOT WORKING
NOT WORKING
Low water level
BELT ISSUE
PATIENT PLATE LOOSE CONTACT
LEAK
Need to arrange
Not working properly
Check circuit down
Lead off
E13 Error
Low water level

need pappe
need to chek
PRESSURE ADJUST
ECG NOT WORKING
CHANGING CARTIAGE
CHEKING
NEED WATER
BIPOLAR NOT WORKING
NOT WORKING
NEED WATER
NEED TO BIPAP
NIBP not working
Not ready mode
Need to arrange for emergency purpose.
Not working
Showing ECG variation
Catridge expired
Thermostat compartment temperature out of range
NOT WORKING
LAMP NOT WORKING
NOT WORKING
NOT WORKING
CASET JAM
NOT WORKING
MEMORY FULL
RO WATER LEVEL LOW
CASSETTE jam
CASSETTE jam
RO WATER LEVEL LOW
NOT WORKING
CASSETTE jam
MACHINE DOWN
NEED TO CHANGE CATRIDGE
ECG NOT GETTING
ALIT LAMP CARRIER ALIGNMENT NOT WORKING PROPERLY
CHIN SET TABLE ALIGNMENT NOT WORKING PROPERLY
NEED TO CALLIBRATE
O2 CELL NOT SHOWING VALUES
NEED TO SET
Cassatte jam
Not working
catrige expired
UPS DOWN
NO ECG WAVE
LEAD NOT READY

NEED TO SET
LOW WATER LEVEL
WAVE NOT GETTING
NEED TO ADJUST PRESSURE
BELT LOOSE
SHOWING ERROR
SHOWING ERROR
LOW WATER LEVEL
NOT WORKING
TEMP ISSUE
O2 LEAK
MEMORY FULL
REMOTE NOT WORKING
WAVE NOT GETTING
Low water level
NIBP ISSUE
NEED CALIBRATION
V1 LEAD NOT GETTING
ERROR 050
Air alarm
C3 ,E12 ERROR
Continuousn expose
Need to change bulbb
ARRANGE FROM ICCU
ARRANGED FROM NSOT
ARRANGED FROM MOT
NOT WORKING
NOT WORKING
lamp not working
NEED TO ARRAGED
Low water level
NEED TO ARRAGED
NEED TO ARRAGED
ECG&SPO2 not on display
O2 leak
Need to set
Caliberation failure
NEED TO ARRAGED
Not working
not connected to pressure
Low water level
Calibration needed
FUSE NEED TO CHANGE
Lamb fuse
Display problem

Need to set
Need to set
NEED TO ARRANGE AND SET
NOT WORKING
NEED TO ARRANGE AND SET ,NIBP NOT WORKING
Low water level
NOT WORKING
Cassette jam
LINEAR PROBE COMPLAINT
Machine stuck automatically
NEED TO SET PRINTER
Coil damage.need to replace coil
Not working
Cassatte jam
Need to refilling
Water alarm.ro plant issue
E25 Error
Need to refilling
Cassatte jam
Need to set
PRINTER ISSUE
Artifact
Need to set
Diode complaint
Need to insert new printer paper
CN010 ERROR
Need to refilling
Cassatte jam
Need to set
Need to refilling
Lead off
need pappe
need to chek
PRESSSURE ADJUST
ECG NOT WORKING
CHANGING CARTIAGE
CHEKKING
NEED WATER
BIPOLAR NOT WORKING
NOT WORKING
NEED WATER
NEED TO BIPAP
NIBP not working
Not ready mode

Need to arrange for emergency purpose.
Not working
Showing ECG variation
Catridge expired
Thermostat compartment temperature out of range
NOT WORKING
LAMP NOT WORKING
CASET JAM
NOT WORKING
MEMORY FULL
RO WATER LEVEL LOW
CASSETTE jam
CASSETTE jam
RO WATER LEVEL LOW
NOT WORKING
CASSETTE jam
MACHINE DOWN
NEED TO CHANGE CATRIDGE
ECG NOT GETTING
ALIT LAMP CARRIER ALIGNMENT NOT WORKING PROPERLY
CHIN SET TABLE ALIGNMENT NOT WORKING PROPERLY
NEED TO CALLIBRATE
O2 CELL NOT SHOWING VALUES
NEED TO SET
Cassatte jam
Not working
catrige expired
KEYPAD NOT WORKIND
Low water level
Cassette jam
showing error
NOT WORKING
FOR DEMO PURPOSE
Artifact
Need new holder
Not working
ERROR 050
Air alarm
C3 ,E12 ERROR
Continuousn expose
Need to change bulbb
ARRANGE FROM ICCU
ARRANGED FROM NSOT
ARRANGED FROM MOT

NOT WORKING
NOT WORKING
lamp not working
NEED TO ARRAGED
Low water level
NEED TO ARRAGED
NEED TO ARRAGED
ECG&SPO2 not on display
O2 leak
Need to set
Caliberation failure
NEED TO ARRAGED
Not working
not connected to pressure
Low water level
Calibration needed
FUSE NEED TO CHANGE
Lamb fuse
Display problem
Need to set
Need to set
NEED TO ARRANGE AND SET
NOT WORKING
NEED TO ARRANGE AND SET ,NIBP NOT WORKIN
Low water level
NOT WORKING
Cassette jam
LINEAR PROBE COMPLAINT
Machine stuck automatically
NEED TO SET PRINTER
Coil damage.need to replace coil
Not working
Cassatte jam
Need to refilling
Water alarm.ro plant issue
E25 Error
Need to refilling
Cassatte jam
Need to set
PRINTER ISSUE
Artifact
Need to set
Diode complaint
Need to insert new printer paper
CN010 ERROR

Leakage
MEMORY FULL
COLAER INFORMED TO SERVICER
NEED TO WATER
MOT WORKING
FILLTER PROBLEM
SPO2 NOT WORKING
Need to calibration
Need to change carbon brush
ECG cable not working
Cassatte jam
Need to change valve
Speed sensor not working properly
CN010 ERROR
Calibration failure
Need to calibration
Monitor issue
Need to checke
KB010 Disconnected
Loud noise due improper alignment
ECG not working
Wire disconnected
Not working need to check
Need to bipap
ECG not working
Need to calibrate
Water level low
Power failure
To change unit
O2 Level need to check
Not working
Low o2 alarm expiratory mint volume
Low battery
Low water level
Casset jam
showing error
Loud noise
Cassete jam
Power failure
To change unit
O2 Level need to check
Wire burning
Not working
NOT WORKING
NOT CLARITY

NOT WORKING
ECG&SPO2 not on display
O2 leak
Need to set
Caliberation failure
NEED TO ARRAGED
Not working
Low water level
BELT ISSUE
PATIENT PLATE LOOSE CONTACT
LEAK
Need to arrange
Not working properly
Check circuit down
PROBE NOT WORKING
NOT WORKING PROPERLY
ECG NOT WORKING
CATRIDGE COMPLAINT
LEAD NOT WORKING
NIBP NOT WORKING
CASSATTE JAM
NOT WORKING
CASSATTE JAM
NEEDS MONITOR
SOME KEYS IN ALPHA NUMERIC NOT WORKING
SPO2 NOT WORKING
Printer not working
Low water level
SPO2 probe not working
LOW TIDAL VOLUME
CASSATTE JAM
Need to arrange &set
Alarm
Bag leakage
Power isssue
Not working
E1 ERROR
o2 reading tube broken
bipolar not working
Low water level
SPO2 probe not working
temperature cannot set in warmer
need a standby for emergency patient shifting
Need to set and check
Power fuilare need to calibrate

Software issue
E25 error
Gauge damage
Airway pressure alarm
Low water level
Leak from the watertank
R3 lead off
Not getting enough pressure
Low o2 concentration
cassette jam
Need to change cartridge
Cassette jam
SPO2 probe not working
Flow sensor complaint
LOW WATER LEVEL
UPS DOWN
NO ECG WAVE
LEAD NOT READY
NEED TO SET
LOW WATER LEVEL
WAVE NOT GETTING
NEED TO ADJUST PRESSURE
BELT LOOSE
SHOWING ERROR
SHOWING ERROR
LOW WATER LEVEL
NOT WORKING
TEMP ISSUE
O2 LEAK
MEMORY FULL
REMOTE NOT WORKING
WAVE NOT GETTING
Low water level
NIBP ISSUE
NEED CALIBRATION
V1 LEAD NOT GETTING
NOT WORKING
TECHNICAL ISSUE
NOT WORKING
LOW PRESSURE
NEED TO SETUP
NEED TO REFILL
SHOW ERROR
WANT TO SETUP
TEST FAILED

SPO2 NOT WORKING
LEAK
NOT WORKING
NOT WORKING
NEED TO SETUP
NIBP NOT WORKING
NOT WORKING
Low water level
BELT ISSUE
PATIENT PLATE LOOSE CONTACT
LEAK
Need to arrange
Not working properly
Check circuit down
Lead off
E13 Error
Low water level
need pappe
need to chek
PRESSSURE ADJEST
ECG NOT WORKING
CHANGING CARTIAGE
CHEKKING
NEED WATER
BIPOLAR NOT WORKING
NOT WORKING
NEED WATER
NEED TO BIPAP
NIBP not working
Not ready mode
Need to arrange for emergency purpose.
Not working
Showing ECG variation
Catridge expired
Thermostat compartment temperature out of ra
NOT WORKING
LAMP NOT WORKING
NOT WORKING
NOT WORKING
CASET JAM
NOT WORKING
MEMORY FULL
RO WATER LEVEL LOW

CASSETTE jam
CASSETTE jam
RO WATER LEVEL LOW
NOT WORKING
CASSETTE jam
MACHINE DOWN
NEED TO CHANGE CATRIDGE
ECG NOT GETTING
ALIT LAMP CARRIER ALLIGNMENT NOT WORKING PRO
CHIN SET TABLE ALLIGNMENT NOT WORKING PROPE
NEED TO CALLIBRATE
O2 CELL NOT SHOWING VALUES
NEED TO SET
Cassatte jam
Not working
catrige expired
KEYPAD NOT WORKIND
Low water level
Cassette jam
showing error
NOT WORKING
FOR DEMO PURPOSE
Artifact
Need new holder
Not working
ERROR 050
Air alarm
C3 ,E12 ERROR
Continuousn expose
Need to change bulbb
ARRANGE FROM ICCU
ARRANGED FROM NSOT
ARRANGED FROM MOT
NOT WORKING
NOT WORKING
lamp not working
NEED TO ARRAGED
Low water level
NEED TO ARRAGED
NEED TO ARRAGED
ECG&SPO2 not on display
O2 leak
Need to set
Caliberation failure
NEED TO ARRAGED

Not working
not connected to pressure
Low water level
Calibration needed
FUSE NEED TO CHANGE
Lamb fuse
Display problem
Need to set
Need to set
NEED TO ARRANGE AND SET
NOT WORKING
NEED TO ARRANGE AND SET ,NIBP NOT WORKING
Low water level
NOT WORKING
Cassette jam
LINEAR PROBE COMPLAINT
Machine stuck automatically
NEED TO SET PRINTER
Coil damage.need to replace coil
Not working
Cassatte jam
Need to refilling
Water alarm.ro plant issue
E25 Error
Need to refilling
Cassatte jam
Need to set
PRINTER ISSUE
Artifact
Need to set
Diode complaint
Need to insert new printer paper
CN010 ERROR
Need to refilling
Cassatte jam
Need to set
Need to refilling
Lead off
need pappe
need to chek
PRESSURE ADJUST
ECG NOT WORKING
CHANGING CARTIAGE
CHEKKING

NEED WATER
BIPOLAR NOT WORKING
NOT WORKING
NEED WATER
NEED TO BIPAP
NIBP not working
Not ready mode
Need to arrange for emergency purpose.
Not working
Showing ECG variation
Catridge expired
Thermostat compartment temperature out of range
NOT WORKING
LAMP NOT WORKING
CASET JAM
NOT WORKING
MEMORY FULL
RO WATER LEVEL LOW
CASSETTE jam
CASSETTE jam
RO WATER LEVEL LOW
NOT WORKING
CASSETTE jam
MACHINE DOWN
NEED TO CHANGE CATRIDGE
ECG NOT GETTING
ALIT LAMP CARRIER ALLIGNMENT NOT WORKING PROPERLY
CHIN SET TABLE ALLIGNMENT NOT WORKING PROPERLY
NEED TO CALLIBRATE
O2 CELL NOT SHOWING VALUES
NEED TO SET
Cassatte jam
Not working
catrige expired
KEYPAD NOT WORKIND
Low water level
Cassette jam
showing error
NOT WORKING
FOR DEMO PURPOSE
Artifact
Need new holder
Not working
ERROR 050

Air alarm
C3 ,E12 ERROR
Continuousn expose
Need to change bulbb
ARRANGE FROM ICCU
ARRANGED FROM NSOT
ARRANGED FROM MOT
NOT WORKING
NOT WORKING
lamp not working
NEED TO ARRAGED
Low water level
NEED TO ARRAGED
NEED TO ARRAGED
ECG&SPO2 not on display
O2 leak
Need to set
Caliberation failure
NEED TO ARRAGED
Not working
not connected to pressure
Low water level
Calibration needed
FUSE NEED TO CHANGE
Lamb fuse
Display problem
Need to set
Need to set
NEED TO ARRANGE AND SET
NOT WORKING
NEED TO ARRANGE AND SET ,NIBP NOT WORKIN
Low water level
NOT WORKING
Cassette jam
LINEAR PROBE COMPLAINT
Machine stuck automatically
NEED TO SET PRINTER
Coil damage..need to replace coil
Not working
Cassatte jam
Need to refilling
Water alarm.ro plant issue
E25 Error
Need to refilling
Cassatte jam

Need to set
PRINTER ISSUE
Artifact
Need to set
Diode complaint
Need to insert new printer paper
CN010 ERROR
Leakage
MEMORY FULL
COLAER INFORMED TO SERVICER
NEED TO WATER
MOT WORKING
FILLTER PROBLEM
SPO2 NOT WORKING
Need to calibration
Need to change carbon brush
ECG cable not working
Cassatte jam
Need to change valve
Speed sensor not working properly
CN010 ERROR
Calibration failure
Need to calibration
Monitor issue
Need to checke
KB010 Disconnected
Loud noise due improper alignment
ECG not working
Wire disconnected
Not working need to check
Need to bipap
ECG not working
Need to calibrate
Water level low
Power failure
To change unit
O2 Level need to check
Not working
Low o2 alarm expiratory mint volume
Low battery
Low water level
Casset jam
showing error
Loud noise
Cassete jam

[illegible]

[illegible]

[illegible]

[illegible]

ACTION TAKEN	PART REPLACE SUGGESTED	SPARE USED	ATTENDED BY
CHECKED AND RECTIFIED	NA	NA	CS,AG
checked&rectified by the sensor complaint	NA	NA	AS,RR
checked&rectified	NA	NA	CT,JY
checked&rectified	NA	NA	RF
Caliberated	NA	NA	CS,AG
checked	NA	NA	CS,AG
refiled	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
done	NA	NA	AS,RR
done the same	NA	NA	AS,RR
Checked and rectified	NA	NA	AG
done the same	NA	NA	AG
checked and rectified	NA	NA	AG
Checked and rectified	NA	NA	AS,RR
Checked and rectified	NA	NA	AS,RR
Checked and rectified	NA	NA	AS,RR
done the same	NA	NA	CS,AG
checked by changing transducer.now machine is ok	NA	NA	CS,AG
Done	NA	NA	RR,CT
refill the water	NA	NA	CT,JY
refill the water	NA	NA	RR
informed to servicer.he came and completely changed the machine	NA	NA	AS,RR
Checked ,self test passed now working good	NA	NA	CS,AG
Done	NA	NA	AS,RR
Cassatte problem and serviced need to change sensor	NA	NA	AG
Done	NA	NA	AG
Checked and rectified	NA	NA	CT,JY
Done	NA	NA	CT,JY
CHECKED, UNDER OBSERVATION	NA	NA	RF
CHECKED	NA	NA	RF
Checked and rectified	NA	NA	RF
NEW PAPER INSERTED	NA	NA	CT,JY
Rectified	NA	NA	AS,RR
checked&rectified.	NA	NA	AS,CS
Done full calibration	NA	NA	AS,RR
Checked,inform to servicer.servicer visited&rectified	NA	NA	AS,RR
checked&rectified.	NA	NA	CS,AG
Soldered&now working	NA	NA	CS,AG
checked&rectified.	NA	NA	RR,CT
Loose connection.soldered &now working	NA	NA	CT,JY
refilled	NA	NA	RR
Checked¬ repairable.so return to store	NA	NA	AS,RR

REFILLED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
NON REPAIRABLE, PHYSICAL DAMAGE	NA	NA	AG
SOLVED	NA	NA	AG
RECONNECTED THE PROBE	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	CT,JY
SOLVED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
RECTIFIED	NA	NA	CS,AG
GIVEN	NA	NA	CS,AG
CALL LOGGED	NA	NA	CT,JY
RECTIFIED	NA	NA	RR
NON REPAIRABLE, PHYSICAL DAMAGE	NA	NA	AS,RR
refilled	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
RECTIFIED	NA	NA	AG
done the same	NA	NA	AS,RR
checked&rectified	NA	NA	AS,RR
changed bag	NA	NA	AS,RR
need new port	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG
Checked&rectified	NA	NA	CS,AG
checked not repairable,told to send to department for changing	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
checked the water plant and rectified the iissue	NA	NA	RR
checked and rectified	NA	NA	AS,RR
servicer come and checked	NA	NA	CS,AG
arranged from department	NA	NA	AS,RR
Done	NA	NA	AG
Done	NA	NA	AG
Informed to servicer he will visit tommorow,catridge locked	NA	NA	CT,JY
Checked and rectified	NA	NA	CT,JY
gauge changed.now working good	NA	NA	RF
checked and rectified	NA	NA	RF
refilled	NA	NA	RF
Sensor compliant.checked&rectified with the help of the service	NA	NA	CT,JY
checked and rectified	NA	NA	AS,RR
checked and rectified	NA	NA	AG
Done	NA	NA	CT,JY
Cassatte problem and serviced need to change sensor	NA	NA	CT,JY

Done	NA	NA	RF
Checked and rectified	NA	NA	RF
Checked and rectified	NA	NA	RF
Change flow sensor & rectified	NA	NA	CT,JY
REFILLED	NA	NA	AS,RR
INFORMED SERVICER	NA	NA	AS,CS
CHECKED AND RECTIFIED	NA	NA	AS,RR
RECTIFIED	NA	NA	AS,RR
SET	NA	NA	CS,AG
REFILLED	NA	NA	CS,AG
SOLVED	NA	NA	RR,CT
DONE	NA	NA	CT,JY
RECTIFIED	NA	NA	RR
INFORMED TO SERVICER	NA	NA	AS,RR
INFORMED SERVICER	NA	NA	CS,AG
REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
HORCE CLIP CHANGED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
REFILLED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
checked&rectified.	NA	NA	CT,JY
checked&rectified.	NA	NA	RF
checked&rectified.	NA	NA	CS,AG
FLOW SENSOR AND DIAPHRAM CHANGED	NA	NA	CS,AG
UNDER OBSERVATION	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
SETUPED	NA	NA	AS,RR
REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
SETUPED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
SETUPED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHANGED THE LAMP	NA	NA	RR,CT
REFILLED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR

CHECKED AND RECTIFIED	NA	NA	CS,AG
Done	NA	NA	AS,RR
checked,not rectified	NA	NA	AG
checked&rectified	NA	NA	RF
checked&rectified	NA	NA	RF
checked&rectified	NA	NA	CT,JY
Done	NA	NA	AS,RR
done	NA	NA	AG
done	NA	NA	CT,JY
DONE	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
DONE	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,CS
DONE	NA	NA	AS,RR
DONE	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG
Restart the machine&rectified	NA	NA	CS,AG
DONE	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
	NA	NA	RR
Installed new one	NA	NA	AS,RR
checked&restart the machine.now working good	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
SOLVED	NA	NA	AS,RR
RO WATER REFILLED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	RR,CT
RO WATER REFILLED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
SOLVED	NA	NA	CS,AG
CHANGED THE CATRIDGE	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
TIGHTNED THE SCREWS	NA	NA	RF
SOLVED	NA	NA	RF

DONE	NA	NA	CT,JY
CHECKED & O2 CELL NEED TO CHANGE	NA	NA	CT,JY
DONE	NA	NA	RR
checked and rectifieed	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG
insert neal catrigen	NA	NA	AG
CHECKED & NEED TO COMPANY SERVICE	NA	NA	AG
Refilled	NA	NA	AG
CHECKED & RECTIFIED	NA	NA	AS,RR
CHECKED & RECTIFIED	NA	NA	AS,RR
checked and rectified	NA	NA	AS,RR
DONE THE SAME	NA	NA	CT,JY
checked and rectified	NA	NA	RR
insert new holder	NA	NA	AS
Checked&rectified	NA	NA	CS
CHECKED&RECTIFIED	NA	NA	AG
checked&rectified	NA	NA	AG
Checked&rectified	NA	NA	CT,JY
restart the machine.now ok	NA	NA	CT
Bulb changed.now ok	BULB	BULB	CT
DONE	NA	NA	RR
DONE	NA	NA	RR
DONE	NA	NA	CS
checked and rectified	NA	NA	CS
checked and rectified	NA	NA	RF
checked and rectified	NA	NA	RF
DONE	NA	NA	RF
REFILLED	NA	NA	RR
DONE	NA	NA	AS,RR
DONE	NA	NA	CS,AG
Checked&rectified	NA	NA	AG
properly set o2 connection	NA	NA	AG
done the same	NA	NA	AG
checked,o2 cell compliant.temporarily arranged to anothor	NA	NA	AS,RR
DONE	NA	NA	AS,RR
seviced	NA	NA	AS,RR
done	NA	NA	AS,RR
Refilled	NA	NA	CS,AG
done	NA	NA	CS,AG
fuse changed.now working good	NA	NA	RR,CT
lamb changed now ok	NA	NA	CT,JY
checked and rectified	NA	NA	RR
done	NA	NA	AS,RR
done	NA	NA	CS,AG
DONE	NA	NA	AS,RR

DISPLAY DAMAGED NEEDED NEW DISPLAY	NA	NA	AG
checked and rectified	NA	NA	CT,JY
REFILLED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	AS,RR
DONE THE SAME	NA	NA	AS,CS
coil replaced.now working good	COMPANY	COMPANY	AS,RR
checked & found that motor was damaged.change motor n	MOTOR	MOTOR	AS,RR
checked&rectified	NA	NA	CS,AG
refilled	NA	NA	CS,AG
checked&rectified	NA	NA	RR,CT
checked&rectified	NA	NA	CT,JY
done the same	NA	NA	RR
checked&rectified	NA	NA	AS,RR
done	NA	NA	CS,AG
checked&rectified	NA	NA	AS,RR
checked&rectified	NA	NA	AG
done	NA	NA	AG
inform to servicer.part orderd.part changed now working fine	COMPANY	COMPANY	CT,JY
done the same	ABG PAPER	ABG PAPER	CT,JY
Rearrange the remote connection.now working	NA	NA	RF
refilled	NA	NA	RF
checked&rectified	NA	NA	CT
Done the same	NA	NA	AG
refilled	NA	NA	AS
checked&rectified	NA	NA	CS
done	NA	NA	CS,AG
done	NA	NA	AS,RR
DONE	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	CT,JY
DONE	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CS,AG
DONE	NA	NA	AS,RR
DONE	NA	NA	CT,JY
checked and rectified	NA	NA	RF
Restart the machine&rectified	NA	NA	CS,AG
DONE	NA	NA	CS,AG
checked and rectified	NA	NA	CT,JY
	NA	NA	RR

Installed new one	NA	NA	AS,RR
checked&restart the machine.now working good	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
SOLVED	NA	NA	AS,RR
RO WATER REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
RO WATER REFILLED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	RR,CT
CHECKED AND RECTIFIED	NA	NA	CT,JY
SOLVED	NA	NA	RR
CHANGED THE CATRIDGE	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
TIGHTNED THE SCREWS	NA	NA	AS,RR
SOLVED	NA	NA	AG
DONE	NA	NA	AG
CHECKED & O2 CELL NEED TO CHANGE	NA	NA	CT,JY
DONE	NA	NA	CT,JY
checked and rectifieed	NA	NA	RF
checked and rectified	NA	NA	RF
insert neal catrigen	NA	NA	RF
CHECKED & NEED TO COMPANY SERVICE	NA	NA	CT,JY
Refilled	NA	NA	AS,RR
CHECKED & RECTIFIED	NA	NA	AG
CHECKED & RECTIFIED	NA	NA	CT,JY
checked and rectified	NA	NA	CT,JY
DONE THE SAME	NA	NA	RF
checked and rectified	NA	NA	RF
insert new holder	NA	NA	RF
Checked&rectified	NA	NA	CT,JY
CHECKED&RECTIFIED	NA	NA	AS,RR
checked&rectified	NA	NA	AS,CS
Checked&rectified	NA	NA	AS,RR
restart the machine.now ok	NA	NA	AS,RR
Bulb changed.now ok	BULB	BULB	CS,AG
DONE	NA	NA	CS,AG
DONE	NA	NA	RR,CT
DONE	NA	NA	CT,JY
checked and rectified	NA	NA	RR
checked and rectified	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG

DONE	NA	NA	AS,RR
REFILLED	NA	NA	AG
DONE	NA	NA	AG
DONE	NA	NA	CT,JY
Checked&rectified	NA	NA	CT,JY
properly set o2 connection	NA	NA	RF
done the same	NA	NA	RF
checked,o2 cell compliant.temporarily arranged to another	NA	NA	CS,AG
DONE	NA	NA	AS,RR
sevised	NA	NA	CT,JY
done	NA	NA	RF
Refilled	NA	NA	CS,AG
done	NA	NA	CS,AG
fuse changed.now working good	NA	NA	CT,JY
lamb changed now ok	NA	NA	RR
checked and rectified	NA	NA	AS,RR
done	NA	NA	AS,RR
done	NA	NA	AG
DONE	NA	NA	AG
DISPLAY DAMAGED NEEDED NEW DISPLAY	NA	NA	AG
checked and rectified	NA	NA	AS,RR
REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	CS,AG
DONE THE SAME	NA	NA	RR,CT
coil replaced.now working good	COMPANY	COMPANY	CT,JY
checked &found that motor was damaged.change motor n	MOTOR	MOTOR	RR
checked&rectified	NA	NA	AS,RR
refilled	NA	NA	CS,AG
checked&rectified	NA	NA	AS,RR
checked&rectified	NA	NA	AG
done the same	NA	NA	RF
checked&rectified	NA	NA	RF
done	NA	NA	CT,JY
checked&rectified	NA	NA	AS,RR
checked&rectified	NA	NA	AG
done	NA	NA	CT,JY
inform to servicer.part orderd.part changed now working fine	COMPANY	COMPANY	CT,JY
done the same	ABG PAPER	ABG PAPER	RF
Rearrange the remote connection.now working	NA	NA	RF
checked&rectified	NA	NA	RF
solved	NA	NA	CT,JY
SOLVED	NA	NA	AS,RR

WATER REFILLED	NA	NA	AS,CS
solved	NA	NA	AS,RR
checked&rectified	NA	NA	AS,RR
checked&rectified			CS,AG
calibration done	NA	NA	CS,AG
changed	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
checked and rectified self test passed	NA	NA	RR
change new one	NA	NA	AS,RR
Checked and rectified	NA	NA	CS,AG
Checked and rectified now working good	NA	NA	AG
checked o2 sensor and flow sensor problem rectified now	NA	NA	AG
done	NA	NA	AG
checked and rectified	NA	NA	AS,RR
done	NA	NA	AS,RR
soldered now ok	NA	NA	AS,RR
	NA	NA	AS,RR
checked now working good	NA	NA	CS,AG
soldered now ok	NA	NA	CS,AG
checked and rectified	NA	NA	RR,CT
arranged from department	NA	NA	CT,JY
checked and rectified	NA	NA	RR
calibration done	NA	NA	AS,RR
refiled	NA	NA	CS,AG
replace another mechine	NA	NA	AS,RR
mode change	NA	NA	AG
checked and rectified	NA	NA	RF
checked and rectified	NA	NA	RF
checked and informed servicer	NA	NA	CT,JY
changed the battery	NA	NA	CT,JY
refiled	NA	NA	RR
checked and rectified self test passed	NA	NA	AS,RR
calibration done and now working good	NA	NA	CS,AG
adjusted screws	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
replace another mechine	NA	NA	AG
checked and rectified	NA	NA	AS,RR
checked and rectified	NA	NA	AS,RR
checked and rectified	NA	NA	AS,RR
checked and rectified	NA	NA	CT,JY
checked and rectified	NA	NA	RR
checked and rectified	NA	NA	AS
Checked&rectified	NA	NA	CS
properly set o2 connection	NA	NA	AG

done the same	NA	NA	AG
checked,o2 cell compliant.temporarily arranged to another	NA	NA	CT,JY
DONE	NA	NA	CT
sevised	NA	NA	CT
REFILLED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	CS
CHECKED AND RECTIFIED	NA	NA	CS
Done	NA	NA	RF
checked,not rectified	NA	NA	RF
checked&rectified	NA	NA	RF
Checked ,self test passed now working good	NA	NA	CT
Done	NA	NA	AS,RR
DONE	NA	NA	CT
Done	COMPANY	COMPANY	AG
Checked and rectified	NA	NA	AG
Checked and rectified	NA	NA	AG
REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
SETUPED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	RR,CT
SETUPED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
Done	NA	NA	AS,RR
checked&rectified.	NA	NA	CS,AG
checked and rectified	NA	NA	AS,RR
checked&rectified.	NA	NA	AG
checked and rectifued	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
DONE	NA	NA	RF
DONE	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,CS
CHECKED AND RECTIFIED	NA	NA	AS,RR
BULT CONNECTED	NA	NA	AS,RR
CHANGE TRANSUCTION	NA	NA	CS,AG
DONE	NA	NAA	CS,AG
CHECKED AND RECTIFIED	NA	NAA	RR,CT

CHECKED AND RECTIFIED	NA	NAA	CT,JY
CHECKED AND RECTIFIED	NA	NAA	RR
DONE	NA	NAA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
Done	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
INSERTED NEW ONE NOW WORKING GOOD COND	NA	NA	AG
checked and rectified	NA	NA	CT,JY
solved	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
Done	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
checked and rectified	NA	NA	RR
checked and rectified	NA	NA	AS,RR
DONE THE SAME	NA	NA	CS,AG
REFILLED	NA	NA	AS,RR
checked and rectified	NA	NA	AG
1 MICRON & 5 MICRON FILTER CHANGED	NA	NA	AG
Parameters are not accurate,taken to department	NA	NA	CT,JY
SOLVED	NA	NA	CT,JY
DONE THE SAME	NA	NA	RF
REFILLED	NA	NA	RF
SLOVED	NA	NA	CS,AG
Checked and rectified.	NA	NA	AS,RR
Extend with reading barcode	NA	NA	CT,JY
Changed the battery and now working fine	NA	NA	RF
Checked and rectified.	NA	NA	CS,AG
checked&rectified	NA	NA	CS,AG
refilled	NA	NA	CT,JY
Done	NA	NA	RR
Checked,temparature shows low when select the baby model n	NA	NA	AS,RR
Done	NA	NA	AS,RR
checked and rectified	NA	NA	AG
checked and rectified	NA	NA	AG
Checked and temparature calibration completed	NA	NA	AG
checked&found that SMPS have problem	NA	NA	AS,RR
Checked and rectified	NA	NA	AS,RR
changed lamp now working fine	NA	NA	AS,RR
servicer visit&checked.now working good	NA	NA	AS,RR
refilled	NA	NA	CS,AG
checked&reset the cable.	NA	NA	CS,AG
checked&changedthe settings.found ok	NA	NA	RR,CT

checked&rectified	NA	NA	CT,JY
checked&rectified	NA	NA	RR
changed the filter.now working good	NA	NA	AS,RR
checked&found cable has issue.temporarily adjust it	NA	NA	CS,AG
Rectified	NA	NA	AS,RR
checked&reset the cable	NA	NA	AG
Checked&rectified	NA	NA	RF
Checked&rectified	NA	NA	RF
refilled	NA	NA	CT,JY
Done the same	NA	NA	AS,RR
refilled	NA	NA	AG
Checked&rectified	NA	NA	CT,JY
Refilled	NA	NA	CT,JY
checked&changed paper.now working fine	NA	NA	RF
Adjusted	NA	NA	RF
checked & rectified	NA	NA	RF
checked & rectified	NA	NA	CT,JY
changed	NA	NA	AS,RR
checked & inform to sevicer	NA	NA	AS,CS
checked & rectified	NA	NA	AS,RR
checked & rectified	NA	NA	AS,RR
Refilled	NA	NA	CS,AG
done	NA	NA	CS,AG
checked & rectified	NA	NA	RR,CT
checked & rectified	NA	NA	CT,JY
change bulb	NA	NA	RR
checked & rectified	NA	NA	AS,RR
checked & rectified	NA	NA	CS,AG
checked & rectified	NA	NA	AG
refilled	NA	NA	AG
change power cord	NA	NA	AG
calibration done	NA	NA	AS,RR
checked & rectified	NA	NA	AS,RR
checked & rectified	NA	NA	AS,RR
REFILLED	NA	NA	AS,RR
checked & rectified	NA	NA	CS,AG
REFILLED	NA	NA	CS,AG
DONE	NA	NA	RR,CT
checked & rectified	NA	NA	CT,JY
refilled	NA	NA	RR
PART INSTALLED	NA	NA	AS,RR
Rectified	NA	NA	CS,AG
DONE THE SAME	NA	NA	AS,RR
refilled	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	RF

CHECKED AND RECTIFIED	NA	NA	RF
clear the cassette and rectified	NA	NA	CT,JY
refilled	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
refilled	NA	NA	AS,RR
done	NA	NA	CS,AG
vinyl sheet changed	NA	NA	AG
done	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
refilled	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHANGE CARBON BRUSH	NA	NA	AS,RR
change filter and rectified	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
cleaned	NA	NA	AS
move to hard drive	NA	NA	CS
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	CT,JY
READ CATRIDGE	NA	NA	CT
CHECKED AND RECTIFIED	NA	NA	CT
CHECKED AND RECTIFIED	NA	NA	RR
CHECKED AND CALIBRATED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	CS
CHECKED CHANGE PAIENT PLATE AND RECTIFIED	NA	NA	CS
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHANGE Bulb AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RR
refilled	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
DONE	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND INFORM TO SERVICER	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	RR,CT
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
Bulb changed now working good	BULB	BULB	CS,AG

CHECKED AND RECTIFIED	NA	NA	AS,RR
checked & inform to sevicer	COMPANY	COMPANY	AG
CHECKED AND RECTIFIED	NA	NA	CT,JY
refilled	NA	NA	RF
PAPER CHANGED	PRINTER PAPER	PRINTER PAPER	RF
checked and rectified	NA	NA	RF
connected air valve correctly,now working	NA	NA	CT,JY
checked and working good	NA	NA	AS,RR
done	NA	NA	AS,CS
found that wire is burned,changed it now working ok	NA	NA	AS,RR
checked and rectified	NA	NA	AS,RR
lamp complaint , battery has no charge,changed it now working	NA	NA	CS,AG
reconnected all the parts and cleaned, now working good	NA	NA	CS,AG
Checked and rectified	NA	NA	RR,CT
found that the wire is burned,power plug also burned,changed it	NA	NA	CT,JY
Changed it	HINGES [L]	NA	RR
catridge changed	NA	NA	AS,RR
informed to servicer	NA	NA	CS,AG
changed by servicer	NA	NA	AS,RR
refilled	NA	NA	AG
done	NA	NA	AG
checked the plant and switch ON the liquid O2	NA	NA	CT,JY
pack changed and calibrated.now working	NA	NA	CT,JY
done	NA	NA	RF
checked&now working good	NA	NA	RF
Checked and serviced	NA	NA	RR,CT
checked and rectified.	NA	NA	CT,JY
CARBON BRUSS CHANGED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
Checked&rectified	NA	NA	AG
Checked&rectified	NA	NA	AG
Checked&rectified	NA	NA	CT,JY
change 3 nos battery &rectified	NA	NA	CT,JY
BACTERIAL FILTER CLEANED	NA	NA	RF
REFILLED	NA	NA	RF
Checked&rectified	NA	NA	CS,AG
Checked&rectified	NA	NA	AS,RR
DONE THE SAME	NA	NA	CT,JY
BACTERIAL FILTER CLEANED	NA	NA	RF
REFILLED	NA	NA	CS,AG
Checked&rectified	NA	NA	CS,AG
Checked&rectified	NA	NA	CT,JY
DONE THE SAME	NA	NA	RR

Checked&rectified	NA	NA	AS,RR
REFILLED	NA	NA	AS,RR
changed the bag	BP BAG	BP BAG	AG
Rearrange the remote.now working ok	NA	NA	AG
checked and rectified	NA	NA	AG
checked and rectified	NA	NA	AS,RR
done the same	NA	NA	AS,RR
CHECKED&RECTIFIED	NA	NA	AS,RR
checked and resoldered the cable and check all combinations.n	NA	NA	AS,RR
checked&rectified.	NA	NA	CS,AG
put it in standby	NA	NA	CS,AG
checked and rectified	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
found that filter is dirt and need to change it. Changed it and no	NA	NA	RR
checked and rectified	NA	NA	AS,RR
done	NA	NA	CS,AG
done	NA	NA	AS,RR
DONE	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CT,JY
checked and rectified	NA	NA	AS,RR
DONE	NA	NA	AG
REFILLED	NA	NA	CT,JY
DONE	NA	NA	CT,JY
checked and rectified	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	AS,RR
DONE THE SAME	NA	NA	AS,CS
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
DONE THE SAME	NA	NA	CS,AG
fuse changed.now working good	NA	NA	CS,AG
DONE	NA	NA	RR,CT
RESET THE MACHINE. NOW WORKING GOOD	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
CALIBRATION DONE	NA	NA	AS,RR
soldered now ok	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AG
Refilled	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AS,RR

CHECKED & ADJUSTED	NA	NA	AS,RR
DONE	NA	NA	AS,RR
checked&rectified	NA	NA	AS,RR
checked&rectified	NA	NA	CS,AG
checked and rectified	NA	NA	CS,AG
Need to change all battery	NA	NA	RR,CT
Done	NA	NA	CT,JY
Adapter complaint	ADAPTER	COMPANY	RR
checked and rectified	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG
suggested battery replacement now working ok	BATTERY 3V	consumable	AS,RR
checked and rectified	NA	NA	AG
checked and rectified	NA	NA	RF
set printer time and rectified	NA	NA	RF
Done	NA	NA	CT,JY
check with simulator and rectified	NA	NA	CT,JY
machine now work on battery	NA	NA	RR
checked and rectified	SS VALVE	SS VALVE	AS,RR
Checked and rectified	NA	NA	CS,AG
Checked and rectified	NA	NA	AG
INSTALLATION COMPLETED	NA	NA	AG
Read the barcode once again and informed servicer	NA	NA	AG
Checked and rectified	NA	NA	AS,RR
checked and rectified	NA	NA	AS,RR
Done	NA	NA	AS,RR
Clear out the cassette	NA	NA	CT,JY
Water refilling	NA	NA	RR
Set correct mode	NA	NA	AS
Done	NA	NA	CS
Change cartridge and rectified	CATRIDGE	COMPANY	AG
Calibration done	NA	NA	AG
Checked and set speed	NA	NA	CT,JY
Done	NA	NA	CT
Water refilling	NA	NA	CT
Monitor display changed now working good	MONITOR DISPL	MONITOR DIS	RR
Done	NA	NA	RR
Water refilling	NA	NA	CS
Replace bulb and cuff	BP bulb and cuff	BP bulb and c	CS
Done	NA	NA	RF
Jacket valve opened.And RO Water refilling	NA	NA	RF
Done	NA	NA	RF
Done	ECG cable	consumable	RR
Water refilling	NA	NA	AS,RR
Done	NA	NA	CS,AG

changed the carbon brush&tightend the sensor	CARBON BRUSH	CARBON BRU	AG
Bilb changed.now working good	BULB	BULB	AG
Clear out the cassatte.now working good	NA	NA	AG
checked&rectified.	NA	NA	AS,RR
checked&rectified.	NA	NA	AS,RR
checked&rectified.	NA	NA	AS,RR
changed filter and now working good	NA	NA	AS,RR
checked&insert new new screw	NA	NA	CS,AG
Refilling	NA	NA	CS,AG
checked and converted to inveter mode	NA	NA	RR,CT
checked and change fuse now working good	FUSE	FUSE	CT,JY
done	NA	NA	RR
Done	NA	NA	AS,RR
done	NA	NA	CS,AG
checked and inform to servicer	NA	NA	AS,RR
Refilling	NA	NA	AG
Refilling	NA	NA	CT,JY
Done	NA	NA	RF
checked and inform to servicer	NA	NA	RF
Checked,machine restart and manually take sample .	NA	NA	RF
to refill	NA	NA	CT,JY
Done	ABG PAPER	ABG PAPER	AS,RR
Done	NA	NA	AS,CS
to refill	NA	NA	AS,RR
checked and rectified	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG
checked and inform to servicer	NA	NA	CS,AG
checked and rectified	NA	NA	RR,CT
checked and reset	NA	NA	CT,JY
Done	NA	NA	RR
checked&rectified.	NA	NA	AS,RR
coupler changed	COUPLER	COMPANY	CS,AG
to refill	NA	NA	AS,RR
checked and rectified	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	RF
checked&rectified.	NA	NA	AG
arrange from NMICU	NA	NA	CT,JY
checked&rectified.	NA	NA	CT,JY
dOne	NA	NA	RF
done	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
DONE	NA	NA	RR
REFILLED	NA	NA	AS,RR
DONE	NA	NA	CS,AG

checked and rectified	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	CT,JY
DONE THE SAME	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CS,AG
DONE THE SAME	NA	NA	AS,RR
fuse changed.now working good	NA	NA	CT,JY
DONE	NA	NA	RF
RESET THE MACHINE. NOW WORKING GOOD	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	CS,AG
CALIBRATION DONE	NA	NA	CT,JY
soldered now ok	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
Refilled	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED & ADJUSTED	NA	NA	AG
DONE	NA	NA	AS,RR
checked&rectified	NA	NA	AS,RR
checked&rectified	NA	NA	AS,RR
checked&rectified	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	RR,CT
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND CHANGED	NA	NA	RF
CHECKED AND UNDER OBSERVATION	NA	NA	RF
CALIBRATED	NA	NA	CT,JY
CHECKED	NA	NA	AS,RR
MACHINE BELT CHANGED	NA	NA	AG
CALIBRATED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	CT,JY
done the same	NA	NA	CT,JY
checked by changing transducer.now machine is ok	NA	NA	RF

Done	NA	NA	RF
refill the water	NA	NA	RF
refill the water	NA	NA	CT,JY
informed to servicer.he came and completely changed the machine	NA	NA	AS,RR
Checked ,self test passed now working good	NA	NA	AS,CS
Done	NA	NA	AS,RR
Cassatte problem and serviced need to change sensor	NA	NA	AS,RR
Done	NA	NA	CS,AG
Checked and rectified	NA	NA	CS,AG
Done	NA	NA	RR,CT
CHECKED, UNDER OBSERVATION	NA	NA	CT,JY
CHECKED	NA	NA	RR
Checked and rectified	NA	NA	AS,RR
NEW PAPER INSERTED	NA	NA	CS,AG
Rectified	NA	NA	AG
checked&rectified.	NA	NA	AG
Done full calibration	NA	NA	AG
Checked,inform to servicer.servicer visited&rectified	NA	NA	AS,RR
Done full calibration	NA	NA	AG
checked&rectified.	NA	NA	AS,RR
Soldered&now working	NA	NA	AS,RR
checked&rectified.	NA	NA	AS,RR
Loose connection.soldered &now working	NA	NA	CS,AG
refilled	NA	NA	CS,AG
Checked¬ repairable.so return to store	NA	NA	RR,CT
REFILLED	NA	NA	CT,JY
Loose connection.soldered &now working	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	RR
NON REPAIRABLE, PHYSICAL DAMAGE	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	RR
SOLVED	NA	NA	CS,AG
RECONNECTED THE PROBE	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
refilled	NA	NA	RF
PART INSTALLED	NA	NA	RF
Rectified	NA	NA	CT,JY
DONE THE SAME	NA	NA	CT,JY
refilled	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
clear the cassette and rectified	NA	NA	AG
clear the cassette and rectified	NA	NA	AG
refilled	NA	NA	AG

CHECKED AND RECTIFIED	NA	NA	AG
refilled	NA	NA	AS,RR
done	NA	NA	AS,RR
vinyl sheet changed	NA	NA	AS,RR
done	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
refilled	NA	NA	AS
CHECKED AND RECTIFIED	NA	NA	CS
CHANGE CARBON BRUSH	NA	NA	AG
change filter and rectified	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED, NOT RECTIFIED,INFORMED TO FSE,HE WILL COME NEXT WEAK	NA	NA	CT
CHECKED,INFORMED TO FSE	NA	NA	CT
DONE	NA	NA	RY
CHECKED, REPLACED NEW FOOT SWITCH	YES	YES	RR,AG
CHECKED & RECTIFIED	NA	NA	RR
CHECKED & RECTIFIED	NA	NA	
CHECKED & CLEANED THE SENSOR	NA	NA	
CHECKED & RECTIFIED	NA	NA	JY
CHECKED & RECTIFIED	NA	NA	JY
CHECKED & INFORMED TO FSE	NA	NA	JY
CHECKED, TEMPERERLY PROBLEM SOLVED	NA	NA	CT
CHANGED NEW ONE,CALIBRATION PASSED	YES	YES	AS
CHECKED,INFORMED TO FSE	NA	NA	AS,RY
CHECKED & RECTIFIED	NA	NA	JY
CHANGED & ISSUE SOLVED	NA	NA	JY
CHECKED & RECTIFIED	NA	NA	JY
CHECKED & RECTIFIED	NA	NA	JY
CHECKED & RECTIFIED	NA	NA	JY
CHECKED & CHANGED	NA	NA	CS
checked&rectified	NA	NA	RF
Checked ,self test passed now working good	NA	NA	CT
Done	NA	NA	AS,RR
DONE	NA	NA	CT
Done	COMPANY	COMPANY	AG
Checked and rectified	NA	NA	AG
Checked and rectified	NA	NA	AG
REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
SETUPED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG

CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	RR,CT
SETUPED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
Done	NA	NA	AS,RR
checked&rectified.	NA	NA	CS,AG
checked and rectified	NA	NA	AS,RR
checked&rectified.	NA	NA	AG
checked and rectifued	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
DONE	NA	NA	RF
DONE	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,CS
CHECKED AND RECTIFIED	NA	NA	AS,RR
BULT CONNECTED	NA	NA	AS,RR
CHANGE TRANSUCTION	NA	NA	CS,AG
DONE	NA	NAA	CS,AG
CHECKED AND RECTIFIED	NA	NAA	RR,CT
CHECKED AND RECTIFIED	NA	NAA	CT,JY
CHECKED AND RECTIFIED	NA	NAA	RR
DONE	NA	NAA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
Done	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
INSERTED NEW ONE NOW WORKING GOOD COND	NA	NA	AG
checked and rectified	NA	NA	CT,JY
solved	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
Done	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
checked and rectified	NA	NA	RR
checked and rectified	NA	NA	AS,RR
DONE THE SAME	NA	NA	CS,AG
REFILLED	NA	NA	AS,RR
checked and rectified	NA	NA	AG
1 MICRON & 5 MICRON FILTER CHANGED	NA	NA	AG
Parameters are not accurate,taken to department	NA	NA	CT,JY
SOLVED	NA	NA	CT,JY

DONE THE SAME	NA	NA	RF
REFILLED	NA	NA	RF
SLOVED	NA	NA	CS,AG
Checked and rectified.	NA	NA	AS,RR
Extend with reading barcode	NA	NA	CT,JY
Changed the battery and now working fine	NA	NA	RF
Checked and rectified.	NA	NA	CS,AG
checked&rectified	NA	NA	CS,AG
refilled	NA	NA	CT,JY
Done	NA	NA	RR
Checked,temparature shows low when select the baby model n	NA	NA	AS,RR
Done	NA	NA	AS,RR
checked and rectified	NA	NA	AG
checked and rectified	NA	NA	AG
Checked and temparature calibration completed	NA	NA	AG
checked&found that SMPS have problem	NA	NA	AS,RR
Checked and rectified	NA	NA	AS,RR
changed lamp now working fine	NA	NA	AS,RR
servicer visit&checked.now working good	NA	NA	AS,RR
refilled	NA	NA	CS,AG
checked&reset the cable.	NA	NA	CS,AG
checked&changedthe settings.found ok	NA	NA	RR,CT
checked&rectified	NA	NA	CT,JY
checked&rectified	NA	NA	RR
changed the filter.now working good	NA	NA	AS,RR
checked&found cable has issue.temporarily adjust it	NA	NA	CS,AG
Rectified	NA	NA	AS,RR
checked&reset the cable	NA	NA	AG
Checked&rectified	NA	NA	RF
Checked&rectified	NA	NA	RF
refilled	NA	NA	CT,JY
Done the same	NA	NA	AS,RR
refilled	NA	NA	AG
Checked&rectified	NA	NA	CT,JY
Refilled	NA	NA	CT,JY
checked&changed paper.now working fine	NA	NA	RF
Adjusted	NA	NA	RF
checked & rectified	NA	NA	RF
checked & rectified	NA	NA	CT,JY
changed	NA	NA	AS,RR
checked & inform to sevicer	NA	NA	AS,CS
checked & rectified	NA	NA	AS,RR
checked & rectified	NA	NA	AS,RR
Refilled	NA	NA	CS,AG
done	NA	NA	CS,AG

checked & rectified	NA	NA	RR,CT
checked & rectified	NA	NA	CT,JY
change bulb	NA	NA	RR
checked & rectified	NA	NA	AS,RR
checked & rectified	NA	NA	CS,AG
checked & rectified	NA	NA	AG
refilled	NA	NA	AG
change power cord	NA	NA	AG
calibration done	NA	NA	AS,RR
checked & rectified	NA	NA	AS,RR
checked & rectified	NA	NA	AS,RR
REFILLED	NA	NA	AS,RR
checked & rectified	NA	NA	CS,AG
REFILLED	NA	NA	CS,AG
DONE	NA	NA	RR,CT
checked & rectified	NA	NA	CT,JY
refilled	NA	NA	RR
PART INSTALLED	NA	NA	AS,RR
Rectified	NA	NA	CS,AG
DONE THE SAME	NA	NA	AS,RR
refilled	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
clear the cassette and rectified	NA	NA	CT,JY
refilled	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
refilled	NA	NA	AS,RR
done	NA	NA	CS,AG
vinyl sheet changed	NA	NA	AG
done	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
refilled	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHANGE CARBON BRUSH	NA	NA	AS,RR
change filter and rectified	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
cleaned	NA	NA	AS
move to hard drive	NA	NA	CS
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	CT,JY
READ CATRIDGE	NA	NA	CT
CHECKED AND RECTIFIED	NA	NA	CT
CHECKED AND RECTIFIED	NA	NA	RR
CHECKED AND CALIBRATED	NA	NA	RR

CHECKED AND RECTIFIED	NA	NA	CS
CHECKED CHANGE PAIENT PLATE AND RECTIFIED	NA	NA	CS
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHANGE Bulb AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RR
refilled	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
DONE	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND INFORM TO SERVICER	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	RR,CT
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
Bulb changed now working good	BULB	BULB	CS,AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
checked & inform to sevicer	COMPANY	COMPANY	AG
CHECKED AND RECTIFIED	NA	NA	CT,JY
refilled	NA	NA	RF
PAPER CHANGED	PRINTER PAPER	PRINTER PAPER	RF
checked and rectified	NA	NA	RF
connected air valve correctly,now working	NA	NA	CT,JY
checked and working good	NA	NA	AS,RR
done	NA	NA	AS,CS
found that wire is burned,changed it now working ok	NA	NA	AS,RR
checked and rectified	NA	NA	AS,RR
lamp complaint , battery has no charge,changed it now working	NA	NA	CS,AG
reconnected all the parts and cleaned, now working good	NA	NA	CS,AG
Checked and rectified	NA	NA	RR,CT
found that the wire is burned,power plug also burned,changed it	NA	NA	CT,JY
Changed it	HINGES [L]	NA	RR
catridge changed	NA	NA	AS,RR
informed to servicer	NA	NA	CS,AG
changed by servicer	NA	NA	AS,RR
refilled	NA	NA	AG
done	NA	NA	AG
checked the plant and switch ON the liquid O2	NA	NA	CT,JY
pack changed and calibrated.now working	NA	NA	CT,JY

done	NA	NA	RF
checked&now working good	NA	NA	RF
Checked and serviced	NA	NA	RR,CT
checked and rectified.	NA	NA	CT,JY
CARBON BRUSS CHANGED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
Checked&rectified	NA	NA	AG
Checked&rectified	NA	NA	AG
Checked&rectified	NA	NA	CT,JY
change 3 nos battery &rectified	NA	NA	CT,JY
BACTERIAL FILTER CLEANED	NA	NA	RF
REFILLED	NA	NA	RF
Checked&rectified	NA	NA	CS,AG
Checked&rectified	NA	NA	AS,RR
DONE THE SAME	NA	NA	CT,JY
BACTERIAL FILTER CLEANED	NA	NA	RF
REFILLED	NA	NA	CS,AG
Checked&rectified	NA	NA	CS,AG
Checked&rectified	NA	NA	CT,JY
DONE THE SAME	NA	NA	RR
Checked&rectified	NA	NA	AS,RR
REFILLED	NA	NA	AS,RR
changed the bag	BP BAG	BP BAG	AG
Rearrange the remote.now working ok	NA	NA	AG
checked and rectified	NA	NA	AG
checked and rectified	NA	NA	AS,RR
done the same	NA	NA	AS,RR
CHECKED&RECTIFIED	NA	NA	AS,RR
checked and resoldered the cable and check all combinations.n	NA	NA	AS,RR
checked&rectified.	NA	NA	CS,AG
put it in standby	NA	NA	CS,AG
checked and rectified	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
found that filter is dirt and need to change it. Changed it and no	NA	NA	RR
checked and rectified	NA	NA	AS,RR
done	NA	NA	CS,AG
done	NA	NA	AS,RR
DONE	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CS,AG
rchecked&rectified by the sensor complaint	NA	NA	AS,RR
checked&rectified	NA	NA	CT,JY
checked&rectified	NA	NA	RF

Caliberated	NA	NA	CS,AG
checked	NA	NA	CS,AG
refiled	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
done	NA	NA	AS,RR
done the same	NA	NA	AS,RR
Checked and rectified	NA	NA	AG
done the same	NA	NA	AG
checked and rectified	NA	NA	AG
Checked and rectified	NA	NA	AS,RR
Checked and rectified	NA	NA	AS,RR
Checked and rectified	NA	NA	AS,RR
done the same	NA	NA	CS,AG
checked by changing transducer.now machine is ok	NA	NA	CS,AG
Done	NA	NA	RR,CT
refill the water	NA	NA	CT,JY
refill the water	NA	NA	RR
informed to servicer.he came and completely changed the machine	NA	NA	AS,RR
Checked ,self test passed now working good	NA	NA	CS,AG
Done	NA	NA	AS,RR
Cassatte problem and serviced need to change sensor	NA	NA	AG
Done	NA	NA	AG
Checked and rectified	NA	NA	CT,JY
Done	NA	NA	CT,JY
CHECKED, UNDER OBSERVATION	NA	NA	RF
CHECKED	NA	NA	RF
Checked and rectified	NA	NA	RF
NEW PAPER INSERTED	NA	NA	CT,JY
Rectified	NA	NA	AS,RR
checked&rectified.	NA	NA	AS,CS
Done full calibration	NA	NA	AS,RR
Checked,inform to servicer.servicer visited&rectified	NA	NA	AS,RR
checked&rectified.	NA	NA	CS,AG
Soldered&now working	NA	NA	CS,AG
checked&rectified.	NA	NA	RR,CT
Loose connection.soldered &now working	NA	NA	CT,JY
refilled	NA	NA	RR
Checked¬ repairable.so return to store	NA	NA	AS,RR
REFILLED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
NON REPAIRABLE, PHYSICAL DAMAGE	NA	NA	AG
SOLVED	NA	NA	AG
RECONNECTED THE PROBE	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	CT,JY
SOLVED	NA	NA	RF

CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
RECTIFIED	NA	NA	CS,AG
GIVEN	NA	NA	CS,AG
CALL LOGGED	NA	NA	CT,JY
RECTIFIED	NA	NA	RR
NON REPAIRABLE, PHYSICAL DAMAGE	NA	NA	AS,RR
refilled	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
RECTIFIED	NA	NA	AG
done the same	NA	NA	AS,RR
checked&rectified	NA	NA	AS,RR
changed bag	NA	NA	AS,RR
need new port	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG
Checked&rectified	NA	NA	CS,AG
checked not repairable,told to send to department for changing	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
checked the water plant and rectified the iissue	NA	NA	RR
checked and rectified	NA	NA	AS,RR
servicer come and checked	NA	NA	CS,AG
arranged from department	NA	NA	AS,RR
Done	NA	NA	AG
Done	NA	NA	AG
Informed to servicer he will visit tommorow,catridge locked	NA	NA	CT,JY
Checked and rectified	NA	NA	CT,JY
gauge changed.now working good	NA	NA	RF
checked and rectified	NA	NA	RF
refilled	NA	NA	RF
Sensor compliant.checked&rectified with the help of the service	NA	NA	CT,JY
checked and rectified	NA	NA	AS,RR
checked and rectified	NA	NA	AG
Done	NA	NA	CT,JY
Cassatte problem and serviced need to change sensor	NA	NA	CT,JY
Done	NA	NA	RF
Checked and rectified	NA	NA	RF
Checked and rectified	NA	NA	RF
Change flow sensor & rectified	NA	NA	CT,JY
REFILLED	NA	NA	AS,RR
INFORMED SERVICIER	NA	NA	AS,CS
CHECKED AND RECTIFIED	NA	NA	AS,RR

RECTIFIED	NA	NA	AS,RR
SET	NA	NA	CS,AG
REFILLED	NA	NA	CS,AG
SOLVED	NA	NA	RR,CT
DONE	NA	NA	CT,JY
RECTIFIED	NA	NA	RR
INFORMED TO SERVICER	NA	NA	AS,RR
INFORMED SERVICER	NA	NA	CS,AG
REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
HORCE CLIP CHANGED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
REFILLED	NA	NA	CS,AG
CHECKED AND RECTIFIOED	NA	NA	AS,RR
checked&rectified.	NA	NA	CT,JY
checked&rectified.	NA	NA	RF
checked&rectified.	NA	NA	CS,AG
FLOW SENSOR AND DIAPHRAM CHANGED	NA	NA	CS,AG
UNDER OBSERVATION	NA	NA	CT,JY
CHECKED AND RECTIFIOED	NA	NA	RR
SETUPED	NA	NA	AS,RR
REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
SETUPED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
SETUPED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHANGED THE LAMP	NA	NA	RR,CT
REFILLED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
Done	NA	NA	AS,RR
checked,not rectified	NA	NA	AG
checked&rectified	NA	NA	RF
checked&rectified	NA	NA	RF
checked&rectified	NA	NA	CT,JY
Done	NA	NA	AS,RR

done	NA	NA	AG
done	NA	NA	CT,JY
DONE	NA	NA	CT,JY
Done	NA	NA	RF
checked,not rectified	NA	NA	RF
checked&rectified	NA	NA	RF
checked&rectified	NA	NA	CT,JY
checked&rectified	NA	NA	AS,RR
Done	NA	NA	AS,CS
DONE	NA	NA	AS,RR
DONE	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG
Restart the machine&rectified	NA	NA	CS,AG
DONE	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
	NA	NA	RR
Installed new one	NA	NA	AS,RR
checked&restart the machine.now working good	NA	NA	CS,AG
checked,not rectified	NA	NA	AG
checked&rectified	NA	NA	AG
checked&rectified	NA	NA	AG
checked&rectified	NA	NA	AS,RR
Done	NA	NA	AS,RR
DONE	NA	NA	AS,RR
SOLVED	NA	NA	AS,RR
RO WATER REFILLED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	RR,CT
RO WATER REFILLED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
SOLVED	NA	NA	CS,AG
CHANGED THE CATRIDGE	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
TIGHTNED THE SCREWS	NA	NA	RF
SOLVED	NA	NA	RF
DONE	NA	NA	CT,JY
CHECKED & O2 CELL NEED TO CHANGE	NA	NA	CT,JY
DONE	NA	NA	RR
checked and rectifieed	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG
insert neal catrigen	NA	NA	AG
INFORMED SERVCIER	NA	NA	AS,CS
CHECKED AND RECTIFIED	NA	NA	AS,RR
RECTIFIED	NA	NA	AS,RR

SET	NA	NA	CS,AG
REFILLED	NA	NA	CS,AG
SOLVED	NA	NA	RR,CT
DONE	NA	NA	CT,JY
RECTIFIED	NA	NA	RR
INFORMED TO SERVICER	NA	NA	AS,RR
INFORMED SERVICER	NA	NA	CS,AG
REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
HORCE CLIP CHANGED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
REFILLED	NA	NA	CS,AG
CHECKED AND RECTIFIOED	NA	NA	AS,RR
checked&rectified.	NA	NA	CT,JY
checked&rectified.	NA	NA	RF
CHECKED&RECTIFIED	NA		
checked&rectified	NA		
Checked&rectified	NA		
restart the machine.now ok	NA		
Bulb changed.now ok	BULB		
DONE	NA		
DONE	NA		
DONE	NA		
checked and rectified	NA		
checked and rectified	NA		
checked and rectified	NA		
DONE	NA		
REFILLED	NA		
DONE	NA		
DONE	NA		
Checked&rectified	NA		
properly set o2 connection	NA		
done the same	NA		
checked,o2 cell compliant.temporarily arranged to anothor	NA		
DONE	NA		
seviced	NA		
done	NA		
Refilled	NA		
done	NA		
fuse changed.now working good	NA		
lamb changed now ok	NA		
checked and rectified	NA		

done	NA
done	NA
DONE	NA
DISPLAY DAMAGED NEEDED NEW DISPLAY	NA
checked and rectified	NA
REFILLED	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
DONE THE SAME	NA
coil replaced.now working good	COMPANY
checked & found that motor was damaged.change motor n	MOTOR
checked&rectified	NA
refilled	NA
checked&rectified	NA
checked&rectified	NA
done the same	NA
checked&rectified	NA
done	NA
checked&rectified	NA
checked&rectified	NA
done	NA
inform to servicer.part orderd.part changed now working fine	COMPANY
done the same	ABG PAPER
Rearrange the remote connection.now working	NA
refilled	NA
checked&rectified	NA
Done the same	NA
refilled	NA
checked&rectified	NA
done	NA
done	NA
DONE	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
DONE	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
DONE	NA
DONE	NA
checked and rectified	NA
Restart the machine&rectified	NA

DONE	NA
checked and rectified	NA
	NA
Installed new one	NA
checked&restart the machine.now working good	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
SOLVED	NA
RO WATER REFILLED	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
RO WATER REFILLED	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
SOLVED	NA
CHANGED THE CATRIDGE	NA
CHECKED AND RECTIFIED	NA
TIGHTNED THE SCREWS	NA
SOLVED	NA
DONE	NA
CHECKED & O2 CELL NEED TO CHANGE	NA
DONE	NA
checked and rectifieed	NA
checked and rectified	NA
insert neal catrigen	NA
CHECKED & NEED TO COMPANY SERVICE	NA
Refilled	NA
CHECKED & RECTIFIED	NA
CHECKED & RECTIFIED	NA
checked and rectified	NA
DONE THE SAME	NA
checked and rectified	NA
insert new holder	NA
Checked&rectified	NA
CHECKED&RECTIFIED	NA
checked&rectified	NA
Checked&rectified	NA
restart the machine.now ok	NA
Bulb changed.now ok	BULB
DONE	NA
DONE	NA
DONE	NA

checked and rectified	NA
checked and rectified	NA
checked and rectified	NA
DONE	NA
REFILLED	NA
DONE	NA
DONE	NA
Checked&rectified	NA
properly set o2 connection	NA
done the same	NA
checked,o2 cell compliant.temporarily arranged to anothor	NA
DONE	NA
seviced	NA
done	NA
Refilled	NA
done	NA
fuse changed.now working good	NA
lamb changed now ok	NA
checked and rectified	NA
done	NA
done	NA
DONE	NA
DISPLAY DAMAGED NEEDED NEW DISPLAY	NA
checked and rectified	NA
REFILLED	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
CHECKED AND RECTIFIED	NA
DONE THE SAME	NA
coil replaced.now working good	COMPANY
checked &found that motor was damaged.change motor n	MOTOR
checked&rectified	NA
refilled	NA
checked&rectified	NA
checked&rectified	NA
done the same	NA
checked&rectified	NA
done	NA
checked&rectified	NA
checked&rectified	NA
done	NA
inform to servicer.part orderd.part changed now working fine	COMPANY
done the same	ABG PAPER
Rearrange the remote connection.now working	NA

checked&rectified	NA
solved	NA
SOLVED	NA
WATER REFILLED	NA
solved	NA
checked&rectified	NA
checked&rectified	
calibration done	NA
changed	NA
checked and rectified	NA
checked and rectified self test passed	NA
change new one	NA
Checked and rectified	NA
Checked and rectified now working good	NA
checked o2 sensor and flow sensor problem rectified now	NA
done	NA
checked and rectified	NA
done	NA
soldered now ok	NA
	NA
checked now working good	NA
soldered now ok	NA
checked and rectified	NA
arranged from department	NA
checked and rectified	NA
calibration done	NA
refiled	NA
replace another mechine	NA
mode change	NA
checked and rectified	NA
checked and rectified	NA
checked and informed servicer	NA
changed the battery	NA
refiled	NA
checked and rectified self test passed	NA
calibration done and now working good	NA
adjusted screws	NA
CHECKED AND RECTIFIED	NA
replace another mechine	NA
checked and rectified	NA
checked and rectified	NA
checked and rectified	NA
checked and rectified	NA
checked and rectified	NA

checked and rectified	NA		
Checked&rectified	NA		
properly set o2 connection	NA		
done the same	NA		
checked,o2 cell compliant.temporarily arranged to anothor	NA		
DONE	NA		
sevised	NA		
REFILLED	NA		
CHECKED AND RECTIFIED	NA		
CHECKED AND RECTIFIED	NA		
CHECKED AND RECTIFIED	NA		
Done	NA		
checked,not rectified	NA		
checked&rectified	NA		
RECONNECTED THE PROBE	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	CT,JY
SOLVED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
RECTIFIED	NA	NA	CS,AG
GIVEN	NA	NA	CS,AG
CALL LOGGED	NA	NA	CT,JY
RECTIFIED	NA	NA	RR
NON REPAIRABLE, PHYSICAL DAMAGE	NA	NA	AS,RR
refilled	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
RECTIFIED	NA	NA	AG
done the same	NA	NA	AS,RR
checked&rectified	NA	NA	AS,RR
changed bag	NA	NA	AS,RR
need new port	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG
Checked&rectified	NA	NA	CS,AG
checked not repairable,told to send to department for changing	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
checked the water plant and rectified the iisue	NA	NA	RR
checked and rectified	NA	NA	AS,RR
servicer come and checked	NA	NA	CS,AG
arranged from department	NA	NA	AS,RR
Done	NA	NA	AG
Done	NA	NA	AG

Informed to servicer he will visit tomorrow, cartridge locked	NA	NA	CT,JY
Checked and rectified	NA	NA	CT,JY
gauge changed.now working good	NA	NA	RF
checked and rectified	NA	NA	RF
refilled	NA	NA	RF
Sensor compliant.checked&rectified with the help of the service	NA	NA	CT,JY
checked and rectified	NA	NA	AS,RR
checked and rectified	NA	NA	AG
Done	NA	NA	CT,JY
Cassette problem and serviced need to change sensor	NA	NA	CT,JY
Done	NA	NA	RF
Checked and rectified	NA	NA	RF
Checked and rectified	NA	NA	RF
Change flow sensor & rectified	NA	NA	CT,JY
REFILLED	NA	NA	AS,RR
INFORMED SERVICER	NA	NA	AS,CS
CHECKED AND RECTIFIED	NA	NA	AS,RR
RECTIFIED	NA	NA	AS,RR
SET	NA	NA	CS,AG
REFILLED	NA	NA	CS,AG
SOLVED	NA	NA	RR,CT
DONE	NA	NA	CT,JY
RECTIFIED	NA	NA	RR
INFORMED TO SERVICER	NA	NA	AS,RR
INFORMED SERVICER	NA	NA	CS,AG
REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
HORSE CLIP CHANGED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
REFILLED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
checked&rectified.	NA	NA	CT,JY
checked&rectified.	NA	NA	RF
checked&rectified.	NA	NA	CS,AG
FLOW SENSOR AND DIAPHRAM CHANGED	NA	NA	CS,AG
UNDER OBSERVATION	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
SETUPED	NA	NA	AS,RR
REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
SETUPED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG

CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
SETUPED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHANGED THE LAMP	NA	NA	RR,CT
REFILLED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
Done	NA	NA	AS,RR
checked,not rectified	NA	NA	AG
checked&rectified	NA	NA	RF
checked&rectified	NA	NA	RF
checked&rectified	NA	NA	CT,JY
Done	NA	NA	AS,RR
done	NA	NA	AG
done	NA	NA	CT,JY
DONE	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
DONE	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,CS
DONE	NA	NA	AS,RR
DONE	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG
Restart the machine&rectified	NA	NA	CS,AG
DONE	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
	NA	NA	RR
Installed new one	NA	NA	AS,RR
checked&restart the machine.now working good	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
SOLVED	NA	NA	AS,RR
RO WATER REFILLED	NA	NA	CS,AG

CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	RR,CT
RO WATER REFILLED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
SOLVED	NA	NA	CS,AG
CHANGED THE CATRIDGE	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
TIGHTNED THE SCREWS	NA	NA	RF
SOLVED	NA	NA	RF
DONE	NA	NA	CT,JY
CHECKED & O2 CELL NEED TO CHANGE	NA	NA	CT,JY
DONE	NA	NA	RR
checked and rectifieed	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG
insert neal catrigen	NA	NA	AG
CHECKED & NEED TO COMPANY SERVICE	NA	NA	AG
Refilled	NA	NA	AG
CHECKED & RECTIFIED	NA	NA	AS,RR
CHECKED & RECTIFIED	NA	NA	AS,RR
checked and rectified	NA	NA	AS,RR
DONE THE SAME	NA	NA	CT,JY
checked and rectified	NA	NA	RR
insert new holder	NA	NA	AS
Checked&rectified	NA	NA	CS
CHECKED&RECTIFIED	NA	NA	AG
checked&rectified	NA	NA	AG
Checked&rectified	NA	NA	CT,JY
restart the machine.now ok	NA	NA	CT
Bulb changed.now ok	BULB	BULB	CT
DONE	NA	NA	RR
DONE	NA	NA	RR
DONE	NA	NA	CS
checked and rectified	NA	NA	CS
checked and rectified	NA	NA	RF
checked and rectified	NA	NA	RF
DONE	NA	NA	RF
REFILLED	NA	NA	RR
DONE	NA	NA	AS,RR
DONE	NA	NA	CS,AG
Checked&rectified	NA	NA	AG
properly set o2 connection	NA	NA	AG
done the same	NA	NA	AG
checked,o2 cell compliant.temporarily arranged to anothor	NA	NA	AS,RR
DONE	NA	NA	AS,RR

sevised	NA	NA	AS,RR
done	NA	NA	AS,RR
Refilled	NA	NA	CS,AG
done	NA	NA	CS,AG
fuse changed.now working good	NA	NA	RR,CT
lamb changed now ok	NA	NA	CT,JY
checked and rectified	NA	NA	RR
done	NA	NA	AS,RR
done	NA	NA	CS,AG
DONE	NA	NA	AS,RR
DISPLAY DAMAGED NEEDED NEW DISPLAY	NA	NA	AG
checked and rectified	NA	NA	CT,JY
REFILLED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	AS,RR
DONE THE SAME	NA	NA	AS,CS
coil replaced.now working good	COMPANY	COMPANY	AS,RR
checked & found that motor was damaged.change motor n	MOTOR	MOTOR	AS,RR
checked&rectified	NA	NA	CS,AG
refilled	NA	NA	CS,AG
checked&rectified	NA	NA	RR,CT
checked&rectified	NA	NA	CT,JY
done the same	NA	NA	RR
checked&rectified	NA	NA	AS,RR
done	NA	NA	CS,AG
checked&rectified	NA	NA	AS,RR
checked&rectified	NA	NA	AG
done	NA	NA	AG
inform to servicer.part orderd.part changed now working fine	COMPANY	COMPANY	CT,JY
done the same	ABG PAPER	ABG PAPER	CT,JY
Rearrange the remote connection.now working	NA	NA	RF
refilled	NA	NA	RF
checked&rectified	NA	NA	CT
Done the same	NA	NA	AG
refilled	NA	NA	AS
checked&rectified	NA	NA	CS
done	NA	NA	CS,AG
done	NA	NA	AS,RR
DONE	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	CT,JY
CHECKED AND RECTIFIED	NA	NA	CT,JY

DONE	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	RF
CHECKED AND RECTIFIED	NA	NA	CS,AG
DONE	NA	NA	AS,RR
DONE	NA	NA	CT,JY
checked and rectified	NA	NA	RF
Restart the machine&rectified	NA	NA	CS,AG
DONE	NA	NA	CS,AG
checked and rectified	NA	NA	CT,JY
	NA	NA	RR
Installed new one	NA	NA	AS,RR
checked&restart the machine.now working good	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AS,RR
SOLVED	NA	NA	AS,RR
RO WATER REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
RO WATER REFILLED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	RR,CT
CHECKED AND RECTIFIED	NA	NA	CT,JY
SOLVED	NA	NA	RR
CHANGED THE CATRIDGE	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
TIGHTNED THE SCREWS	NA	NA	AS,RR
SOLVED	NA	NA	AG
DONE	NA	NA	AG
CHECKED & O2 CELL NEED TO CHANGE	NA	NA	CT,JY
DONE	NA	NA	CT,JY
checked and rectified	NA	NA	RF
checked and rectified	NA	NA	RF
insert neal catrigen	NA	NA	RF
CHECKED & NEED TO COMPANY SERVICE	NA	NA	CT,JY
Refilled	NA	NA	AS,RR
CHECKED & RECTIFIED	NA	NA	AG
CHECKED & RECTIFIED	NA	NA	CT,JY
checked and rectified	NA	NA	CT,JY
DONE THE SAME	NA	NA	RF
checked and rectified	NA	NA	RF
insert new holder	NA	NA	RF
Checked&rectified	NA	NA	CT,JY
CHECKED&RECTIFIED	NA	NA	AS,RR

checked&rectified	NA	NA	AS,CS
Checked&rectified	NA	NA	AS,RR
restart the machine.now ok	NA	NA	AS,RR
Bulb changed.now ok	BULB	BULB	CS,AG
DONE	NA	NA	CS,AG
DONE	NA	NA	RR,CT
DONE	NA	NA	CT,JY
checked and rectified	NA	NA	RR
checked and rectified	NA	NA	AS,RR
checked and rectified	NA	NA	CS,AG
DONE	NA	NA	AS,RR
REFILLED	NA	NA	AG
DONE	NA	NA	AG
DONE	NA	NA	CT,JY
Checked&rectified	NA	NA	CT,JY
properly set o2 connection	NA	NA	RF
done the same	NA	NA	RF
checked,o2 cell compliant.temporarily arranged to anothor	NA	NA	CS,AG
DONE	NA	NA	AS,RR
sevised	NA	NA	CT,JY
done	NA	NA	RF
Refilled	NA	NA	CS,AG
done	NA	NA	CS,AG
fuse changed.now working good	NA	NA	CT,JY
lamb changed now ok	NA	NA	RR
checked and rectified	NA	NA	AS,RR
done	NA	NA	AS,RR
done	NA	NA	AG
DONE	NA	NA	AG
DISPLAY DAMAGED NEEDED NEW DISPLAY	NA	NA	AG
checked and rectified	NA	NA	AS,RR
REFILLED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	AS,RR
CHECKED AND RECTIFIED	NA	NA	CS,AG
CHECKED AND RECTIFIED	NA	NA	CS,AG
DONE THE SAME	NA	NA	RR,CT
coil replaced.now working good	COMPANY	COMPANY	CT,JY
checked &found that motor was damaged.change motor n	MOTOR	MOTOR	RR
checked&rectified	NA	NA	AS,RR
refilled	NA	NA	CS,AG
checked&rectified	NA	NA	AS,RR
checked&rectified	NA	NA	AG
done the same	NA	NA	RF
checked&rectified	NA	NA	RF

done	NA	NA	CT,JY
checked&rectified	NA	NA	AS,RR
checked&rectified	NA	NA	AG
done	NA	NA	CT,JY
inform to servicer.part orderd.part changed now working fine	COMPANY	COMPANY	CT,JY
done the same	ABG PAPER	ABG PAPER	RF
Rearrange the remote connection.now working	NA	NA	RF
checked&rectified	NA	NA	RF
solved	NA	NA	CT,JY
SOLVED	NA	NA	AS,RR
WATER REFILLED	NA	NA	AS,CS
solved	NA	NA	AS,RR
checked&rectified	NA	NA	AS,RR
checked&rectified			CS,AG
calibration done	NA	NA	CS,AG
changed	NA	NA	RR,CT
checked and rectified	NA	NA	CT,JY
checked and rectified self test passed	NA	NA	RR
change new one	NA	NA	AS,RR
Checked and rectified	NA	NA	CS,AG
Checked and rectified now working good	NA	NA	AG
checked o2 sensor and flow sensor problem rectified now	NA	NA	AG
done	NA	NA	AG
checked and rectified	NA	NA	AS,RR
done	NA	NA	AS,RR
soldered now ok	NA	NA	AS,RR
	NA	NA	AS,RR
checked now working good	NA	NA	CS,AG
soldered now ok	NA	NA	CS,AG
checked and rectified	NA	NA	RR,CT
arranged from department	NA	NA	CT,JY
checked and rectified	NA	NA	RR
calibration done	NA	NA	AS,RR
refiled	NA	NA	CS,AG
replace another mechine	NA	NA	AS,RR
mode change	NA	NA	AG
checked and rectified	NA	NA	RF
checked and rectified	NA	NA	RF
checked and informed servicer	NA	NA	CT,JY
changed the battery	NA	NA	CT,JY
refiled	NA	NA	RR
checked and rectified self test passed	NA	NA	AS,RR
calibration done and now working good	NA	NA	CS,AG
adjusted screws	NA	NA	AG
CHECKED AND RECTIFIED	NA	NA	AG

[illegible]

[illegible]

[illegible]

[illegible]

COMPLAINT ATTENDED DATE	COMPLAIN T ATTENDED	COMPLAIN T CLOSING DATE	COMPLAINT CLOSING TIME
6/1/2023	9:30AM	6/1/2023	11:30AM
6/1/2023	9:00AM	6/1/2023	9:30AM
6/1/2023	4:20PM	6/1/2023	4:30PM
6/1/2023	3:20PM	6/1/2023	3:30PM
6/1/2023	3:00PM	6/1/2023	3:20PM
6/1/2023	2:30PM	6/1/2023	2:50PM
6/1/2023	2:00PM	6/2/2023	2:30PM
6/2/2023	12:30PM	6/2/2023	1:00PM
6/2/2023	12:00PM	6/2/2023	12:30PM
6/2/2023	11:30AM	6/2/2023	12:00PM
6/2/2023	9:45AM	6/2/2023	10:00AM
6/3/2023	9:30PM	6/3/2023	10:20PM
6/3/2023	9:30AM	6/3/2023	9:30AM
6/3/2023	9:00AM	6/3/2023	9:30AM
6/3/2023	8:50PM	6/3/2023	9:20PM
6/4/2023	8:15PM	6/4/2023	8:30PM
6/4/2023	7:30PM	6/4/2023	7:50PM
6/4/2023	6:30PM	6/4/2023	7:30PM
2/4/2023	8:00AM	2/4/2023	8:15AM
2/4/2023	6:00AM	2/4/2023	6:20AM
6/5/2023	4:00PM	6/5/2023	4:30PM
6/5/2023	10:00AM	6/5/2023	11:00AM
6/5/2023	10:30AM	6/5/2023	10:30AM
6/5/2023	10:00AM	6/5/2023	10:15AM
6/5/2023	1:00PM	6/5/2023	1:30PM
6/5/2023	9:30AM	6/5/2023	9:50AM
6/5/2023	9:00PM	6/5/2023	9:15PM
6/5/2023	2:30PM	6/5/2023	3:00PM
6/6/2023	2:00PM	6/6/2026	2:45PM
6/6/2023	10:03AM	6/6/2023	10:10AM
6/6/2023	2:00PM	6/6/2023	2:00PM
6/7/2023	2:00PM	6/7/2023	2:05PM
6/7/2023	6:00PM	6/7/2023	6:30PM
6/7/2023	9:00AM	6/7/2023	09:15AM
6/7/2023	10:00AM	6/7/2023	10:20AM
6/8/2023	10:30AM	6/8/2023	10:50AM
6/8/2023	11:00AM	6/8/2023	11:30AM
6/8/2023	12:00PM	6/8/2023	12:15PM
6/8/2023	1:00PM	6/8/2023	01:14PM
6/8/2023	1:30PM	6/8/2023	01:40PM
6/9/2023	3:00PM	6/9/2023	03:10PM
6/9/2023	3:30PM	6/9/2023	03:50PM

6/9/2023	09:15AM	6/9/2023	09:30AM
6/9/2023	09:30AM	6/9/2023	09:45AM
6/9/2023	11:20AM	6/9/2023	11:55AM
6/9/2023	12:00PM	6/9/2023	12:20PM
6/10/2023	03:00PM	6/10/2023	03:10PM
6/10/2023	03:45PM	6/10/2023	03:55PM
6/10/2023	04:00PM	6/10/2023	04:50PM
6/10/2023	09:00AM	6/10/2023	09:00AM
6/10/2023	10:30AM	6/10/2023	10:30AM
6/11/2023	01:30PM	6/11/2023	01:30PM
6/11/2023	03:45PM	6/11/2023	03:45PM
6/11/2023	09:00AM	6/11/2023	09:00AM
6/11/2023	10:30AM	6/11/2023	10:30AM
6/11/2023	11:15AM	6/11/2023	11:15AM
6/11/2023	01:30PM	6/11/2023	01:30PM
6/12/2023	03:45PM	6/12/2023	03:45PM
6/12/2023	04:15PM	6/12/2023	04:15PM
6/12/2023	04:45PM	6/12/2023	04:45PM
6/12/2023	09:00AM	6/12/2023	09:00AM
6/12/2023	12:00PM	6/12/2023	12:00PM
6/12/2023	02:45PM	6/12/2023	02:45PM
6/12/2023	06:00PM	6/12/2023	06:00PM
6/12/2023	09:00AM	6/12/2023	09:00AM
6/12/2023	10:05AM	6/12/2023	10:05AM
6/13/2023	11:12AM	6/13/2023	11:12AM
6/13/2023	12:50PM	6/13/2023	12:50PM
6/13/2023	09:15AM	6/13/2023	09:15AM
6/13/2023	10:10AM	6/13/2023	10:10AM
6/14/2023	11:20AM	6/14/2023	11:20AM
6/14/2023	01:50PM	6/14/2023	01:50PM
6/14/2023	02:45PM	6/14/2023	02:45PM
6/14/2023	03:45PM	6/14/2023	03:45PM
6/14/2023	09:20AM	6/14/2023	09:20AM
6/14/2023	10:20PM	6/14/2023	10:20PM
6/15/2023	12:35PM	6/15/2023	12:35PM
6/15/2023	02:50PM	6/15/2023	02:50PM
6/15/2023	04:00PM	6/15/2023	04:00PM
6/15/2023	09:30AM	6/15/2023	09:30AM
6/15/2023	10:20PM	6/15/2023	10:20PM
6/15/2023	11:35AM	6/15/2023	11:35AM
6/15/2023	01:40PM	6/15/2023	01:40PM
6/15/2023	03:00PM	6/15/2023	03:00PM
6/16/2023	04:20PM	6/16/2023	04:20PM
6/16/2023	09:15PM	6/16/2023	09:15PM
6/16/2023	10:20PM	6/16/2023	10:20PM

6/16/2023	11:30PM	6/16/2023	11:30PM
6/16/2023	02:10PM	6/16/2023	02:10PM
6/16/2023	03:20PM	6/16/2023	03:20PM
6/16/2023	06:00PM	6/16/2023	06:00PM
6/17/2023	09:30AM	6/17/2023	09:30AM
6/17/2023	11:30AM	6/17/2023	11:30AM
6/17/2023	01:00PM	6/17/2023	01:00PM
6/17/2023	03:50PM	6/17/2023	02:10PM
6/17/2023	04:55PM	6/17/2023	03:20PM
6/17/2023	7:10PM	6/17/2023	06:00PM
6/17/2023	09:10AM	6/17/2023	09:30AM
6/18/2023	10:15AM	6/18/2023	7:20M
6/18/2023	11:30AM	6/18/2023	6:45AM
6/18/2023	12:30PM	6/18/2023	2:45PM
6/18/2023	01:45PM	6/18/2023	12:40PM
6/18/2023	02:30PM	6/18/2023	12:30PM
6/18/2023	03:40PM	6/18/2023	12:15PM
6/18/2023	04:50PM	6/18/2023	2:00PM
6/19/2023	09:50AM	6/19/2023	1:20PM
6/19/2023	11:30AM	6/19/2023	11:00PM
6/19/2023	02:20PM	6/19/2023	9:45PM
6/19/2023	04:30PM	6/19/2023	8:10PM
6/19/2023	10:10AM	6/19/2023	7:20M
6/19/2023	11:20AM	6/19/2023	6:45AM
6/19/2023	01:50PM	6/19/2023	2:45PM
6/19/2023	02:45PM	6/19/2023	12:40PM
6/20/2023	03:45PM	6/20/2023	3:50PM
6/20/2023	09:20AM	6/20/2023	9:45PM
6/20/2023	10:20PM	6/20/2023	9:00PM
6/20/2023	12:35PM	6/20/2023	7:05 PM
6/20/2023	02:50PM	6/20/2023	3:00PM
6/20/2023	04:00PM	6/20/2023	11:15AM
6/20/2023	09:30AM	6/20/2023	6:15PM
6/20/2023	10:20PM	6/20/2023	10:40 AM
6/21/2023	11:35AM	6/21/2023	9:45 AM
6/21/2023	01:40PM	6/21/2023	9:35 AM
6/21/2023	03:00PM	6/21/2023	9:35
6/21/2023	04:20PM	6/21/2023	8:10AM
6/21/2023	09:15PM	6/21/2023	7:50AM
6/21/2023	10:20PM	6/21/2023	8:15AM
6/21/2023	11:30PM	6/21/2023	7:20M
6/21/2023	02:10PM	6/21/2023	6:45AM
6/22/2023	03:20PM	6/22/2023	2:45PM
6/22/2023	06:00PM	6/22/2023	12:40PM
6/22/2023	09:30AM	6/22/2023	12:30PM

6/22/2023	10:20AM	6/22/2023	12:15PM
6/22/2023	11:30AM	6/22/2023	2:00PM
6/22/2023	01:00PM	6/22/2023	1:20PM
6/23/2023	02:20PM	6/23/2023	11:00PM
6/23/2023	3:00pm	6/23/2023	9:45PM
6/23/2023	55:0pm	6/23/2023	8:10PM
6/23/2023	5:30pm	6/23/2023	8:20PM
6/23/2023	9:00am	6/23/2023	8:25PM
6/23/2023	10:00am	6/23/2023	7:35PM
6/23/2023	11:00AM	6/23/2023	7:40PM
6/23/2023	12:00PM	6/23/2023	7:20PM
6/24/2023	2:00PM	6/24/2023	6:20PM
6/24/2023	3:00PM	6/24/2023	7:10PM
6/24/2023	5:00PM	6/24/2023	6:05PM
6/24/2023	9:00PM	6/24/2023	5:30PM
6/24/2023	10:00AM	6/24/2023	3:40PM
6/24/2023	11:00AM	6/24/2023	2:20PM
6/24/2023	12:00PM	6/24/2023	12:30PM
6/24/2023	10:00AM	6/24/2023	11;50PM
6/25/2023	11:00AM	6/25/2023	11:10AM
6/25/2023	12:00PM	6/25/2023	11:05PM
6/25/2023	3:00PM	6/25/2023	9:50AM
6/25/2023	5:00PM	6/25/2023	1:45PM
6/25/2023	10:00AM	6/25/2023	8:45 PM
6/25/2023	11:00AM	6/25/2023	5:45 PM
6/25/2023	12:00PM	6/25/2023	12:20 PM
6/25/2023	1:00PM	6/25/2023	8:50AM
6/26/2023	2:00PM	6/26/2023	8:40AM
6/26/2023	2:30PM	6/26/2023	8:25AM
6/26/2023	3:00PM	6/26/2023	11:10AM
6/26/2023	4:30PM	6/26/2023	10:15PM
6/26/2023	02:10AM	6/26/2023	9:50AM
6/26/2023	05:00AM	6/26/2023	12:30PM
6/26/2023	07:00AM	6/26/2023	9:10PM
6/27/2023	10:30AM	6/27/2023	8:45PM
6/27/2023	01:00AM	6/27/2023	8:25PM
6/27/2023	11:30AM	6/27/2023	8:00PM
6/27/2023	10:15AM	6/27/2023	7:45PM
6/27/2023	11:15PM	6/27/2023	4;10PM
6/27/2023	10:00AM	6/27/2023	3:10PM
6/27/2023	9:30AM	6/27/2023	3:00PM
6/28/2023	12:10PM	6/28/2023	12:40PM
6/28/2023	4:15PM	6/28/2023	11:00PM

6/28/2023	9:45AM	6/28/2023	10:40AM
6/28/2023	9:30AM	6/28/2023	10:40PM
6/28/2023	9:00AM	6/28/2023	10:30AM
6/28/2023	4:20PM	6/28/2023	12:30PM
6/28/2023	4:00PM	6/28/2023	1:25PM
6/28/2023	3:20PM	6/28/2023	9:00AM
3/29/2024	2:30PM	3/29/2024	8:30AM
3/29/2024	1:42PM	3/29/2024	7:40AM
3/29/2024	1:15PM	3/29/2024	6:50AM
3/29/2024	1:10PM	3/29/2024	6:25AM
3/29/2024	9:30AM	3/29/2024	9:35AM
3/29/2024	9:00PM	3/29/2024	9:10PM
3/29/2024	9:00AM	3/29/2024	8:55PM
3/29/2024	8:00PM	3/29/2024	8:20PM
3/30/2024	7:00PM	3/30/2024	7:40PM
3/30/2024	6:00PM	3/30/2024	7:50PM
3/30/2024	5:15PM	3/30/2024	7:20PM
3/30/2024	5:00PM	3/30/2024	5:45PM
3/30/2024	4:00PM	3/30/2024	4:15PM
3/30/2024	3:10PM	3/30/2024	12:05PM
3/30/2024	12:00PM	3/30/2024	9:55PM
3/30/2024	11:30PM	3/30/2024	9:40AM
4/1/2024	11:00PM	4/1/2024	9:45AM
4/1/2024	10:0AM	4/1/2024	9:30AM
4/1/2024	10:0AM	4/1/2024	8:20AM
4/1/2024	10:00PM	4/1/2024	7:40PM
4/1/2024	8:40PM	4/1/2024	4:45PM
4/1/2024	6:55PM	4/1/2024	11:30AM
4/1/2024	5:00AM	4/1/2024	10:55AM
4/1/2024	2:00AM	4/1/2024	10:40AM
4/3/2024	12:30PM	4/3/2024	10:30PM
4/3/2024	12:00PM	4/3/2024	1:45PM
4/3/2024	11:00AM	4/3/2024	10:00PM
4/3/2024	10:00AM	4/3/2024	8:45PM
4/3/2024	1:00PM	4/3/2024	8:05PM
4/3/2024	9:00AM	4/3/2024	7:55PM
4/3/2024	7:35PM	4/3/2024	7:35PM
4/3/2024	7:00AM	4/3/2024	7:10PM
4/6/2024	3:00PM	4/6/2024	6:50PM
4/6/2024	3:00PM	4/6/2024	6:35PM
4/6/2024	9:30AM	4/6/2024	6:20PM
4/6/2024	9:00AM	4/6/2024	10:45PM
4/6/2024	8:10PM	4/6/2024	10:40PM
4/6/2024	7:00PM	4/6/2024	10:25PM
4/6/2024	2:00PM	4/6/2024	10:15PM

4/8/2024	10:30AM	4/8/2024	6:20PM
4/8/2024	1:00PM	4/8/2024	7:45PM
4/8/2024	12:00PM	4/8/2024	5:00PM
4/8/2024	11:00AM	4/8/2024	4:00PM
4/8/2024	10:30AM	4/8/2024	11:20AM
4/8/2024	10:00AM	4/8/2024	11:15AM
4/8/2024	9:30AM	4/8/2024	11:45AM
4/11/2024	9:30AM	4/11/2024	9:15PM
4/11/2024	10:30AM	4/11/2024	9:30AM
4/11/2024	1:30PM	4/11/2024	3:00PM
4/11/2024	9:30PM	4/11/2024	8:55PM
4/11/2024	3:30PM	4/11/2024	6:5PM
4/11/2024	11:30AM	4/11/2024	5:40PM
4/11/2024	10:30PM	4/11/2024	5:55PM
4/15/2024	3:00PM	4/15/2024	12:15PM
4/15/2024	12:30PM	4/15/2024	11:40PM
4/15/2024	11:00AM	4/15/2024	11:00PM
4/15/2024	10:00AM	4/15/2024	11:00PM
4/15/2024	4:00PM	4/15/2024	10:20PM
4/15/2024	3:50PM	4/15/2024	1:15PM
4/15/2024	12:30PM	4/15/2024	10:15PM
4/15/2024	11:30AM	4/15/2024	9:20AM
4/18/2024	10:00AM	4/18/2024	7:15PM
4/18/2024	2:00PM	4/18/2024	7:00PM
4/18/2024	11:00AM	4/18/2024	6:35PM
4/18/2024	10:00PM	4/18/2024	6:20PM
4/18/2024	6:30AM	4/18/2024	3:30PM
4/18/2024	11:00AM	4/18/2024	3:00PM
4/18/2024	9:32PM	4/18/2024	10:50PM
4/18/2024	8:05PM	4/18/2024	10:30PM
4/18/2024	6:35PM	4/18/2024	4:00PM
4/18/2024	12:35PM	4/18/2024	2:20 PM
4/21/2024	11:05AM	4/21/2024	1:45 PM
4/21/2024	6:05 PM	4/21/2024	1:25 PM
4/21/2024	10:32 AM	4/21/2024	11:15 AM
4/21/2024	9:35 AM	4/21/2024	10:20 AM
4/21/2024	9:25 AM	4/21/2024	9:30 AM
4/21/2024	9:17 am	4/21/2024	9:35 AM
4/21/2024	8:00AM	4/21/2024	8:35AM
4/21/2024	7:35AM	4/21/2024	7:30PM
4/24/2024	8:10AM	4/24/2024	12:30PM
4/24/2024	7:00AM	4/24/2024	12:15PM
4/24/2024	6:20AM	4/24/2024	12:30PM
4/24/2024	2:20PM	4/24/2024	11:45PM

4/24/2024	12:30PM	4/24/2024	3:00PM
4/24/2024	12:25PM	4/24/2024	3:30 PM
4/24/2024	11:15AM	4/24/2024	2:45 PM
4/24/2024	1:40PM	4/24/2024	1:10 PM
4/29/2024	1:15PM	4/29/2024	10:30 AM
4/29/2024	10:00PM	4/29/2024	10:00 AM
4/29/2024	9:30AM	4/29/2024	9:30PM
4/29/2024	8:06PM	4/29/2024	9:00PM
4/29/2024	8:15PM	4/29/2024	8:40PM
4/29/2024	8:20PM	4/29/2024	8:45AM
4/29/2024	7:30PM	4/29/2024	8:50AM
4/29/2024	7:20PM	4/29/2024	8:20PM
4/29/2024	7:00PM	4/29/2024	7:45PM
4/29/2024	6:00PM	4/29/2024	6:45PM
4/29/2024	7:05PM	4/29/2024	6:50PM
4/29/2024	5:50PM	4/29/2024	7:20AM
4/29/2024	5:10PM	4/29/2024	6:30PM
4/30/2024	3:20PM	4/30/2024	6:10PM
4/30/2024	2:20PM	4/30/2024	2:10PM
4/30/2024	12:00PM	4/30/2024	11:55PM
4/30/2024	11:45PM	4/30/2024	11:00AM
4/30/2024	11:00AM	4/30/2024	10:10AM
4/30/2024	10:55PM	4/30/2024	8:10PM
4/30/2024	9:40AM	4/30/2024	7:45PM
4/30/2024	1:35PM	4/30/2024	7:15PM
4/30/2024	8:30 PM	4/30/2024	6:20PM
4/30/2024	5:15 PM	4/30/2024	6:20AM
4/30/2024	12:10 PM	4/30/2024	6:00PM
4/30/2024	8:45AM	4/30/2024	5:00PM
4/30/2024	8:30AM	4/30/2024	4:00PM
4/30/2024	8;10AM	4/30/2024	3:40AM
4/30/2024	11:00AM	4/30/2024	3:20AM
5/1/2024	9:35PM	5/1/2024	2:25PM
5/1/2024	9:35AM	5/1/2024	1:30PM
5/1/2024	9:30AM	5/1/2024	12:55AM
5/1/2024	9:05PM	5/1/2024	12:25AM
5/1/2024	8:35PM	5/1/2024	12:30PM
5/1/2024	8:10PM	5/1/2024	11:20PM
5/1/2024	7:50PM	5/1/2024	12:30PM
5/1/2024	7:05PM	5/1/2024	10:45AM
5/3/2024	4:05PM	5/3/2024	11:30AM
5/3/2024	3:00PM	5/3/2024	10:30AM
5/3/2024	2:35PM	5/3/2024	10:30AM
5/3/2024	12:30PM	5/3/2024	2:00PM

5/3/2024	10:55PM	5/3/2024	9:45PM
5/3/2024	10:30AM	5/3/2024	9:40PM
5/3/2024	10:20PM	5/3/2024	9:30PM
5/3/2024	10:15AM	5/3/2024	7:00PM
5/5/2024	10:00AM	5/5/2024	1:45AM
5/5/2024	1:18PM	5/5/2024	2:30PM
5/5/2024	8:45AM	5/5/2024	11:00PM
5/5/2024	8:05AM	5/5/2024	2:00PM
5/5/2024	7:20AM	5/5/2024	9:00PM
5/5/2024	6:35AM	5/5/2024	8:45AM
5/8/2024	6:05AM	5/8/2024	8:15 PM
5/8/2024	9:35AM	5/8/2024	7:15PM
5/8/2024	8:55PM	5/8/2024	6:30PM
5/8/2024	8:50PM	5/8/2024	3:30PM
5/8/2024	8:05PM	5/8/2024	12:00 AM
5/8/2024	7:25PM	5/8/2024	9:30 AM
5/10/2024	7:10PM	5/10/2024	8:30 AM
5/10/2024	6:45PM	5/10/2024	7:45AM
5/10/2024	5:00PM	5/10/2024	9:30AM
5/10/2024	4:05PM	5/10/2024	7:50PM
5/10/2024	12:05PM	5/10/2024	7:45AM
5/13/2024	9:40PM	5/13/2024	11:30PM
5/13/2024	9:35AM	5/13/2024	7:15PM
5/13/2024	9:35AM	5/13/2024	6:10PM
5/13/2024	9:00AM	5/13/2024	11:20PM
5/13/2024	8:00AM	5/13/2024	4:00PM
5/13/2024	7:30PM	5/13/2024	5:00PM
5/13/2024	4:00PM	5/13/2024	3:30PM
5/15/2024	11:00AM	5/15/2024	11:50AM
5/15/2024	10:45AM	5/15/2024	12:30PM
5/15/2024	10:35AM	5/15/2024	11:10 AM
5/15/2024	10:00PM	5/15/2024	11:20AM
5/15/2024	1:00PM	5/15/2024	10:30 AM
5/15/2024	8:53PM	5/15/2024	1:35PM
5/15/2024	8:40PM	5/15/2024	1:30 PM
5/17/2024	7:59PM	5/17/2024	10:55 AM
5/17/2024	7:40PM	5/17/2024	12:20 AM
5/17/2024	7:20PM	5/17/2024	12:15 AM
5/17/2024	7:00PM	5/17/2024	9:45AM
5/17/2024	6:40PM	5/17/2024	11:30PM
5/17/2024	6:24PM	5/17/2024	7:00PM
5/19/2024	6:10PM	5/19/2024	6:40PM
5/19/2024	10:40PM	5/19/2024	6:15PM
5/19/2024	10:25PM	5/19/2024	5:30PM
5/19/2024	10:15PM	5/19/2024	5:00PM

5/19/2024	10:10PM	5/19/2024	1:00PM
5/19/2024	6:00PM	5/19/2024	12:00PM
5/21/2024	5:00PM	5/21/2024	2:00PM
5/21/2024	4:30PM	5/21/2024	12:00PM
5/21/2024	3:00PM	5/21/2024	12:00PM
5/21/2024	11:09AM	5/21/2024	11:55PM
5/21/2024	11:05AM	5/21/2024	1:30PM
5/25/2024	11:00AM	5/25/2024	10:40AM
5/25/2024	9:00PM	5/25/2024	10:20AM
5/25/2024	9:00AM	5/25/2024	5:00PM
5/25/2024	9:00AM	5/25/2024	10:10AM
5/25/2024	8:35PM	5/25/2024	4:15PM
5/25/2024	6:00PM	5/25/2024	11:30PM
5/25/2024	5:35PM	5/25/2024	9:50AM
5/25/2024	5:25PM	5/25/2024	8:00PM
5/25/2024	12:00PM	5/25/2024	7:00PM
5/25/2024	11:00PM	5/25/2024	6:15PM
5/25/2024	10:25PM	5/25/2024	5:00PM
5/25/2024	10:25PM	5/25/2024	3:45PM
5/25/2024	10:00PM	5/25/2024	3:50PM
5/25/2024	1:00PM	5/25/2024	2:50PM
5/25/2024	9:45PM	5/25/2024	3:00PM
5/28/2024	8:55AM	5/28/2024	1:20PM
5/28/2024	7:05PM	5/28/2024	3:30PM
5/28/2024	6:50PM	5/28/2024	12:50PM
5/28/2024	6:25PM	5/28/2024	11:45PM
5/28/2024	6:10PM	5/28/2024	11:45AM
5/28/2024	3:00PM	5/28/2024	11:30AM
5/28/2024	2:30PM	5/28/2024	10:55AM
5/29/2024	10:40PM	5/29/2024	1:45PM
5/29/2024	10:20PM	5/29/2024	1:50 PM
5/29/2024	1:30PM	5/29/2024	10:55AM
5/29/2024	2:10 PM	5/29/2024	9:45AM
5/29/2024	1:30 PM	5/29/2024	3:30PM
5/29/2024	1:15 PM	5/29/2024	2:30PM
5/29/2024	10:50 AM	5/29/2024	12:45AM
5/29/2024	10:10AM	5/29/2024	12:00PM
5/29/2024	9:15 AM	5/29/2024	2:30PM
5/29/2024	9:20 AM	5/29/2024	2:00PM
5/30/2024	8:15AM	5/30/2024	11:15 AM
5/30/2024	5:30PM	5/30/2024	8:30PM
5/30/2024	12:00PM	5/30/2024	12:30PM
5/30/2024	12:00PM	5/30/2024	1:00PM
5/30/2024	12:00 PM	5/30/2024	11:45PM

5/30/2024	11:30AM	5/30/2024	11:30AM
5/30/2024	2:45PM	5/30/2024	9:50AM
5/31/2024	2:35 PM	5/31/2024	4:00PM
5/31/2024	2:30 PM	5/31/2024	3:00PM
5/31/2024	12:45 PM	5/31/2024	
5/31/2024	10:00AM	5/31/2024	12:00PM
5/31/2024	9:30 AM	5/31/2024	11:40AM
5/31/2024	9:10PM	5/31/2024	12:30PM
5/31/2024	8:45PM	5/31/2024	11:00AM
5/31/2024	8:30PM	5/31/2024	9:05 PM
2/1/2024	8:30AM	2/1/2024	7:10 PM
2/1/2024	7:45AM	2/1/2024	6:50 PM
2/1/2024	7:30PM	2/1/2024	6:20 PM
2/1/2024	10:00AM	2/1/2024	2:00PM
2/1/2024	6:40PM	2/1/2024	6:40 PM
2/1/2024	6:35PM	2/1/2024	3:55PM
2/1/2024	6:30AM	2/1/2024	3:30PM
2/1/2024	6:25PM	2/1/2024	2:40PM
8/2/2023	6:05PM	8/2/2023	12:30PM
8/2/2023	2:00PM	8/2/2023	11:00PM
8/2/2023	11:45PM	8/2/2023	10:00AM
8/2/2023	11:00AM	8/2/2023	1:00PM
8/2/2023	10:00AM	8/2/2023	1:30PM
8/2/2023	7:50PM	8/2/2023	7:35 PM
8/2/2023	7:25PM	8/2/2023	6:10 PM
8/3/2023	7:10PM	8/3/2023	6:15 PM
8/3/2023	5:50PM	8/3/2023	8:50 AM
8/3/2023	5:40AM	8/3/2023	8:00 AM
8/3/2023	5:30PM	8/3/2023	10:00AM
8/3/2023	4:00PM	8/3/2023	9:40AM
8/3/2023	3:30PM	8/3/2023	9:40AM
8/3/2023	3:30AM	8/3/2023	9:40AM
8/3/2023	3:15AM	8/3/2023	
8/3/2023	2:20PM	8/3/2023	6:25 AM
8/3/2023	12:30PM	8/3/2023	4:3PM
8/4/2023	12:30AM	8/4/2023	2:10PM
8/4/2023	12:20AM	8/4/2023	2:20PM
8/4/2023	11:30AM	8/4/2023	12:20PM
8/4/2023	11:15PM	8/4/2023	12:10PM
8/4/2023	10:55AM	8/4/2023	12:30PM
8/4/2023	10:30AM	8/4/2023	10:00AM
8/4/2023	10:30AM	8/4/2023	1:10PM

8/5/2023	10:20AM	8/5/2023	10:00AM
8/5/2023	10:00AM	8/5/2023	6:55 PM
8/5/2023	1:30PM	8/5/2023	6:25 PM
8/5/2023	9:35pm	8/5/2023	9:00 AM
8/5/2023	9:30PM	8/5/2023	10:15AM
8/6/2023	9:20PM	8/6/2023	9:40AM
8/6/2023	6:45PM	8/6/2023	9:10PM
8/6/2023	11:40PM	8/6/2023	10:00AM
8/6/2023	10:50PM	8/6/2023	8:00AM
8/6/2023	10:50PM	8/6/2023	
8/6/2023	10:30PM	8/6/2023	4:40PM
8/6/2023	8:45PM	8/6/2023	4:30PM
8/6/2023	8:20 PM	8/6/2023	4:00PM
8/6/2023	7:30 PM	8/6/2023	3:30PM
8/7/2023	7:00PM	8/7/2023	2:30PM
8/7/2023	6:15PM	8/7/2023	11:35AM
8/7/2023	3:00 PM	8/7/2023	11:10AM
8/7/2023	11:30 AM	8/7/2023	10:50AM
8/7/2023	9:10 AM	8/7/2023	10:40AM
8/7/2023	8:00 AM	8/7/2023	11:00AM
8/7/2023	7:30 AM	8/7/2023	10:20AM
8/7/2023	9:20AM	8/7/2023	3:00PM
8/8/2023	7:45PM	8/8/2023	3:45AM
8/8/2023	7:40AM	8/8/2023	8:50AM
8/8/2023	7:20PM	8/8/2023	6:55PM
8/8/2023	6:30AM	8/8/2023	6:25PM
8/8/2023	6:00AM	8/8/2023	4:20PM
8/8/2023	5:50AM	8/8/2023	12:50PM
8/8/2023	3:30AM	8/8/2023	11:20AM
8/8/2023	3:30AM	8/8/2023	11:10PM
8/8/2023	3:00AM	8/8/2023	10:50AM
8/8/2023	11:40AM	8/8/2023	10:25AM
8/9/2023	12:00PM	8/9/2023	10:30AM
8/9/2023	11:00 AM	8/9/2023	1:30PM
8/9/2023	11:00AM	8/9/2023	1:10AM
8/9/2023	10:10 AM	8/9/2023	9:30AM
8/9/2023	1:15PM	8/9/2023	6:40AM
8/9/2023	12:55 PM	8/9/2023	6:20PM
8/9/2023	10:40AM	8/9/2023	4:15PM
8/9/2023	12:10 AM	8/9/2023	12:45PM
8/10/2023	12:00 AM	8/10/2023	12:20PM
8/10/2023	9:15AM	8/10/2023	11:15AM
8/10/2023	11:15PM	8/10/2023	11:30AM

8/10/2023	6:30PM	8/10/2023	11:00AM
8/10/2023	6:30PM	8/10/2023	10:30AM
8/10/2023	5:40PM	8/10/2023	1:15PM
8/10/2023	5:10PM	8/10/2023	1:30PM
8/11/2023	4:35PM	8/11/2023	1:10AM
8/11/2023	12:30PM	8/11/2023	9:30PM
8/11/2023	11:50PM	8/11/2023	9:00PM
8/11/2023	11:30AM	8/11/2023	8:30PM
8/11/2023	11:30AM	8/11/2023	8:45PM
8/11/2023	11:30AM	8/11/2023	7:30PM
8/11/2023	11:50PM	8/11/2023	6:30PM
8/11/2023	10:30AM	8/11/2023	3:45PM
8/12/2023	10:20AM	8/12/2023	10:15PM
8/12/2023	10:00AM	8/12/2023	1:20PM
8/12/2023	10:00AM	8/12/2023	9:00PM
8/12/2023	9:45AM	8/12/2023	7:30PM
8/12/2023	3:50PM	8/12/2023	6:20PM
8/12/2023	9:30AM	8/12/2023	6:00PM
8/12/2023	9:00AM	8/12/2023	4:00PM
8/12/2023	7:30PM	8/12/2023	3:30PM
8/13/2023	6:30PM	8/13/2023	2:30PM
8/13/2023	6:00PM	8/13/2023	5:00PM
8/13/2023	4:30PM	8/13/2023	12:30PM
8/13/2023	3:30PM	8/13/2023	12:30PM
8/13/2023	3:30PM	8/13/2023	11:30AM
8/13/2023	2:30PM	8/13/2023	11:00AM
8/13/2023	2:00PM	8/13/2023	10:30AM
8/14/2023	1:00PM	8/14/2023	10:30PM
8/14/2023	12:30PM	8/14/2023	10:30AM
8/14/2023	12:30PM	8/14/2023	1:30PM
8/14/2023	11:30PM	8/14/2023	10:00AM
8/14/2023	11:30AM	8/14/2023	4:00PM
8/14/2023	11:00AM	8/14/2023	12:30PM
8/14/2023	10:30AM	8/14/2023	12:00PM
8/15/2023	1:30PM	8/15/2023	11:10AM
8/15/2023	1:30 PM	8/15/2023	10:30AM
8/15/2023	10:45 AM	8/15/2023	8:00PM
8/15/2023	9:30AM	8/15/2023	7:00PM
8/15/2023	3:10 PM	8/15/2023	4:30PM
8/15/2023	2:00 PM	8/15/2023	4:00PM
8/15/2023	12:30PAM	8/15/2023	3:00PM
8/15/2023	11:40AM	8/15/2023	2:20PM
8/16/2023	1:50PM	8/16/2023	11:30AM
8/16/2023	1:30PM	8/16/2023	11:30AM
8/16/2023	11:00AM	8/16/2023	10:30AM

8/16/2023	8:00PM	8/16/2023	2:30PM
8/16/2023	12:30PM	8/16/2023	1:20PM
8/16/2023	11:40PM	8/16/2023	
8/16/2023	11:30PM	8/16/2023	5:30PM
8/17/2023	11:30AM	8/17/2023	4:50PM
8/17/2023	9:40AM	8/17/2023	
8/17/2023	3:30PM	8/17/2023	2:30PM
8/17/2023	2:00PM	8/17/2023	1:00PM
8/17/2023	11:40AM	8/17/2023	10:45AM
8/17/2023	11:50AM	8/17/2023	10:20AM
8/17/2023	11:30AM	8/17/2023	10:30AM
8/18/2023	10:00AM	8/18/2023	1:30PM
8/18/2023	10:00AM	8/18/2023	9:00PM
8/18/2023	9:00 PM	8/18/2023	6:30PM
8/18/2023	7:00PM	8/18/2023	4:30PM
8/18/2023	6:45 PM	8/18/2023	3:45PM
8/18/2023	6:00 PM	8/18/2023	3:50PM
8/18/2023	9:30AM	8/18/2023	3:30PM
8/18/2023	6:35 PM	8/18/2023	2:50PM
8/19/2023	3:45PM	8/19/2023	12:30PM
8/19/2023	2:45PM	8/19/2023	3:00PM
8/19/2023	2:21PM	8/19/2023	10:00AM
8/19/2023	11:45AM	8/19/2023	5:00PM
8/19/2023	10:30AM	8/19/2023	9:25AM
8/19/2023	9:30AM	8/19/2023	9:00PM
8/19/2023	12:30PM	8/19/2023	8:30PM
8/19/2023	1:00PM	8/19/2023	7:30PM
8/19/2023	7:10 PM	8/19/2023	6:30PM
8/20/2023	6:05 PM	8/20/2023	11:35PM
8/20/2023	6:10 PM	8/20/2023	12:00PM
8/20/2023	8:35 AM	8/20/2023	3/13/2022
8/20/2023	7:50 AM	8/20/2023	3/13/2022
8/20/2023	9:40AM	8/20/2023	6:00PM
8/20/2023	9:30AM	8/20/2023	5:15PM
8/20/2023	9:30AM	8/20/2023	12:15PM
8/21/2023	9:20AM	8/21/2023	11:40PM
8/21/2023	8:00PM	8/21/2023	11:50AM
8/21/2023	6:10 AM	8/21/2023	11:30PM
8/21/2023	4:15PM	8/21/2023	11:15AM
8/21/2023	2:00PM	8/21/2023	11:00PM
8/21/2023	2:10PM	8/21/2023	10:40AM
8/21/2023	12:00PM	8/21/2023	9:50PM
8/21/2023	12:00PM	8/21/2023	8:30AM
8/22/2023	12:00PM	8/22/2023	7:20AM
8/22/2023	10:00AM	8/22/2023	12:00AM

8/22/2023	1:00PM	8/22/2023	11:45AM
8/22/2023	9:45 PM	8/22/2023	8:40PM
8/22/2023	6:50 PM	8/22/2023	9:30PM
8/22/2023	6:05 PM	8/22/2023	10:20AM
8/23/2023	8:50 AM	8/23/2023	7:20PM
8/23/2023	9:45AM	8/23/2023	1:30PM
8/23/2023	9:20AM	8/23/2023	12:45PM
8/23/2023	9:00PM	8/23/2023	12:10PM
8/23/2023	9:00AM	8/23/2023	3:00AM
8/23/2023	7:45AM	8/23/2023	8:55AM
8/23/2023		8/23/2023	8:40AM
8/23/2023	4:20PM	8/23/2023	5:50AM
8/23/2023	4:00PM	8/23/2023	5:10PM
8/24/2023	3:30PM	8/24/2023	5:10PM
8/24/2023	3:10PM	8/24/2023	
8/24/2023	2:00PM	8/24/2023	3:40PM
8/24/2023	11:15AM	8/24/2023	3:45PM
8/24/2023	10:50AM	8/24/2023	3:00PM
8/24/2023	10:40AM	8/24/2023	12:10AM
8/25/2023	10:30AM	8/25/2023	6:00PM
8/25/2023	10:00AM	8/25/2023	10:50PM
8/25/2023	10:00AM	8/25/2023	11:30AM
8/25/2023	1:00PM	8/25/2023	10:30AM
8/25/2023	9:30AM	8/25/2023	10:30AM
8/25/2023	8:30AM	8/25/2023	5:00PM
8/25/2023	6:45PM	8/25/2023	1:50AM
8/26/2023	6:00PM	8/26/2023	1:30PM
8/26/2023	3:50PM	8/26/2023	11:00PM
8/26/2023	12:00PM	8/26/2023	9:30AM
8/26/2023	11:00AM	8/26/2023	8:00AM
8/26/2023	11:00PM	8/26/2023	
8/26/2023	10:30AM	8/26/2023	4:45PM
8/26/2023	10:15AM	8/26/2023	2:30PM
8/27/2023	10:00AM	8/27/2023	12:15PM
8/27/2023	1:00PM	8/27/2023	12:30AM
8/27/2023	1:00AM	8/27/2023	10:30PM
8/27/2023	9:00AM	8/27/2023	10:15AM
8/27/2023	6:30AM	8/27/2023	11:20AM
8/27/2023	6:00PM	8/27/2023	10:3AM
8/27/2023	4:00PM	8/27/2023	4:50PM
8/27/2023	12:30POM	8/27/2023	4:20PM
8/28/2023	12:00PM	8/28/2023	3:25PM
8/28/2023	11:00AM	8/28/2023	2:45PM
8/28/2023	11:00AM	8/28/2023	4:00PM
8/28/2023	10:00AM	8/28/2023	12:50PM

8/28/2023	10:00AM	8/28/2023	12:20PM
8/28/2023	1:00PM	8/28/2023	12:00PM
8/28/2023	1:00PM	8/28/2023	11:10AM
8/28/2023	1:00AM	8/28/2023	10:30AM
8/29/2023	9:15PM	8/29/2023	7:00PM
8/29/2023	8:30PM	8/29/2023	10:30AM
8/29/2023	8:00PM	8/29/2023	1:20PM
8/29/2023	8:15PM	8/29/2023	7:00PM
8/29/2023	7:15PM	8/29/2023	
8/29/2023	6:00PM	8/29/2023	12:30PM
8/29/2023	3:35PM	8/29/2023	3:00PM
8/30/2023	10:00PM	8/30/2023	1:15PM
8/30/2023	1:15PM	8/30/2023	11:30PAM
8/30/2023	8:30PM	8/30/2023	11:00AM
8/30/2023	7:00PM	8/30/2023	10:45AM
8/30/2023	6:00PM	8/30/2023	11:00AM
8/30/2023	5:30PM	8/30/2023	10:30AM
8/30/2023	3:30PM	8/30/2023	10:20AM
8/30/2023	3:00PM	8/30/2023	4:20PM
9/1/2023	2:00PM	9/1/2023	8:30AM
9/1/2023	12:45PM	9/1/2023	9:00PM
9/1/2023	12:15PM	9/1/2023	7:20PM
9/1/2023	12:00PM	9/1/2023	10:00PM
9/1/2023	11:00AM	9/1/2023	4:00PM
9/1/2023	10:30AM	9/1/2023	3:15PM
9/1/2023	10:15AM	9/1/2023	3:00PM
9/1/2023	10:00PM	9/1/2023	2:45PM
9/1/2023	10:00AM	9/1/2023	2:30PM
9/2/2023	1:00PM	9/2/2023	12:45PM
9/2/2023	9:00AM	9/2/2023	3:00PM
9/2/2023	3:30PM	9/2/2023	11:40AM
9/2/2023	12:00PM	9/2/2023	11:25AM
9/2/2023	11:15AM	9/2/2023	11:10AM
9/2/2023	11:00AM	9/2/2023	11:20AM
9/2/2023	10:00AM	9/2/2023	10:50AM
9/2/2023	7:00PM	9/2/2023	11:00AM
9/2/2023	6:00PM	9/2/2023	10:20AM
9/3/2023	4:00PM	9/3/2023	2:30PM
9/3/2023	3:00PM	9/3/2023	3:00PM
9/3/2023	2:30PM	9/3/2023	1:20PM
9/3/2023	2:00PM	9/3/2023	1:15PM
9/3/2023	11:40AM	9/3/2023	9:40AM
9/3/2023	11:00AM	9/3/2023	4:45PM
9/3/2023	10:30AM	9/3/2023	2:30PM

9/4/2023	1:30PM	9/4/2023	1:30PM
9/4/2023	1:00PM	9/4/2023	12:30PM
9/4/2023	9:00AM	9/4/2023	12:30AM
9/4/2023	4:50PM	9/4/2023	11:00AM
9/4/2023	4:30PM	9/4/2023	10:30AM
9/4/2023	2:30PM	9/4/2023	9:50AM
9/4/2023	1:30PM	9/4/2023	9:00PM
9/4/2023	12:00PM	9/4/2023	4:10PM
9/5/2023	10:30AM	9/5/2023	3:30PM
9/5/2023	10:00AM	9/5/2023	4:00PM
9/5/2023	10:00AM	9/5/2023	3:00PM
9/5/2023	1:00PM	9/5/2023	12:30PM
9/5/2023	8:30PM	9/5/2023	11:30PM
9/5/2023	6:00PM	9/5/2023	11:30AM
9/5/2023	4:00PM	9/5/2023	10:30AM
9/5/2023	3:30PM	9/5/2023	10:30PM
9/5/2023	3:30PM	9/5/2023	10:20AM
9/6/2023	3:00PM	9/6/2023	2:00PM
9/6/2023	2:30PM	9/6/2023	1:50AM
9/6/2023	12:00PM	9/6/2023	10:00AM
9/6/2023	2:30PM	9/6/2023	9:20PM
9/6/2023	9:50AM	9/6/2023	9:15AM
9/6/2023	9:30AM	9/6/2023	8:15PM
9/6/2023	9:10AM	9/6/2023	4:00PM
9/6/2023	8:30PM	9/6/2023	9:45PM
9/6/2023	7:30PM	9/6/2023	10:00AM
9/7/2023	7:00PM	9/7/2023	8:50PM
9/7/2023	5:30PM	9/7/2023	10:30PM
9/7/2023	11:30PM	9/7/2023	12:00PM
9/7/2023	11:30AM	9/7/2023	7:15AM
9/7/2023	3/13/2022	9/7/2023	6:00PM
9/7/2023	3/13/2022	9/7/2023	2:30PM
9/7/2023	3/13/2022	9/7/2023	1:00PM
9/8/2023	3/13/2022	9/8/2023	11:20PM
9/8/2023	3/13/2022	9/8/2023	11:15PM
9/8/2023	3/13/2022	9/8/2023	12:00PM
9/8/2023	3/13/2022	9/8/2023	10:30AM
9/8/2023	5:30PM	9/8/2023	1:15PM
9/8/2023	5:00PM	9/8/2023	10:00AM
9/8/2023	12:00PM	9/8/2023	4:00PM
9/8/2023	11:30PM	9/8/2023	3:30PM
9/8/2023	11:40AM	9/8/2023	3:20PM
9/9/2023	11:10PM	9/9/2023	3:00PM
9/9/2023	11:00AM	9/9/2023	12:30PM

9/9/2023	10:30PM	9/9/2023	2:00PM
9/9/2023	10:30AM	9/9/2023	11:40AM
9/9/2023	9:30PM	9/9/2023	11:30PM
9/9/2023	8:10AM	9/9/2023	11:20PM
9/10/2023	7:10AM	9/10/2023	12:00PM
9/10/2023	11:50PM	9/10/2023	11:00AM
9/10/2023	11:30AM	9/10/2023	1:50PM
9/10/2023	8:30PM	9/10/2023	2:00PM
9/10/2023	9:00AM	9/10/2023	1:20PM
9/10/2023	10:00AM	9/10/2023	10:00AM
9/10/2023	7:00PM	9/10/2023	4:30PM
9/10/2023	1:15PM	9/10/2023	4:30PM
9/11/2023	12:10PM	9/11/2023	3:00PM
9/11/2023	11:40AM	9/11/2023	1:30PM
9/11/2023	11:00PM	9/11/2023	12:30AM
9/11/2023	8:45AM	9/11/2023	12:20PM
9/11/2023	8:20AM	9/11/2023	12:00PM
9/11/2023	5:30AM	9/11/2023	9:50AM
9/11/2023	5:00PM	9/11/2023	9:00PM
9/11/2023	4:40PM	9/11/2023	7:30PM
9/11/2023		9/11/2023	6:30AM
9/12/2023	3:15PM	9/12/2023	3:40PM
9/12/2023	3:15PM	9/12/2023	2:20PM
9/12/2023	2:30PM	9/12/2023	3:30PM
9/12/2023	11:40AM	9/12/2023	12:20AM
9/12/2023	11:00PAM	9/12/2023	1:00PM
9/12/2023	10:30PM	9/12/2023	11:30PM
9/12/2023	10:30AM	9/12/2023	11:15AM
9/12/2023	10:00AM	9/12/2023	10:20AM
9/12/2023	10:00AM	9/12/2023	11:00AM
9/13/2023	11:00AM	9/13/2023	11:00AM
9/13/2023	1:30AM	9/13/2023	1:20PM
9/13/2023	1:10PM	9/13/2023	1:30PM
9/13/2023	9:30PM	9/13/2023	1:30PM
15--09-2023	12:30PM	15--09-2023	10:15AM
9/13/2023	9:15AM	9/13/2023	1:10PM
9/13/2023	7:50AM	9/13/2023	10:00PM
9/13/2023	4:15PM	9/13/2023	8:40PM
9/13/2023	4:00PM	9/13/2023	7:30PM
9/13/2023	2:00PM	9/13/2023	5:35PM
9/14/2023	12:00PM	9/14/2023	3:05PM
9/14/2023	12:00AM	9/14/2023	3:15PM
9/14/2023	10:15PM	9/14/2023	3:00PM
9/14/2023	10:00AM	9/14/2023	2:30PM
9/14/2023	11:00AM	9/14/2023	1:00PM

9/14/2023	9:45AM	9/14/2023	12:35PM
9/14/2023	4:30PM	9/14/2023	11:50AM
9/14/2023	4:00PM	9/14/2023	11:15AM
9/14/2023	3:15PM	9/14/2023	11:15AM
9/14/2023	2:30PM	9/14/2023	11:20AM
15--09-2023	2:00PM	15--09-2023	11:00AM
15--09-2023	12:30PM	15--09-2023	10:15AM
15--09-2023	12:15PM	15--09-2023	2:00PM
15--09-2023	11:20AM	15--09-2023	9:50PM
15--09-2023	10:45AM	15--09-2023	9:30AM
15--09-2023	10:15AM	15--09-2023	8:00PM
15--09-2023	6:00PM	15--09-2023	12:45PM
16--09-2023	1:00PM	16--09-2023	4:15PM
15--09-2023	10:20PM	15--09-2023	11:30AM
15--09-2023	1:00PM	15--09-2023	10:30PM
16--09-2023	3:00PM	16--09-2023	2:00PM
16--09-2023		16--09-2023	9:30AM
16--09-2023	12:00PM	16--09-2023	9:30PM
16--09-2023	12:00APM	16--09-2023	6:50AM
16--09-2023	1:00PM	16--09-2023	4:15PM
16--09-2023	11:00AM	16--09-2023	4:30PM
9/17/2023	10:40AM	9/17/2023	4:00PM
9/17/2023	10:30AM	9/17/2023	3:30PM
9/17/2023	10:30PAM	9/17/2023	4:00PM
9/17/2023	10:00AM	9/17/2023	11:30PM
9/18/2023	3:00PM	9/18/2023	11:30AM
9/17/2023	10:00AM	9/17/2023	11:30PM
9/17/2023	4:00PM	9/17/2023	12:20PM
9/17/2023	8:00AM	9/17/2023	10:40PM
9/17/2023	7:20PM	9/17/2023	2:00PM
9/18/2023	7:00PM	9/18/2023	9:30PM
9/18/2023	6:00PM	9/18/2023	9:20AM
9/18/2023	3:30PM	9/18/2023	3:45PM
9/18/2023	3:00PM	9/18/2023	11:30AM
9/18/2023	2:45PM	9/18/2023	10:30PM
9/18/2023	2:30PM	9/18/2023	9:35PM
9/18/2023	2:00PM	9/18/2023	10:30AM
9/18/2023	12:30PM	9/18/2023	9:00PM
9/19/2023	12:00PM	9/19/2023	8:30PM
9/19/2023	11:25AM	9/19/2023	6:00PM
9/18/2023	12:30PM	9/18/2023	9:00PM
9/19/2023	11:15AM	9/19/2023	4:10PM
9/19/2023	11:00AM	9/19/2023	3:50PM
9/19/2023	11:00AM	9/19/2023	3:40PM

9/19/2023	10:40AM	9/19/2023	12:20PM
9/19/2023	10:30AM	9/19/2023	11:50PM
9/19/2023	10:00AM	9/19/2023	12:00PM
9/20/2023	1:30PM	9/20/2023	10:50PM
9/20/2023	1:00PM	9/20/2023	10:30PM
9/20/2023	1:00PM	9/20/2023	11:00AM
9/20/2023	1:00PM	9/20/2023	9:55PM
9/20/2023	9:20AM	9/20/2023	10:15PM
9/20/2023	4:00PM	9/20/2023	10:00PM
9/20/2023	2:00PM	9/20/2023	9:20AM
9/20/2023	12:50PM	9/20/2023	8:45PM
9/21/2023	12:00PM	9/21/2023	6:50PM
9/21/2023	11:00AM	9/21/2023	7:10AM
9/21/2023	10:30AM	9/21/2023	6:30PM
9/21/2023	10:00AM	9/21/2023	6:25AM
9/21/2023	9:30AM	9/21/2023	6:15PM
9/21/2023	8:40PM	9/21/2023	5:55AM
9/21/2023	3:30PM	9/21/2023	5:40PM
9/21/2023	8:40PM	9/21/2023	5:55AM
9/21/2023	3:00PM	9/21/2023	5:45AM
9/22/2023	2:50PM	9/22/2023	4:30PM
9/22/2023	12:00PM	9/22/2023	1:00AM
9/22/2023	11:40PM	9/22/2023	3:30PM
9/22/2023	11:00AM	9/22/2023	12:35AM
9/22/2023	10:30AM	9/22/2023	12:20PM
9/22/2023	10:20AM	9/22/2023	11:45PM
9/22/2023	11:40PM	9/22/2023	3:30PM
9/22/2023	10:00PM	9/22/2023	12:00PM
9/22/2023	10:10AM	9/22/2023	12:00PM
9/22/2023	10:00PM	9/22/2023	12:00PM
9/23/2023	1:40AM	9/23/2023	11:20AM
9/23/2023	1:30AM	9/23/2023	11:10PM
9/23/2023	9:30AM	9/23/2023	2:00PM
9/23/2023	9:00PM	9/23/2023	1:35PM
9/23/2023	9:00AM	9/23/2023	2:00PM
9/23/2023	9:00AM	9/23/2023	10:50PM
9/23/2023	8:45PM	9/23/2023	10:30PM
9/23/2023	8:45AM	9/23/2023	11:00AM
9/24/2023	8:00PM	9/24/2023	9:55PM
9/24/2023	3:00PM	9/24/2023	10:15PM
9/24/2023	9:30PM	9/24/2023	10:00PM
9/24/2023	9:30PM	9/24/2023	10:00PM
9/24/2023	9:30AM	9/24/2023	9:20AM

9/24/2023	8:40PM	9/24/2023	8:45PM
9/24/2023	10:00PM	9/24/2023	6:50PM
9/24/2023	10:00AM	9/24/2023	7:10AM
9/25/2023	6:45AM	9/25/2023	6:30PM
9/25/2023	5:45PM	9/25/2023	6:25AM
9/25/2023	2:00PM	9/25/2023	6:15PM
9/25/2023	12:00PM	9/25/2023	5:55AM
9/25/2023	11:15PM	9/25/2023	5:40PM
9/25/2023	11:00PM	9/25/2023	5:45AM
9/25/2023	11:00AM	9/25/2023	4:30PM
9/25/2023	10::00AM	9/25/2023	1:00AM
9/26/2023	9:45 AM		
9/26/2023	10:30 AM		
9/26/2023	1:45 PM	9/26/2023	5:00PM
9/26/2023	10:20AM	9/27/2023	4:30PM
9/26/2023	11:00 AM	9/26/2023	11:15AM
9/26/2023	12:03PM	9/26/2023	12:15PM
9/26/2023	2:00PM	9/26/2023	2:13PM
9/26/2023	6:10PM	9/26/2023	6:20
9/26/2023	9:05PM	9/26/2023	9:40PM
9/26/2023	8:05AM	9/26/2023	8:30 AM
9/27/2023	11:30AM	9/27/2023	12:00PM
9/27/2023	9:45 AM	9/27/2023	10:15AM
27/58/2023	12:00PM		
9/27/2023	7:05PM	9/27/2023	7:20PM
9/27/2023	10:05PM	9/27/2023	11:10PM
9/27/2023	2:10AM	9/27/2023	2:25AM
9/28/2023	7:10AM	9/28/2023	7:30AM
9/28/2023	8:10AM	9/28/2023	8:30AM
9/28/2023	10:00AM	9/28/2023	10:20AM
10/1/2023	8:30AM	10/1/2023	7:10 PM
10/1/2023	7:45AM	10/1/2023	6:50 PM
10/1/2023	7:30PM	10/1/2023	6:20 PM
10/1/2023	10:00AM	10/1/2023	2:00PM
10/1/2023	6:40PM	10/1/2023	6:40 PM
10/1/2023	6:35PM	10/1/2023	3:55PM
10/1/2023	6:30AM	10/1/2023	3:30PM
10/1/2023	6:25PM	10/1/2023	2:40PM
10/2/2023	6:05PM	10/2/2023	12:30PM
10/2/2023	2:00PM	10/2/2023	11:00PM
10/2/2023	11:45PM	10/2/2023	10:00AM
10/2/2023	11:00AM	10/2/2023	1:00PM

10/2/2023	10:00AM	10/2/2023	1:30PM
10/2/2023	7:50PM	10/2/2023	7:35 PM
10/2/2023	7:25PM	10/2/2023	6:10 PM
10/3/2023	7:10PM	10/3/2023	6:15 PM
10/3/2023	5:50PM	10/3/2023	8:50 AM
10/3/2023	5:40AM	10/3/2023	8:00 AM
10/3/2023	5:30PM	10/3/2023	10:00AM
10/3/2023	4:00PM	10/3/2023	9:40AM
10/3/2023	3:30PM	10/3/2023	9:40AM
10/3/2023	3:30AM	10/3/2023	9:40AM
10/3/2023	3:15AM	10/3/2023	
10/3/2023	2:20PM	10/3/2023	6:25 AM
10/3/2023	12:30PM	10/3/2023	4:3PM
10/4/2023	12:30AM	10/4/2023	2:10PM
10/4/2023	12:20AM	10/4/2023	2:20PM
10/4/2023	11:30AM	10/4/2023	12:20PM
10/4/2023	11:15PM	10/4/2023	12:10PM
10/4/2023	10:55AM	10/4/2023	12:30PM
10/4/2023	10:30AM	10/4/2023	10:00AM
10/4/2023	10:30AM	10/4/2023	1:10PM
10/5/2023	10:20AM	10/5/2023	10:00AM
10/5/2023	10:00AM	10/5/2023	6:55 PM
10/5/2023	1:30PM	10/5/2023	6:25 PM
10/5/2023	9:35pm	10/5/2023	9:00 AM
10/5/2023	9:30PM	10/5/2023	10:15AM
10/6/2023	9:20PM	10/6/2023	9:40AM
10/6/2023	6:45PM	10/6/2023	9:10PM
10/6/2023	11:40PM	10/6/2023	10:00AM
10/6/2023	10:50PM	10/6/2023	8:00AM
10/6/2023	10:50PM	10/6/2023	
10/6/2023	10:30PM	10/6/2023	4:40PM
10/6/2023	8:45PM	10/6/2023	4:30PM
10/6/2023	8:20 PM	10/6/2023	4:00PM
10/6/2023	7:30 PM	10/6/2023	3:30PM
10/7/2023	7:00PM	10/7/2023	2:30PM
10/7/2023	6:15PM	10/7/2023	11:35AM
10/7/2023	3:00 PM	10/7/2023	11:10AM
10/7/2023	11:30 AM	10/7/2023	10:50AM
10/7/2023	9:10 AM	10/7/2023	10:40AM
10/7/2023	8:00 AM	10/7/2023	11:00AM
10/7/2023	7:30 AM	10/7/2023	10:20AM

10/7/2023	9:20AM	10/7/2023	3:00PM
10/8/2023	7:45PM	10/8/2023	3:45AM
10/8/2023	7:40AM	10/8/2023	8:50AM
10/8/2023	7:20PM	10/8/2023	6:55PM
10/8/2023	6:30AM	10/8/2023	6:25PM
10/8/2023	6:00AM	10/8/2023	4:20PM
10/8/2023	5:50AM	10/8/2023	12:50PM
10/8/2023	3:30AM	10/8/2023	11:20AM
10/8/2023	3:30AM	10/8/2023	11:10PM
10/8/2023	3:00AM	10/8/2023	10:50AM
10/8/2023	11:40AM	10/8/2023	10:25AM
10/9/2023	12:00PM	10/9/2023	10:30AM
10/9/2023	11:00 AM	10/9/2023	1:30PM
10/9/2023	11:00AM	10/9/2023	1:10AM
10/9/2023	10:10 AM	10/9/2023	9:30AM
10/9/2023	1:15PM	10/9/2023	6:40AM
10/9/2023	12:55 PM	10/9/2023	6:20PM
10/9/2023	10:40AM	10/9/2023	4:15PM
10/9/2023	12:10 AM	10/9/2023	12:45PM
10/10/2023	12:00 AM	10/10/2023	12:20PM
10/10/2023	9:15AM	10/10/2023	11:15AM
10/10/2023	11:15PM	10/10/2023	11:30AM
10/10/2023	6:30PM	10/10/2023	11:00AM
10/10/2023	6:30PM	10/10/2023	10:30AM
10/10/2023	5:40PM	10/10/2023	1:15PM
10/10/2023	5:10PM	10/10/2023	1:30PM
10/11/2023	4:35PM	10/11/2023	1:10AM
10/11/2023	12:30PM	10/11/2023	9:30PM
10/11/2023	11:50PM	10/11/2023	9:00PM
10/11/2023	11:30AM	10/11/2023	8:30PM
10/11/2023	11:30AM	10/11/2023	8:45PM
10/11/2023	11:30AM	10/11/2023	7:30PM
10/11/2023	11:50PM	10/11/2023	6:30PM
10/11/2023	10:30AM	10/11/2023	3:45PM
10/12/2023	10:20AM	10/12/2023	10:15PM
10/12/2023	10:00AM	10/12/2023	1:20PM
10/12/2023	10:00AM	10/12/2023	9:00PM
10/12/2023	9:45AM	10/12/2023	7:30PM
10/12/2023	3:50PM	10/12/2023	6:20PM
10/12/2023	9:30AM	10/12/2023	6:00PM
10/12/2023	9:00AM	10/12/2023	4:00PM
10/12/2023	7:30PM	10/12/2023	3:30PM
10/13/2023	6:30PM	10/13/2023	2:30PM
10/13/2023	6:00PM	10/13/2023	5:00PM
10/13/2023	4:30PM	10/13/2023	12:30PM

10/13/2023	3:30PM	10/13/2023	12:30PM
10/13/2023	3:30PM	10/13/2023	11:30AM
10/13/2023	2:30PM	10/13/2023	11:00AM
10/13/2023	2:00PM	10/13/2023	10:30AM
10/14/2023	1:00PM	10/14/2023	10:30PM
10/14/2023	12:30PM	10/14/2023	10:30AM
10/14/2023	12:30PM	10/14/2023	1:30PM
10/14/2023	11:30PM	10/14/2023	10:00AM
10/14/2023	11:30AM	10/14/2023	4:00PM
10/14/2023	11:00AM	10/14/2023	12:30PM
10/14/2023	10:30AM	10/14/2023	12:00PM
10/15/2023	1:30PM	10/15/2023	11:10AM
10/15/2023	1:30 PM	10/15/2023	10:30AM
10/15/2023	10:45 AM	10/15/2023	8:00PM
10/15/2023	9:30AM	10/15/2023	7:00PM
10/15/2023	3:10 PM	10/15/2023	4:30PM
10/15/2023	2:00 PM	10/15/2023	4:00PM
10/15/2023	12:30PAM	10/15/2023	3:00PM
10/15/2023	11:40AM	10/15/2023	2:20PM
10/16/2023	1:50PM	10/16/2023	11:30AM
10/16/2023	1:30PM	10/16/2023	11:30AM
10/16/2023	11:00AM	10/16/2023	10:30AM
10/16/2023	8:00PM	10/16/2023	2:30PM
10/16/2023	12:30PM	10/16/2023	1:20PM
10/16/2023	11:40PM	10/16/2023	
10/16/2023	11:30PM	10/16/2023	5:30PM
10/17/2023	11:30AM	10/17/2023	4:50PM
10/17/2023	9:40AM	10/17/2023	
10/17/2023	3:30PM	10/17/2023	2:30PM
10/17/2023	2:00PM	10/17/2023	1:00PM
10/17/2023	11:40AM	10/17/2023	10:45AM
10/17/2023	11:50AM	10/17/2023	10:20AM
10/17/2023	11:30AM	10/17/2023	10:30AM
10/18/2023	10:00AM	10/18/2023	1:30PM
10/18/2023	10:00AM	10/18/2023	9:00PM
10/18/2023	9:00 PM	10/18/2023	6:30PM
10/18/2023	7:00PM	10/18/2023	4:30PM
10/18/2023	6:45 PM	10/18/2023	3:45PM
10/18/2023	6:00 PM	10/18/2023	3:50PM
10/18/2023	9:30AM	10/18/2023	3:30PM
10/18/2023	6:35 PM	10/18/2023	2:50PM
10/19/2023	3:45PM	10/19/2023	12:30PM
10/19/2023	2:45PM	10/19/2023	3:00PM
10/19/2023	2:21PM	10/19/2023	10:00AM
10/19/2023	11:45AM	10/19/2023	5:00PM

10/19/2023	10:30AM	10/19/2023	9:25AM
10/19/2023	9:30AM	10/19/2023	9:00PM
10/19/2023	12:30PM	10/19/2023	8:30PM
10/19/2023	1:00PM	10/19/2023	7:30PM
10/19/2023	7:10 PM	10/19/2023	6:30PM
10/20/2023	6:05 PM	10/20/2023	11:35PM
10/20/2023	6:10 PM	10/20/2023	12:00PM
10/20/2023	8:35 AM	10/20/2023	3/13/2022
10/20/2023	7:50 AM	10/20/2023	3/13/2022
10/20/2023	9:40AM	10/20/2023	6:00PM
10/20/2023	9:30AM	10/20/2023	5:15PM
10/20/2023	9:30AM	10/20/2023	12:15PM
10/21/2023	9:20AM	10/21/2023	11:40PM
10/21/2023	8:00PM	10/21/2023	11:50AM
10/21/2023	6:10 AM	10/21/2023	11:30PM
10/21/2023	4:15PM	10/21/2023	11:15AM
10/21/2023	2:00PM	10/21/2023	11:00PM
10/21/2023	2:10PM	10/21/2023	10:40AM
10/21/2023	12:00PM	10/21/2023	9:50PM
10/21/2023	12:00PM	10/21/2023	8:30AM
10/22/2023	12:00PM	10/22/2023	7:20AM
10/22/2023	10:00AM	10/22/2023	12:00AM
10/22/2023	1:00PM	10/22/2023	11:45AM
10/22/2023	9:45 PM	10/22/2023	8:40PM
10/22/2023	6:50 PM	10/22/2023	9:30PM
10/22/2023	6:05 PM	10/22/2023	10:20AM
10/23/2023	8:50 AM	10/23/2023	7:20PM
10/23/2023	9:45AM	10/23/2023	1:30PM
10/23/2023	9:20AM	10/23/2023	12:45PM
10/23/2023	9:00PM	10/23/2023	12:10PM
10/23/2023	9:00AM	10/23/2023	3:00AM
10/23/2023	7:45AM	10/23/2023	8:55AM
10/23/2023		10/23/2023	8:40AM
10/23/2023	4:20PM	10/23/2023	5:50AM
10/23/2023	4:00PM	10/23/2023	5:10PM
10/24/2023	3:30PM	10/24/2023	5:10PM
10/24/2023	3:10PM	10/24/2023	
10/24/2023	2:00PM	10/24/2023	3:40PM
10/24/2023	11:15AM	10/24/2023	3:45PM
10/24/2023	10:50AM	10/24/2023	3:00PM
10/24/2023	10:40AM	10/24/2023	12:10AM
10/25/2023	10:30AM	10/25/2023	6:00PM
10/25/2023	10:00AM	10/25/2023	10:50PM
10/25/2023	10:00AM	10/25/2023	11:30AM
10/25/2023	1:00PM	10/25/2023	10:30AM

10/25/2023	9:30AM	10/25/2023	10:30AM
10/25/2023	8:30AM	10/25/2023	5:00PM
10/25/2023	6:45PM	10/25/2023	1:50AM
10/26/2023	6:00PM	10/26/2023	1:30PM
10/26/2023	3:50PM	10/26/2023	11:00PM
10/26/2023	12:00PM	10/26/2023	9:30AM
10/26/2023	11:00AM	10/26/2023	8:00AM
10/26/2023	11:00PM	10/26/2023	
10/26/2023	10:30AM	10/26/2023	4:45PM
10/26/2023	10:15AM	10/26/2023	2:30PM
10/27/2023	10:00AM	10/27/2023	12:15PM
10/27/2023	1:00PM	10/27/2023	12:30AM
10/27/2023	1:00AM	10/27/2023	10:30PM
10/27/2023	9:00AM	10/27/2023	10:15AM
10/27/2023	6:30AM	10/27/2023	11:20AM
10/27/2023	6:00PM	10/27/2023	10:3AM
10/27/2023	4:00PM	10/27/2023	4:50PM
10/27/2023	12:30POM	10/27/2023	4:20PM
10/28/2023	12:00PM	10/28/2023	3:25PM
10/28/2023	11:00AM	10/28/2023	2:45PM
10/28/2023	11:00AM	10/28/2023	4:00PM
10/28/2023	10:00AM	10/28/2023	12:50PM
10/28/2023	10:00AM	10/28/2023	12:20PM
10/28/2023	1:00PM	10/28/2023	12:00PM
10/28/2023	1:00PM	10/28/2023	11:10AM
10/28/2023	1:00AM	10/28/2023	10:30AM
10/29/2023	9:15PM	10/29/2023	7:00PM
10/29/2023	8:30PM	10/29/2023	10:30AM
10/29/2023	8:00PM	10/29/2023	1:20PM
10/29/2023	8:15PM	10/29/2023	7:00PM
10/29/2023	7:15PM	10/29/2023	
10/29/2023	6:00PM	10/29/2023	12:30PM
10/29/2023	3:35PM	10/29/2023	3:00PM
10/30/2023	10:00PM	10/30/2023	1:15PM
10/30/2023	1:15PM	10/30/2023	11:30PAM
10/30/2023	8:30PM	10/30/2023	11:00AM
10/30/2023	7:00PM	10/30/2023	10:45AM
10/30/2023	6:00PM	10/30/2023	11:00AM
10/30/2023	5:30PM	10/30/2023	10:30AM
10/30/2023	3:30PM	10/30/2023	10:20AM
10/30/2023	3:00PM	10/30/2023	4:20PM
11/1/2023	9:30AM	11/1/2023	11:30AM
11/1/2023	9:00AM	11/1/2023	9:30AM
11/1/2023	4:20PM	11/1/2023	4:30PM
11/1/2023	3:20PM	11/1/2023	3:30PM

11/1/2023	3:00PM	11/1/2023	3:20PM
11/1/2023	2:30PM	11/1/2023	2:50PM
11/1/2023	2:00PM	11/2/2023	2:30PM
11/2/2023	12:30PM	11/2/2023	1:00PM
11/2/2023	12:00PM	11/2/2023	12:30PM
11/2/2023	11:30AM	11/2/2023	12:00PM
11/2/2023	9:45AM	11/2/2023	10:00AM
11/3/2023	9:30PM	11/3/2023	10:20PM
11/3/2023	9:30AM	11/3/2023	9:30AM
11/3/2023	9:00AM	11/3/2023	9:30AM
11/3/2023	8:50PM	11/3/2023	9:20PM
11/4/2023	8:15PM	11/4/2023	8:30PM
11/4/2023	7:30PM	11/4/2023	7:50PM
11/4/2023	6:30PM	11/4/2023	7:30PM
2/4/2023	8:00AM	2/4/2023	8:15AM
2/4/2023	6:00AM	2/4/2023	6:20AM
11/5/2023	4:00PM	11/5/2023	4:30PM
11/5/2023	10:00AM	11/5/2023	11:00AM
11/5/2023	10:30AM	11/5/2023	10:30AM
11/5/2023	10:00AM	11/5/2023	10:15AM
11/5/2023	1:00PM	11/5/2023	1:30PM
11/5/2023	9:30AM	11/5/2023	9:50AM
11/5/2023	9:00PM	11/5/2023	9:15PM
11/5/2023	2:30PM	11/5/2023	3:00PM
11/6/2023	2:00PM	6/6/2026	2:45PM
11/6/2023	10:03AM	11/6/2023	10:10AM
11/6/2023	2:00PM	11/6/2023	2:00PM
11/7/2023	2:00PM	11/7/2023	2:05PM
11/7/2023	6:00PM	11/7/2023	6:30PM
11/7/2023	9:00AM	11/7/2023	09:15AM
11/7/2023	10:00AM	11/7/2023	10:20AM
11/8/2023	10:30AM	11/8/2023	10:50AM
11/8/2023	11:00AM	11/8/2023	11:30AM
11/8/2023	12:00PM	11/8/2023	12:15PM
11/8/2023	1:00PM	11/8/2023	01:14PM
11/8/2023	1:30PM	11/8/2023	01:40PM
11/9/2023	3:00PM	11/9/2023	03:10PM
11/9/2023	3:30PM	11/9/2023	03:50PM
11/9/2023	09:15AM	11/9/2023	09:30AM
11/9/2023	09:30AM	11/9/2023	09:45AM
11/9/2023	11:20AM	11/9/2023	11:55AM
11/9/2023	12:00PM	11/9/2023	12:20PM
4/24/2024	03:00PM	4/24/2024	03:10PM
4/24/2024	03:45PM	4/24/2024	03:55PM
4/24/2024	04:00PM	4/24/2024	04:50PM

4/24/2024	09:00AM	4/24/2024	09:00AM
4/24/2024	10:30AM	4/24/2024	10:30AM
4/24/2024	01:30PM	4/24/2024	01:30PM
4/24/2024	03:45PM	4/24/2024	03:45PM
4/24/2024	09:00AM	4/24/2024	09:00AM
4/24/2024	10:30AM	4/24/2024	10:30AM
4/24/2024	11:15AM	4/24/2024	11:15AM
4/24/2024	01:30PM	4/24/2024	01:30PM
4/29/2024	03:45PM	4/29/2024	03:45PM
4/29/2024	04:15PM	4/29/2024	04:15PM
4/29/2024	04:45PM	4/29/2024	04:45PM
4/29/2024	09:00AM	4/29/2024	09:00AM
4/29/2024	12:00PM	4/29/2024	12:00PM
4/29/2024	02:45PM	4/29/2024	02:45PM
4/29/2024	06:00PM	4/29/2024	06:00PM
4/29/2024	09:00AM	4/29/2024	09:00AM
4/29/2024	10:05AM	4/29/2024	10:05AM
11/13/2023	11:12AM	11/13/2023	11:12AM
11/13/2023	12:50PM	11/13/2023	12:50PM
11/13/2023	09:15AM	11/13/2023	09:15AM
11/13/2023	10:10AM	11/13/2023	10:10AM
11/14/2023	11:20AM	11/14/2023	11:20AM
11/14/2023	01:50PM	11/14/2023	01:50PM
11/14/2023	02:45PM	11/14/2023	02:45PM
11/14/2023	03:45PM	11/14/2023	03:45PM
11/14/2023	09:20AM	11/14/2023	09:20AM
11/14/2023	10:20PM	11/14/2023	10:20PM
11/5/2023	12:35PM	11/5/2023	12:35PM
11/5/2023	02:50PM	11/5/2023	02:50PM
11/5/2023	04:00PM	11/5/2023	04:00PM
11/5/2023	09:30AM	11/5/2023	09:30AM
11/5/2023	10:20PM	11/5/2023	10:20PM
11/5/2023	11:35AM	11/5/2023	11:35AM
11/5/2023	01:40PM	11/5/2023	01:40PM
11/5/2023	03:00PM	11/5/2023	03:00PM
11/16/2023	04:20PM	11/16/2023	04:20PM
11/16/2023	09:15PM	11/16/2023	09:15PM
11/16/2023	10:20PM	11/16/2023	10:20PM
11/16/2023	11:30PM	11/16/2023	11:30PM
11/16/2023	02:10PM	11/16/2023	02:10PM
11/16/2023	03:20PM	11/16/2023	03:20PM
11/16/2023	06:00PM	11/16/2023	06:00PM
11/17/2023	09:30AM	11/17/2023	09:30AM
11/17/2023	11:30AM	11/17/2023	11:30AM
11/17/2023	01:00PM	11/17/2023	01:00PM

11/17/2023	03:50PM	11/17/2023	02:10PM
11/17/2023	04:55PM	11/17/2023	03:20PM
11/17/2023	7:10PM	11/17/2023	06:00PM
11/17/2023	09:10AM	11/17/2023	09:30AM
11/30/2023	10:15AM	11/30/2023	7:20M
11/30/2023	11:30AM	11/30/2023	6:45AM
11/30/2023	12:30PM	11/30/2023	2:45PM
11/30/2023	01:45PM	11/30/2023	12:40PM
11/30/2023	02:30PM	11/30/2023	12:30PM
11/30/2023	03:40PM	11/30/2023	12:15PM
11/30/2023	04:50PM	11/30/2023	2:00PM
12/1/2023	09:50AM	12/1/2023	1:20PM
12/1/2023	11:30AM	12/1/2023	11:00PM
12/1/2023	02:20PM	12/1/2023	9:45PM
12/1/2023	04:30PM	12/1/2023	8:10PM
12/1/2023	10:10AM	12/1/2023	7:20M
12/1/2023	11:20AM	12/1/2023	6:45AM
12/1/2023	01:50PM	12/1/2023	2:45PM
12/1/2023	02:45PM	12/1/2023	12:40PM
5/1/2024	03:45PM	5/1/2024	3:50PM
5/1/2024	09:20AM	5/1/2024	9:45PM
5/1/2024	10:20PM	5/1/2024	9:00PM
5/1/2024	12:35PM	5/1/2024	7:05 PM
5/1/2024	02:50PM	5/1/2024	3:00PM
5/1/2024	04:00PM	5/1/2024	11:15AM
5/1/2024	09:30AM	5/1/2024	6:15PM
5/1/2024	10:20PM	5/1/2024	10:40 AM
5/3/2024	11:35AM	5/3/2024	9:45 AM
5/3/2024	01:40PM	5/3/2024	9:35 AM
5/3/2024	03:00PM	5/3/2024	9:35
5/3/2024	04:20PM	5/3/2024	8:10AM
5/3/2024	09:15PM	5/3/2024	7:50AM
5/3/2024	10:20PM	5/3/2024	8:15AM
5/3/2024	11:30PM	5/3/2024	7:20M
5/3/2024	02:10PM	5/3/2024	6:45AM
5/5/2024	03:20PM	5/5/2024	2:45PM
5/5/2024	06:00PM	5/5/2024	12:40PM
5/5/2024	09:30AM	5/5/2024	12:30PM
5/5/2024	10:20AM	5/5/2024	12:15PM
5/5/2024	11:30AM	5/5/2024	2:00PM
5/5/2024	01:00PM	5/5/2024	1:20PM
11/23/2023	02:20PM	11/23/2023	11:00PM
11/23/2023	3:00pm	11/23/2023	9:45PM
11/23/2023	55:0pm	11/23/2023	8:10PM
11/23/2023	5:30pm	11/23/2023	8:20PM

11/23/2023	9:00am	11/23/2023	8:25PM
11/23/2023	10:00am	11/23/2023	7:35PM
11/23/2023	11:00AM	11/23/2023	7:40PM
11/23/2023	12:00PM	11/23/2023	7:20PM
5/8/2024	2:00PM	5/8/2024	6:20PM
5/8/2024	3:00PM	5/8/2024	7:10PM
5/8/2024	5:00PM	5/8/2024	6:05PM
5/8/2024	9:00PM	5/8/2024	5:30PM
5/8/2024	10:00AM	5/8/2024	3:40PM
5/8/2024	11:00AM	5/8/2024	2:20PM
5/8/2024	12:00PM	5/8/2024	12:30PM
5/8/2024	10:00AM	5/8/2024	11:50PM
11/25/2023	11:00AM	11/25/2023	11:10AM
11/25/2023	12:00PM	11/25/2023	11:05PM
11/25/2023	3:00PM	11/25/2023	9:50AM
11/25/2023	5:00PM	11/25/2023	1:45PM
11/25/2023	10:00AM	11/25/2023	8:45 PM
11/25/2023	11:00AM	11/25/2023	5:45 PM
11/25/2023	12:00PM	11/25/2023	12:20 PM
11/25/2023	1:00PM	11/25/2023	8:50AM
5/10/2024	2:00PM	5/10/2024	8:40AM
5/10/2024	2:30PM	5/10/2024	8:25AM
5/10/2024	3:00PM	5/10/2024	11:10AM
5/10/2024	4:30PM	5/10/2024	10:15PM
5/10/2024	02:10AM	5/10/2024	9:50AM
5/10/2024	05:00AM	5/10/2024	12:30PM
5/10/2024	07:00AM	5/10/2024	9:10PM
11/27/2023	10:30AM	11/27/2023	8:45PM
11/27/2023	01:00AM	11/27/2023	8:25PM
11/27/2023	11:30AM	11/27/2023	8:00PM
11/27/2023	10:15AM	11/27/2023	7:45PM
11/27/2023	11:15PM	11/27/2023	4:10PM
11/27/2023	10:00AM	11/27/2023	3:10PM
11/27/2023	9:30AM	11/27/2023	3:00PM
11/28/2023	12:10PM	11/28/2023	12:40PM
11/28/2023	4:15PM	11/28/2023	11:00PM
11/28/2023	9:45AM	11/28/2023	10:40AM
11/28/2023	9:30AM	11/28/2023	10:40PM
11/28/2023	9:00AM	11/28/2023	10:30AM
11/28/2023	4:20PM	11/28/2023	12:30PM
11/28/2023	4:00PM	11/28/2023	1:25PM
11/28/2023	3:20PM	11/28/2023	9:00AM
11/17/2023	11:30AM	11/17/2023	11:30AM
11/17/2023	01:00PM	11/17/2023	01:00PM
11/17/2023	03:50PM	11/17/2023	02:10PM

11/17/2023	04:55PM	11/17/2023	03:20PM
11/17/2023	7:10PM	11/17/2023	06:00PM
11/17/2023	09:10AM	11/17/2023	09:30AM
11/30/2023	10:15AM	11/30/2023	7:20M
11/30/2023	11:30AM	11/30/2023	6:45AM
11/30/2023	12:30PM	11/30/2023	2:45PM
11/30/2023	01:45PM	11/30/2023	12:40PM
11/30/2023	02:30PM	11/30/2023	12:30PM
11/30/2023	03:40PM	11/30/2023	12:15PM
11/30/2023	04:50PM	11/30/2023	2:00PM
12/1/2023	09:50AM	12/1/2023	1:20PM
12/1/2023	11:30AM	12/1/2023	11:00PM
12/1/2023	02:20PM	12/1/2023	9:45PM
12/1/2023	04:30PM	12/1/2023	8:10PM
12/1/2023	10:10AM	12/1/2023	7:20M
12/1/2023	11:20AM	12/1/2023	6:45AM
12/1/2023	01:50PM	12/1/2023	2:45PM
12/1/2023	02:45PM	12/1/2023	12:40PM

2/2/2024	03:00PM	2/2/2024	03:10PM
2/2/2024	03:45PM	2/2/2024	03:55PM
2/2/2024	04:00PM	2/2/2024	04:50PM
2/2/2024	09:00AM	2/2/2024	09:00AM
2/2/2024	10:30AM	2/2/2024	10:30AM
2/3/2024	01:30PM	2/3/2024	01:30PM
2/3/2024	03:45PM	2/3/2024	03:45PM
2/3/2024	09:00AM	2/3/2024	09:00AM
2/3/2024	10:30AM	2/3/2024	10:30AM
2/3/2024	11:15AM	2/3/2024	11:15AM
2/3/2024	01:30PM	2/3/2024	01:30PM
2/6/2024	03:45PM	2/6/2024	03:45PM
2/6/2024	04:15PM	2/6/2024	04:15PM
2/6/2024	04:45PM	2/6/2024	04:45PM
2/6/2024	09:00AM	2/6/2024	09:00AM
2/6/2024	12:00PM	2/6/2024	12:00PM
2/6/2024	02:45PM	2/6/2024	02:45PM
2/6/2024	06:00PM	2/6/2024	06:00PM
2/6/2024	09:00AM	2/6/2024	09:00AM
2/6/2024	10:05AM	2/6/2024	10:05AM
2/10/2024	11:12AM	2/10/2024	11:12AM
2/10/2024	12:50PM	2/10/2024	12:50PM
2/10/2024	09:15AM	2/10/2024	09:15AM
2/10/2024	10:10AM	2/10/2024	10:10AM
2/14/2024	11:20AM	2/14/2024	11:20AM
2/14/2024	01:50PM	2/14/2024	01:50PM
2/14/2024	02:45PM	2/14/2024	02:45PM
2/14/2024	03:45PM	2/14/2024	03:45PM
2/14/2024	09:20AM	2/14/2024	09:20AM
2/14/2024	10:20PM	2/14/2024	10:20PM
2/18/2024	12:35PM	2/18/2024	12:35PM

2/18/2024	02:50PM	2/18/2024	02:50PM
2/18/2024	04:00PM	2/18/2024	04:00PM
2/18/2024	09:30AM	2/18/2024	09:30AM
2/18/2024	10:20PM	2/18/2024	10:20PM
2/18/2024	11:35AM	2/18/2024	11:35AM
2/18/2024	01:40PM	2/18/2024	01:40PM
2/18/2024	03:00PM	2/18/2024	03:00PM
2/20/2024	04:20PM	2/20/2024	04:20PM
2/20/2024	09:15PM	2/20/2024	09:15PM
2/20/2024	10:20PM	2/20/2024	10:20PM
2/20/2024	11:30PM	2/20/2024	11:30PM
2/20/2024	02:10PM	2/20/2024	02:10PM
2/20/2024	03:20PM	2/20/2024	03:20PM
2/20/2024	06:00PM	2/20/2024	06:00PM
2/23/2024	09:30AM	2/23/2024	09:30AM
2/23/2024	11:30AM	2/23/2024	11:30AM
2/23/2024	01:00PM	2/23/2024	01:00PM
2/23/2024	03:50PM	2/23/2024	02:10PM
2/23/2024	04:55PM	2/23/2024	03:20PM
2/23/2024	7:10PM	2/23/2024	06:00PM
2/23/2024	09:10AM	2/23/2024	09:30AM
2/26/2024	10:15AM	2/26/2024	7:20M
2/26/2024	11:30AM	2/26/2024	6:45AM
2/26/2024	12:30PM	2/26/2024	2:45PM
2/26/2024	01:45PM	2/26/2024	12:40PM
2/26/2024	02:30PM	2/26/2024	12:30PM
2/26/2024	03:40PM	2/26/2024	12:15PM
2/26/2024	04:50PM	2/26/2024	2:00PM
2/27/2024	09:50AM	2/27/2024	1:20PM
2/27/2024	11:30AM	2/27/2024	11:00PM
2/27/2024	02:20PM	2/27/2024	9:45PM
2/27/2024	04:30PM	2/27/2024	8:10PM
2/27/2024	10:10AM	2/27/2024	7:20M
2/27/2024	11:20AM	2/27/2024	6:45AM
2/27/2024	01:50PM	2/27/2024	2:45PM
2/27/2024	02:45PM	2/27/2024	12:40PM
3/1/2024	03:45PM	3/1/2024	3:50PM
3/1/2024	09:20AM	3/1/2024	9:45PM
3/1/2024	10:20PM	3/1/2024	9:00PM
3/1/2024	12:35PM	3/1/2024	7:05 PM
3/1/2024	02:50PM	3/1/2024	3:00PM
3/1/2024	04:00PM	3/1/2024	11:15AM
3/1/2024	09:30AM	3/1/2024	6:15PM
3/1/2024	10:20PM	3/1/2024	10:40 AM
3/6/2024	11:35AM	3/6/2024	9:45 AM

3/6/2024	01:40PM	3/6/2024	9:35 AM
3/6/2024	03:00PM	3/6/2024	9:35
3/6/2024	04:20PM	3/6/2024	8:10AM
3/6/2024	09:15PM	3/6/2024	7:50AM
3/6/2024	10:20PM	3/6/2024	8:15AM
3/6/2024	11:30PM	3/6/2024	7:20M
3/6/2024	02:10PM	3/6/2024	6:45AM
3/8/2024	03:20PM	3/8/2024	2:45PM
3/8/2024	06:00PM	3/8/2024	12:40PM
3/8/2024	09:30AM	3/8/2024	12:30PM
3/8/2024	10:20AM	3/8/2024	12:15PM
3/8/2024	11:30AM	3/8/2024	2:00PM
3/8/2024	01:00PM	3/8/2024	1:20PM
3/10/2024	02:20PM	3/10/2024	11:00PM
3/10/2024	3:00pm	3/10/2024	9:45PM
3/10/2024	55:0pm	3/10/2024	8:10PM
3/10/2024	5:30pm	3/10/2024	8:20PM
3/10/2024	9:00am	3/10/2024	8:25PM
3/10/2024	10:00am	3/10/2024	7:35PM
3/10/2024	11:00AM	3/10/2024	7:40PM
3/10/2024	12:00PM	3/10/2024	7:20PM
3/14/2024	2:00PM	3/14/2024	6:20PM
3/14/2024	3:00PM	3/14/2024	7:10PM
3/14/2024	5:00PM	3/14/2024	6:05PM
3/14/2024	9:00PM	3/14/2024	5:30PM
3/14/2024	10:00AM	3/14/2024	3:40PM
3/14/2024	11:00AM	3/14/2024	2:20PM
3/14/2024	12:00PM	3/14/2024	12:30PM
3/14/2024	10:00AM	3/14/2024	11;50PM
3/17/2024	11:00AM	3/17/2024	11:10AM
3/17/2024	12:00PM	3/17/2024	11:05PM
3/17/2024	3:00PM	3/17/2024	9:50AM
3/17/2024	5:00PM	3/17/2024	1:45PM
3/17/2024	10:00AM	3/17/2024	8:45 PM
3/17/2024	11:00AM	3/17/2024	5:45 PM
3/17/2024	12:00PM	3/17/2024	12:20 PM
3/17/2024	1:00PM	3/17/2024	8:50AM
3/19/2024	2:00PM	3/19/2024	8:40AM
3/19/2024	2:30PM	3/19/2024	8:25AM
3/19/2024	3:00PM	3/19/2024	11:10AM
3/19/2024	4:30PM	3/19/2024	10:15PM
3/19/2024	02:10AM	3/19/2024	9:50AM
3/19/2024	05:00AM	3/19/2024	12:30PM

3/19/2024	07:00AM	3/19/2024	9:10PM
3/21/2024	10:30AM	3/21/2024	8:45PM
3/21/2024	01:00AM	3/21/2024	8:25PM
3/21/2024	11:30AM	3/21/2024	8:00PM
3/21/2024	10:15AM	3/21/2024	7:45PM
3/21/2024	11:15PM	3/21/2024	4;10PM
3/21/2024	10:00AM	3/21/2024	3:10PM
3/21/2024	9:30AM	3/21/2024	3:00PM
3/25/2024	12:10PM	3/25/2024	12:40PM
3/25/2024	4:15PM	3/25/2024	11:00PM
3/25/2024	9:45AM	3/25/2024	10:40AM
3/25/2024	9:30AM	3/25/2024	10:40PM
3/25/2024	9:00AM	3/25/2024	10:30AM
3/25/2024	4:20PM	3/25/2024	12:30PM
3/25/2024	4:00PM	3/25/2024	1:25PM
3/25/2024	3:20PM	3/25/2024	9:00AM
3/29/2024	2:30PM	3/29/2024	8:30AM
3/29/2024	1:42PM	3/29/2024	7:40AM
3/29/2024	1:15PM	3/29/2024	6:50AM
3/29/2024	1:10PM	3/29/2024	6:25AM
3/29/2024	9:30AM	3/29/2024	9:35AM
3/29/2024	9:00PM	3/29/2024	9:10PM
3/29/2024	9:00AM	3/29/2024	8:55PM
3/29/2024	8:00PM	3/29/2024	8:20PM
3/30/2024	7:00PM	3/30/2024	7:40PM
3/30/2024	6:00PM	3/30/2024	7:50PM
3/30/2024	5:15PM	3/30/2024	7:20PM
3/30/2024	5:00PM	3/30/2024	5:45PM
3/30/2024	4:00PM	3/30/2024	4:15PM
3/30/2024	3:10PM	3/30/2024	12:05PM
3/30/2024	12:00PM	3/30/2024	9:55PM
3/30/2024	11:30PM	3/30/2024	9:40AM
4/1/2024	11:00PM	4/1/2024	9:45AM
4/1/2024	10:0AM	4/1/2024	9:30AM
4/1/2024	10:0AM	4/1/2024	8:20AM
4/1/2024	10:00PM	4/1/2024	7:40PM
4/1/2024	8:40PM	4/1/2024	4:45PM
4/1/2024	6:55PM	4/1/2024	11:30AM
4/1/2024	5:00AM	4/1/2024	10:55AM
4/1/2024	2:00AM	4/1/2024	10:40AM
4/3/2024	12:30PM	4/3/2024	10:30PM
4/3/2024	12:00PM	4/3/2024	1:45PM
4/3/2024	11:00AM	4/3/2024	10:00PM
4/3/2024	10:00AM	4/3/2024	8:45PM
4/3/2024	1:00PM	4/3/2024	8:05PM

4/3/2024	9:00AM	4/3/2024	7:55PM
4/3/2024	7:35PM	4/3/2024	7:35PM
4/3/2024	7:00AM	4/3/2024	7:10PM
4/6/2024	3:00PM	4/6/2024	6:50PM
4/6/2024	3:00PM	4/6/2024	6:35PM
4/6/2024	9:30AM	4/6/2024	6:20PM
4/6/2024	9:00AM	4/6/2024	10:45PM
4/6/2024	8:10PM	4/6/2024	10:40PM
4/6/2024	7:00PM	4/6/2024	10:25PM
4/6/2024	2:00PM	4/6/2024	10:15PM
4/8/2024	10:30AM	4/8/2024	6:20PM
4/8/2024	1:00PM	4/8/2024	7:45PM
4/8/2024	12:00PM	4/8/2024	5:00PM
4/8/2024	11:00AM	4/8/2024	4:00PM
4/8/2024	10:30AM	4/8/2024	11:20AM
4/8/2024	10:00AM	4/8/2024	11:15AM
4/8/2024	9:30AM	4/8/2024	11:45AM
4/11/2024	9:30AM	4/11/2024	9:15PM
4/11/2024	10:30AM	4/11/2024	9:30AM
4/11/2024	1:30PM	4/11/2024	3:00PM
4/11/2024	9:30PM	4/11/2024	8:55PM
4/11/2024	3:30PM	4/11/2024	6:5PM
4/11/2024	11:30AM	4/11/2024	5:40PM
4/11/2024	10:30PM	4/11/2024	5:55PM
4/15/2024	3:00PM	4/15/2024	12:15PM
4/15/2024	12:30PM	4/15/2024	11:40PM
4/15/2024	11:00AM	4/15/2024	11:00PM
4/15/2024	10:00AM	4/15/2024	11:00PM
4/15/2024	4:00PM	4/15/2024	10:20PM
4/15/2024	3:50PM	4/15/2024	1:15PM
4/15/2024	12:30PM	4/15/2024	10:15PM
4/15/2024	11:30AM	4/15/2024	9:20AM
4/18/2024	10:00AM	4/18/2024	7:15PM
4/18/2024	2:00PM	4/18/2024	7:00PM
4/18/2024	11:00AM	4/18/2024	6:35PM
4/18/2024	10:00PM	4/18/2024	6:20PM
4/18/2024	6:30AM	4/18/2024	3:30PM
4/18/2024	11:00AM	4/18/2024	3:00PM
4/18/2024	9:32PM	4/18/2024	10:50PM
4/18/2024	8:05PM	4/18/2024	10:30PM
4/18/2024	6:35PM	4/18/2024	4:00PM
4/18/2024	12:35PM	4/18/2024	2:20 PM
4/21/2024	11:05AM	4/21/2024	1:45 PM
4/21/2024	6:05 PM	4/21/2024	1:25 PM

4/21/2024	10:32 AM	4/21/2024	11:15 AM
4/21/2024	9:35 AM	4/21/2024	10:20 AM
4/21/2024	9:25 AM	4/21/2024	9:30 AM
4/21/2024	9:17 am	4/21/2024	9:35 AM
4/21/2024	8:00AM	4/21/2024	8:35AM
4/21/2024	7:35AM	4/21/2024	7:30PM
4/24/2024	8:10AM	4/24/2024	12:30PM
4/24/2024	7:00AM	4/24/2024	12:15PM
4/24/2024	6:20AM	4/24/2024	12:30PM
4/24/2024	2:20PM	4/24/2024	11:45PM
4/24/2024	12:30PM	4/24/2024	3:00PM
4/24/2024	12:25PM	4/24/2024	3:30 PM
4/24/2024	11:15AM	4/24/2024	2:45 PM
4/24/2024	1:40PM	4/24/2024	1:10 PM
4/29/2024	1:15PM	4/29/2024	10:30 AM
4/29/2024	10:00PM	4/29/2024	10:00 AM
4/29/2024	9:30AM	4/29/2024	9:30PM
4/29/2024	8:06PM	4/29/2024	9:00PM
4/29/2024	8:15PM	4/29/2024	8:40PM
4/29/2024	8:20PM	4/29/2024	8:45AM
4/29/2024	7:30PM	4/29/2024	8:50AM
4/29/2024	7:20PM	4/29/2024	8:20PM
4/29/2024	7:00PM	4/29/2024	7:45PM
4/29/2024	6:00PM	4/29/2024	6:45PM
4/29/2024	7:05PM	4/29/2024	6:50PM
4/29/2024	5:50PM	4/29/2024	7:20AM
4/29/2024	5:10PM	4/29/2024	6:30PM
4/30/2024	3:20PM	4/30/2024	6:10PM
4/30/2024	2:20PM	4/30/2024	2:10PM
4/30/2024	12:00PM	4/30/2024	11:55PM
4/30/2024	11:45PM	4/30/2024	11:00AM
4/30/2024	11:00AM	4/30/2024	10:10AM
4/30/2024	10:55PM	4/30/2024	8:10PM
4/30/2024	9:40AM	4/30/2024	7:45PM
4/30/2024	1:35PM	4/30/2024	7:15PM
4/30/2024	8:30 PM	4/30/2024	6:20PM
4/30/2024	5:15 PM	4/30/2024	6:20AM
4/30/2024	12:10 PM	4/30/2024	6:00PM
4/30/2024	8:45AM	4/30/2024	5:00PM
4/30/2024	8:30AM	4/30/2024	4:00PM
4/30/2024	8;10AM	4/30/2024	3:40AM
4/30/2024	11:00AM	4/30/2024	3:20AM
5/1/2024	9:35PM	5/1/2024	2:25PM
5/1/2024	9:35AM	5/1/2024	1:30PM

5/1/2024	9:30AM	5/1/2024	12:55AM
5/1/2024	9:05PM	5/1/2024	12:25AM
5/1/2024	8:35PM	5/1/2024	12:30PM
5/1/2024	8:10PM	5/1/2024	11:20PM
5/1/2024	7:50PM	5/1/2024	12:30PM
5/1/2024	7:05PM	5/1/2024	10:45AM
5/3/2024	4:05PM	5/3/2024	11:30AM
5/3/2024	3:00PM	5/3/2024	10:30AM
5/3/2024	2:35PM	5/3/2024	10:30AM
5/3/2024	12:30PM	5/3/2024	2:00PM
5/3/2024	10:55PM	5/3/2024	9:45PM
5/3/2024	10:30AM	5/3/2024	9:40PM
5/3/2024	10:20PM	5/3/2024	9:30PM
5/3/2024	10:15AM	5/3/2024	7:00PM
5/5/2024	10:00AM	5/5/2024	1:45AM
5/5/2024	1:18PM	5/5/2024	2:30PM
5/5/2024	8:45AM	5/5/2024	11:00PM
5/5/2024	8:05AM	5/5/2024	2:00PM
5/5/2024	7:20AM	5/5/2024	9:00PM
5/5/2024	6:35AM	5/5/2024	8:45AM
5/8/2024	6:05AM	5/8/2024	8:15 PM
5/8/2024	9:35AM	5/8/2024	7:15PM
5/8/2024	8:55PM	5/8/2024	6:30PM
5/8/2024	8:50PM	5/8/2024	3:30PM
5/8/2024	8:05PM	5/8/2024	12:00 AM
5/8/2024	7:25PM	5/8/2024	9:30 AM
5/10/2024	7:10PM	5/10/2024	8:30 AM
5/10/2024	6:45PM	5/10/2024	7:45AM
5/10/2024	5:00PM	5/10/2024	9:30AM
5/10/2024	4:05PM	5/10/2024	7:50PM
5/10/2024	12:05PM	5/10/2024	7:45AM
5/13/2024	9:40PM	5/13/2024	11:30PM
5/13/2024	9:35AM	5/13/2024	7:15PM
5/13/2024	9:35AM	5/13/2024	6:10PM
5/13/2024	9:00AM	5/13/2024	11:20PM
5/13/2024	8:00AM	5/13/2024	4:00PM
5/13/2024	7:30PM	5/13/2024	5:00PM
5/13/2024	4:00PM	5/13/2024	3:30PM
5/15/2024	11:00AM	5/15/2024	11:50AM
5/15/2024	10:45AM	5/15/2024	12:30PM
5/15/2024	10:35AM	5/15/2024	11:10 AM
5/15/2024	10:00PM	5/15/2024	11:20AM
5/15/2024	1:00PM	5/15/2024	10:30 AM
5/15/2024	8:53PM	5/15/2024	1:35PM
5/15/2024	8:40PM	5/15/2024	1:30 PM

5/17/2024	7:59PM	5/17/2024	10:55 AM
5/17/2024	7:40PM	5/17/2024	12:20 AM
5/17/2024	7:20PM	5/17/2024	12:15 AM
5/17/2024	7:00PM	5/17/2024	9:45AM
5/17/2024	6:40PM	5/17/2024	11:30PM
5/17/2024	6:24PM	5/17/2024	7:00PM
5/19/2024	6:10PM	5/19/2024	6:40PM
5/19/2024	10:40PM	5/19/2024	6:15PM
5/19/2024	10:25PM	5/19/2024	5:30PM
5/19/2024	10:15PM	5/19/2024	5:00PM
5/19/2024	10:10PM	5/19/2024	1:00PM
5/19/2024	6:00PM	5/19/2024	12:00PM
5/21/2024	5:00PM	5/21/2024	2:00PM
5/21/2024	4:30PM	5/21/2024	12:00PM
5/21/2024	3:00PM	5/21/2024	12:00PM
5/21/2024	11:09AM	5/21/2024	11:55PM
5/21/2024	11:05AM	5/21/2024	1:30PM
5/25/2024	11:00AM	5/25/2024	10:40AM
5/25/2024	9:00PM	5/25/2024	10:20AM
5/25/2024	9:00AM	5/25/2024	5:00PM
5/25/2024	9:00AM	5/25/2024	10:10AM
5/25/2024	8:35PM	5/25/2024	4:15PM
5/25/2024	6:00PM	5/25/2024	11:30PM
5/25/2024	5:35PM	5/25/2024	9:50AM
5/25/2024	5:25PM	5/25/2024	8:00PM
5/25/2024	12:00PM	5/25/2024	7:00PM
5/25/2024	11:00PM	5/25/2024	6:15PM
5/25/2024	10:25PM	5/25/2024	5:00PM
5/25/2024	10:25PM	5/25/2024	3:45PM
5/25/2024	10:00PM	5/25/2024	3:50PM
5/25/2024	1:00PM	5/25/2024	2:50PM
5/25/2024	9:45PM	5/25/2024	3:00PM
5/28/2024	8:55AM	5/28/2024	1:20PM
5/28/2024	7:05PM	5/28/2024	3:30PM
5/28/2024	6:50PM	5/28/2024	12:50PM
5/28/2024	6:25PM	5/28/2024	11:45PM
5/28/2024	6:10PM	5/28/2024	11:45AM
5/28/2024	3:00PM	5/28/2024	11:30AM
5/28/2024	2:30PM	5/28/2024	10:55AM
5/29/2024	10:40PM	5/29/2024	1:45PM
5/29/2024	10:20PM	5/29/2024	1:50 PM
5/29/2024	1:30PM	5/29/2024	10:55AM
5/29/2024	2:10 PM	5/29/2024	9:45AM
5/29/2024	1:30 PM	5/29/2024	3:30PM
5/29/2024	1:15 PM	5/29/2024	2:30PM

[illegible]

[illegible]

[illegible]

[illegible]

B/12 03/06/23 Physiology, 2000's ~~exam~~

05/06/23 automatic door closure is not working in
molecular lab. ✓

8/6/23 Friday - door is not closed Microbiology
9/6/23. Microbiology HOD Room Ballroom & Fresh tank
complaint.

10/06/23 wall mounting of projector in skill lab. - Amal

13/06/23 Forensic medicine. — Bathroom Blocked (Gens bathroom)

13/6/23 - community medicine
(adies) toilet-Bathroom blocked

13/6/23 - Community medicine (too room Fan
(working slowly))

20/06/23 - Projector setting in Skill Lab up

21-6-23 - CONV അമ്മയുടെ ബാത്ത് മുറയ്ക്കുക. ✓
 Fresh water നൽകി അമ്മയെ CONV Bath Room

21/6/23

- Bookroom Flesh Cank leak
- ^{demonstrator} ~~e~~ Curstain change to Liberator.

21-06-2023 KMIT college of Architecture, Manassery to change
a tube electrical maintenance / example

23/06/2023 - Dept of pharmacology -
Demo room - Tube (2 nos) Not working

24/6/23 - Plug fitting. ~~not~~

3/7/23 bathroom floor complaint - Hob doors micro biology

4/7/2023 - KMCT college of Alvaring - ~~switches~~ not working in
Hersel.

4/7/23 Tap Logging NCP in Cards

07/07/23 Physiology: 2 problems room number 2000

11/17/23 Anatomy Associate intern (2 room) tube net working

15-7-23 CON TV ^{Thy} year Class Room Plug ~~not~~ working.

20/7/23 girls Common room bathroom closet
fresh tank complaint

21/7/23 - Department of PHARMACOLOGY
PA Room - Fan not working. *[Signature]*
21/7/2023

24/7/23 - Not getting water supply in cooler
- Biochemistry

24/4/23 - Electrical & maintenance work in kmr Architecture college. ✓ Hand N N

25/7/23 please change the Lap For. ^{pathology} ~~pathology~~
(pathology)

26/7/23 Department of forensic medicine & Toxicology
Department office - Fan not working.

26/7/23 Community medicine boys toilet
water leakage

26/7/23 Pathology department Toilet Flush not working

26/7/23 Pathology dept system computer system
not working

27/7/23 Forensic medicine = girls toilet —
The wash basin in the
bathroom is broken and urgently
needs to be repaired.

31/7/23 Kmer Nany college: Bleached
were lack of whiskey can
urgent

31.7.23 STAFF QUARTERS NO. 12
DR. SATHEESH 8129594840
Fix a power point socket (Three pin
Large & Small 2 in one)

31/07/23 Biochemstry Department — Research Lab - Pipe Complaint

28-7-23 K.M.C.T. College of Nursing IIIrd year Classroom
Smith board complaint.

11/8/23 office Motion complaint. Babu. K.T.C.

16/08/23 Lecture hall 4 AC is Not working,
Program on 20/08/23.

16/08/23 Washing Room (Bathroom) Pipe complaint - Pathology
Projector cable change - NCP

26/08/23 P.C. Hostel tube mirror in mygalid. (F4 R.No.)

01/09/2023 24.08.2023 electrical maintenance work at kmct
College of Architecture, Manassery.

01-09-2023 Electrical & plumbing maintenance work at
kmct college of Architecture, Manassery.

4/9/23 L.H.I. - Room No: 35 fans working
Room No: 51 - Fan sound

11-9-2023 CON 2nd 2nd 2nd 2nd 2nd 2nd 2nd 2nd
2nd 2nd 2nd 2nd 2nd 2nd 2nd 2nd

4/9/23 Community medicine 1100 room tube light
not working

13/09/23 Physiology Demo room 3 fans not working.

14/09/23 Table fan not working in the Room - 11th floor of
Buchanan's Building

14/9/23 Crossing Room - Pipe change. Crossing station

16/9/23 A/c new plug. Anatomy dept (Tutor room)

17 16/9/23 Tap and bulb compliance (pathology)

2 20/9/23 Tube is not working (micro) Students Lab

20/09/23

Dept of pharmacology - PCR room - Fan, Tube not working
Experimental pharmacology Tube not working

30-09-2023

Tube fitting - Diploma class Lab, Exam hall
NCP

30/9

3/10/23 - 1) Community medicine Gents toilet (near) Research Lab
Light not working

2) Community medicine Lab - Tube light not working

3) Community medicine Lab first half no-water supply

03/10/23

6/10/23

Tube complaint

micro

6/10/2023

Associate professor board bldg Community medicine

13/10/23

Sink leak in Students Lab - microbiology

16/10/23

Pharmacology tutor room Tube light and fan fitting

19/10/2023 Boys Hostel water pipe complaint

19/10/23 - For Molecular lab, water leakage in Sterilization room

L.H-1

Room-69 - Light - complaint

Room-53 - Light - complaint

Room-61 - bulb - complaint

3/11/2023 - Community Medicine Dr Anupama room
Tube light complaint

21/11/23 - Sink leak - Students Lab - micro

3/11/23 M.B.S Boys hostel switch complaint

4/11/23 M.B.S Boys hostel Bath Room Tube Complaint

4.11.2023. കോളിംഗ് കോട്ടേജ് Room No. 17-ൽ
അടവ് ലിജ്. നഷ്ടം.

4.11.2023. C.O.N വിതരണ ക്ലബ്ബ് ഓഫ് റിപ്പോർട്ട്

6/11/23 M.B.S Boys Hostel Bath Room Complaint

11/11/23

15/11/23 - Music work urgent - pharmacology

16/11/20 M.B.S Boys Hostel San Sand. complaint

11F 21 Room

22/11/23 Boys Hostel
#3 Tube complaint

22/11/23 - Community Medicine Dr Anupama - Room
Light not working